

Laboratory EMC Division, SAMEER – Centre for Microwave Research, Sector – 7,
 Rain Tree Marg, CBD – Belapur, Navi Mumbai, Maharashtra
Accreditation Standard ISO/IEC 17025: 2005
Certificate Number CC-2273 **Page** 1 of 10
Validity 18.05.2018 to 17.05.2020 **Last Amended on** 11.01.2019

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
<u>ELECTRO-TECHNICAL CALIBRATION</u>				
I.	SOURCE			
1.	Insertion Loss/RF Attenuation [§] <ul style="list-style-type: none"> • Transient Limiter • Line Impedance Stabilization Network • Bulk Current Injection Probe • Directional Coupler • Attenuator • Cable or similar equipment • 50 Ω conical adapter line 	9 kHz to 7 GHz 0.005 dB to 40 dB	0.6 dB to 0.7 dB	Using Vector Network Analyser by Direct method
2.	Return Loss [§] (VSWR) <ul style="list-style-type: none"> • Attenuator • Antenna • Amplifier • Termination/RF Load • Directional Coupler or Similar Equipment • 50 Ω conical adapter line 	9 kHz to 7 GHz 1 to 46 dB	0.7 dB to 3.3 dB	Using Vector Network Analyser by Direct method

Laboratory EMC Division, SAMEER – Centre for Microwave Research, Sector – 7,
Rain Tree Marg, CBD – Belapur, Navi Mumbai, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2273

Page 2 of 10

Validity 18.05.2018 to 17.05.2020

Last Amended on 11.01.2019

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
3.	Voltage Division Factor ^s <ul style="list-style-type: none"> Coupling & Decoupling Network (CDN) Impedance Stabilization Network (ISN) 	0.15 MHz to 230 MHz (0 dB to 70 dB)	0.7 dB	Using Vector Network Analyser by Direct method as per IEC 610004-6: 2013-10 Ed.4.0, CISPR22: 2008-09 Ed.6.0, CISPR32: 2015-03 Ed.2.0
4.	Impedance ^s <ul style="list-style-type: none"> Line Impedance Stabilization Network Coupling & Decoupling Network, Impedance Stabilization Network 	(1 to 210 Ω) 9 kHz to 30 MHz 30 MHz to 230 MHz	11.44% to 10.8% 10.8 % to 9 %	Using Vector Network Analyser by Direct method as per CISPR 16-1-2:2014-03 Ed.2.0, CISPR22: 2008-09 Ed.6.0
5.	Phase Angle ^s <ul style="list-style-type: none"> Line Impedance Stabilization Network 	9 kHz to 30 MHz (-90° to +90°)	11.44 % to 11.58 %	Using Vector Network Analyser Direct method as per CISPR 16-1-2:2014-03 Ed.2.0
6.	Coupling Factor/Coupling Loss ^s <ul style="list-style-type: none"> Directional Coupler EM Clamp 	9 kHz to 1 GHz 1 to 70 dB	0.5 dB	Using Vector Network Analyser by Direct method

Vishal Shukla
Convenor

Avijit Das
Program Manager

Laboratory EMC Division, SAMEER – Centre for Microwave Research, Sector – 7,
Rain Tree Marg, CBD – Belapur, Navi Mumbai, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2273 **Page** 3 of 10

Validity 18.05.2018 to 17.05.2020 **Last Amended on** 11.01.2019

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
7.	Decoupling of Common Mode Disturbance [§] <ul style="list-style-type: none"> • Coupling & Decoupling Network • Impedance Stabilization Network 	0.15 MHz to 230 MHz (1 dB to 70 dB)	0.5 dB	Using Vector Network Analyser by Direct method
8.	Isolation [§] <ul style="list-style-type: none"> • Line Impedance Stabilization Network • Dual Directional Coupler 	9 kHz to 1000 MHz (10 dB to 90 dB)	0.5 dB	Using Vector Network Analyser CISPR by Direct method 16-1-2:2014-03 Ed.2.0
9.	Combination Wave Surge Test Systems [§]			
	Voltage Amplitude	± 0.5 kV to ± 4.0 kV	5.57 %	Using Digital Storage Oscilloscope HV differential probe & Current Transformer by Direct method as per IEC 61000-4-5 Ed.3.0 2017-08
	Voltage Front Time	$1.2\mu\text{s} \pm 30\%$	8.53 %	
	Voltage Pulse Width	$50\mu\text{s} \pm 20\%$	4.0 %	
	Current Amplitude	± 0.25 kA to ± 3.0 kA	5.52 %	
	Current Front Time	$8\mu\text{s} \pm 20\%$	4.22 %	
	Current Pulse Width	$20\mu\text{s} \pm 20\%$	4.1 %	

Vishal Shukla
Convenor

Avijit Das
Program Manager

Laboratory EMC Division, SAMEER – Centre for Microwave Research, Sector – 7,
Rain Tree Marg, CBD – Belapur, Navi Mumbai, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2273

Page 4 of 10

Validity 18.05.2018 to 17.05.2020

Last Amended on 11.01.2019

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
10.	Electrical Fast Transient Test Systems ^s			
	Pulse Amplitude	± 125 V to ± 4.0 kV	6.38 %	Using Digital Storage Oscilloscope by Direct method as per IEC 61000-4-4 Ed.2012-04 Ed.3.0
	Pulse Rise Time	5 ± 1.5 ns	4.17 %	
	Pulse Width	50 ± 15 ns	4.1 %	
	Repetition Frequency	5 kHz ± 20 %	4.00 %	
		100 kHz ± 20 %	4.00 %	
	Burst Duration	15 ± 3 ms (5 kHz)	4.00 %	
Burst Period	300 ± 60 ms	4.00 %		
11.	Electrostatic Discharge Generator Contact Discharge ^s			
	First Peak Current	7.5- 56.25 A (± 15 %) (± 2 kV to ± 15 kV)	4.6 %	Using Digital Storage Oscilloscope (DSO) along with current target attenuator cable chain by Direct method as per IEC61000-4-2: 2008-12, Ed.2.0
	Rise Time	0.8 ns ± 15 %	7.7 %	
	Current at 30 ns	4.0 - 30 A (± 30 %) (± 2 kV to ± 15 kV)	4.8 %	
	Current at 60 ns	2.0 - 15 A (± 30 %) (± 2 kV to ± 15 kV)	5.7 %	
	Air discharge (DC voltage)			
	Voltage level	± 2 kV (± 5 %)	1.7 %	Using Brandenburg make HV volt meter by Direct method
± 4 kV (± 5 %)				
± 8 kV (± 5 %)				
± 15 kV (± 5 %)				
± 30 kV (± 5 %)				

Vishal Shukla
Convenor

Avijit Das
Program Manager

Laboratory EMC Division, SAMEER – Centre for Microwave Research, Sector – 7,
Rain Tree Marg, CBD – Belapur, Navi Mumbai, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2273

Page 5 of 10

Validity 18.05.2018 to 17.05.2020

Last Amended on 11.01.2019

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
12.	• Electrostatic Discharge Generator ^s (as per automotive standard ISO 10605:2008 E)			
	Contact Discharge (RC Module: 150pF/330 Ω)			
	First Peak Current	7.5- 75 A (± 15 %) (± 2 kV to ± 20 kV)	9.0 %	Using Digital Storage Oscilloscope (DSO) along with current target attenuator cable chain Direct method as per ISO 10605 Second Edition 2014-04-15
	Rise Time	0.8 ns ± 15 %	10.7 %	
	Current at 30 ns	4.0 -40 A (± 30 %) (± 2 kV to ± 20 kV)	9.1 %	
	Current at 60 ns	2.0 -20 A (± 30 %) (± 2 kV to ± 20 kV)	9.4 %	
	Contact Discharge (RC Module: 330pF/330 Ω)			
	First Peak Current	7.5- 75 A (± 15 %) (± 2 kV to ± 20 kV)	9.0 %	Using Digital Storage Oscilloscope (DSO) along with current target attenuator cable chain by Direct method as per ISO 10605: Second Edition 2014-04-15
	Rise Time	0.8 ns ± 25 %	10.6 %	
	Current at 65 ns	4.0 -40 A (± 30 %) (± 2 kV to ± 20 kV)	9.1 %	
	Current at 130 ns	2.0 -20 A (± 30 %) (± 2 kV to ± 20 kV)	9.2 %	

Vishal Shukla
Convenor

Avijit Das
Program Manager

Laboratory

EMC Division, SAMEER – Centre for Microwave Research, Sector – 7,
Rain Tree Marg, CBD – Belapur, Navi Mumbai, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number

CC-2273

Page

6 of 10

Validity

18.05.2018 to 17.05.2020

Last Amended on 11.01.2019

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
	Contact Discharge (RC Module: 150pF/2000Ω)			
	First Peak Current	7.5- 75 A (± 30 %) (±2kV to ± 20 kV)	9.0 %	Using Digital Storage Oscilloscope (DSO) along with current target attenuator cable chain by Direct method as per ISO 10605: Second Edition 2014-04-15
	Rise Time	0.8 ns ± 25 %	10.6 %	
	Current at 180 ns	0.55 -5.55 A (± 30 %) (±2 kV to ± 20 kV)	9.3 %	
	Current at 360 ns	0.3 – 3.0 A (± 50 %) (±2 kV to ± 20 kV)	9.2 %	
	Contact Discharge (RC Module: 330pF/2000Ω)			
	First Peak Current	7.5- 75 A (±30 %) (±2kV to ± 20 kV)	9.0 %	Using Digital Storage Oscilloscope (DSO) along with current target attenuator cable chain by Direct method as per ISO 10605: Second Edition 2014-04-15
	Rise Time	0.8 ns ± 25 %	10.6 %	
	Current at 180 ns	0.55 -5.55 A (± 30 %) (±2 kV to ± 20 kV)	9.3 %	
	Current at 360 ns	0.3 – 3.0 A (± 50 %) (±2 kV to ± 20 kV)	9.2 %	
	Air discharge (DC voltage)			
	Voltage Level	± 2 kV (± 5 %)	0.6 %	Using Brandenburg make HV volt meter by Direct method
		± 4 kV (± 5 %)	0.6 %	
		± 8 kV (± 5 %)	0.6 %	
		± 15 kV (± 5 %)	0.6 %	
		± 25 kV (± 5 %)	0.6 %	
		± 30 kV (± 5 %)	0.5 %	

Vishal Shukla
Convenor

Avijit Das
Program Manager

Laboratory EMC Division, SAMEER – Centre for Microwave Research, Sector – 7, Rain Tree Marg, CBD – Belapur, Navi Mumbai, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2273

Page 7 of 10

Validity 18.05.2018 to 17.05.2020

Last Amended on 11.01.2019

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
13.	Target – attenuator-cable chain ^s			
	Insertion Loss variation(S21)	± 0.5 dB up to 1 GHz	0.7 dB	Using R&S Make VNA, Model No. ZVB 8, Sr. No.16232 and Keithly Make DMM by Direct method as per IEC61000-4-2: 2008-12, Ed.2.0
		± 1.2 dB 1 to 4 GHz	0.7 dB	
	Low Frequency Transfer Impedance	0.603 Ω	0.24%	
DC Resistance	< 2.1 Ω	0.6%		
14.	Ring Wave Generator ^b			
	Voltage Rise Time (Open circuit)	0.5 μ s \pm 30 %	4.64 %	Using Digital Storage Oscilloscope (DSO), HV differential probe & Current Transformer by Direct method as per IEC 61000-4-12: Ed.3.0 2017-07
	Current Rise Time (Short Circuit)	≤ 1 μ s	4.09 %	
	Voltage Oscillation Frequency	100 kHz \pm 10 %	4.0 %	
	Decay in voltage	0.4 < Ratio of Pk2 to Pk1 < 1.1 0.4 < Ratio of Pk3 to Pk2 < 0.8 0.4 < Ratio of Pk4 to Pk3 < 0.8	6.03 %	
	Open Circuit Voltage	250 V - 4 kV \pm 10%	5.82 %	
	Short Circuit Current	133 - 333 A \pm 10%	5.56 %	

Laboratory EMC Division, SAMEER – Centre for Microwave Research, Sector – 7, Rain Tree Marg, CBD – Belapur, Navi Mumbai, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2273

Page 8 of 10

Validity 18.05.2018 to 17.05.2020

Last Amended on 11.01.2019

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
15.	Damped Oscillatory Wave Generator [§]			
	Voltage Amplitude	±0.25 to ±4.0 kV	5.74 %	Using Digital Storage Oscilloscope (DSO), HV differential probe & Current Transformer as per IEC 61000-4-12: 2001-04 Ed. 1.1 & by Direct method as per IEC 61000-4-18: 2011-03 Ed.1.1
	Voltage Rise Time	75 ns ±20 %	4.32 %	
	Oscillation period	100 kHz ±10 %	4.0 %	
	Current Amplitude	1.25-80 A, ± 10 %	6.1 %	
Voltage Decay	$P_{k5} > 0.5 P_{k1}$ and $P_{k10} < 0.5 P_{k1}$	5.69 %		
16.	Voltage Dips & Interruption [§]			
	Output Voltage at no load	Nominal line Voltage (0-100 %)	7.24 %	Using Digital Storage Oscilloscope (DSO), HV differential probe Direct method as per IEC 61000-4-11 Ed. 2.0 2017-05 IEC 61000-4-29:2000-08 first edition
	Dips / Interruption Time	10 ms to 5 s	4.0 %	
RF Signal Generator [§]				
17.	Frequency Accuracy	100 kHz to 8 GHz	0.1 %	Using Spectrum Analyzer (SA), Power Meter, & Frequency Counter by Direct method
	Level Accuracy	-30 dBm to +10 dBm (10 MHz to 18 GHz)	0.4 dB	
	Harmonic Level (upto third harmonic)	Upto -90 dBc (10 Hz to 13.6 GHz)	0.2 dB	
	Amplitude Modulation (AM) Depth	10 % to 90 % (10 Hz to 13.6 GHz)	3.48%	

Laboratory

EMC Division, SAMEER – Centre for Microwave Research, Sector – 7, Rain Tree Marg, CBD – Belapur, Navi Mumbai, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number

CC-2273

Page

9 of 10

Validity

18.05.2018 to 17.05.2020

Last Amended on 11.01.2019

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
18.	RF Amplifiers ^s (Pre amplifier)			
	Gain	50 dB (10 Hz to 6 GHz)	1.7 dB	Using Vector Network Analyzer (VNA), Signal Generator & Spectrum Analyzer (SA) byDirect method
	Harmonics Distortion	Upto (-)90 dBc (10 Hz to 13.6 GHz)	0.2 dB	
	Return Loss of Power Input Port & Output Port	1 to 45 dB (9kHz to 8.5 GHz)	0.5 dB	
	Power Amplifier ^s			
	Gain	50 dB (10 Hz to 3.0 GHz, 1000 W)	1.7 dB	Using Vector Network Analyzer (VNA), Signal Generator & Spectrum Analyzer (SA) byDirect method
	Harmonics Distortion	Upto - 90 dBc (10 Hz to 13.6 GHz, upto 1000 W)	0.2 dB	
	Return Loss of Power Input Port & Output Port	1 - 45 dB (9 kHz to 8.5 GHz)	0.5 dB	
19.	EMI Test Receiver ^s			
	Return (VSWR)	100 kHz to 3 GHz		Using <ul style="list-style-type: none">R & S make VNASchwarzbeck make Calibration Pulse GeneratorAgilent make signal generator As per CISPR 16-1-1 Ed. 3.2 2014-06 byDirect method
	Display error of Quasi Peak Detector	Band A 100 kHz 100 Hz PRF	0.7 dB 0.8 dB	
		Band B 100 Hz PRF MG 100 Hz PRF AG	0.8 dB 0.8 dB	
		Band C/D 100 Hz PRF MG 100 Hz PRF AG	0.8 dB 0.8 dB	

Vishal Shukla
Convenor

Avijit Das
Program Manager

Laboratory EMC Division, SAMEER – Centre for Microwave Research, Sector – 7,
Rain Tree Marg, CBD – Belapur, Navi Mumbai, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2273

Page 10 of 10

Validity 18.05.2018 to 17.05.2020

Last Amended on 11.01.2019

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	Display error of detector with sinusoidal signal (Pk, Qp, Av)	QP detector 9KHz to 1 GHz	0.6 dB	
		Peak Detector 9KHz to 1 GHz	0.6 dB	
		Average Detector 9KHz to 1 GHz	0.6 dB	
	Amplitude Relationship (Peak/QP & QP/AV)	Peak/QP 9KHz to 1 GHz	0.8 dB	
		QP/AV 9KHz to 1 GHz	0.8 dB	
20.	Conducted RF Test System ^{\$}			
	Harmonics Level (up to 3 rd Harmonics)	Upto -90 dBc (10 Hz to 13.6 GHz)	0.6 dB	Using Spectrum Analyzer or Equivalent by Direct Method
	Gain	50 dB (10 Hz to 6 GHz)	1.7 dB	
	AM Depth	10 % to 90 % (10 Hz to 13.6 GHz)	3.48 %	
	Level Accuracy	-30 dBm to +10 dBm (10 MHz to 18 GHz)	0.3 dB	Using Power Meter or Equivalent by Direct Method
	Error of Monitor Input	9 kHz to 1 GHz (-) 40 dBm to (+) 13 dBm	0.6 dB	Using Signal Generator & Power Meter or Equivalent by Direct Method
	Frequency Accuracy	100 kHz to 8 GHz	0.1 %	Using Frequency Counter or Equivalent by Direct Method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

^{\$}Only in Permanent Laboratory

^{*}Only for Site Calibration

Vishal Shukla
Convenor

Avijit Das
Program Manager