

Laboratory Eppendorf Calibration Services, No. 32, 1st Main, Temple Road,
Mallechwaram, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2710 (in lieu of C-1383) Page 1 of 1

Validity 30.05.2018 to 29.05.2020 Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>MECHANICAL CALIBRATION</u>				
I.	VOLUME			
1.	Micro Pipettes - Piston Operated [§]	10 μ l to 100 μ l 100 μ l to 200 μ l 200 μ l to 1000 μ l 1000 μ l to 5000 μ l 5 ml to 10 ml	0.16 μ l 0.18 μ l 0.51 μ l 3.99 μ l 3.99 μ l	Using Weighing Balance (Readability: 0.01 mg) by Gravimetric Method as per ISO 8655 (Part-6): 2002 and ISO 20461:2000

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

[§]Only in Permanent Laboratory

Dheeraj Chawla
Convenor

Avijit Das
Program Manager