

Laboratory Government Electrical Standards Laboratory, Office of The Chief Electrical Inspector to Government, Thiru. Vi. Ka. Industrial Estate, Guindy, Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025: 2005

Discipline Electro-Technical Calibration **Issue Date** 19.07.2014

Certificate Number C-0832 **Valid Until** 18.07.2016

Last Amended on - **Page** 1 of 4

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
1. SOURCE DC VOLTAGE ^{\$}	10 mV to 100 mV 100 mV to 1 V 1 V to 100 V 100 V to 1000 V	0.057 % to 0.012 % 0.012 % 0.12 % to 0.013 % 0.013 % to 0.0093 %	Using Universal Calibration System Fluke 9100E By Direct Method
2. AC VOLTAGE ^{\$}	50 Hz 10 mV to 100 mV 100 mV to 1 V 1 V to 100 V 100 V to 1000 V	4.48 % to 0.077 % 0.077 % to 0.070 % 0.70 % to 0.056 % 0.056 % to 0.074 %	Universal Calibration System Fluke 9100E By Direct Method
3. DC CURRENT ^{\$}	100 μ A to 100 mA 100 mA to 1 A 1 A to 10 A 10 A to 100 A 100 A to 990 A	0.031 % to 0.038 % 0.038 % to 0.36 % 0.36 % to 0.11 % 0.11 % to 0.39 % 0.39 % to 0.41 %	Using Universal Calibration System Fluke 9100E. By Direct Method Using Universal Calibration System Fluke 9100E, Current Coil By Direct Method
4. AC CURRENT ^{\$}	50Hz 1 mA to 100 mA 100 mA to 1 A 1 A to 10 A 10 A to 100 A 100 A to 990 A	0.68 % to 0.16 % 0.16 % to 0.21 % 0.21 % to 0.29 % 0.29 % to 0.86 % 0.86 % to 0.82 %	Using Universal Calibration System Fluke 9100E. By Direct Method Using Universal Calibration System Fluke 9100E, with Current Coil by Direct Method

Laboratory Government Electrical Standards Laboratory, Office of The Chief Electrical Inspector to Government, Thiru. Vi. Ka. Industrial Estate, Guindy, Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025: 2005

Discipline Electro-Technical Calibration **Issue Date** 19.07.2014

Certificate Number C-0832 **Valid Until** 18.07.2016

Last Amended on - **Page** 2 of 4

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
5. RESISTANCE ^{\$}	0.1 Ω to 1 Ω	0.61 % to 0.18 %	Using Vaiseshika High Precision DRB 7400 by Direct Method
	1 Ω to 10 Ω	0.18 % to 0.70 %	Using 9100 E by Direct Method
	10 Ω to 100 Ω	0.70 % to 0.16 %	
	100 Ω to 100 k Ω	0.16 % to 0.070 %	
	100k Ω to 100 M Ω	0.070 % to 1.60 %	
	100 M Ω to 1 G Ω	1.60 % to 2.61 %	Using Vaiseshika with Electronic Megger by Direct Method
	1 G Ω to 10 G Ω	2.61 % to 5.95 %	
	10 G Ω to 100 G Ω	5.95 % to 6.16 %	
100 G Ω to 1000 G Ω	6.16 %		
6. FREQUENCY ^{\$}	10 Hz to 30 kHz	0.009 % to 0.01 %	Using Universal Calibration System Fluke 9100E. By Direct Method
<u>MEASURE</u>			
1. DC VOLTAGE ^{\$}	100 mV to 1 V	0.054 % to 0.0040 %	Using 8.5 Digit Reference Multimeter Fluke 8508A By Direct Method
	1 V to 100 V	0.0040 % to 0.0010 %	
	100 V to 1000 V	0.0010 %	
2. AC VOLTAGE ^{\$}	50 Hz		Using 8.5 Digit Reference Multimeter Fluke 8508A By Direct Method
	100 mV to 1 V	0.30 %	
	1 V to 100 V	0.30 % to 0.018 %	
3. DC RESISTANCE ^{\$}	10 Ω to 100 Ω	3.30 % to 0.0011 %	Using 8.5 Digit Reference Multimeter Fluke 8508A By Direct Method
	100 Ω to 10 k Ω	0.0011 % to 0.0010 %	
	10 k Ω to 100 k Ω	0.0010 %	
	100 k Ω to 10 M Ω	0.0010 % to 0.017 %	
	10 M Ω to 400 M Ω	0.017 % to 1.55 %	

Laboratory Government Electrical Standards Laboratory, Office of The Chief Electrical Inspector to Government, Thiru. Vi. Ka. Industrial Estate, Guindy, Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025: 2005

Discipline Electro-Technical Calibration **Issue Date** 19.07.2014

Certificate Number C-0832 **Valid Until** 18.07.2016

Last Amended on - **Page** 3 of 4

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks	
4. AC POWER ^s 3 PHASE ACTIVE PF: 0.5 to UPF	240V @ 50 Hz			
	5 mA to 50 mA	0.025 %	Using ZERA Com 3003 By Comparison Method	
	>50 mA to 10 A	0.015 %		
	>10 A to 120 A	0.017 %		
		63.5 V @ 50 Hz		
	5 mA to 50 mA	0.054 %	Using ZERA COM 3003 by Comparison Method	
	50 mA to 10 A	0.015 %		
	ACTIVE POWER 0.5 PT TO UPF	240V@47.5 Hz & 52.5 Hz		Using ZERA COM 3003 by Comparison Method
		100mA to 10 A	0.015 %	
		>10A to 100 A	0.017 %	
		63.5V @47.5 Hz & 52.5Hz		Using ZERA COM 3003 by Comparison Method
		100 mA to 10A	0.015 %	
REACTIVE POWER 0.5 PF TO UPF	240 V @ 50 Hz		Using ZERA COM 3003 by Comparison Method	
	5 mA to 50 mA	0.023 %		
	50 mA to 10 A	0.015 %		
	>10 A to 120 A	0.018 %		
	63.5 V@50 Hz		Using ZERA COM 3003 by Comparison Method	
	5 mA to 50 mA	0.056 %		
	>50 mA to 10 A	0.015 %		
APPARENT POWER	240 V @ 50 Hz		Using ZERA COM 3003 by Comparison Method	
	5 mA to 50 mA	0.014 %		
	>50mA to 10 A	0.012 %		
	>10 A to 120 A	0.013 %		

Laboratory	Government Electrical Standards Laboratory, Office of The Chief Electrical Inspector to Government, Thiru. Vi. Ka. Industrial Estate, Guindy, Chennai, Tamil Nadu		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	19.07.2014
Certificate Number	C-0832	Valid Until	18.07.2016
Last Amended on	-	Page	4 of 4

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
5. AC POWER AND ENERGY^{\$} 1PHASE ACTIVE PF: 0.5PF to UPF	63.5V@ 50Hz		
	5 mA to 50 mA	0.036 %	
	>50 mA to 10 A	0.018 %	
	240V@ 50 Hz		
	5mA to 50 mA	0.055 %	Using ZERA COM 3003 by Comparison Method
	>50mA to 10 A	0.015 %	
	10 A to 120 A	0.017 %	

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

^{\$}Only in Permanent Laboratory