

Laboratory HEICO Dynamics Private Limited, B-59/4, Naraina Industrial Area,
 Phase - II, New Delhi
Accreditation Standard ISO/IEC 17025: 2005
Certificate Number CC-2406 **Page** 1 of 1
Validity 12.03.2019 to 05.10.2019 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>MECHANICAL CALIBRATION</u>				
I.	UTM, TENSION CREEP AND TORSION TESTING MACHINE			
1.	Force* UTM CTM Under Compression Mode (Class-1)	5 kN to 2000 kN	0.62 %	Using Calibrated Proving Ring/ Load Cells as per IS 1828 (Part-I)

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

*Only for Site Calibration

Pankaj Varshney
Convenor

Avijit Das
Program Manager