

Laboratory Hi-Tech Meter Laboratory, UGVCL, Nr. Rly. Over Bridge, Sabarmati, Ahmedabad, Gujarat
Accreditation Standard ISO/IEC 17025: 2005
Discipline Electro-Technical Calibration **Issue Date** 24.06.2015
Certificate Number C-0550 **Valid Until** 23.06.2017
Last Amended on - **Page** 1 of 2

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
MEASURE			
1. AC ENERGY ^{\$} (Without ICT)	45 Hz to 65 Hz 40 V to 300 V 10 mA to 120 A PF: -1.0 to +1.0	0.0650 %	Using PRS 121.3 Three Phase Electronic Reference Standard; Voltage Power Source PSU 10; Current Power Source PSI 10 by Comparison Method
2. AC ENERGY ^{\$} (With ICT)	45 Hz to 65 Hz 40 V to 300 V 10 mA to 120 A PF: -1.0 to +1.0	0.0850 %	Using PRS 121.3 Three Phase Electronic Reference Standard; Voltage Power Source PSU 10; Current Power Source PSI 10 by Comparison Method
3. AC POWER ^{\$} (Without ICT)	45 Hz to 65 Hz 40 V to 300 V 10 mA to 120 A PF: -1.0 to +1.0	0.0650 %	Using PRS 121.3 Three Phase Electronic Reference Standard; Voltage Power Source PSU 10; Current Power Source PSI 10 by Comparison Method
4. AC POWER ^{\$} (With ICT)	45 Hz to 65 Hz 40 V to 300 V 10 mA to 120 A PF: -1.0 to +1.0	0.0850 %	Using PRS 121.3 Three Phase Electronic Reference Standard; Voltage Power Source PSU 10; Current Power Source PSI 10 by Comparison Method

Bibin Philip
Convenor

Avijit Das
Program Manager

Laboratory	Hi-Tech Meter Laboratory, UGVCL, Nr. Rly. Over Bridge, Sabarmati, Ahmedabad, Gujarat		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	24.06.2015
Certificate Number	C-0550	Valid Until	23.06.2017
Last Amended on	-	Page	2 of 2

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
5. AC VOLTAGE \$	50 Hz 40 V to 480 V	0.660 %	Using PRS 121.3 Three Phase Electronic Reference Standard; Voltage Power Source PSU 10; Current Power Source PSI 10 by Comparison Method
6. AC CURRENT \$	50 Hz 10 mA to 120 A	0.11 %	Using PRS 121.3 Three Phase Electronic Reference Standard; Voltage Power Source PSU 10; Current Power Source PSI 10 by Comparison Method
7. POWER FACTOR \$	50 Hz 40 V to 300 V 10 mA to 120 A PF: -1.0 to +1.0	0.01 %	Using PRS 121.3 Three Phase Electronic Reference Standard; Voltage Power Source PSU 10; Current Power Source PSI 10 by Comparison Method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

\$Only in Permanent Laboratory

Bibin Philip
Convenor

Avijit Das
Program Manager