

Laboratory Horiba India Private Limited, D- 225, Chakan MIDC Phase-II,
Bhamboli Village, Pune, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2825 Page 1 of 1

Validity 30.08.2018 to 29.08.2020 Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
FLUID FLOW CALIBRATION				
1.	Gas Flow Measurement [§]	0 to 4000 ml/min	0.32 % of MV	Using Laminar Flow Element (Molblocs) By Comparison Method with Reference Standard / Reference Material
2.	Gas Analyzer for THC [§]	0 to 1000 ppmC	0.36 % of MV	Using Reference Material (Gas Propane) By Comparison Method with Reference Standard / Reference Material

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

[§]Only in Permanent Laboratory

Shally Sharma
Convenor

Anuja Anand
Program Manager