

Laboratory HTA Instrumentation (P) Ltd., #73, Ramachandra Agrahara,
Chamarajpet, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2166 Page 1 of 13

Validity 06.08.2018 to 05.08.2020 Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>ELECTRO-TECHNICAL CALIBRATION</u>				
I.	SOURCE			
	DC Voltage [§]	1 mV to 10 mV 10 mV to 1000 V	0.36 % to 0.05 % 0.05 % to 0.009 %	Using Fluke 5500A MPC by Direct Method
	DC Voltage *	100 mV to 30 V	0.04 % to 0.07 %	Using Yokogawa CA 71 Handy Calibrator by Direct Method
		0.2 mA to 2.2 A 2.2 A to 10 A	0.05 % to 0.09 % 0.09 % to 0.07 %	Using Fluke 5500A MPC by Direct Method
		10 A to 500 A	0.67 % to 0.31 %	Using Fluke 5500A & Current Coil
	DC Current *	0.1 mA to 24 mA	6.28 % to 0.04 %	Using Fluke 725 Process Calibrator by Direct Method
		45 Hz to 10 kHz 1 mV to 33 V	2.49 % to 0.24 %	Using Fluke 5500A MPC by Direct Method
		45 Hz to 1 kHz 33 V to 750 V	0.24 % to 0.07 %	Using Fluke 5500A MPC by Direct Method
		45 Hz to 1 kHz 100 μ A to 3.3 mA 3.3 mA to 330 mA 0.33 A to 10 A	0.45 % to 0.23 % 0.23 % to 0.34 % 0.34 % to 0.41 %	Using Fluke 5500A MPC by Direct Method
		50 Hz 10 A to 550 A	0.68 % to 0.36 %	Using Fluke 5500A & Current Coil

Pankaj Varshney
Convenor

Avijit Das
Program Manager

Laboratory

HTA Instrumentation (P) Ltd., #73, Ramachandra Agrahara,
Chamarajpet, Bangalore, Karnataka

Accreditation Standard

ISO/IEC 17025: 2005

Certificate Number

CC-2166

Page

2 of 13

Validity

06.08.2018 to 05.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
5.	Resistance [§]	2 Ω to 11 Ω 11 Ω to 1 K Ω 1 K Ω to 100 K Ω 100 K Ω to 1 M Ω 1 M Ω to 20 M Ω 20 M Ω to 100 M Ω	0.48 % to 0.17 % 0.17 % to 0.01 % 0.01 % to 0.007 % 0.007 % to 0.02 % 0.02 % to 0.12 % 0.12 % to 0.58 %	Using Fluke 5500A MPC by Direct Method
	Resistance *	15 Ω to 500 Ω 500 Ω to 3.2 k Ω	1.28 % to 0.22 % 0.22 %	Using Fluke 725 Process Calibrator by Direct Method
		1 kHz 350 pF to 330 nF	3.88 % to 0.61 %	
		100 Hz 1 μ F to 300 μ F	0.61 % to 0.95 %	
	Frequency [§]	3 V 45 Hz to 1 MHz	0.01 % to 0.06 %	Using Fluke 5500A MPC by Direct Method
	Frequency *	45 Hz to 1 kHz 1 kHz to 10 kHz	0.14 % to 0.06 % 0.06 % to 1.2 %	Using Fluke 725 Process Calibrator by Direct Method
8.	AC Power [§] (Single Phase, UPF, 50 Hz)	12 W to 2400 W 120 V (0.1 A to 10 A) 240 V (1 A to 10 A)	0.18 %	Using Fluke 5500A MPC by Direct Method
	Temperature Simulation [§]			
	RTD (Pt 100) Thermocouple :	(-) 200 $^{\circ}$ C to 600 $^{\circ}$ C	0.16 $^{\circ}$ C	Using Fluke 5500A MPC by Direct Method
	K Type	(-) 200 $^{\circ}$ C to 1300 $^{\circ}$ C	0.48 $^{\circ}$ C	
	J Type	(-) 200 $^{\circ}$ C to 900 $^{\circ}$ C	0.33 $^{\circ}$ C	
	R Type	0 $^{\circ}$ C to 1700 $^{\circ}$ C	0.67 $^{\circ}$ C	
	S Type	0 $^{\circ}$ C to 1700 $^{\circ}$ C	0.57 $^{\circ}$ C	
	T Type	(-) 200 $^{\circ}$ C to 400 $^{\circ}$ C	0.75 $^{\circ}$ C	
	N Type	(-) 150 $^{\circ}$ C to 1300 $^{\circ}$ C	0.49 $^{\circ}$ C	

Pankaj Varshney
Convenor

Avijit Das
Program Manager

Laboratory

HTA Instrumentation (P) Ltd., #73, Ramachandra Agrahara,
Chamarajpet, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number

CC-2166

Page

3 of 13

Validity

06.08.2018 to 05.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
10.	Temperature Simulation*			
	RTD (Pt 100)	(-) 200 °C to 800 °C	0.39 °C	Using Fluke 725 Process Calibrator by Direct Method
	Thermocouple :			
	K Type	(-) 200 °C to 1350 °C	1.40 °C	
	J Type	(-) 200 °C to 1200 °C	1.16 °C	
	T Type	(-) 200 °C to 400 °C	1.46 °C	
	R Type	0 °C to 1750 °C	2.09 °C	
S Type	0 °C to 1750 °C	2.11 °C		
N Type	(-) 200 °C to 1300 °C	1.77 °C		
II.	MEASURE			
	DC Voltage [§]	1 mV to 10 V 10 V to 1000 V	0.42 % to 0.005 % 0.005 % to 0.006 %	Using Fluke 8845A DMM by Direct Method
	DC Voltage*	1 mV to 100 mV 100 mV to 1000 V	0.28 % to 0.011 % 0.011 %	Using UNI-T UT-805 5½ DMM by Direct Method
	DC Current [§]	100 µA to 100 mA 100 mA to 10 A	0.03 % to 0.06 % 0.06 % to 0.20 %	Using Fluke 8845A DMM by Direct Method
	DC Current*	100 µA to 100 mA 100 mA to 1 A 1 A to 10 A	0.15 % to 0.05 % 0.05 % to 1.62 % 1.62 % to 0.99 %	Using UNI-T UT-805 5½ DMM by Direct Method
	AC Voltage [§]	45 Hz to 1 kHz 100 mV to 10 V 10V to 750V	0.12 % to 0.1 % 0.1 % to 0.11 %	Using Fluke 8845A DMM by Direct Method
	AC Voltage*	45 Hz 10 mV to 1 V 1 kHz 1 V to 750 V	1.75 % to 0.24 % 0.24 % to 0.50 %	Using UNI-T UT-805 5½ DMM by Direct Method

Pankaj Varshney
Convenor

Avijit Das
Program Manager

Laboratory HTA Instrumentation (P) Ltd., #73, Ramachandra Agrahara,
 Chamarajpet, Bangalore, Karnataka
Accreditation Standard ISO/IEC 17025: 2005
Certificate Number CC-2166 **Page** 4 of 13
Validity 06.08.2018 to 05.08.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
4.	AC Current [§]	1 kHz 10 mA to 1 A 1 A to 10 A	0.24 % to 0.17 % 0.17 % to 0.26 %	Using Fluke 8845A DMM by Direct Method
		45 Hz 1 mA to 10 mA	0.48 % to 1.55 %	
		1 kHz 10 mA to 1 A 1 A to 10 A	1.55 % to 4.63 % 4.63 % to 2.54 %	
	Frequency [§]	1 V 10 Hz to 100 kHz	0.58 % to 0.01 %	Using Fluke 8845A DMM by Direct Method
	Frequency*	100 Hz to 2 MHz	3.67 % to 0.53 %	Using UNI-T UT-805 5½ DMM by Direct Method
	Resistance [§]	100 Ω to 100 M Ω	0.02 % to 0.94 %	Using Fluke 8845A DMM by Direct Method
	Resistance*	10 Ω to 10 M Ω	0.01% to 0.14%	Using UNI-T UT-805 5½ DMM by Direct Method
	Timer [§]	60 s to 3600 s	0.29 s to 0.72 s	Using Digital Stop Watch by Comparison Method
	Timer*	60 s to 3600 s	0.17 s to 0.72 s	Using Digital Stop Watch by Comparison Method

Pankaj Varshney
 Convenor

Avijit Das
 Program Manager

Laboratory HTA Instrumentation (P) Ltd., #73, Ramachandra Agrahara,
Chamarajpet, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2166 **Page** 5 of 13

Validity 06.08.2018 to 05.08.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>MECHANICAL CALIBRATION</u>				
1. DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)				
1.	Pressure (Hydraulic)/ Pressure Gauges, Pressure Transmitters [§]	0 bar to 700 bar	0.03 % rdg	Using Digital Pressure Calibrator with External Transducer by Comparison Method as per DKD R-6-1
2.	Pressure (Pneumatic) / Pressure Gauges, Pressure Transmitters [§]	0 bar to 20 bar	0.06 % rdg	Using Digital Pressure Calibrator by Comparison Method as per DKD R-6-1
3.	Vacuum, Vacuum Gauges, Vacuum Transmitters [§]	(-) 0.85 bar to 0 bar	0.20 % rdg	Using Digital Pressure Calibrator by Comparison Method as per ISO 3567
4.	Differential Pressure (Pneumatic)/Gauges, Transmitters, Manometers [§]	(-) 25 mbar to 25 mbar	0.15 % rdg	Using Digital Pressure Calibrator with Low Pressure Transducer by Comparison Method as per Euramet/ cg-17/v 2.0
5.	Pressure (Hydraulic)/ Pressure Gauges , Pressure Transmitters*	0 bar to 700 bar >700 bar to 1000 bar	0.03 % rdg 0.07 % rdg	Using Digital Pressure Gauge by Comparison Method as per DKD R-6-1
6.	Pressure(Pneumatic)/ Pressure Gauges, Pressure Transmitters*	0 bar to 3 bar >3 bar to 30 bar	0.24 % rdg 0.20 % rdg	Using Digital pressure Gauge by Comparison Method as per DKD R-6-1

Pankaj Varshney
Convenor

Avijit Das
Program Manager

Laboratory HTA Instrumentation (P) Ltd., #73, Ramachandra Agrahara,
Chamarajpet, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2166 **Page** 6 of 13

Validity 06.08.2018 to 05.08.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
7.	Vacuum, Vacuum Gauges, Vacuum Transmitters*	(-) 0.85 bar to 0 bar	0.70 % rdg	Using Digital Pressure Gauge by Comparison Method as per ISO 3567
8.	Differential Pressure (Pneumatic)/Gauges, Transmitters, Manometers*	(-) 20 mbar to 20 mbar	0.32 % rdg	Using Digital Manometer by Comparison Method as per Euramet/cg-17/v 2.0
9.	RPM Tachometer - Contact Type [§]	100 rpm to 10000 rpm	1.8 % to 0.17% rdg.	Using Digital Tachometer by Comparison Method - Based on SANAS TR 45-05
10.	RPM Tachometer – Non-Contact Type [§]	60 rpm to 30000 rpm	1.25 % to 0.01% rdg.	Using Digital Tachometer by Comparison Method – Based on SANAS TR 45-05
11.	RPM Tachometer / Centrifuge/Rotating Machines (Non – Contact Type) *	60 rpm to 15000 rpm	1.7 % to 0.06 % rdg.	Using Digital Tachometer by Comparison Method - SANAS TR 45-05
12.	Vibration Meter [§] (Velocity)	1 kHz 5.5 mm/s to 140 mm/s	3.5 % rdg	Using Digital Vibration Meter by Comparison Method Based on OIML R 103
13.	Sound Level Meter [§]	94 dB & 114 dB	0.77 dB	Using Sound Level Calibrator by Comparison Method – Based on OIML R 58 Annexure B & C

Pankaj Varshney
Convenor

Avijit Das
Program Manager

Laboratory HTA Instrumentation (P) Ltd., #73, Ramachandra Agrahara,
 Chamarajpet, Bangalore, Karnataka
Accreditation Standard ISO/IEC 17025: 2005
Certificate Number CC-2166 **Page** 7 of 13
Validity 06.08.2018 to 05.08.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
14.	Calipers ^s (Analog/Digital) L.C.: 0.01 mm	0 to 150 mm	8.9 μ m	Using Gauge Block Comparison Method - based on IS: 3651 (Part-1 & 2)
15.	Micrometer ^s (Analog/Digital) L.C.: 0.001 mm	0 to 25 mm	1.7 μ m	Using Gauge Block Comparison Method - based on IS:2967

Laboratory HTA Instrumentation (P) Ltd., #73, Ramachandra Agrahara,
Chamarajpet, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2166 Page 8 of 13

Validity 06.08.2018 to 05.08.2020 Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>THERMAL CALIBRATION</u>				
I.	TEMPERATURE			
		(-) 100 °C to 140 °C	0.14 °C	Using Reference Thermometer Fluke 1523 with Fluke 5626 PRT Sensor, DMM Fluke 8845A and Field Metrology Dry Well Fluke 9190A as a Source by Comparison Method
		140 °C to 550 °C	0.5 °C	Using Reference Thermometer Fluke 1523 with Fluke 5626 PRT Sensor, DMM Fluke 8845A and Dry Block Calibrator as a Source By Comparison Method
		550 °C to 1000 °C	3.48 °C	Using Thermocouple Thermometer Delta Ohm HD 2108.1 with "R" Type Thermocouple, DMM Fluke 8845A and Dry Block Calibrator as a Source By Comparison Method

Pankaj Varshney
Convenor

Avijit Das
Program Manager

Laboratory HTA Instrumentation (P) Ltd., #73, Ramachandra Agrahara,
 Chamarajpet, Bangalore, Karnataka
Accreditation Standard ISO/IEC 17025: 2005
Certificate Number CC-2166 **Page** 9 of 13
Validity 06.08.2018 to 05.08.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
2.	Temperature Indicator With Sensor [§]	(-) 100 °C to 140 °C	0.12 °C	Using Reference Thermometer Fluke 1523 with Fluke 5626 PRT Sensor, Field Metrology Dry Well Fluke 9190A as a Source By Comparison Method
		140 °C to 550 °C	0.51 °C	Using Reference Thermometer Fluke 1523 with Fluke 5626 PRT Sensor, Dry Block Calibrator as a Source By Comparison Method
		550 °C to 1000 °C	3.44 °C	Using Thermocouple Thermometer Delta Ohm HD 2108.1 with "R" type Thermocouple and Dry Block Calibrator as a Source By Comparison Method
3.	Temperature Gauge [§]	0 °C to 550 °C	1.25 °C	Using Reference Thermometer Fluke 1523 with Fluke 5626 PRT Sensor, Dry Block Calibrator & Field Metrology Dry Well Fluke 9190A as a source By Comparison Method
4.	Infrared Thermometer [§]	50 °C to 500 °C	3.42 °C	Using Precision RTD Thermometer Delta ohm HD 2107.1 By Comparison Method

Pankaj Varshney
 Convenor

Avijit Das
 Program Manager

Laboratory HTA Instrumentation (P) Ltd., #73, Ramachandra Agrahara,
Chamarajpet, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2166 **Page** 10 of 13

Validity 06.08.2018 to 05.08.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
5.	Temperature & Humidity Meter/ Temperature & Humidity Data Logger/ Thermo Hygrometer ^s	Temp: 10 °C to 40 °C @ 50 %RH RH: 30 %RH to 90 %RH @25 °C	Temp : 0.513 °C Relative Humidity : 1.815 %	Using Temperature Humidity Meter Rotronic & HP22-A with portable Temperature Humidity Generator as a Source By Comparison Method
		(-) 40 °C to 140 °C	0.45 °C	Using Digital Thermometer Delta Ohm HD 2307.0 & HD 2107.1 with SPRT Sensor, DMM UNI-T UT-805 and Metrology Well Fluke 9170 as a Source By Comparison Method
		140 °C to 500 °C	1.00 °C	Using Digital Thermometer Delta Ohm HD 2307.0 and HD 2107.1, with SPRT Sensor, DMM UNI-T UT-805 and Dry Block Calibrator and DMM as a Source By Comparison Method
		500 °C to 1100 °C	3.48 °C	Using Thermocouple Thermometer Delta Ohm HD 2128.2 with "S" Type Thermocouple, DMM UNI-T UT-805 and Dry Block Calibrator as a Source By Comparison Method

Pankaj Varshney
Convenor

Avijit Das
Program Manager

Laboratory HTA Instrumentation (P) Ltd., #73, Ramachandra Agrahara,
Chamarajpet, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2166 **Page** 11 of 13

Validity 06.08.2018 to 05.08.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
7.	Temperature Indicator with Sensor*	(-) 40 °C to 140 °C	0.35 °C	Using Digital Thermometer Delta Ohm HD 2307.0 & HD 2107.1, with SPRT Sensor and Metrology Well Fluke 9170 as a Source By Comparison Method
		140 °C to 500 °C	0.92 °C	Using Digital Thermometer Delta Ohm HD 2307.0 & HD 2107.1 with SPRT Sensor and Dry Block Calibrator as a Source By Comparison Method
		500 °C to 1100 °C	4.09 °C	Using Thermocouple Thermometer Delta Ohm HD 2128.2 with "S" Type Thermocouple and Dry Block Calibrator as a Source By Comparison Method
8.	Indicator of Oven, Incubator (for Non Medical Applications), Freezer, Autoclave, Cold Room, Liquid Bath, Dry Block Calibrator (Single Point) *	(-)100 °C to 300 °C	1.2 °C	Using Digital Thermometer Delta Ohm HD 2307.0 & HD 2107.1 with SPRT Sensor/RTD PT 100 Sensor along with Multichannel Paperless Recorder By Direct Method

Pankaj Varshney
Convenor

Avijit Das
Program Manager

Laboratory HTA Instrumentation (P) Ltd., #73, Ramachandra Agrahara,
Chamarajpet, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2166 **Page** 12 of 13

Validity 06.08.2018 to 05.08.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
9.	Indicator of Muffle Furnace, Dry Block Calibrator (Single Point) *	300 °C to 1100 °C	2.2 °C	Using Thermo couple Thermometer with "S" Type Thermocouple / "N"/T Type Thermocouple along with Multi-Channel Paperless Recorder By Direct Method
10.	Indicator of Climatic Chambers/ Humidity Chambers (Single Point) *	30 % RH to 90 % RH @ ~25 °C	1.0 % RH @ 25 °C	Using Temperature Humidity Meter EQ-321S by direct method
11.	Oven, Incubator (for Non Medical Applications), Climatic Chamber, Freezer, Autoclave, Cold Room, Liquid Bath (Min 9 Point) *	(-)100 °C to 150 °C	3.0 °C	Using Multi Points RTD PT100 Sensor along with Multi-Channel Paperless Recorder. Spatial Mapping by Comparison Method & As per DKD-R-5-7, AMS 2750E
12.	Oven, Incubator (for Non Medical Applications), Chamber, Freezer, Autoclave, Cold Room, Liquid Bath (Min 9 Point) *	150 °C to 300 °C	4.0 °C	Using Multi-Channel 'T' Type Thermocouple along with Multi-Channel Paperless Recorder. Spatial Mapping by Comparison Method & As per DKD-R-5-7, AMS 2750E

Laboratory HTA Instrumentation (P) Ltd., #73, Ramachandra Agrahara,
 Chamarajpet, Bangalore, Karnataka
Accreditation Standard ISO/IEC 17025: 2005
Certificate Number CC-2166 **Page** 13 of 13
Validity 06.08.2018 to 05.08.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
13.	Muffle Furnace, Furnace (Min 9 Point) *	300 °C to 1100 °C	10.2 °C	Using Multi-Channel 'N' Type Thermocouple along with Multi-Channel Paperless Recorder. Spartial Mapping by Comparison Method & As per DKD-R-5-7, AMS 2750E

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

§Only in Permanent Laboratory

*Only for Site Calibration