

Laboratory Index Lab & Consultants, B-11, Navlakha Complex, Agrasen Square, Indore, Madhya Pradesh

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2643 (In lieu of C-1364) Page 1 of 1

Validity 22.04.2018 to 21.04.2020 Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>MECHANICAL CALIBRATION</u>				
I.	UTM, TENSION CREEP AND TORSION TESTING MACHINE			
1.	Force Measuring System of CTM/UTM* (Compression Mode) Class 1 device only	10 kN to 2000 kN	0.55 %	Using Force Proving Instruments as per IS: 1828 part 1

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

*Only for Site Calibration

Dheeraj Chawla
Convenor

Avijit Das
Program Director