

Laboratory Indian Technical Services, 28, Electronic Complex, Industrial Area,
 Pardeshipura, Indore, Madhya Pradesh
Accreditation Standard ISO/IEC 17025: 2005
Certificate Number CC-2580 (in lieu of C-1045) **Page** 1 of 3
Validity 03.03.2018 to 02.03.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
<u>ELECTRO-TECHNICAL CALIBRATION</u>				
I.	SOURCE			
1.	DC Voltage ^o	1mV to 100 mV	1.2 % to 0.13 %	Using Temperature Calibrator Source and 5½ MFC by Direct Method
		100 mV to 990 V	0.13 %	Using 5½ MFC by Direct Method
	DC Voltage [*]	1mV to 100 mV 100mV to 5V	1.2 % to 0.6 % 0.6 % to 0.3 %	Using Temperature Calibrator Source by Direct Method
2.	AC Voltage ^o	50 Hz 20 mV to 100 mV 100 mV to 10 V 10 V to 900 V	0.5 % to 0.23 % 0.23 % 0.23 % to 0.25 %	Using 5½ MFC by Direct Method
3.	DC Current ^o	0.2 mA to 100 mA 100 mA to 2.0A 2.0 A to 9.9 A	0.20 % to 0.14 % 0.15 % 0.19 % to 0.48 %	Using 5½ MFC by Direct Method
4.	AC Current ^o	50 Hz 0.2 mA to 100 mA 100 mA to 2.0A 2.0 A to 9.9 A	0.34 % to 0.26 % 0.27 % 0.27 % to 0.53 %	Using 5½ MFC by Direct Method
5.	Resistance ^s (2 Wire)	1Ω to 10 Ω 10Ω to 1kΩ 1kΩ to 1 MΩ	1.52 % to 1.20 % 1.20 % 1.21 %	Using Decade Resistance Box by Direct Method

Rajeshwar Kumar
 Convenor

Avijit Das
 Program Director

Laboratory Indian Technical Services, 28, Electronic Complex, Industrial Area,
Pardeshipura, Indore, Madhya Pradesh

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2580 (in lieu of C-1045) **Page** 2 of 3

Validity 03.03.2018 to 02.03.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
6.	Capacitance DC ^s	0.1 nF to 10 nF 10 nF to 10 µF	1.63 % to 1.3 % 1.3 %	Using Capacitance Box & 8846 DMM by Direct Method
7.	Frequency ^s	45 Hz to 1000 Hz	0.3 % to 0.28 %	Using 5½ MFC by Direct Method
8.	Temperature Simulation* (Indicator, Recorder, Controller, Data Logger)			
	RTD Type	(-) 200 °C to 800 °C	1.5 °C to 1.00 °C	Using Process Source/ Temperature Calibrator by Direct Method
	Thermocouple			
	R Type	300 °C to 1700 °C	2 °C	
	S Type	300 °C to 1700 °C	2 °C	
	J Type	(-) 100 °C to 600 °C	1.3 °C	
	K Type	(-) 100 °C to 1300 °C	1.3 °C	
	B Type	600 °C to 1700 °C	2 °C	
	T Type	(-) 200 °C to 350 °C	1.2 °C	
II.	MEASURE			
1.	DC Voltage ^s	1 mV to 100 mV 0.1 V to 900 V	0.9 % to 0.012 % 0.012 % to 0.02 %	Using Fluke 6½ DMM by Direct Method
2.	AC Voltage ^s	50 Hz 20 mV to 100 mV 0.1 V to 900 V	0.65 % 0.65 % to 0.10 %	Using Fluke 6½ DMM by Direct Method
3.	DC Current ^s	0.2mA to 100 mA 0.1A to 9 A	0.12 % to 0.10 % 0.1 % to 0.2 %	Using Fluke 6½ DMM by Direct Method
4.	AC Current ^s	50 Hz 0.2 mA to 100 mA 0.1A to 9 A	0.76 % to 0.65 % 0.65 % to 0.4 %	Using Fluke 6½ DMM by Direct Method
5.	DC Resistance ^s	1Ω to 10 MΩ	0.4 % to 0.46 %	Using Fluke 6½ DMM by Direct Method

Laboratory Indian Technical Services, 28, Electronic Complex, Industrial Area,
Pardeshipura, Indore, Madhya Pradesh

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2580 *(in lieu of C-1045)* **Page** 3 of 3

Validity 03.03.2018 to 02.03.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
6.	Frequency ^{\$}	10 Hz to 1000 kHz	0.6 % to 0.12 %	Using Fluke 6½ DMM by Direct Method
7.	Temperature Simulation ^{\$} RTD Type	(-) 200 °C to 600 °C	0.41 °C	Using Fluke 6½ DMM by Direct Method
8.	Time- Stop Watch [#]	10.00 s to 60.00 min.	0.35 to 2.3 s	Using Digital Stop Watch by Comparison Method

* Measurement Capability is expressed as an uncertainty ($\pm$) at a confidence probability of 95%

^{\$}Only in Permanent Laboratory

^{*}Only for Site Calibration

[#] The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Rajeshwar Kumar
Convenor

Avijit Das
Program Director