

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Taty Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 1 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>ELECTRO-TECHNICAL CALIBRATION</u>				
I.	SOURCE			
	DC Voltage [§]	10 μ V to 10 mV 10 mV to 10 V 10 V to 1050 V	500 ppm to 15 ppm 15 ppm to 0.1 ppm 0.1 ppm to 2.5 ppm	Using Binary Voltage Divider with Range Extender, Reference Voltage Divider, DC Voltage Standard, Reference Multimeter & Multifunction Calibrator Direct / Comparison Method
	DC Voltage*	0.1mV to 1000 V	0.3% to 0.001 %	Using Multifunction Calibrator with 8½ Digital Multimeter by Direct / Comparison Method
	DC High Voltage [§]	1 kV to 100 kV	0.6 %	Using DC High Voltage Source with DC High Voltage Divider & 6½ Digit Multimeter by Comparison Method
	DC Current [§]	1 pA to 1mA 1 mA to 20 A 20 A to 1000 A 1000 A to 3000 A	500 ppm to 4 ppm 4 ppm to 20 ppm 20 ppm to 0.2 % 0.2 % to 0.5 %	Using High Precision Multi-Tap Shunt with Standard Multimeter by Comparison / V/R / Direct Method
		10 μ A to 300 mA 300 mA to 20 A	0.25 % to 0.02 % 0.02 % to 0.15 %	Using Multifunction Calibrator with 8½ Digital Multimeter by

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Taty Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 2 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
				Direct / Comparison Method
		20 A to 1000 A	0.6 %	Using Multifunction Calibrator with Current Coil by Direct Method
		10 $\mu\Omega$ (3000A/30 mV) 40 $\mu\Omega$ (1500A/60 mV) 60 $\mu\Omega$ (1000A/60 mV) 100 $\mu\Omega$ (300A/30 mV) 150 $\mu\Omega$ (400A/60 mV) 200 $\mu\Omega$ (500A/100 mV) 240 $\mu\Omega$ (250A/60 mV) 600 $\mu\Omega$ (100A/60 mV)	0.008 % 0.005 % 0.12 % 0.005 % 0.12 % 0.003 % 0.12 % 0.12 %	Using Standard Shunts by Direct Method
		1 m Ω 8 m Ω 10 m Ω 16 m Ω 100 m Ω	0.006 % 0.002 % 0.0008 % 0.002 % 0.0005 %	Using Precision Standard Shunts/ Standard Resistors by Direct Method
		1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω 1 G Ω 10 G Ω 1 T Ω 100 T Ω	2 ppm 10 ppm 10 ppm 2 ppm 2 ppm 15 ppm 15 ppm 16 ppm 30 ppm 0.03 % 1.2 % 2 % 3 %	Using Standard Resistors by Direct Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Taty Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 3 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
		0.1 Ω to 190 k Ω 190 k Ω to 100 M Ω	60 ppm to 12 ppm 12 ppm to 100 ppm	Using Multifunction Calibrator & Decade Resistance Box with Reference Multimeter by Comparison Method
		100 M Ω to 1 T Ω	100 ppm to 1.2 %	Using High Value Decade Resistance Box with Dual Voltage Source High value Resistance Bridge by Comparison Method
		40 $\mu\Omega$ (1500A/60 mV) 60 $\mu\Omega$ (1000A/60 mV) 100 $\mu\Omega$ (600A/60 mV) 150 $\mu\Omega$ (400A/60 mV) 240 $\mu\Omega$ (250A/60 mV) 600 $\mu\Omega$ (100A/60 mV)	0.15 % 0.15 % 0.15 % 0.15 % 0.15 % 0.15 %	
		1m Ω 10m Ω 100m Ω 1 Ω 10 Ω 100 Ω 1k Ω 10k Ω 100k Ω 1M Ω 10 M Ω 100 M Ω 1 T Ω	0.01 % 0.005 % 0.005% 0.002 % 0.002 % 0.002 % 0.002 % 0.002 % 0.002% 0.002% 0.002% 0.005% 3%	
		0.1 Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 100 M Ω	0.06% to 0.005% 0.005 % to 0.05% 0.05% to 0.1%	

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 4 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
		100 M Ω to 1 T Ω	0.1% to 2.5%	Using Multifunction Calibrator & Decade Meg Ohm Box by Direct Method
		100 T Ω	5%	Using Standard Resistor by Direct Method
		40 Hz to 70 Hz 1 V to 480 V	5 ppm to 20 ppm	Using Precision Power Calibration System by Direct Method
		10 Hz to 1 KHz 1 mV to 2 V 2 V to 200 V 200 V to 1000 V	0.64 % to 20 ppm 20 ppm to 200 ppm 200 ppm to 30 ppm	
		1 KHz to 1 MHz 1 mV to 20 V	0.62 % to 20 ppm	
		1 kHz to 300 kHz 20 V to 60 V	20 ppm to 50 ppm	
		1 kHz to 20 kHz 60 V to 1000 V	20 ppm to 200 ppm	
		500 kHz to 1.1 GHz 5 mV to 5.5 V (p-p)	4% to 10%	Using Multifunction Calibrator by Direct Method
	AC Voltage*	40 Hz to 1 kHz 1m V to 1000 V	0.7 % to 0.008 %	Using Multifunction Calibrator with AC Measurement Standard by Comparison Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Taty Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 5 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	AC High Voltage ^s	At 50 Hz to 60 Hz 1 kV to 2.4 kV	0.35%	Using Precision AC Divider with 6½ DMM by Comparison Method
		At 50 Hz 2.4 kV to 100 kV 100 kV to 200 kV	0.6 % 1.5 %	Using AC High Voltage Source with AC High Voltage Divider with 6½ Digit Multimeter and KV Meter by Comparison Method
		40 Hz to 70 Hz 1 mA to 10 mA 10 mA to 100 A 100 A to 160A	250 ppm to 10 ppm 10 ppm to 30 ppm 30 ppm to 80 ppm	Using Precision Power Calibration System, Power Comparator by Direct/Comparison Method
		10 Hz to 1 kHz 10 µA to 1 mA 1 mA to 120 A	0.08 % to 0.006 % 0.006 % to 0.25 %	
		1 kHz to 10 kHz 10 µA to 1 mA 1 mA to 100 A	0.05 % to 0.006 % 0.006 % to 5.5 %	
		50 Hz 120 A to 5000 A	0.03 % to 0.15 %	Using AC Current Source with Standard CT & Power Meter by Direct Method
		40 Hz to 1 kHz 30 µA to 1 A 1 A to 20 A	0.5 % to 0.07 % 0.07 % to 0.12 %	Using Multifunction Calibrator by Direct Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 6 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
		40 to 70 Hz 1 mA to 120 A	0.03 % to 0.01 %	Using Three Phase Power Calibrator with Power / Energy Comparator by Comparison Method
		50 Hz 120 A to 1000 A	0.6 %	Using Multifunction Calibrator with 50 Turn Current Coil by Direct Method
		40 Hz to 70 Hz 1 V to 480 V PF: 0.01 to 1 10 mA to 100 A	20 ppm to 40 ppm / PF	Using Precision Power Calibration System (PPCS) By Direct Method
		40 Hz to 70 Hz 25 V to 480 V PF: 0.01 to 1 1 mA to 10 mA 100 A to 160 A	0.04% to 0.01% /PF 0.01% / PF	Using Three Phase Power Calibrator /Tester with Power/Energy Comparator by Comparison Method
	Active / Reactive Apperent Power / Energy (Single Phase / Three Phase)*	40 Hz to 70 Hz 25 V to 480 V PF: 0.01 to 1 1 mA to 10 mA 10 mA to 120 A	0.04% / PF to 0.013%/PF 0.013% /PF	Using Three Phase Power Calibrator with Power / Energy Comparator by Comparison Method
	Active / Reactive / Apparent Power /Energy (1 Phase) ^s	50 Hz to 60 Hz 480 V to 1050 V PF: 0.01 to 1 1 mA to 10 mA 10 mA to 120 A	0.04% to 0.011% /PF 0.011% / PF	Using Multifunction Calibrator, Trans Conductance Amplifier, Power Energy Comparator with Precision AC Voltage Divider by Comparison Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Taty Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 7 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
7.	Power Factor / Phase Angle [§]	40 Hz to 70 Hz 0 to UPF (0° to 360°) 1 V to 480 V 10 mA to 100 A	0.0012°	Using Precision Power Calibration System (PPCS) by Comparison Method
		40 Hz to 70 Hz 25 V to 480 V 0 to UPF (0° to 360°) 1 mA to 10 mA 100 A to 160 A	0.008°	Using Three Phase Power Calibrator /Tester with Reference Meter by Comparison Method
		50 Hz to 60 Hz 480 V to 1050 V 0 to UPF (0° to 360°) 1 mA to 120 A	0.008°	Using Multifunction Calibrator, Trans Conductance Amplifier , Power Energy Comparator with Precision AC Voltage Divider by Comparison Method
	Power Factor/ Phase Angle*	40 Hz to 70 Hz 0 to UPF 0 to 360° 25 V to 480 V 1 mA to 120 A	0.008°	Using Three Phase Power Calibrator with Power / Energy Comparator by Comparison Method
	Frequency / Time Period [§]	40 mHz to 4 GHz 250 ps to 25 s	1.5×10^{-11} to 1.3×10^{-12}	Using GPS Controlled Rubidium Standard with RF Reference Source by Direct Method
	Frequency / Time Period*	1 Hz to 4 GHz 1s to 0.25 ns	6 ppm to 0.1 ppm	Using RF Reference by Source Direct Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 8 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
9.	Time Interval [§]	1 μ s to 1000 s 1000 s to 24 Hrs & multiple of 24 hrs.	1.1×10^{-3} to 1.1×10^{-7} 1×10^{-7} to 0.023 %	Using Multifunction Calibrator and Frequency Counter/ Timer/Analyzer by Comparison Method
	Time Interval [*]	1 μ s to 1000 s 1000 s to 24 Hours & Above	1.1×10^{-3} to 1.1×10^{-7} 1×10^{-7} to 0.023 %	Using Frequency Counter/Timer/Analyzer & Time Interval Meter by Direct Method
		50 Hz 200 $\mu\Omega$ 1 m Ω 10 m Ω 100 m Ω 1 Ω	0.012 % 0.03 % 0.03 % 0.02 % 0.02 %	
		1 kHz 0.01 Ω 0.1 Ω 1 Ω 10 Ω 100 Ω	0.07 % 0.04 % 0.005 % 0.005 % 0.005 %	
		50 Hz, 100 Hz & 1 kHz 1 k Ω 10 k Ω	0.36 % 0.36 %	
		AC Resistance* 1 kHz 0.01 Ω 0.1 Ω 1 Ω 10 Ω 100 Ω	0.1% 0.05% 0.01% 0.01% 0.01%	

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 9 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
11.	Inductance [§]	1 kHz 100 μ H 1 mH 10 mH 100 mH 1 H 10 H	0.3 % 0.15 % 0.06 % 0.05 % 0.05 % 0.06 %	Using Standard Inductors by Direct Method
	Inductance*	1kHz 100 μ H 1 mH 10 mH 100 mH 1 H 10 H	0.3 % 0.2 % 0.1 % 0.05 % 0.05 % 0.06 %	Using Standard Inductors by Direct Method
		50 Hz 1 pF 10 pF 100 pF	0.06% 0.06% 0.06%	
		100 Hz 1 pF 10 pF 100 pF 1 μ F 10 μ F 100 μ F 1 mF 10 mF 100 mF 1 F	6 ppm 3 ppm 2.5 ppm 0.015 % 0.03 % 0.03 % 0.05 % 0.05 % 0.05 % 0.2 %	

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 10 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
		1 kHz 1 pF 10 pF 100 pF 1 nF 10 nF 100 nF 1 μ F 10 μ F 100 μ F 1 mF 10 mF 100 mF 1 F	4 ppm 2.5 ppm 2.3 ppm 0.015 % 0.015 % 0.015 % 0.015 % 0.03 % 0.03 % 0.03 % 0.03 % 0.04 % 0.05 % 0.07 %	
		1 nF to 110 mF	0.4 % to 1.5 %	Using Multifunction Calibrator by Direct Method
		1kHz 1pF 10 pF 100 pF 0.001 μ F 0.01 μ F 0.1 μ F 1 μ F 10 μ F 100 μ F 1mF 10 mF 100 mF 1 F	0.4 % 0.15 % 0.08 % 0.02 % 0.02 % 0.02 % 0.03 % 0.04 % 0.04 % 0.04 % 0.05 % 0.06 % 0.08 %	Using Standard Capacitors & Four Terminal Capacitance Standard by Direct Method
		1 nF to 110 mF	0.5 % to 2 %	Using Multifunction Calibrator by Direct Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 11 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
13.	Capacitance at High Voltage \$	Up to 2 kV, 50 Hz 125 pF & 2000 pF	0.05 %	Using Standard Capacitor with High Precision Capacitance Bridge (Schering Bridge) by Comparison Method
		Up to 30 kV, 50 Hz 1000 pF Up to 100 kV, 50 Hz 100 pF	0.012 % 0.02 %	Using Standard Gas Filled Capacitor by Direct Method
		Up to 2 kV, 50 Hz 125 pF & 2000 pF	0.05 %	Using Standard Capacitor with High Precision Capacitance Bridge by Comparison Method
		Up to 30 kV, 50 Hz 1000 pF Up to 100 kV, 50 Hz 100 pF	0.012 % 0.02 %	Using Standard Gas Filled Capacitor by Direct Method
		5 x 10 ⁻⁵ to 5 x 10 ⁻² Up to 25 kV AC, 50 Hz	2.5 x 10 ⁻⁵ to 4.2 x 10 ⁻⁴	Using Standard Gas Filled Capacitor with Dissipation Boxes by Direct Method
		1x 10 ⁻¹ 1.5 x 10 ⁻¹ 1.8 x 10 ⁻¹ Up to 2 kV, 50 Hz	1.5%	Using Standard Tan Delta Calibrator with C & Tan Delta Measurement System by Comparison Method
		5 x 10 ⁻⁵ to 5 x 10 ⁻² Up to 25 kV AC, 50 Hz	2.5 x 10 ⁻⁵ to 4.2 x 10 ⁻⁴	Using Standard Gas Filled Capacitor with Dissipation Boxes by Direct Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 12 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
		1×10^{-1} 1.5×10^{-1} 1.8×10^{-1} Up to 2 kV, 50 Hz	1.5%	Using Standard Tan Delta Calibrator with C & Tan Delta Measurement System by Comparison Method
	DC Power / Energy [#]	33 mV to 1000 V 3.3 mA to 20.5 A	0.015 % to 0.15 %	Using Multifunction Calibrator by Direct Method
	DC Power DC Clamp-on-Power Meter [§]	33 mV to 1000 V 3.3 mA to 1000 A	0.015 % to 0.8 %	Using Multifunction Calibrator & Current Coil by Direct Method
	DC Power DC Clamp-On-Power Meter [*]	33 mV to 1000 V 3.3 mA to 1000 A	0.015 % to 0.8 %	Using Multifunction Calibrator & Current Coil by Direct Method
	Temperature Simulation [§] (For Temperature Indicators, Recorders, Controllers)			
	RTD	(-) 200 °C to 850 °C	0.005 °C to 0.02 °C	
	T/C – K, J, N, E, T, R, S, B, C, L & U	(-) 270 °C to 2300 °C	0.01 °C to 0.26 °C	
	Temperature Simulation [*] (For Temperature Indicators, Recorders, Controllers)			
	RTD	(-) 200 °C to 850 °C	0.005 °C to 0.02 °C	
	T/C – K, J, N, E, T, R, S, B, C, L & U	(-) 270 °C to 2300 °C	0.01 °C to 0.3 °C	
	Oscilloscope [§]			
	Bandwidth	50 kHz to 1.1 GHz 5 mV to 5.5 V(p-p)	2.4 % to 5 %	
	Amplitude (Deflection Factor)	1 mV to 130 V (1M Ω) 1 mV to 6.6 V (50 Ω)	5 % to 0.15 % 5 % to 0.3 %	

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 13 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	Time Base (Marker)	1 ns to 20 ms 50 ms to 5 s	3 ppm 80 ppm to 0.6 %	
	Oscilloscope Calibration *			
	Bandwidth	50 kHz to 1.1 GHz 5 mV to 5.5 V p-p	3 % to 5 %	
	Amplitude (Deflection Factor)	1 mV to 130V (1M Ω) 1 mV to 6.6 V (50 Ω)	5 % to 0.2% 5 % to 0.3 %	
	Time Base (Marker)	1 ns to 20 ms 50 ms to 5 s	3 ppm 80 ppm to 0.6 %	
	Power Quality § Harmonics	1 to 40 th	0.2%	Using Multifunction Calibrator with PQ Option by Direct Method
	Power Quality* Harmonics	1 to 40 th Order	0.2 %	
	Transformer Turns Ratio Meter §	At 50 Hz 0.8 to 10,000	0.05 %	Using TTR Calibrator with Digital Multimeter by Comparison Method
	Transformer Turns Ratio Meter*	At 50 Hz 0.8 to 10,000	0.05 %	Using TTR Calibrator with Digital Multimeters by Comparison Method
	RF Power § (With 50 Ω Level Head)	10 Hz to 128 MHz +20 dBm to +24 dBm 100 mW to 251 mW 10 Hz to 1.4 GHz (-) 48 dBm to +20 dBm 15.85 nW to 100 mW	2.2 % 6 %	

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatyta Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 14 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
		1.4 GHz to 4GHz (-) 48 dBm to +14 dBm 15.85 nW to 25.12mW 100 KHz to 4 GHz (-) 94 dBm to (-) 48dBm 0.398 pW to 15.85 nW	15 % 30 %	
	RF Power [§] (With 75 Ω Level Head)	10 Hz to 125 MHz +18dBm to (-) 54 dBm 63.10 mW to 3.98 mW 125 MHz to 4 GHz +14 dBm to (-) 54 dBm 25.12 mW to 3.98 mW 100 KHz to 4GHz -90 dBm to (-) 54 dBm 1 pW to 3.98 mW 100 KHz to 3 GHz (-) 100 dBm to (-)90dBm 100 fW to 1 pW	5 % 16 % 20 % 30 %	
	RF Power* (With 50 Ω Level head)	10 Hz to 128 MHz +20 dBm to +24 dBm 100 mW to 251 mW 10 Hz to 1.4 GHz (-) 48 dBm to 20 dBm 15.85 nW to 100 mW 1.4 GHz to 4GHz (-) 48 dBm to +14 dBm 15.85 nW to 25.12mW 100 KHz to 4 GHz (-) 94 dBm to (-) 48dBm 0.398 pW to 15.85 nW	2.2 % 6 % 15 % 30 %	Using RF Reference Source by Direct Method
	RF Power* (With 75 Ω Level head)	10 Hz to 125MHz +18 dBm to (-) 54 dBm 63.10 mW to 3.98 mW 125 MHz to 4 GHz	5 %	Using RF Reference Source by Direct Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 15 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
		+14 dBm to (-) 54 dBm 25.12 mW to 3.98 mW 100 KHz to 4GHz (-) 90 dBm to (-) 54 dBm 1 pW to 3.98mW 100 KHz to 3 GHz (-) 100 dBm to (-)90dBm 100 fW to 1 pW	16 % 20 % 30 %	
	RF Attenuation [§]	200 Hz to 4 GHz 1 dB to 110 dB	0.03 dB to 0.2 dB	Using RF Calibrator by Direct Method
	RF Attenuation*	At 200 Hz to 4 GHz 1 dB to 110 dB	0.03 dB to 0.2 dB	Using RF Reference Source by Direct Method
	Amplitude Modulation [§]	Carrier Frequency 125 MHz to 1 GHz Modulation Depth 10 % to 99 % Modulation Rate 1 Hz to 100 kHz at 125 MHz 1 Hz to 20 kHz at 1GHz	5 %	Using RF Calibrator by Direct Method
	Amplitude Modulation*	Carrier Frequency 125 MHz to 1 GHz Modulation Depth 10 % to 99 % Modulation Rate 1 Hz to 100 kHz at 125 MHz 1 Hz to 20 kHz at 1GHz	5 %	Using RF Reference Source by Direct Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory

Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number

CC-2287

Page 16 of 48

Validity

03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
23.	Frequency Modulation [§]	Carrier Frequency 125MHz to 1GHz Modulation Rate 400Hz to 200 kHz Deviation 1 Hz to 300 kHz at 125MHz 300kHz to 1 MHz at 1GHz	4 %	Using RF Calibrator by Direct Method
	Frequency Modulation*	Carrier Frequency 125MHz to 1GHz Modulation Rate 400Hz to 200 kHz Deviation 1 Hz to 300 kHz at 125 MHz 300 kHz to 1 MHz at 1 GHz	4 %	Using RF Reference Source by Direct Method
	Impulse Calibration [§] Load: >250 k Ω 100 pF to 300 pF			
	Lightning Impulse Voltage (LI)	80 V to 1600 V (\pm Polarity)	0.6 %	
	Time Parameters T1 (Front Time) T2 (Time to Half Value)	0.84 μ s 60 μ s	2.5 % 2.5 %	
	Lightning Impulse Chopped Voltage (LIC)	400 V to 1250 V (\pm Polarity)	1.3 %	
	Time Parameters Tc (Time to Chop)	0.50 μ s	2.5 %	

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 17 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	Switching Impulse Voltage (SI)	80 V to 1600 V (\pm Polarity)	0.6 %	
	Time parameters Tp (Time to Peak) T2 (Time to Half Value)	20 μ s 4000 μ s	2.5 % 2.5 %	
	Impulse Calibration* Load: >250 k Ω , 100 pF to 300 pF			
	Lightning Impulse voltage (LI)	80 V to 1600 V (\pm Polarity)	0.6 %	
	Time parameters T1(Front Time) T2(Time to Half Value)	0.84 μ s 60 μ s	2.5 % 2.5 %	
	Lightning Impulse Chopped Voltage (LIC)	400 V to 1250 V (\pm Polarity)	1.3 %	
	Time parameters Tc (Time to Chop)	0.50 μ s	2.5 %	
	Switching Impulse Voltage (SI)	80 V to 1600 V (\pm Polarity)	0.6%	
	Time parameters Tp(Time to Peak) T2(Time to Half Value)	20 μ s 4000 μ s	2.5 % 2.5 %	
	Partial Discharge Calibration ^s			
	Apparent Charge q ₀	1 pC to 50 nC	6 %	

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 18 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	Partial Discharge Calibration*			
	Apparent Charge q_0	1 pC to 50 nC	6 %	Using Partial Discharge Calibrator by Comparison Method.
II.	MEASURE			
		10 μ V to 10 mV	500 ppm to 15 ppm	
		10 mV to 10 V 10 V to 1050 V	0.6 ppm to 0.1 ppm 0.1 ppm to 2.5 ppm	
	DC Voltage *	0.1 mV to 1050 V	0.3% to 0.001%	Using Reference Multimeter by Direct Method
	DC High Voltage [§]	1 kV to 100 kV	0.03%	Using DC High Voltage Divider & 6½ Digit Multimeter by Direct Method
	DC High Voltage *	1 kV to 150 kV	1.4 %	Using DC High Voltage Divider with kV Meter by Direct Method
	DC Current [§]	1 pA to 1mA 1 mA to 20 A 20 A to 1000 A 1000 A to 3000 A	1.5 % to 40 ppm 40 ppm to 10 ppm 10 ppm to 20 ppm 20 ppm to 0.2 %	Using Electrometer Standard Resistor / Standard Shunts with Reference Multimeter by Direct/ V/R Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 19 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	DC Current *	10 μ A to 1 A 1 A to 100 A 100 A to 3000 A	0.015% to 0.002% 0.002% to 0.005% 0.005% to 0.15%	Using Standard DC Resistor / DC Shunt with Digital Multimeter by V / R Method
	DC Resistance [§]	1 μ Ω to 10 Ω 10 Ω to 100 M Ω 100 M Ω to 100 G Ω 100 G Ω to 1 T Ω 1 T Ω to 100 T Ω	500 ppm to 1 ppm 1 ppm to 15 ppm 15 ppm to 200 ppm 200 ppm to 0.1 % 0.1 % to 1.5 %	Using DCC Bridge and Automated Dual Source & High Resistance Ratio Bridge by Comparison Method
		0.1 Ω to 200 k Ω 200 k Ω to 100 T Ω	0.02 % to 0.001 % 0.001 % to 3 %	
		10 Hz to 20 kHz 1 mV	0.3 % to 0.06 %	
		10 Hz to 1 kHz 2 mV to 2 V 2 V to 1000 V	400 ppm to 10 ppm 10 ppm to 36 ppm	
		1 kHz to 1 MHz 2 mV to 20 V	800 ppm to 10 ppm	
		1 kHz to 300 kHz 20 V to 60 V	30 ppm to 10 ppm	
		1 kHz to 100 kHz 60 V to 600 V	400 ppm to 10 ppm	
		1 kHz to 20 kHz 600 V to 1000 V	12 ppm	
		1 MHz to 1100 MHz 5 mV to 5.5 V (p-p)	4 % to 10%	Using Multifunction Calibrator & Digital Storage Oscilloscope by Comparison Method

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Taty Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 20 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	AC Voltage*	10 Hz to 100 kHz 1 mV to 1000 V	0.15 % to 0.005%	Using AC Measurement Standard by Direct Method
		100 kHz to 300 kHz 2mV to 60 V	0.03% to 0.5%	
		300 kHz to 1 MHz 2 mV to 20 V	0.03% to 1%	
		1 MHz to 1100 MHz 5 mV to 5.5 V (p-p)	4% to 10%	Using Multifunction Calibrator & Digital Storage Oscilloscope by Comparison Method
		At 50 Hz to 60 Hz 1 kV to 2.4 kV	0.01%	Using Precision AC Divider with 6½ DMM by Direct Method
		At 50 Hz 2.4 kV to 100 kV 100 kV to 200 kV	0.22 % 1.4 %	Using AC High Voltage Divider & 6½ Digit Multimeter by Direct Method
	AC High Voltage*	50 Hz 1 kV to 200 kV	1.4%	Using AC High Voltage Divider with kV Meter by Direct Method
		10 Hz to 1 kHz 10 μ A to 1 mA 1 mA to 100 A	0.025 % to 0.006 % 0.006 % to 0.01 %	
		1 kHz to 10 kHz 10 μ A to 100 A	0.015 %	
		40 Hz to 70 Hz 100 A to 160 A	90 ppm	Using Power / Energy Comparator by Direct Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Taty Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 21 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
		50 Hz 100 A to 10,000 A	0.25%	Using Standard CT with Power / Energy Reference Meter by Direct Method
		40 Hz to 70 Hz 1mA to 50 mA 50 mA to 160 A	300 ppm to 70 ppm 70 ppm to 90 ppm	Using Power / Energy Comparator by Direct Method
		10 Hz to 1 kHz 10 μ A to 200 μ A 200 μ A to 20 A	0.3 % to 0.05 % 0.05% to 0.3 %	
		1 kHz to 10 kHz 10 μ A to 20 A	0.1 % to 0.3%	
		50 Hz 100 A to 3500 A	0.1%	Using Standard CT with Power / Energy Reference Meter by Direct Method
	Frequency / Period [§]	40 mHz to 20 GHz 50 ps to 25 s	4×10^{-8} to 1.5×10^{-11}	Using GPS Controlled Frequency Standard & Counter by Direct Method
	Frequency /Time Period *	40 mHz to 20 GHz 25 s to 50 ps	4×10^{-7} to 1.5×10^{-8}	Using Frequency Counter by Direct Method
	Time Interval [§]	1 μ s to 1000 s 1000 s to 24 Hrs & multiple of 24 hours	1.1×10^{-3} to 1.1×10^{-7} 1×10^{-7} to 0.023 %	Using GPS Controlled Frequency Standard & Counter by Comparison Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 22 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	Time Interval *	1 μ s to 1000 s 1000 s to 24 Hrs. & Multiple of 24 Hrs.	1.1×10^{-3} to 1.1×10^{-7} 1×10^{-7} to 0.023 %	Using Frequency Counter & Time Interval Meter by Direct Method
		40 Hz to 70 Hz 25 V to 480 V PF: 0.01 to 1 1 mA to 120 A	0.04 % to 0.01 % / PF	Using Power/ Energy Comparator by Direct Method
		40 Hz to 70 Hz 25 V to 480 V PF: 0.01 to 1 120 A to 300 A	0.01 % to 0.25 % / PF	Using Power/ Energy Meter Test System by Direct Method
		40 Hz to 70 Hz 25 V to 480 V PF: 0.01 to 1 1 mA to 120 A	0.04% to 0.01% / PF	Using Power/ Energy Comparator by Direct Method
		40 Hz to 70 Hz 25 V to 480 V PF: 0.01 to 1 120 A to 300 A	0.01 % to 0.25 % / PF	Using Power/ Energy Meter Test System by Direct Method
	Active / Reactive / Apparent Power /Energy (1 Phase) [§]	50 Hz & 60 Hz 480 V to 1050V PF: 0.01 to 1 1 mA to 10 mA 10 mA to 120A	0.04 % to 0.011 % /PF 0.011 % / PF	Using Power Energy Comparator with Precision AC Voltage Divider by Direct Method
	Active / Reactive / Apparent Power / Energy (1 Phase) *	50 Hz & 60 Hz 480 V to 1000V 1 mA to 10 mA 10 mA to 120A PF: 0.01 to 1	110 ppm /PF	Using Power Energy Comparator with Precision AC Voltage Divider by Direct Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 23 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
9.	Power Factor [§] (Phase Angle)	40 to 70 Hz 0 to UPF (0 to 360°) 25 V to 1050 V, 1 mA to 160 A	0.008°	Using Power/ Energy Comparator & Precision AC Voltage Divider by Direct Method
	Power Factor / Phase Angle *	40Hz to 70 Hz 0 to UPF (0 to 360°) 25 V to 1050 V 1 mA to 160 A	0.008°	Using Power/ Energy Comparator & Precision AC Voltage Divider by Direct Method
		1 kHz 1 pF to 100 pF 100 pF to 1 μ F 1 μ F to 1 F	5 ppm 5 ppm to 150 ppm 150 ppm to 800 ppm	
		100 Hz 1 pF to 100 pF 1 μ F to 1 F	7 ppm to 3 ppm 200 ppm to 800 ppm	
		50 Hz 1 pF 10 pF 100 pF	60 ppm 7 ppm 4 ppm	
	Capacitance *	1 kHz 1 pF to 100 nF 100 nF to 1 F	0.5% to 0.03% 0.03% to 0.1 %	Using Standard Capacitor & Four Terminal Capacitance Standard with RLC Dig bridge by Comparison Method
	Inductance [§]	1 kHz 100 μ H to 100 mH 100 mH to 10 H	0.07 % to 0.02 % 0.02 % to 0.04 %	Using Precision Component Analyzer by Direct Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 24 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	Inductance *	1 kHz 100 μ H to 100 mH 100 mH to 10 H	0.07 % to 0.02 % 0.02 % to 0.04 %	Using Precision Component Analyzer by Direct Method
	AC Resistance [§]	1 kHz 0.001 Ω to 10 k Ω	0.02 % to 0.0025 %	Using Precision Component Analyzer by Direct Method
	AC Resistance *	1 kHz 0.001 Ω to 10 k Ω	0.02 % to 0.0025 %	Using Precision Component Analyzer by Direct Method
	DC Power / Energy [§]	10 V to 1000 V 1 A to 30 A	0.006 % to 0.05 %	Using Digital Multimeters by Direct Method
	DC Power/ Energy*	10 V to 1000 V 1 A to 30 A	0.15 %	Using Digital Power Meter by Direct Method
	Temperature By Simulation Method : For Temperature Indicators, Recorders, Controllers [§]			
	RTD	(-) 200 $^{\circ}$ C to 850 $^{\circ}$ C	0.001 $^{\circ}$ C to 0.02 $^{\circ}$ C	
	T/C – K, J, N, E, T, R, S, B, C, L & U	(-) 270 $^{\circ}$ C to 2300 $^{\circ}$ C	0.01 $^{\circ}$ C to 0.26 $^{\circ}$ C	
	Temperature Simulation By Simulation Method For Temperature Indicators, Recorders, Controllers *			
	RTD [§]	(-) 200 $^{\circ}$ C to 850 $^{\circ}$ C	0.005 $^{\circ}$ C to 0.02 $^{\circ}$ C	
	T/C – K, J, N, E, T, R, S, B, C, L & U [§]	(-) 270 $^{\circ}$ C to 2300 $^{\circ}$ C	0.01 $^{\circ}$ C to 0.3 $^{\circ}$ C	
	Tan Delta [§] (Absolute Value) & Capacitance at High Voltage	At 50 Hz 200 V to 100kV 10 pF to 1.5 μ F 1 x 10 ⁻⁵ to 5 x 10 ⁻²	0.14 % to 0.025% 1.6 x 10 ⁻⁵ to 5.7 x 10 ⁻³	Using C & Tan Delta Measurement System & Standard Capacitor by Comparison Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Taty Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 25 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	Tan Delta * (Absolute Value) & Capacitance at High Voltage	At 50 Hz 200 V to 100kV 10 pF to 1.5 μ F 1×10^{-5} to 1	0.14 % to 0.025% 1.6×10^{-5} to 5.7×10^{-3}	Using C & Tan Delta Measurement System & Standard Capacitor by Direct Method
	Harmonics § Fundamental Frequency 50 Hz	1 to 40 th Order with Fundamental Frequency	0.5 %	Using Power / Energy Comparator by Direct Method
	Harmonics Order * Fundamental Frequency 50 Hz	1 to 40 th Order with Fundamental Frequency	0.5 %	Using Power / Energy Test System by Direct Method
	CT / PT Burden §	At 50 Hz 1 VA to 100 VA	0.05 %	Using Power / Energy Meter by Direct Method
	CT / PT Burden *	At 50 Hz 1 VA to 100 VA	0.05 %	Using Power / Energy Test System by Direct Method
18.	Transformer Turns Ratio Meter Calibrator §	At 50 Hz 0.8 to 2100	0.03 %	Using Multifunction Calibrator with Digital Multimeter by Comparison Method
	Ratio Error and Phase Displacement Error of CT/ PT Comparator § (AITTS)	50 Hz 0.05 A to 6A 50 Hz 25 V to 150 V	For CT Ratio Error: 0.003% to 0.02% Phase Displacement Error: 0.1 min to 0.6 min For PT Ratio Error: 0.008% to 0.013% Phase Displacement	Using Power / Energy Test System with AITTS by Comparison Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 26 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
			Error: 0.25 min	
	Ratio Error and Phase Displacement Error of CT/ PT Comparator * (AITTS)	50 Hz 0.05 A to 6A 50 Hz 25 V to 150 V	CT Ratio Error= 0.003% to 0.02% Phase Displacement Error=0.1 min to 0.6 min PT Ratio Error =0.008% to 0.013% Phase Displacement Error =0.25 min	Using Power / Energy Test System with AITTS by Comparison Method
	CT/PT Calibration [§]			
	Current Transformer Ratio Error & Phase Displacement Error	At 50 Hz 1-5,000A / 1-5A (Direct) 5000-10,000A / 1-5A (By Turns)	Ratio Error: 0.004% to 0.025% Phase Displacement Error: 0.15 min to 0.65 min	Using Standard CT , Instrument Transformer Measuring Bridge by Comparison Method
	Voltage Transformer Ratio Error & Phase Displacement Error (Inductive)	At 50 Hz 1.1 kV - 100 kV / 110 V 1.1 kV - 132 kV/ $\sqrt{3}$ / 110 V/ $\sqrt{3}$	Ratio Error: 0.01% Phase Displacement Error: 0.32 min	Using Standard Capacitor, EPD, Instrument Transformer Measuring Bridge by Comparison Method
	CT / PT Calibration*			
	Current Transformer Ratio Error & Phase Displacement Error	At 50 Hz 1-10,000A / 1-5A	Ratio Error = 0.02% to 0.06% Phase Displacement Error =0.5 min to 2.5 min	Using Standard CT , Instrument Transformer Measuring Bridge by Comparison Method
	Voltage Transformer Ratio Error & Phase Displacement Error (Inductive)	At 50 Hz 1.1 kV to 66 kV / 110 V 1.1 kV to 66 kV/ $\sqrt{3}$ /	Ratio Error = 0.06% Phase Displacement	Using Standard Capacitor, EPD , Instrument Transformer Measuring Bridge by

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 27 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
		110 V/ $\sqrt{3}$	Error = 2 min	Comparison Method
		At 50 Hz & 60 Hz 1 mA to 100 A	Ratio Error 0.0076% Phase Displacement Error: 0.30 min to 0.25 min	Using ICT Calibration System by Direct Method
		At 50 Hz & 60 Hz 100A to 120 A	0.01% 0.25 min	Using Power/ Energy Comparator by Comparison Method
	Isolation Current Transformer*	At 40 Hz to 70 Hz 1 mA to 120 A	Ratio Error =0.008% Phase Displacement Error =0.30 min	Using Power/ Energy Comparator by Comparison Method
	Electrical Fast Transient (50 Ω & 1k Ω) [§] As per IEC 61000-4-4:2012 Ed.3.0			
	a. Amplitude	\pm 0.25 kV to 4.0kV	6.7%	
	b. Rise Time	5ns	6.5%	
	c. Pulse Width	50 ns \pm 30% at 50 Ω 50 ns, -15ns to +100ns at 1k Ω	6.5%	
	d. Repetition Rate	200 μ s (5kHz) 10 μ s (100kHz)	6.5%	
	e. Burst Period	300 ms	6.5%	
	f. Burst Duration	15ms (5 kHz) 0.75ms (100kHz)	6.5%	
	Surge In Open Circuit Voltage [§] As per IEC 61000-4-5:2017 Ed.3.1			
	a. Amplitude	\pm 0.5kV to \pm 7.0 kV	6.6%	

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 28 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	b. Front Time	1.2 μ s	7.9%	Differential probe model 4241 & Current Monitor Model 411 by Direct Method
	c. Pulse Width	50 μ s	6.5%	
	In short Circuit Current			
	a. Current Amplitude	\pm 0.25kA to \pm 4 kA (\pm 0.5kV to \pm 7.0 kV)	6.3%	
	b. Front Time	8 μ s	6.1%	
	c. Pulse Width	20 μ s	6.0%	
	Telecom Surge In Open Circuit Voltage ^s As per IEC 61000-4-5:2017 Ed.3.1			
	a. Amplitude	\pm 0.5kV to \pm 10.0 kV	6.6%	
	b. Front Time	10 μ s	7.9%	
	c. Pulse Width	700 μ s	6.5%	
	In short Circuit Current			
	a. Current Amplitude	\pm 12.5 A to \pm 250 A (\pm 0.5kV to \pm 10.0 kV)	6.3%	
	b. Front Time	5 μ s	6.1%	
	c. Pulse Width	320 μ s	6.0%	
	Electrostatic Discharge ^s (\pm 2kV to \pm 15 kV) As per IEC 61000-4-2:2008 Ed.2.0			
	a. Peak current	\pm 7.5 to \pm 60 A	7.2%	
	b. Rise Time	0.8ns	7.0%	
	c. Current at 30ns	4 A to 30 A	7.2%	
	d. Current at 60ns	2A to 15 A	7.2%	
		a. 0 % to 80% of Voltage	5 %	
		b. 10ms – 5 sec	6 %	

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 29 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	As per IEC 61000-4-11 : 2017 IEC 61000-4-29 : 2000 ^s			model 4241 by Direct Method
	Power frequency Magnetic Field ^s As per IEC 61000-4-8:2009 Ed.2.0)			
	Current	1 A to 100A	1%	
	Pulse Magnetic Field ^s As per IEC 61000-4-9 :2016 Ed.2.0			
	a. Pulse Level	100 A to 1000 A	6.3%	
	b. Rise Time	8 μ s	6.1%	
	c. Pulse Duration	20 μ s	6.1%	
	Damped Oscillatory Generator ^s As per IEC 61000-4-18:2011 Ed.1.1			
	I. Slow Damped Oscillatory (In open Circuit)			
	a. Amplitude	\pm 0.25kV to \pm 2.5 kV	6.6%	
	b. Rise Time	75ns	6.5%	
	c. Repetition Rate	40/s for 100kHz & 400/s for 1 MHz	6.0%	
	d. Voltage Decay	Pk 5 must be > 50% of the Pk1 Pk10 must be < 50% of the Pk1	6.7%	
	(Short circuit)			
	a. Current Amplitude	1.25 to 12.5A	6.3%	
	II. Fast Damped Oscillatory (Open Circuit)			
	a. Amplitude	\pm 0.25kV to \pm 4 kV	6.5%	
	b. Rise Time	5ns	6.1%	
	c. Voltage oscillation Frequency	3MHz, 10MHz, 30MHz	6.0 %	
	d. Decaying	Pk 5 must be > 50% of	6.7 %	

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Taty Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 30 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
		the Pk1 Pk10 must be < 50% of the Pk1		
	e. Bust Duration	3 MHz to 50ms 10 MHz to 15ms 30 MHz to 5 ms	6.0 %	
	f. Burst Period (Short Circuit)	300ms	6.0 %	
	a. Current Amplitude	5 A to 80A	6.2 %	
	b. Current Rise Time	3 MHz to < 330 ns 10 MHz to < 100 ns 30 MHz to < 33 ns	6.0 %	
	c. Current Oscillation Frequency	3, 10, 30MHz	6.1 %	
	d. Decaying	Pk 5 must be > 25% of the Pk1 Pk10 must be < 25% of the Pk1	6.2 %	
	Ring Wave Generator Open Circuit ^s As per IEC 61000-4-12:2017 Ed.3.0			
	a. Amplitude	± 0.25 kV to ± 4 kV	6.6%	
	b. Rise Time	0.5 μ s	12.1%	
	c. Oscillation Frequency	100kHz	6.5%	
	d. Decaying	Pk2 40% to 110% of Pk1 Pk3 40% to 80% of Pk2 Pk4 40% to 80% of Pk3	6.7%	
	Short Circuit			
	a. Current Amplitude	20.8 to 333.3 A at 12 Ω 8.3 to 133.3 A at 30 Ω	6.2%	

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 31 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	b. Rise Time	$\leq 1 \mu\text{s}$	11.8%	
	Damped Oscillatory Magnetic Field [§] As per IEC 61000-4-10:2016 Ed.2.0			
	a. Peak current	11.1 A to 111 A \pm 20%	6.3%	
	b. Oscillation Period	10 μs \pm 1 μs at 100kHz 1 μs \pm 0.1 μs at 1MHz	6.0%	
	c. Repetition Time	25ms \pm 2.5 ms at 100kHz 2.5ms \pm 0.25 ms at 1MHz	6.0%	
	d. Decay Rate	Pk5 shall be >50% of the Pk1 Value Pk10 shall be < 50% of the Pk1 Value	6.3%	
	Impulse Voltage [§]	1 kV to 15 kV 10 ns to 100 μs	6.3% 0.1%	By Using Impulse Probe with DSO by Direct Method
	Impulse Voltage [*]	1 kV to 15 kV 10 ns to 100 μs	6.3% 0.1%	Using Impulse Probe with DSO by Direct Method

FLUID FLOW CALIBRATION

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Taty Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 32 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
1.	Volumetric Flow Rate [§]	1.5 m ³ /h to 240 m ³ /h	0.20 %	By Comparison Method as per ISO-4185
2.	Mass Flow Rate [§]	1500 kg/h to 240000 kg/h	0.20 %	By Comparison Method as per ISO-4185
3.	Volumetric Flow Rate [*]	5 m ³ /h to 240 m ³ /h	1.5 %	Using Ultrasonic Flow Meter By Comparison Method

MECHANICAL CALIBRATION

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 33 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
1.	DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)			
1.	Calipers (Analog, Dial, Digital) [§] L.C.: 10 μ m	Up to 600 mm Up to 1000 mm	10.1 μ m 13.2 μ m	Using Caliper Checker & By Comparison IS 3651 Using Check Master By Comparison IS 3651
2.	Height Gauges [§] (Analog, Dial, Digital) L.C.: 10 μ m	Up to 600mm	12.8 μ m	Using Caliper Checker, & Surface plate By Comparison IS 2921
3.	External Micrometer [§] L.C.: 1 μ m	Up to 25mm Up to 300mm	1.2 μ m 3.1 μ m	Using Tungsten Carbide Slip Gauges Grade '0' & 'I', Long Slip Gauge Grade 'K' & Optical Flat By Comparison IS 2967
4.	Inside Micrometer [§] L.C.: 10 μ m	25 to 200mm	11 μ m	Using Universal Measuring System Microrep By Comparison IS 2961
5.	Bevel Protractor [§] L.C.: 1 min.	0° to 180 °	0.71 min.	Using Steel Angle Gauge Set By Comparison IS 4239
6.	Dial Gauge (Plunger / Lever Type) [§] L.C.: 1 μ m L.C.: 10 μ m	Up to 25 mm Up to 50 mm	3.9 μ m 5.1 μ m	Using Electronic Dial Calibrator Tester & Universal Measuring System Microrep By Comparison IS 2092 & IS11498
7.	Bore Dial Gauge (Transmission Movement) [§]	Up to 2 mm	3.9 μ m	Using Electronic Dial Calibrator Tester

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 34 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
8.	Feeler Gauges Set/ Step Wedge [§]	Up to 2 mm Up to 40 mm	1.9 μ m 5.7 μ m	Using Ext. Micrometer L.C. 1 μ m By Comparison IS 3179
9.	Measuring Scale [§] L.C.: 0.5 mm	Up to 1000mm Up to 2000mm	75 μ m 100 \sqrt{L} μ m (L in m)	Using Tape & Scale Calibrator
10.	Measuring Tape (Oven Metallic Steel, Glass Fiber, Cloth) [§]	Up to 50000 mm	108 \sqrt{L} μ m (L in m)	Using Tape & Scale Calibrator
11.	Depth Gauge [§] (Analog, Dial, Digital) L.C. 10 μ m	Up to 300 mm	7.8 μ m	Using Tungsten Carbide Slip Gauges Grade '0' & Long Slip Gauge Grade 'K' & Surface Plate
12.	Depth Micrometer [§] (Analog, Dial, Digital) L.C.: 1 μ m	Up to 300 mm	5.0 μ m	Using Tungsten Carbide Slip Gauges Grade '0' & 'I', Long Slip Gauge Grade 'K' & surface plate By Comparison IS 2967
13.	Micrometer Head [§] L.C.: 1 μ m	Up to 50 mm	2.69 μ m	Using Tungsten Carbide Slip Gauges Grade '0' & 'I', & Optical Flat By Comparison IS 9483
14.	Length Gauge / Setting Rod / Length Bar / Height Block [§]	25 mm to 500 mm	4.4 μ m	Using Electronic Probe with DRO & Comparator stand, Tungsten Carbide Slip Gauges Grade '0' & Long Slip Gauge Grade 'K' By Comparison IS

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 35 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
				7014
15.	Dial Thickness Gauge [§] L.C.: 1 μ m	Up to 10mm	7.1 μ m	Using Tungsten Carbide Slip Gauge Grade '0' & Lever dial gauge By Comparison IS 2092
16.	Cylindrical Measuring Pins [§]	0.1 to 20 mm	1.3 μ m	Using Electronic Probe with DRO & Comparator stand, Tungsten Carbide Slip Gauges Grade '0' By Comparison IS 11103
17.	Plain Plug Gauge / Cylindrical Setting Master / Plain Mandrill [§]	Up to 100mm	3.3 μ m	Using Electronic Probe with DRO & Comparator stand, Tungsten Carbide Slip Gauges Grade '0' By Comparison IS 3455, IS 2220, IS 4349
18.	Snap Gauge / Dial Snap Gauge (Parameter - Flatness, Parallelism) [§]	Up to 300mm	2.6 μ m	Using Tungsten Carbide Slip Gauge Grade '0', Long Slip Gauge Grade 'K' & Lever dial gauge By Comparison IS 8023 & IS 7606
19.	Test Sieve Aperture Size / Pitch [§]	30 μ m to 4mm	6.1 μ m	Using Universal Measuring System Microrep By Comparison IS 460 (Part I, II & III)
20.	Ultrasonic Thickness Gauge [§] L.C. 0.1 mm	Up to 300 mm	71 μ m	Using Tungsten Carbide Slip Gauge Grade '0', Long Slip Gauge Grade

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 36 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
				'K'
21.	Wire Gauge (Parameter - Diameter) [§]	Up to 10 mm	5.1 μ m	Using Universal Measuring System
22.	Ring Gauge [§]	Up to 100 mm	6.5 μ m	Using CMM
23.	Industrial Gauge Test Probe (Electrical) (Parameter - Length, Diameter, Angle) [§]	Up to 450 mm Up to 90°	5 μ m 1.2 min.	Using Universal Measuring System Microrep By comparison IEC 61032
24.	Industrial Gauge Go - Nogo Gauge (Electrical Verification) (Parameter - Length, Diameter, Width, Thickness, Angle) [§]	Up to 100mm Up to 90°	5.7 μ m 1.2 min.	Using CMM By comparison IS 15518:2004 (Part I)
25.	Electronic Height [§] L.C.: 0.1 μ m	Up to 600mm Up to 1000mm	4.4 μ m 5.2 μ m	Using Check Master
26.	Universal Measuring System / Length Measuring Machine [§] L.C.: 1 μ m	X Axis 300mm Y Axis 200mm	1.8 μ m	Using Tungsten Carbide Slip Gauges Grade '0' & Long Slip Gauge Grade 'K'
II.	DIMENSION (PRECISION INSTRUMENTS)			
1.	CMM [§]			Using Check Master

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 37 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	L.C.: 0.01 μm	Up to 1500mm	6 μm	By Comparison Method ISO10360/IS15635
2.	Caliper Checker [§]	Up to 1000mm	4.6 μm	Using Check Master & Lever Dial Gauge
3.	Profile Projector Linear/Magnification / Angle [§] L.C.: 1 μm	300 mm x 300 mm 10X to 100X 0° to 360°	1.8 μm 0.16 % 2.7 min.	Using Glass Scale, Steel Angle Gauge Set & Digital Caliper
4.	Tape & Scale Calibrator [§] L.C.: 1 μm	Up to 1000 mm	16 μm	Using Tungsten Carbide Slip Gauges Grade '0' & Long Slip Gauge Grade 'K'
5.	Angle Gauge Error in Angle [§]	Up to 90°	1 min.	Using CMM
6.	CNC Machine Positioning Accuracy Rotary / Indexing Table [§]	Up to 15000 mm 0° to 360°	5.7 μm 2.3 min.	Using Laser Measuring System Using Laser Measuring System with Rotary Indexer
7.	Electronic Height* L.C.: 0.1 μm	Up to 600 mm Up to 1000 mm	4.4 μm 5.2 μm	Using Check Master
8.	Universal Measuring System / Length Measuring Machine*	X Axis 300 mm	1.8 μm	Using Tungsten Carbide Slip Gauges Grade '0' & Long Slip Gauge

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 38 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	L.C.: 1 μ m	Y Axis 200 mm		Grade 'K'
9.	CMM* L.C.: 0.01 μ m	Up to 1500 mm	6 μ m	Using Check Master By comparison ISO10360/IS15635
10.	Profile Projector Linear/Magnification / Angle* L.C. 1 μ m	300 mm X 300 mm 10X - 100X 0° - 360°	1.8 μ m 0.16% 2.7 min.	Using Glass Scale, Steel Angle Gauge Set & Digital Caliper
11.	Tape & Scale Calibrator* L.C. 1 μ m	Up to 1000mm	16 μ m	Using Tungsten Carbide Slip Gauges Grade '0' & Long Slip Gauge Grade 'K'
12.	CNC Machine Positioning Accuracy	Up to 15000mm	5.7 μ m	Using Laser Measuring System
	Rotary / Indexing Table*	0° to 360°	2.3 min.	Using Laser Measuring System with Rotary Indexer
III.	TORQUE GENERATING DEVICES			
1.	Torque Screw Driver [§]	0 to 10 Nm	2.98 %	Using Torque Calibration Systems ISO 6789
	Torque Wrench Driver [§]	0 to 20 Nm	2.98 %	
		20 Nm to 1000 Nm	2.36 %	
		1000 Nm to 2000 Nm	2.0 %	
IV.	ACCELERATION AND SPEED			
1.	Speed Calibration of	10 RPM to 1,000 RPM	0.83 RPM	Using Tachometer

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Taty Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 39 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	Tachometer / RPM Measurement / Stroboscope / Tachometer (Contact Mode) #	$\geq 1,000$ RPM to 12,000 RPM	3 RPM	By Comparison Method
2.	Non Contact Mode RPM Measurement / Stroboscope#	10 RPM to 30 RPM ≥ 30 RPM to 12,000 RPM $\geq 12,000$ RPM to 1,00,000 RPM $\geq 1,00,000$ RPM to 1,20,000 RPM	0.83 RPM 3 RPM 6 RPM 24 RPM	Using Tachometer By Comparison Method
V.	ACCOUSTICS			
1.	Sound – Source (Sound Measuring Instruments) #	74 dB, 84 dB, 94 dB, 104 dB, 114 dB (at 125 Hz, 250 Hz, 500 Hz, 1 kHz, 2 kHz & 4 kHz)	0.39 dB	Using Sound Level Meter with Anechoic Chamber By Direct Method
2.	Sound-Measure (Sound Level Calibrator / Sound Generator) #	74 dB to 114 dB (at 125 Hz to 4 kHz)	0.39 dB	Using Sound Level Calibrator By Comparison Method

VI.	WEIGHTS			
------------	----------------	--	--	--

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 40 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
1.	Weights ^s	1 mg	0.003 mg	Using E1 class standard Weights 1mg-20kg and Balance of d:0.001mg & 0.01mg Calibration of weights as Per OIML R111-1 of Accuracy class E2
		2 mg	0.003 mg	
		5 mg	0.003 mg	
		10 mg	0.003 mg	
50 mg	0.003 mg			
100 mg	0.003 mg			
200 mg	0.004 mg			
500 mg	0.005 mg			
1 g	0.006 mg			
2 g	0.008 mg			
		5 g	0.009 mg	
		10 g	0.02 mg	
		20 g	0.03 mg	
		50 g	0.03 mg	
		100 g	0.04 mg	
		200 g	0.07 mg	
		500 g	0.2 mg	Using Mass Comparator of d:0.1mg
		1 kg	0.5 mg	
		2 kg	1 mg	
		5 kg	1.5 mg	
		10 kg	4 mg	Using Mass Comparator of d:1mg
		20 kg	9 mg	

VII.	WEIGHING SCALE & BALANCE		
-------------	-------------------------------------	--	--

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 41 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
1.	Mass / Electronic* Weighing Balance d : 0.001mg d : 0.1mg d : 0.01mg d : 1mg d : 10mg d : 100mg	0 to 5g 0 to 5g 0 to 220g 0 to 5kg 0 to 10kg 0 to 34kg	0.005mg 0.2mg 0.06mg 3mg 3mg 220mg	Using E1 & E2 class Standard weights 1mg to 20kg & Electronics Weighing Balances of accuracy Class I & Coarser as per OIML R76-1
	d : 1g d : 10g d : 10g d : 20g d : 0.1kg	0 to 100kg 0 to 100kg 0 to 200kg 0 to 200kg 500g to 5000kg	2g 15g 15g 30g 0.35kg	Using F1 class weights 20kg & Electronics Weighing Balances of accuracy Class III & Coarser as per OIML R76-1 & OIML R 47 & ISO 4185:1990
VIII.	VOLUME			
1.	Micropipettes / Piston Operated Pipettes ^{\$}	5 μ l to 100 μ l >100 μ l to 1000 μ l	0.4 μ l 0.4 μ l	Using weighing balance with d:0.001mg and distilled water & Calibration of Micro Pipettes based on Gravimetric method as per ISO 8655 Part 6
2.	Glassware Pipette/ Burette Measuring Cylinder / Volumetric Flask/Graduated Jar/ Can etc. ^{\$}	1ml to 10ml >10ml to 100ml >100ml to 2000ml >2000ml to 20000ml	12 μ l 0.07ml 0.62ml 4ml	Using weighing balance with d:0.1mg & 1mg distilled water & Calibration of glassware Based on Gravimetric method as per ISO 4787
IX.	PRESSURE INDICATING DEVICES			

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Taty Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 42 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
1.	Pressure Indicator/ Calibrator/ Controller, Manometer, Magnehelic Gauge, Low Pressure Gauge Low Pressure Instruments (Pneumatic) §	0 to 75 mbar g	0.02% of Rdg.	By Comparison method UUC to Standard as per DKD-R-6-1
2.	Pressure Indicator/ Calibrator/ Controller, Manometer, Magnehelic Gauge, Low Pressure Gauge. Low pressure instruments (Pneumatic) §	0 to (-) 75 mbar g	0.02% of Rdg.	By Comparison method UUC to Standard as per DKD-R-6-1
3.	Pressure Indicator/ Calibrator/ Controller, Manometer, Magnehelic Gauge, Low Pressure Gauge Low Pressure Instruments (Pneumatic) §	75 mbar g to 350 mbar g	0.02% of Rdg.	By Comparison method UUC to Standard as per DKD-R-6-1
4.	Pressure Indicator/ Calibrator/ Controller, Manometer, Magnehelic Gauge, Low Pressure Gauge. Low Pressure Instruments (Pneumatic) §	(-) 75 mbar g to (-)350 mbar g	0.02% of Rdg.	By Comparison method UUC to Standard as per DKD-R-6-1
5.	Pneumatic Dead Weight Tester	0.1 bar g to 35 bar g	0.005% of Rdg	By Cross Float Comparison method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 43 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	(Pneumatic) [§]			UUC to Standard as per Euramet-cg-3
6.	Pneumatic Pressure Indicator/Calibrator/ Controller, Digital Pressure Gauge & Pressure Instruments (Pneumatic) [§]	0.1 bar g to 35 bar g	0.005% of Rdg	By Comparison method UUC to Standard as per DKD-R-6-1
7.	Calibration of Pneumatic Pressure Indicator/Calibrator/ Controller, Digital Pressure Gauge & Pressure Instruments (Pneumatic) [§]	2 bar g to 200 bar g	0.005% of Rdg	By Comparison method UUC to Standard as per DKD-R-6-1
8.	Hydraulic Dead Weight Tester (Hydraulic) [§]	2 bar g to 2500 bar g	0.007% of Rdg	By Cross Float Comparison method UUC to Standard as per Euramet-cg-3
9.	Pressure Indicator/ Calibrator/ Controller, Pressure Chart Recorder, Pressure Instruments & Analog Gauge (Hydraulic) [§]	2 bar g to 2800 bar g	0.007% of Rdg	By Direct method UUC to Standard as per DKD-R-6-1
10.	Calibration of Digital Pressure Indicator, Calibrator/Controller Instruments and Gauges (Pneumatic) [§]	(-) 0.9750 bar g to (-)0.1 bar g	0.005 % of Rdg	By Comparison method UUC to Standard as per DKD-R-6-1
11.	Calibration Of Absolute Pressure	0 to 1 bar a	0.02 % Rdg	By Comparison method UUC to Standard as per

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 44 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	Manometer/Indicator, Analog/Digital Barometers (Pneumatic) [§]			DKD-R-6-1
12.	Pressure(Pneumatic) Analog/Digital Pressure Gauge, Pressure Indicator, Transmitter, Pressure Chart Recorder*	0 to 20 bar	0.05 % Rdg	By Comparison method UUC to Standard as per DKD-R-6-1
13.	Pressure(Hydraulic) Analog/Digital Pressure Gauge, Pressure Indicator, Transmitter, Pressure Chart Recorder*	0 to 100 bar	0.05 % Rdg	By Comparison method UUC to Standard as per DKD-R-6-1
14.	Pressure(Hydraulic) Analog/Digital Pressure Gauge, Pressure Indicator, Transmitter, Pressure Chart Recorder*	0 to 1000 bar	0.05 % Rdg	By Comparison method UUC to Standard as per DKD-R-6-1
15.	Negative Pressure Analog/Digital Pressure Gauge, Pressure Indicator, Transmitter, Pressure Chart Recorder*	1 bar to 0.1 bar vacuum	0.1 % Rdg	By Comparison method UUC to Standard as per DKD-R-6-1

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 45 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>THERMAL CALIBRATION</u>				
I.	TEMPERATURE			
	Fixed Point Method			
1.	Liquid Nitrogen Comparator [§]	(-) 195.795 °C	4.4 mK	Using SPRT, Liquid Nitrogen Comparator & Precision Thermometry Bridge by Comparison Method
2.	Triple Point of Mercury [§]	(-) 38.8344 °C	2.9 mK	Using SPRT, Mercury T.P. Cell & Precision Thermometry Bridge & Stirred Liquid Bath by Fixed Point Method
3.	Triple Point of Water [§]	0.01°C	1.5 mK	Using SPRT, Water Triple Point Cell & Precision Thermometry Bridge & Stirred Liquid Bath by Fixed Point Method
4.	Gallium Melting Point [§]	29.7646 °C	3.2 mK	Using SPRT, Optimal Gallium Melting Point Cell & Precision Thermometry Bridge & Stirred Liquid Bath by Fixed Point Method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Taty Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 46 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
5.	Tin Freezing Point [§]	231.928 °C	4.4 mK	Using SPRT, Tin Freezing Point Cell & Precision Thermometry Bridge & Dry Block Calibrator by Fixed Point Method
6.	Zinc Freezing Point [§]	419.527 °C	5.4 mK	Using SPRT, Zinc Freezing Point Cell & Precision Thermometry Bridge & Dry Block Calibrator by Fixed Point Method
7.	Aluminium Freezing Point [§]	660.323°C	6.9 mK	Using SPRT, Aluminum Freezing Point Cell & Precision Thermometry Bridge & Dry Block Calibrator by Fixed Point Method
8.	RTD, Thermocouple (with or without Indicator), Temperature Indicator, Data Loggers etc. with Sensors [§]	(-) 80 °C to 140 °C 140 °C to 660 °C	0.12 °C 0.23 °C	Using SPRT and Read Unit & Fluke Dry Block Calibrators by Comparison method
9.	Liquid-In-Glass Thermometer [§]	(-) 50°C to 250°C	0.17 °C	Using SPRT and Read Unit & Using Julabo / Isotech Liquid Bath Based on Facility Available by Comparison method

Rajeshwar Kumar
Convenor

Avijit Das
Program Manager

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Taty Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 47 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
10.	Thermocouple (with & without Indicator), Temperature Indicators/ Data Logger etc with Sensor [§]	>660 °C to 1000 °C	1.52 °C	Using S-Type Thermocouple Read Unit & Using High Temperature Furnace by Comparison method
11.	IR/Non-Contact Thermometer/ Pyrometer [§]	50 °C to 500 °C	5.7 °C	Using Standard Infra-Red Thermometer, Using Black Body Source Emissivity 0.95 By Comparison Method
12.	Humidity Indicator with Sensor of Humidity Chamber/ Environmental Chamber*	10 % to 95 %RH @25°C 95 %RH @25°C to 70°C	1.3 % RH 1.3 % RH	Using Reference Temp./Humidity Sensor with Indicator by Single Position Calibration at Different Temperature by Comparison Method
13.	Temperature Chambers/ Deep Freezers/Oven*	(-) 80 °C to 0 °C 0 °C to 250 °C	1.2 °C 2.5 °C	Using RTD Sensors With Data Logger by Multi Position Calibration By Comparison Method

Laboratory Institute for Design of Electrical Measuring Instruments,
Swatantryaveer Tatya Tope Marg, Chunabhatti, Sion, Mumbai,
Maharashtra

Accreditation Standard ISO/IEC 17025: 2017

Certificate Number CC-2287

Page 48 of 48

Validity 03.12.2018 to 30.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
14.	Temperature Indicator of Cold Chambers, Ovens, Incubators (for Non-Medical Applications), Furnaces, Baths etc At Single Position*	(-) 80°C to 660 °C 660 °C to 1000 °C	0.23 °C 1.52 °C	Using RTD/SPRT and Read Out Unit By Comparison method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

§Only in Permanent Laboratory

*Only for Site Calibration

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.