

**Laboratory** Instruments Calibration and Test Centre, 88-C, 5th Cross, Barathi Colony, Peelamedu, Coimbatore, Tamil Nadu

**Accreditation Standard** ISO/IEC 17025: 2005

**Discipline** Electro-Technical Calibration **Issue Date** 16.06.2014

**Certificate Number** C-0810 **Valid Until** 15.06.2016

**Last Amended on** - **Page** 1 of 6

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>1. SOURCE</b> <b>DC VOLTAGE<sup>s</sup></b>	1mV to 10mV	0.58% to 0.068%	Using DMM Fluke 8846A by Comparison Method
	10mV to 100mV	0.068% to 0.011%	
	100mV to 1V	0.011% to 0.0071%	Using Fluke 5500A Calibrator by Direct Method
	1V to 10V	0.0071% to 0.0084%	
10V to 100V	0.008%		
	100V to 1000V	0.0084% to 0.007%	
<b>2. DC CURRENT<sup>s</sup></b>	1mA to 10mA	0.021%	Using Fluke 5500A Calibrator by Direct Method
	10mA to 100mA	0.021%	
	100mA to 1A	0.021% to 0.042%	
	1A to 10A	0.042% to 0.073%	
	10A to 100A	0.073% to 0.10%	Using Current Calibrator 2555A by Direct method
	100A to 900A	0.10% to 2.85%	Using 10 Turn Current Coil by Comparison Method
<b>3. RESISTANCE<sup>s</sup></b>	0.5 $\Omega$ to 1 $\Omega$	0.13% to 0.058%	Using Decade Resistance Box by Direct Method
	1 $\Omega$ to 10 $\Omega$	0.058% to 0.11%	
	10 $\Omega$ to 100 $\Omega$	0.11% to 0.058%	
	100 $\Omega$ to 1k $\Omega$	0.058%	
	1k $\Omega$ to 10k $\Omega$	0.058%	
	10k $\Omega$ to 100k $\Omega$	0.058%	Using Decade Meg Ohm Box by Direct Method
	100k $\Omega$ to 1M $\Omega$	0.058% to 2.31%	
	1M $\Omega$ to 10M $\Omega$	2.31%	
	10M $\Omega$ to 100M $\Omega$	2.31%	
	100M $\Omega$ to 1000M $\Omega$	2.31% to 2.36%	

**Avijit Das**  
Program Manager

**Neeraj Verma**  
Convenor

<b>Laboratory</b>	<b>Instruments Calibration and Test Centre, 88-C, 5th Cross, Barathi Colony, Peelamedu, Coimbatore, Tamil Nadu</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro-Technical Calibration</b>	<b>Issue Date</b>	<b>16.06.2014</b>
<b>Certificate Number</b>	<b>C-0810</b>	<b>Valid Until</b>	<b>15.06.2016</b>
<b>Last Amended on</b>	<b>-</b>	<b>Page</b>	<b>2 of 6</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>4. AC VOLTAGE<sup>s</sup></b>	<b>50Hz</b>		Using Fluke 5500A Calibrator by Direct Method
	1mV to 10mV	2.54% to 0.41%	
	10mV to 100mV	0.41% to 0.098%	
	100mV to 1V	0.098% to 0.043%	
	1V to 10V	0.043% to 0.058%	
	10V to 100V	0.058% to 0.0981%	
	100V to 1000V	0.98% to 0.071%	
	<b>10 kHz</b>		Using Fluke 5500A Calibrator by Direct Method
	1mV to 10mV	2.58% to 0.41%	
	10mV to 100mV	0.41% to 0.085%	
	100mV to 1V	0.085% to 0.043%	
	1V to 10V	0.043% to 0.06%	
	10V to 100V	0.06% to 0.098%	
	<b>1kHz</b>		Using Fluke 5500A Calibrator by Direct Method
	100V to 1000V	0.098% to 0.071%	
<b>5. AC CURRENT<sup>s</sup></b>	<b>50 Hz</b>		Using Fluke 5500A Calibrator by Direct Method
	1mA to 10mA	0.36% to 0.15%	
	10mA to 100mA	0.15%	
	100mA to 1A	0.15% to 0.17%	
	1A to 10A	0.17% to 0.11%	
	10A to 70A	0.11% to 0.85%	Using Current Calibrator 2555A by Direct method
70A to 700A	0.85% to 2.08%	Using 10 Turn Current Coil by Comparison method	

**Avijit Das**  
Program Manager

**Neeraj Verma**  
Convenor

<b>Laboratory</b>	<b>Instruments Calibration and Test Centre, 88-C, 5th Cross, Barathi Colony, Peelamedu, Coimbatore, Tamil Nadu</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro-Technical Calibration</b>	<b>Issue Date</b>	<b>16.06.2014</b>
<b>Certificate Number</b>	<b>C-0810</b>	<b>Valid Until</b>	<b>15.06.2016</b>
<b>Last Amended on</b>	<b>-</b>	<b>Page</b>	<b>3 of 6</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
	<b>1kHz</b> 1mA to 10mA 10mA to 100mA 100mA to 1A 1A to 10A	0.19% to 0.15% 0.15% 0.15% to 0.58% 0.58% to 0.41%	Using Fluke 5500A Calibrator by Direct Method
<b>6. AC POWER<sup>\$</sup> Unity Power Factor</b>	<b>50Hz @ UPF</b> <b>120V to 240V</b> <b>10mA to 10A</b> 120W to 1.2kW 1.2kW to 2.4kW	0.31% to 0.18% 0.18%	Using Fluke 5500A Calibrator by Direct Method
<b>7. FREQUENCY<sup>\$</sup></b>	10Hz to 120Hz	0.45% to 0.67%	Using Fluke 8846A by Comparison Method
	120Hz to 1kHz 1kHz to 10kHz 10kHz to 100kHz	0.67% to 0.068% 0.068% to 0.03% 0.03% to 0.021%	Using Fluke 5500A Calibrator by Direct Method
<b>8. CAPACITANCE<sup>\$</sup></b>	<b>1kHz</b> 1nF to 10nF 10nF to 100nF 100nF to 1 $\mu$ F	1.74% to 0.70% 0.70% to 0.41% 0.41%	Using Fluke 5500A Calibrator by Direct Method
	<b>100 Hz</b> 1 $\mu$ F to 10 $\mu$ F 10 $\mu$ F to 100 $\mu$ F 10 $\mu$ F to 300 $\mu$ F	0.41% to 0.52% 0.52% to 0.7% 0.70% to 0.93%	Using Fluke 5500A Calibrator by Direct Method

<b>Laboratory</b>	<b>Instruments Calibration and Test Centre, 88-C, 5th Cross, Barathi Colony, Peelamedu, Coimbatore, Tamil Nadu</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro-Technical Calibration</b>	<b>Issue Date</b>	<b>16.06.2014</b>
<b>Certificate Number</b>	<b>C-0810</b>	<b>Valid Until</b>	<b>15.06.2016</b>
<b>Last Amended on</b>	<b>-</b>	<b>Page</b>	<b>4 of 6</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>9. TEMPERATURE SIMULATION<sup>\$</sup></b>			
<b>J TYPE</b>	-200° C to 740° C	1.115 to 0.41%	Using Temperature Source CC-01 By Direct Method
<b>K TYPE</b>	-200° C to 1300° C	1.12% to 0.26%	
<b>T TYPE</b>	-200° C to 400° C	1.16% to 0.64%	
<b>R TPE</b>	100° C to 1600° C	3.75 to 0.28%	
<b>S TYPE</b>	250° C to 1600° C	1.5% to 0.3 %	
<b>PT-100 TYPE</b>	100° C to 800° C	0.81% to 0.15%	
<b>10. TIME INTERVAL<sup>\$</sup></b>	9 Sec to 90 Sec 90 Sec to 3600Sec 3600Sec to 9000 Sec	0.70% to 0.15% 0.15% to 0.060% 0.06% to 0.077%	Using Time Totalizer by Direct Method
<b>11. POWER<sup>\$</sup>@ UPF</b>	<b>50Hz</b> 0.11kW to 1.5kW 1.5kW to 3.0kW 0.11kW to 1.5kW 1.5kW to 3.0kW 0.11kW to 1.5kW 1.5kW to 3.0kW	0.75% to 0.31% 0.31% to 1.95% 0.75% to 0.31% 0.31% to 1.95% 0.75% to 0.31% 0.31% to 1.95%	Using 3 Phase Calibration by Direct Method
<b>POWER FACTOR<sup>\$</sup> R,Y,B PHASE</b>	<b>LEAD</b> 0.5 0.8 <b>LAG</b> 0.5 0.8 Unity 1.0	0.30% 0.25% 0.30% 0.25% 0.25%	Using 3 Phase Calibration by Direct Method

<b>Laboratory</b>	<b>Instruments Calibration and Test Centre, 88-C, 5th Cross, Barathi Colony, Peelamedu, Coimbatore, Tamil Nadu</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro-Technical Calibration</b>	<b>Issue Date</b>	<b>16.06.2014</b>
<b>Certificate Number</b>	<b>C-0810</b>	<b>Valid Until</b>	<b>15.06.2016</b>
<b>Last Amended on</b>	<b>-</b>	<b>Page</b>	<b>5 of 6</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>1. <u>MEASURE</u> DC VOLTAGE<sup>s</sup></b>	1mV to 10mV 10mV to 100mV 100mV to 1V 1V to 10V 10V to 100V 100V to 1000V 1kV to 10 kV	1.28% to 0.18% 0.18% to 0.008% 0.008% to 0.005% 0.005% 0.0051% to 0.008% 0.008% to 0.006% 1.44% to 1.52%	Using DMM Fluke 8846A by Direct Method  Using High Voltage Probe & DMM Fluke 8846A by Direct Method
<b>2. DC CURRENT<sup>s</sup></b>	1mA to 10mA 10mA to 100mA 100mA to 1A 1A to 2A 2A to 10A	0.064% to 0.081% 0.081% to 0.065% 0.065% to 0.082% 0.082% to 0.18% 0.18% to 0.20%	Using DMM Fluke 8846A by Direct Method
<b>3. AC VOLTAGE<sup>s</sup></b>	<b>50Hz</b> 100mV to 1V 1V to 10V 10V to 100V 100V to 1000V  >1 kV to 5 kV	0.12% to 0.11% 0.11% 0.11% 0.11%  6.19% to 6.64%	Using DMM Fluke 8846A by Direct Method  Using High Voltage Probe & DMM Fluke 8846A by Direct Method
	<b>1kHz</b> 100mV to 1V 1V to 10V 10V to 100V 100V to 1000V	0.13% to 0.11% 0.11% 0.11% 0.11%	Using DMM Fluke 8846A by Direct Method

<b>Laboratory</b>	<b>Instruments Calibration and Test Centre, 88-C, 5th Cross, Barathi Colony, Peelamedu, Coimbatore, Tamil Nadu</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro-Technical Calibration</b>	<b>Issue Date</b>	<b>16.06.2014</b>
<b>Certificate Number</b>	<b>C-0810</b>	<b>Valid Until</b>	<b>15.06.2016</b>
<b>Last Amended on</b>	<b>-</b>	<b>Page</b>	<b>6 of 6</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>4. AC CURRENT<sup>\$</sup></b>	<b>50Hz</b>		
	1mA to 10mA	0.17% to 0.25%	Using DMM Fluke 8846A by Direct Method
	10mA to 100mA	0.25% to 0.18%	
	100mA to 1A	0.18%	
	1A to 2A	0.185 to 0.34%	
	2A to 10A	0.34% to 0.26%	
	<b>1kHz</b>		
	10mA to 100mA	0.26% to 0.18%	Using DMM Fluke 8846A by Direct Method
	100mA to 1A	0.18% to 0.36%	
	1A to 2A	0.36% to 0.48%	
2A to 10A	0.48% to 0.27%		
<b>5. RESISTANCE<sup>\$</sup></b>	0.5 $\Omega$ to 1 $\Omega$	1.35% to 0.36%	Using DMM 8846A by Direct Method
	1 $\Omega$ to 10 $\Omega$	0.36% to 0.05%	
	10 $\Omega$ to 100 $\Omega$	0.05% to 0.017%	
	100 $\Omega$ to 1k $\Omega$	0.017%	
	1k $\Omega$ to 10k $\Omega$	0.017%	
	10k $\Omega$ to 100k $\Omega$	0.017% to 0.019%	
	100k $\Omega$ to 1M $\Omega$	0.019% to 0.060%	
	1M $\Omega$ to 10M $\Omega$	0.060% to 0.074%	
	10M $\Omega$ to 100M $\Omega$	0.074% to 0.94%	
100M $\Omega$ to 1000M $\Omega$	0.94% to 2.33%		

\* Measurement Capability is expressed as an uncertainty ( $\pm$ ) at a confidence probability of 95%

<sup>\$</sup> Only in Permanent Laboratory