

**Laboratory** Intech Calibrations Private Limited (ICPL), No. 7/9, 28<sup>th</sup> Cross, Avvai Nagar, Lawspet, Puducherry  
**Accreditation Standard** ISO/IEC 17025: 2005  
**Certificate Number** CC-2715 (in lieu of C-1071, C-1072 & C-1073) **Page** 1 of 16  
**Validity** 22.05.2018 to 21.05.2020 **Last Amended on** 28.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b><u>ELECTRO-TECHNICAL CALIBRATION</u></b>				
<b>I.</b>	<b>SOURCE</b>			
1.	DC Voltage <sup>s</sup>	1 mV to 330 mV 330 mV to 3.3 V 3.3 V to 1000 V	0.75 % to 0.008 % 0.008 % to 0.019 % 0.019 % to 0.009 %	Using Multi Product Calibrator Fluke 5502E by Direct Method
2.	DC Current <sup>s</sup>	0.2 mA to 330 mA 330 mA to 10A 10 A to 20 A  20 A to 1000 A	0.42 % to 0.007 % 0.007 % to 0.05 % 0.05 % to 0.35 %  1.2 % to 0.94 %	Using Multi Product Calibrator Fluke 5502E by Direct Method  Using Multi Product Calibrator Fluke 5502E with Current coil by Direct Method
3.	AC Voltage <sup>s</sup>	<b>@45 Hz to 10 kHz</b> 3 mV to 33 V 33 V to 1000 V  <b>@10 kHz to 100 kHz</b> 30 mV to 30 V	0.97 % to 0.19 % 0.19 % to 0.11 %  0.53 % to 0.33 %	Using Multi Product Calibrator Fluke 5502E by Direct Method
4.	AC Current <sup>s</sup>	<b>@45 Hz to 1 kHz</b> 190 $\mu$ A to 20 A  <b>@50 Hz</b> 20 A to 1000 A	0.44 % to 0.27 %  1.45 % to 1.0 %	Using Multi Product Calibrator Fluke 5502E by Direct Method  Using Multi Product Calibrator Fluke 5502E with Current coil by Direct Method

Ram Ashray  
 Convenor

Avijit Das  
 Program Director

**Laboratory** Intech Calibrations Private Limited (ICPL), No. 7/9, 28<sup>th</sup> Cross, Avvai Nagar, Lawspet, Puducherry

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2715 (in lieu of C-1071, C-1072 & C-1073) **Page** 2 of 16

**Validity** 22.05.2018 to 21.05.2020 **Last Amended on** 28.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
5.	DC Resistance <sup>s</sup>	2 $\Omega$ to 30 M $\Omega$ 30 M $\Omega$ to 100 M $\Omega$ 100 M $\Omega$ to 300M $\Omega$ 300 M $\Omega$ to 1 G $\Omega$	0.67 % to 0.14 % 0.14 % to 0.70 % 0.70 % to 0.62 % 0.62% to 1.80%	Using Multi Product Calibrator Fluke 5502E by Direct Method
6.	Capacitance <sup>s</sup>	@1 kHz 350 pF to 300 nF  @100 Hz 300 nF to 30 $\mu$ F	3.95 % to 0.31 %  0.31 % to 0.61 %	Using Multi Product Calibrator Fluke 5502E by Direct Method
7.	Frequency <sup>s</sup>	119 Hz to 100 kHz	0.17 % to 0.08 %	Using Multi Product Calibrator Fluke 5502E by Direct Method
8.	Inductance <sup>s</sup>	1 kHz 100 $\mu$ H to 1 mH 1 mH to 10 H	3.6 % 3.6 %	Using Decade Inductance Box by Direct Method
9.	Temperature Simulation <sup>s</sup> (Controllers/ Indicators/Simulators without Sensors)			
	Thermocouple Types			Using Multi Product Calibrator Fluke 5502E by Direct Method
	K	(-) 200 $^{\circ}$ C to 1370 $^{\circ}$ C	0.5 $^{\circ}$ C	
	J	(-) 200 $^{\circ}$ C to 600 $^{\circ}$ C	0.5 $^{\circ}$ C	
	R	0 $^{\circ}$ C to 1760 $^{\circ}$ C	0.7 $^{\circ}$ C	
	S	0 $^{\circ}$ C to 1760 $^{\circ}$ C	0.7 $^{\circ}$ C	
	T	(-) 150 $^{\circ}$ C to 400 $^{\circ}$ C	0.4 $^{\circ}$ C	
	N	(-) 100 $^{\circ}$ C to 1300 $^{\circ}$ C	0.5 $^{\circ}$ C	
	RTD pt-100	(-) 200 $^{\circ}$ C to 800 $^{\circ}$ C	0.08 $^{\circ}$ C to 0.3 $^{\circ}$ C	
10.	Digital Timer/ Stopwatch <sup>s</sup>	1 s to 24 Hrs	0.16 s to 16.4 s	Using Timer Calibrator (Autonics) by Comparison method

Ram Ashray  
Convenor

Avijit Das  
Program Director

**Laboratory** Intech Calibrations Private Limited (ICPL), No. 7/9, 28<sup>th</sup> Cross, Avvai Nagar, Lawspet, Puducherry

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2715 (in lieu of C-1071, C-1072 & C-1073) **Page** 3 of 16

**Validity** 22.05.2018 to 21.05.2020 **Last Amended on** 28.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
11.	Digital Timer / Stopwatch*	1 s to 24 Hrs	0.16 s to 16.4 s	Using Timer Calibrator (Autonics) by Comparison Method
12.	Temperature Simulation* (Controllers/ Indicators/ Simulators without Sensors)			
	Thermocouple Types			Using Temperature/ Electrical Calibrator by Direct Method
	K	(-) 200 °C to 1370 °C	0.61 °C	
	J	(-) 200 °C to 1200 °C	0.5 °C	
	R	0 to 1760 °C	1.3 °C	
	S	0 to 1760 °C	1.8 °C	
	T	(-) 200 °C to 400 °C	0.82 °C	
	RTD pt-100	(-) 200 °C to 800 °C	0.24 °C to 0.41 °C	
<b>II.</b>	<b>MEASURE</b>			
1.	DC Voltage <sup>s</sup>	1 mV to 10 V 10V to 1000 V	0.01 % to 0.004 % 0.004 % to 0.01 %	Using 6½ Digit DMM Fluke 8846 A by Direct Method
2.	DC Current <sup>s</sup>	100 µA to 1 A 1 A to 10 A	0.1 % 0.1 % to 0.2 %	Using 6½ Digit DMM Fluke 8846 A by Direct Method
3.	AC Voltage <sup>s</sup>	<b>10 Hz to 50 kHz</b> 100 mV to 10 V <b>50 Hz to 10 kHz</b> 10 V to 1000 V <b>10 kHz to 50 kHz</b> 1V to 100 V	1.0 %	Using 6½ Digit DMM Fluke 8846 A by Direct Method
4.	AC Current <sup>s</sup>	<b>1 kHz</b> 100 µA to 10 A	0.53 % to 0.3 %	Using 6½ Digit DMM Fluke 8846 A by Direct Method

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Intech Calibrations Private Limited (ICPL), No. 7/9, 28<sup>th</sup> Cross, Avvai Nagar, Lawspet, Puducherry

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2715 (in lieu of C-1071, C-1072 & C-1073) **Page** 4 of 16

**Validity** 22.05.2018 to 21.05.2020 **Last Amended on** 28.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
5.	DC Resistance <sup>s</sup>	1 $\Omega$ to 1 M $\Omega$ 1 M $\Omega$ to 10 M $\Omega$ 10 M $\Omega$ to 100 M $\Omega$ 100 M $\Omega$ to 1 G $\Omega$	0.45 % to 0.014 % 0.014 % to 0.05 % 0.05 % to 0.94 % 0.94 % to 2.32 %	Using 6½ Digit DMM Fluke 8846 A by Direct Method
6.	Frequency <sup>s</sup>	10 Hz to 1 MHz	0.08 % to 0.02 %	Using 6½ Digit DMM Fluke 8846 A by Direct Method
7.	Temperature Simulation <sup>s</sup> (Controllers/ Indicators/ Simulators without Sensors)			
	Thermocouple types			Using Multi Product Calibrator Fluke 5502E (In Measure Mode) by Direct Method
	K	(-) 190 °C to 1300 °C	0.39 °C	
	J	(-) 200 °C to 1100 °C	0.39 °C	
	R	0 °C to 1700 °C	1.0 °C	
	S	0 °C to 1700 °C	1.35 °C	
	T	(-) 150 °C to 300 °C	0.31 °C	
	N	(-) 100 °C to 1000 °C	0.41 °C	
8.	Temperature Simulation <sup>s</sup> (Controllers/ Indicators/ Simulators without Sensors)			
	RTD pt-100	(-) 200 °C to 600 °C	0.1 °C to 0.26 °C	Using 6½ Digit DMM Fluke 8846 A by Direct Method

Ram Ashray  
Convenor

Avijit Das  
Program Director

**Laboratory** Intech Calibrations Private Limited (ICPL), No. 7/9, 28<sup>th</sup> Cross, Avvai Nagar, Lawspet, Puducherry

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2715 (in lieu of C-1071, C-1072 & C-1073) **Page** 5 of 16

**Validity** 22.05.2018 to 21.05.2020 **Last Amended on** 28.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b><u>MECHANICAL CALIBRATION</u></b>				
<b>I. DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)</b>				
1.	Vernier Caliper <sup>s</sup> L.C.: 0.01 mm	0 to 600 mm	9.25 $\mu$ m	Using Slip Gauge Set Grade '0', Caliper Checker IS 3651 (Part 1 & 2)
2.	External Micrometer <sup>s</sup> L.C.: 0.001 mm	0 to 100 mm	1.30 $\mu$ m	Using Slip Gauge Set Grade '0', Optical Flat & Parallels as per IS 2967
3.	Height Gauge <sup>s</sup> LC.: 0.01 mm	0 to 600 mm	8.0 $\mu$ m	Using Caliper Checker as per IS 2921
4.	Plunger Dial Gauge <sup>s</sup> LC.: 0.001 mm LC.: 0.01 mm	0 to 10 mm 0 to 50 mm	1.1 $\mu$ m 5.82 $\mu$ m	Using ULM as per IS 2092
5.	Lever Dial Gauge <sup>s</sup> LC.: 0.001 mm LC.: 0.01 mm	0 to 0.14 mm 0 to 0.8 mm	1.1 $\mu$ m 5.80 $\mu$ m	Using ULM as per IS 11498
6.	Bore Gauge <sup>s</sup> (Only For Transmission error) LC.: 0.001 mm	0 to 2 mm	1.0 $\mu$ m	Using ULM
7.	Plain Ring Gauge <sup>s</sup>	10 mm to 100 mm	2.26 $\mu$ m	Using ULM & Master Setting Ring as per IS 3485

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Intech Calibrations Private Limited (ICPL), No. 7/9, 28<sup>th</sup> Cross, Avvai Nagar, Lawspet, Puducherry

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2715 (in lieu of C-1071, C-1072 & C-1073) **Page** 6 of 16

**Validity** 22.05.2018 to 21.05.2020 **Last Amended on** 28.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
8.	Coating Thickness Gauge <sup>s</sup> L.C.: 0.0001 mm	0 to 1035 $\mu$ m	2.45 $\mu$ m	Using Master Foils by Comparison Method
9.	Thread Ring Gauge <sup>s</sup>	5 mm to 100 mm	2.26 $\mu$ m	Using ULM & Master Setting Ring as per IS 4218, IS 2334
10.	Plain Plug Gauge <sup>s</sup>	4.80 mm to 100 mm	1.05 $\mu$ m	Using ULM as per IS 3455 & IS 7859
11.	Thread Plug Gauge <sup>s</sup>	14 to 100 mm	1.50 $\mu$ m	Using ULM, Thread Measuring Wire as per IS 2334, IS 4218 (Part I to IV)
12.	Feeler Gauge <sup>s</sup>	Upto 2 mm	0.53 $\mu$ m	Using ULM as per IS 3179
13.	Foils <sup>s</sup>	Upto 1035 $\mu$ m	0.53 $\mu$ m	Using ULM
14.	Measuring Scale <sup>s</sup> L.C.: 0.5 mm	0 to 1000 mm	24.46 $\mu$ m	Using Scale & Tape Calibrator as per IS 1481
15.	Measuring Tape <sup>s</sup> L.C.: 1 mm	0 to 30 m	32 $\sqrt{N}$ $\mu$ m	Using Scale & Tape Calibrator as per IS 1269 (Part 1 & 2)
16.	Depth Micrometer <sup>s</sup> L.C.: 0.001 mm	0 to 25 mm	4.91 $\mu$ m	Using Slip Gauge Set Grade '0' as per BS 6468
17.	Dial Thickness Gauge <sup>s</sup> L.C.: 0.01 mm	0 to 10 mm	5.80 $\mu$ m	Using Slip Gauge Set Grade '0' as per IS 2092

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Intech Calibrations Private Limited (ICPL), No. 7/9, 28<sup>th</sup> Cross, Avvai Nagar, Lawspet, Puducherry

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2715 (in lieu of C-1071, C-1072 & C-1073) **Page** 7 of 16

**Validity** 22.05.2018 to 21.05.2020 **Last Amended on** 28.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
18.	Gap / Snap Gauge <sup>s</sup>	Upto 100 mm	1.02 $\mu$ m	Using Slip Gauge Set Grade '0' as per IS 3455, IS 7859 & IS 3477
19.	Test Sieves <sup>s</sup>	150 $\mu$ m to 10 mm	6.6 $\mu$ m	Using Profile Projector as per IS 460
		10 mm to 100 mm	27.2 $\mu$ m	Using Digital Vernier Caliper as per IS 460
20.	Radius Gauge <sup>s</sup>	1 mm to 25 mm	8.9 $\mu$ m	Using Profile Projector as per IS 5273
21.	Surface Table <sup>s</sup>	Upto 2500 mm x 1600 mm	$1.22 \sqrt{\frac{L+W}{100}} \mu$ m	Using Electronic Level as per IS 7327, IS 2285, IS 12937 (L & M is in mm)
22.	Depth Gauge <sup>s</sup> L.C.: 0.01 mm	0 to 300 mm	12.72 $\mu$ m	Using Slip Gauge Set Grade '0', Caliper Checker as per IS 4213
23.	Thread Measuring Wire <sup>s</sup> (2/3 Wire)	0.17 mm to 6.35 mm	1.00 $\mu$ m	Using ULM as per IS 6311
24.	Setting Rod <sup>s</sup>	25 mm to 100 mm	1.27 $\mu$ m	Using Slip Gauge "0" & Electronic comparator
25.	LVDT Probe with DRO <sup>s</sup> L.C.: 0.001 mm	0 to 100 mm	1.00 $\mu$ m	Using Slip Gauge Set Grade '0' by Comparison Method
26.	Extensometer <sup>s</sup>	0 to 3 mm	4.0 $\mu$ m	Using Electronic Probe as per ASTM E83-10a, IS 12872

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Intech Calibrations Private Limited (ICPL), No. 7/9, 28<sup>th</sup> Cross, Avvai Nagar, Lawspet, Puducherry

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2715 (in lieu of C-1071, C-1072 & C-1073) **Page** 8 of 16

**Validity** 22.05.2018 to 21.05.2020 **Last Amended on** 28.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>II. DIMENSION (PRECISION INSTRUMENTS)</b>				
1.	Universal Length Measuring Machine <sup>§</sup> L.C.: 0.0001 mm	0 to 100 mm	0.86 $\mu$ m	Using Slip Gauge Set Grade '0' by Comparison Method
<b>III. ACOUSTICS</b>				
1.	Sound Level Meter <sup>§</sup>	94 dB & 114 dB	0.77 dB	Using Sound Level Calibrator
<b>IV. ACCELERATION &amp; SPEED</b>				
1.	RPM <sup>§</sup> (Contact Type) Tachometers	60 rpm to 3000 rpm	0.13 %	Using Digital Tachometer By Comparison method
2.	RPM <sup>§</sup> (Non Contact Type) Tachometers	120 rpm to 36000 rpm	2.31 %	Using Digital Laser Tachometer By Comparison method
<b>V. PRESSURE INDICATING DEVICES</b>				
1.	Vacuum, Digital And Analog, Gauges And Transmitters <sup>#</sup>	(-) 0.90 bar to 0 bar	1.01%	Using Twin mode Druck, Pneumatic Pump & Digital Indicator by Comparison method
2.	Pneumatic Pressure Digital And Analog, Gauges And Transmitters / Hydraulic Pressure Digital And Analog, Gauges And Transmitters <sup>#</sup>	0 bar to 35 bar  0 bar to 200bar 0 bar to 700 bar	0.16%  0.13% 0.12%	Using Pneumatic Pump and Digital Pressure Gauge by Comparison method  Using Hydraulic Pressure Pump and Digital Pressure Gauge by Comparison method

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Director



**Laboratory** Intech Calibrations Private Limited (ICPL), No. 7/9, 28<sup>th</sup> Cross, Avvai Nagar, Lawspet, Puducherry

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2715 (in lieu of C-1071, C-1072 & C-1073) **Page** 9 of 16

**Validity** 22.05.2018 to 21.05.2020 **Last Amended on** 28.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
3.	Low Pressure Indicator, Magnehelic Gauge, Manometers, BP Apparatus (Electronic, Mercury, Analog) <sup>#</sup>	Upto 10000 mmwc	0.06 %	Using Screw Pump with Pressure Gauge by Comparison method
4.	Pressure Switch <sup>#</sup>	0 bar to 35 bar	3.08 %	Using Digital Pressure Calibrator by Comparison Method
<b>VI.</b>	<b>WEIGHTS</b>			
1.	Calibration of M1 Accuracy Class and coarser <sup>\$</sup>	1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg	0.1 mg 0.1 mg 0.1 mg 0.1 mg 0.2 mg 0.2 mg 0.2 mg 0.2 mg	Using E1 Class Standard Weights and Electronic Weighing Balance (d: 0.1mg, Cap:220g) As per OIML R111-1 : 2004 ABBA Cycle Substitution Method
	Calibration of F2 Accuracy Class and Coarser <sup>\$</sup>	500 mg 1 g 2 g 5 g  10 g 20 g 50 g 100 g 200 g	0.2 mg 0.2 mg 0.2 mg 0.2 mg  0.2 mg 0.2 mg 0.2 mg 0.2 mg 0.2 mg	Using E1 Class Standard Weights and Electronic Weighing Balance (d:0.1mg, Cap:220g) As per OIML R111-1 : 2004 ABBA Cycle Substitution Method  Using E2 Accuracy Class Standard Weights and Balance d: 0.1 mg Cap. 220 g as per OIML R111-1:2004 ABBA Cycle Substitution Method

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Intech Calibrations Private Limited (ICPL), No. 7/9, 28<sup>th</sup> Cross, Avvai Nagar, Lawspet, Puducherry

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2715 (in lieu of C-1071, C-1072 & C-1073) **Page** 10 of 16

**Validity** 22.05.2018 to 21.05.2020 **Last Amended on** 28.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
	Calibration of M1 Accuracy Class and Coarser <sup>\$</sup>	500 g 1 kg 2 kg 5 kg	10 mg 20 mg 20 mg 20 mg	Using E2 Accuracy Class Standard Weights and Balance d: 0.01 g Cap. 5 kg As per OIML R111-1 : 2004 ABBA Cycle Substitution Method
	Calibration of M2 Accuracy Class and Coarser <sup>\$</sup>	10 kg 20 kg 50 kg	1 g 1 g 2 g	Using F1 Accuracy Class Standard Weights and Balance d: 1 g Cap. 50 kg As per OIML R111-1 : 2004 ABBA Cycle Substitution Method
<b>VII.</b>	<b>WEIGHING SCALE AND BALANCE</b>			
1.	Weighing Balance / Scale Weighing Balance* Class I (d: 0.001 mg)	0 to 10 g	0.010 mg	Using E1 Accuracy Class Standard Weights As per OIML R76-1 :2006
2.	Weighing Balance* Class I (d: 0.1 mg)	0 to 200 g	0.1 mg	Using E2 Accuracy Class Standard Weights As per OIML R76-1 :2006
3.	Weighing Balance* Class I (d: 0.01 g)	0 to 5 kg	10 mg	Using E2 Accuracy Class Standard Weights As per OIML R76-1 :2006

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Intech Calibrations Private Limited (ICPL), No. 7/9, 28<sup>th</sup> Cross, Avvai Nagar, Lawspet, Puducherry

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2715 (in lieu of C-1071, C-1072 & C-1073) **Page** 11 of 16

**Validity** 22.05.2018 to 21.05.2020 **Last Amended on** 28.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
4.	Weighing Balance* Class I (d: 1 g)	0 to 50 kg	1 g	Using E2 & F1 Accuracy Class Standard Weights As per OIML R76-1 :2006
5.	Weighing Balance* Class I (d: 10 g)	0 to 300 kg	10 g	Using E2, F1& M1 Accuracy Class Standard Weights as per OIML R76-1 :2006
<b>VIII.</b>	<b>VOLUME</b>			
1.	Glass Pipette <sup>s</sup>	1 ml to 10 ml >10 ml to 25 ml >25 ml to 50 ml	0.004 ml 0.008 ml 0.012 ml	Using Weighing Balance with d: 0.1 mg Cap.220g and distilled water of known density as per IS/ISO 4787 : 2010
2.	Burette <sup>s</sup>	1 ml to 100 ml	0.12 ml	Using Weighing Balance with d:0.1 mg Cap. 220g and distilled water of known density as per IS/ISO4787 : 2010
3.	Volumetric Flask <sup>s</sup>	5 ml to 100 ml >100 ml to 2000 ml	0.04 ml 0.23 ml	Using Weighing Balance with d:0.1 mg Cap. 220g and distilled water of known density as per IS/ISO4787: 2010
4.	Measuring Cylinder <sup>s</sup>	5 ml to 100 ml	0.2 ml	Using Weighing Balance with d: 0.1mg Cap.220 g and distilled water of known density As per IS/ISO4787 : 2010

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Intech Calibrations Private Limited (ICPL), No. 7/9, 28<sup>th</sup> Cross, Avvai Nagar, Lawspet, Puducherry

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2715 (in lieu of C-1071, C-1072 & C-1073) **Page** 12 of 16

**Validity** 22.05.2018 to 21.05.2020 **Last Amended on** 28.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
		>100 ml to 2000 ml	0.7 ml	Using Weighing Balance with d:10mg Cap.5 kg and distilled water of known density as per IS/ISO4787 : 2010
<b>IX.</b>	<b>UTM, TENSION CREEP AND TORSION TESTING MACHINE</b>			
1.	Universal Testing Machine* (Compression and Tension)	200 N to 300 kN	0.96 %	Using Load Cell/Proving Ring with display unit As per IS 1828(Part-I), ISO 7500:(Part I) ASTM E74
2.	Universal Testing Machine* (Compression)	300 kN to 3000 kN	0.66 %	Using Load Cell/Proving Ring with display unit As Per IS 1828 (Part-I), ISO 7500:(Part - 1) ASTM E74
3.	Verification of Displacement Measuring System in Material Testing Machine*	0 to 1000 mm	0.7 mm	Using Height Gauge As per ASTM 2309 -05
<b>X.</b>	<b>HARDNESS TESTING MACHINES</b>			
1.	Verification of Rockwell Hardness Testing Machine*	HRA HRB HRC	1.50 HRA 2.00 HRB 1.86 HRC	Using Standard Hardness Blocks HRA,HRB & HRC as per IS:1586:2012
2.	Verification of Brinell Hardness Testing Machine*	HBW 2.5/187.5 HBW 5/750 HBW 10/3000	3 % 3 % 3 %	Using Standard Hardness Blocks HBW 2.5/ 187.5, HBW 5/750,

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Intech Calibrations Private Limited (ICPL), No. 7/9, 28<sup>th</sup> Cross, Avvai Nagar, Lawspet, Puducherry

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2715 (in lieu of C-1071, C-1072 & C-1073)

**Page** 13 of 16

**Validity** 22.05.2018 to 21.05.2020

**Last Amended on** 28.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
				HBW 10/3000 as per IS:1500:2013
3.	Verification of Vickers Hardness Testing Machine*	HV 5 HV 10 HV 30	3 % 3 % 3 %	Using Standard Hardness Blocks HV 5, HV 10 & HV 30 as per IS:1500:2013
<b>XI.</b>	<b>IMPACT TESTING MACHINES</b>			
1.	Verification of Charpy/ Izod Impact Testing Machine*	300 J	0.60 %	Load Cell, Linear Measuring Instruments & Gauges as per IS 3766, ISO 148 -2 & ASTM E23
<b>XII.</b>	<b>TORQUE GENERATING DEVICES</b>			
1.	Torque Wrench* (Type II-Class A, B & C)	25 Nm to 135 Nm	2.20 %	Torque Transducer with Display Unit with Uncertainty of 0.33 % as per ISO 6789

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Intech Calibrations Private Limited (ICPL), No. 7/9, 28<sup>th</sup> Cross, Avvai Nagar, Lawspet, Puducherry

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2715 (in lieu of C-1071, C-1072 & C-1073) **Page** 14 of 16

**Validity** 22.05.2018 to 21.05.2020 **Last Amended on** 28.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b><u>THERMAL CALIBRATION</u></b>				
<b>I.</b>	<b>TEMPERATURE</b>			
1.	Liquid in Glass Thermometer <sup>§</sup>	(-) 70 °C to 200 °C	1.25 °C	Using PRT Sensor with readout using Alcohol & Oil Bath As per ITS – 90 Scale by Comparison Method
2.	RTD's With or Without Temperature Indicators/ Controllers / Data Loggers /Bath's <sup>§</sup>	(-) 70 °C to 600 °C	0.6 °C	Using PRT Sensor with readout using, Oil Bath & Dry Block Calibrator by Comparison method
3.	Temperature Indicator / Thermostat/Digital Thermometer, Pen Type Thermometer <sup>§</sup>	(-) 40 °C to 300 °C	0.50 °C	Using PRT Sensor with readout using & Dry Block Calibrator by Comparison method
4.	Thermocouple With or Without Temperature Indicators / Controllers/ Data Loggers /Bath's <sup>§</sup>	0 °C to 1200 °C	1.7 °C to 2.12 °C	Using S Type T/C With Thermometer Readout Using Dry block Bath by Comparison method
5.	Temperature Switch/ Temperature Gauge <sup>§</sup>	(-) 40 °C to 600 °C	1.25 °C	Using PRT Sensor with Readout Using Dry Block Calibrator by Comparison method

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Intech Calibrations Private Limited (ICPL), No. 7/9, 28<sup>th</sup> Cross, Avvai Nagar, Lawspet, Puducherry

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2715 (in lieu of C-1071, C-1072 & C-1073) **Page** 15 of 16

**Validity** 22.05.2018 to 21.05.2020 **Last Amended on** 28.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
6.	Temperature / Humidity Indicators With Probe or Inbuilt Sensor <sup>s</sup>	15 °C to 45 °C @ 50 % RH 30 % RH to 85 % RH @ 25 °C	1.35 °C 3.25 % RH	Using Temperature / Humidity indicators With Probe Using Humidity Chamber by Comparison method
7.	RTD's With or Without Temperature Indicators / Controllers / Data Loggers <sup>*</sup>	0°C to 600 °C	0.85 °C	Using PRT Sensor with readout using Dry Block Calibrator by Comparison method
8.	Thermocouple With or Without Temperature Indicators/ Controllers / Data Loggers <sup>*</sup>	0 to 1200 °C	2.2 °C	Using S Type T/C With Thermometer Readout Using Dry block Calibrator by Comparison method
9.	Calibration of Ovens, Furnace, Chambers (Mapping) <sup>*</sup>	(-) 70 °C to 250 °C 250 °C to 1200 °C	2.1 °C 8.0 °C	RTD With Indicator (9 Nos) by multipoint calibration N- Type Thermocouple With Indicator (9 Nos) by Multipoint Calibration
10.	Humidity Indicators of Chamber, Environmental Chamber, Climatic Chamber <sup>*</sup>	30 % RH to 85 % RH @ 25 °C	4.2 %	Using Humidity indicators With Probe in Single Position Calibration (Measured Location in DUC)
11.	Thermostat/ Temperature Indicator with Sensor/ Temperature Switch <sup>*</sup>	(-) 40 °C to 250 °C	0.6 °C	Using PRT Sensor with readout using Dry Block Calibrator by Comparison method

**Ram Ashray**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Intech Calibrations Private Limited (ICPL), No. 7/9, 28<sup>th</sup> Cross, Avvai Nagar, Lawspet, Puducherry

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2715 (in lieu of C-1071, C-1072 & C-1073) **Page** 16 of 16

**Validity** 22.05.2018 to 21.05.2020 **Last Amended on** 28.05.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
12.	Temperature Gauge*	(-) 40 °C to 600 °C	1.5 °C	Using PRT Sensor with readout using Dry Block Calibrator by Comparison method
13.	Environmental Chamber, Cold Chamber, Freezer, Incubator, Autoclave, Oven, Water Bath, Furnace (Single Point Calibration)*	(-) 70 °C to 300 °C 0 to 1200 °C	1 °C 6.5 °C	Using RTD With indicator in Single Position Using N – Type With Indicator

\* Measurement Capability is expressed as an uncertainty ( $\pm$ ) at a confidence probability of 95%

\$ Only in Permanent Laboratory

\* Only for Site Calibration

# The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

---

Ram Ashray  
Convenor

---

Avijit Das  
Program Director