Laboratory Krutam Technosolutions Pvt. Ltd., 64 B, GIDC, Makarpura, Vadodara,

Gujarat

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2671 Page 1 of 2

Validity 18.05.2018 to 17.05.2020 Last Amended on -

SI.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks			
MECHANICAL CALIBRATION							
I.	UTM, TENSION CREE						
1.	Uniaxial Testing Machine (Compression)*	1 kN to 1000 kN	0.53 %	Using Force Proving Instrument of Class 1 or better as per IS 1828 (Part 1) : 2015			
2.	Uniaxial Testing Machine (Tension) *	20 kN to 1000 kN	0.53 %	Using Force Proving Instrument of Class 1 or better as per ASTM E4-16			
		0.5 kN to 50 kN	0.80 %	Using Force Proving Instrument of Class 1 or better as per IS 1828 (Part 1) : 2015			
II.	HARDNESS TESTING MACHINES						
1.	Rockwell Hardness Testing Machine [♣]	Total Force at 588.4 N to 1471 N (60 kgf to 150 kgf)	0.61 %	Using Force Proving Instruments of Class 1/ Class A or better (Load Cells with Indicator) as per Based on IS 1586-2:2012			
2.	Brinell Hardness Testing Machine*	612.9 N to 30 kN (62.5 kgf to 3000 kgf)	0.50 %	Using Force Proving Instruments of Class 1/ Class A or better (Load Cells with Indicator) as per Based on IS 1500-2:2013			

Ashish Kakran Convenor Avijit Das Program Director Laboratory Krutam Technosolutions Pvt. Ltd., 64 B, GIDC, Makarpura, Vadodara,

Gujarat

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2671 Page 2 of 2

Validity 18.05.2018 to 17.05.2020 Last Amended on -

SI.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
3.	Rockwell Hardness Testing Machine (Indirect Verification) *	HRA HRBW HRC	1.56 HRA 2.49 HRBW 2.05 HRC	Using Standard Hardness Block as per IS 1586-2:2012
4.	Superficial Rockwell Hardness Testing Machine (Indirect Verification) *	HR15N HR30N HR45N HR15TW HR30TW HR45TW	1.52 HR15N 1.43 HR30N 1.47 HR45N 1.85 HR15TW 1.72 HR30TW 2.41 HR45TW	Using Standard Hardness Block as per IS 1586-2:2012
5.	Brinell Hardness Testing Machine [*]	HBW 2.5 / 62.5 HBW 2.5 / 187.5 HRW 5/750 HBW 10/3000	9.04 % 6.27 % 2.86 % 2.08 %	Using Standard Hardness Block as per IS 1500-2:2013
6.	Vickers Hardness Testing Machine (Indirect Verification) *	HV 5 HV 10 HV 30	2.40 % 2.40 % 2.40 %	Using Standard Hardness Block as per IS 1501-2:2013

^{*} Measurement Capability is expressed as an uncertainty (±) at a confidence probability of 95%

Ashish Kakran Convenor

^{*}Only for Site Calibration