

Laboratory

KVTEK Power Systems Private Limited, Plot No. 283-286, Sector- 8,
IMT Manesar, Gurgaon, Haryana

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2968

Page 1 of 3

Validity 14.03.2019 to 13.03.2021

Last Amended on -

"In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020"

	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>ELECTRO TECHNICAL CALIBRATION</u>				
I.	SOURCE			
1.	PD Amplifier [#] Frequency Response	1 V 30kHz to 1 MHz	0.1 %	Using Signal Generator 3205 A by Direct Method
2.	Impulse Analyzer [#] Magnitude Rise Time Fall Time	\pm (100V to 1000V) 0.84 μ s 60.0 μ s	3.3 % 5.2% 5.2%	Using Impulse Calibrator by Direct Method
3.	Current Transformer [#] Ratio Error Phase Error	120% to 20% 5% 120% to 20% 5%	0.016 % 0.020 % 0.67 min 0.79 min	Using Current Transformer with Automated Instrument Transformer Test Set 5A-2000A by Direct Method
4.	EVD/PT [#] Ratio Error Phase Error	1kV to 100 kV 1kV to 100 kV	0.09 % 0.67 min.	Using Capacitor with Voltage Channel and Automated Instrument Transformer Test Set by Direct Method
5.	Automated Instrument Transformer Test Set [#] (CT Mode) Ratio Error Phase error	120% to 20% 5% 120% to 20% 5%	0.016% 0.020% 0.67 min. 0.78 min.	Using Current Transformer and Automated Instrument Transformer Test Set by Comparison Method

Shally Sharma
Convenor

Battal Singh
Program Manager

Laboratory

KVTEK Power Systems Private Limited, Plot No. 283-286, Sector- 8,
IMT Manesar, Gurgaon, Haryana

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2968

Page 2 of 3

Validity 14.03.2019 to 13.03.2021

Last Amended on -

"In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020"

	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	(PT Mode) Ratio Error Phase error	1 kV to 100 kV	0.022 % 0.67 min.	Using Voltage Channel and Automated Instrument Transformer Test Set by Comparison Method
6.	Resistance [#]	1 m Ω to 100 Ω	2.3 %	Using High Current Resistance Meter by V/I Method
II.	MEASURE			
1.	AC High Voltage [#]	50 Hz 1kV to 200 kV	1.3%	Using Kilovolt meter with Divider K-21, 011 by Comparison Method
2.	DC High Voltage [#]	1 kV to 100 kV	1.3%	Using Kilovolt meter with Divider K-21, 011 by Comparison Method
3.	AC Current [#]	5 A to 2000 A	0.2%	Using CT with Power Analyzer Infratec 106A by Direct Method
4.	PD Calibrator [#] (1 pC to 1000 pC) a. Amplitude b. Rise Time c. Pulse Frequency d. Duty Cycle	10 mV to 10 V 10ns to 500 ms 50 Hz 50 %	1.2 % 0.12 % 0.08 % 0.1 %	Using Digital Oscilloscope by Direct Method
5.	Capacitance and Tan Delta [#] Capacitance Tan Delta	100 pF to 50 μ F 0.00001 to 0.001	0.17% 0.00008	Using Capacitance Inductance & Tan Delta Test Set/ C & Tan Delta Box Capacitor by Direct / Comparison Method

Shally Sharma
Convenor

Battal Singh
Program Manager

Laboratory KVTEK Power Systems Private Limited, Plot No. 283-286, Sector- 8, IMT Manesar, Gurgaon, Haryana

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2968

Page 3 of 3

Validity 14.03.2019 to 13.03.2021

Last Amended on -

"In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020"

	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
6.	Current Transformer [#] Ratio Error Phase Error	120% to 20% 5% 120% to 20% 5%	0.016 % 0.020 % 0.67 min 0.79 min	Using Current Transformer with AITTS 5A-2000A by Direct Method
7.	Impulse Calibrator [#] Magnitude Rise Time Fall Time	50 Hz \pm (100V to 1000V) 0.84 μ s 60.0 μ s	1.2 % 2.39 % 2.33 %	Using Digital Oscilloscope by Direct Method
8.	Impulse Voltage Divider [#] a. Step response Oscilloscope b. Range for Capacitance(HV/LV) (Impulse Divider Ratio)	50 Hz \geq 20 ns 100pF to 1000 pF (100 to 60000)	2.9 % 0.31 %	Using Digital Oscilloscope and Capacitance & Tan Delta test kit with high and low capacitance measurement by Direct Method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Shally Sharma
Convenor

Battal Singh
Program Manager