

**Laboratory**                      **Lab Equipments Calibration Centre, 1st Floor, B-XXX,1720/B, Focal Point, Ludhiana, Punjab**

**Accreditation Standard**    **ISO/IEC 17025: 2005**

Page 1 of 6

**Last Amended on --**

***“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”***

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b><u>ELECTRO TECHNICAL CALIBRATION</u></b>				
<b>I.</b>	<b>SOURCE</b>			
<b>1.</b>	DC Resistance <sup>#</sup> 4 wire	1 $\Omega$ to 10 $\Omega$ 10 $\Omega$ to 100 $\Omega$ 100 $\Omega$ to 1000 $\Omega$ 1 k $\Omega$ to 10 k $\Omega$ 10 k $\Omega$ to 100 k $\Omega$ 100 k $\Omega$ to 1000 k $\Omega$ 1 M $\Omega$ to 10 M $\Omega$ 10 M $\Omega$ to 100 M $\Omega$ 100 M $\Omega$ to 1 G $\Omega$	0.1 $\Omega$ to 1.2 $\Omega$ 1.2 $\Omega$ to 0.14 $\Omega$ 0.14 $\Omega$ to 0.002 k $\Omega$ 0.013 k $\Omega$ to 0.13 k $\Omega$ 0.13 k $\Omega$ to 0.0013 M $\Omega$ 1.3 k $\Omega$ to 13 k $\Omega$ 13 k $\Omega$ to 580 k $\Omega$ 580 k $\Omega$ to 4640 k $\Omega$	Using Standard Decade Resistance Box
<b>2.</b>	Temperature Simulation <sup>#</sup>			
	Thermocouple 'J' Type 'K' Type 'R' Type 'S' Type	50 $^{\circ}$ C to 750 $^{\circ}$ C 50 $^{\circ}$ C to 1300 $^{\circ}$ C 200 $^{\circ}$ C to 1700 $^{\circ}$ C 200 $^{\circ}$ C to 1700 $^{\circ}$ C	0.2 $^{\circ}$ C to 2.9 $^{\circ}$ C 0.8 $^{\circ}$ C to 5.1 $^{\circ}$ C 1.2 $^{\circ}$ C to 5.4 $^{\circ}$ C 1.5 $^{\circ}$ C to 6.3 $^{\circ}$ C	Using Universal Calibrator Radix by Simulation Method
<b>3.</b>	Frequency <sup>#</sup>	50 Hz to 500 Hz	0.0239 Hz to 0.312 Hz	Using Multifunction Calibrator Zeal By Direct Method
<b>II.</b>	<b>MEASURE</b>			
<b>1.</b>	DC Voltage <sup>#</sup>	5 mV to 180 mV 180 mV to 1 V 1 V to 1000 V	0.06 mV to 0.11mV 0.11 mV to 0.001 V 0.001 V to 0.7 V	Using Digital Multimeter 6 ½ Fluke 8846A By Direct/ Comparison Method

**Shally Sharma**  
**Convenor**

**Avijit Das**  
**Program Manager**

**Laboratory**                      **Lab Equipments Calibration Centre, 1st Floor, B-XXX,1720/B, Focal Point, Ludhiana, Punjab**

**Lab Equipments Calibration Centre, 1st Floor, B-XXX,1720/B, Focal Point, Ludhiana, Punjab**

Accreditation Standard	ISO/IEC 17025: 2005		
Certificate Number	CC-2921	Page	2 of 6

Page 2 of 6

**Last Amended on --**

***“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”***

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
2.	AC Voltage <sup>#</sup>	<b>50 Hz</b> 10 mV to 1V 1 V to 1000 V	1.63 mV to 0.001 V 0.001 V to 1.25 V	Using Digital Multimeter 6 ½ Fluke 8846A By Direct/ Comparison Method
3.	DC Current <sup>#</sup>	100 $\mu$ A to 1000 $\mu$ A 1 mA to 100 mA 100 mA to 1 A 1 A to 10 A	0.12 $\mu$ A to 0.006 mA 0.006 mA to 0.11 mA 0.11 mA to 0.009 A 0.009 A to 0.11 A	Using Digital Multimeter 6 ½ Fluke 8846A By Direct/ Comparison Method
4.	AC Current <sup>#</sup>	<b>50 Hz</b> 100 $\mu$ A to 1 mA 1 mA to 100 mA 100 mA to 1 A 1 A to 10 A	5.8 $\mu$ A to 0.02 mA 0.2 mA to 0.19 mA 0.19 mA to 0.02 A 0.02 A to 0.03 A	Using Digital Multimeter 6 ½ Fluke 8846A By Direct/ Comparison Method
5.	DC Resistance <sup>#</sup>	1 $\Omega$ to 1000 $\Omega$ 1 k $\Omega$ to 1000 k $\Omega$ 1 M $\Omega$ to 100 M $\Omega$ 100M $\Omega$ to 1 G $\Omega$	0.001 $\Omega$ to 0.13 $\Omega$ 0.13 $\Omega$ to 0.13 k $\Omega$ 0.13 k $\Omega$ to 0.94 k $\Omega$ 0.94 k $\Omega$ to 2.2 M $\Omega$	Using Digital Multimeter 6 ½ Fluke 8846A By Direct Method

**Shally Sharma**  
Convenor

**Avijit Das**  
**Program Manager**

**Laboratory**                      **Lab Equipments Calibration Centre, 1st Floor, B-XXX,1720/B, Focal Point, Ludhiana, Punjab**

**Accreditation Standard**    **ISO/IEC 17025: 2005**

Page 3 of 6

**Last Amended on --**

***“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”***

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b><u>MECHANICAL CALIBRATION</u></b>				
<b>I.</b>	<b>DIMENSION ( BASIC MEASURING INSTRUMENT, GAUGES ETC.)</b>			
<b>1.</b>	Vernier Caliper <sup>\$</sup> L.C.: 0.01 mm <sup>φ</sup>	0 to 300 mm 0 to 600 mm	14.50 $\mu$ m 21.00 $\mu$ m	Using Caliper Checker, Slip Gauge Set
<b>2.</b>	External Micrometer <sup>\$</sup> L.C. : 0.001 mm	0 to 100 mm	5.9 $\mu$ m	Using Slip Gauge Set
<b>3.</b>	Plunger Type Dial Gauge <sup>\$</sup> L.C. : 0.01 mm	0 to 10 mm	7.0 $\mu$ m	Using Slip Gauge Set & Comparator Stand
<b>4.</b>	Height Gauge <sup>\$</sup> L.C. : 0.02 mm	0 to 600 mm	17.80 $\mu$ m	Using Caliper Checker & Puppy Dial
<b>5.</b>	Dial Thickness Gauge <sup>\$</sup> L.C. : 0.01 mm	0 to 25 mm	9.0 $\mu$ m	Using Slip Gauge Set
<b>6.</b>	Feeler Gauge <sup>\$</sup>	0 to 1 mm	3.10 $\mu$ m	Using Digital Micrometer
<b>7.</b>	Depth Gauge Vernier <sup>\$</sup> L.C.: 0.02 mm	0 to 200 mm	18.10 $\mu$ m	Using Slip Gauge Set
<b>8.</b>	Coating Thickness Gauge <sup>\$</sup>	0 to 800 $\mu$ m	13.10 $\mu$ m	Using Standard Thickness Foils

**Avijit Das**  
**Program Manager**

**Avijit Das**  
**Program Manager**

**Laboratory**                      **Lab Equipments Calibration Centre, 1st Floor, B-XXX,1720/B, Focal Point, Ludhiana, Punjab**

**Accreditation Standard**    **ISO/IEC 17025: 2005**

**Certificate Number**        **CC-2921**                                      **Page**                      **5 of 6**

**Validity**                      **01.01.2019 to 31.12.2020**                      **Last Amended on --**

*"In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020"*

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b><u>THERMAL CALIBRATION</u></b>				
<b>I.</b>	<b>TEMPERATURE</b>			
<b>1.</b>	RTD, Thermocouple With or without Temperature Indicator/Data Logger/Recorder, Temperature Transmitter, Digital Thermometer <sup>#</sup>	(-)15 °C to 100 °C	0.63 °C	Using RTD Sensor With Temperature Indicator, DMM & Negative Bath By Comparison Method
<b>2.</b>	RTD, Thermocouple With or without Temperature Indicator/Data Logger/Recorder, Temperature Transmitter, Digital Thermometer <sup>#</sup>	100 °C to 250 °C	1.4 °C	Using S-Type Thermocouple with Temp. Indicator, RTD With Temperature Indicator With Oil Bath By Comparison Method
<b>3.</b>	Liquid-in-Glass Thermometer <sup>#</sup>	50 °C to 100 °C 100 °C to 250 °C	0.78 °C 1.4 °C	Using S-Type Thermocouple with Temp. Indicator, RTD with Temperature Indicator With Oil Bath By Comparison Method

---

**Shally Sharma**  
Convenor

---

**Avijit Das**  
Program Manager

**Avijit Das**  
**Program Manager**