Laboratory	AE Low Tension Meter Testing Laboratory, MPPKVV Co. Ltd., Nayagaon, Jabalpur, Madhya Pradesh		
Accreditation Standard	ISO/IEC 17025: 2005		
Certificate Number	CC-2818	Page	1 of 1
Validity	30.08.2018 to 29.08.2020	Last Amended on	ı -

SI.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks			
ELECTRO TECHNICAL CALIBRATION							
l.	MEASURE						
1.	Active Mode 1 Phase 2 Wire/	40 V to 240 V 0.5 A to 40 A UPF, 0.8 lead,0.5 lag, 50Hz 20 W to 28.8 kW	0.6 %	Using ESS Make- KOCOS METES 320 with Test Bench INTRAX By Comparison Method			

 * Measurement Capability is expressed as an uncertainty (±) at a confidence probability of 95% $^{\$}$ Only in Permanent Laboratory