

**Laboratory** Mayuresh Associates, 4080, 4078, 3077A, Bhandup Industrial Estate,  
Pannalal Compound, LBS Marg, Bhandup West, Mumbai,  
Maharashtra

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2584

**Page** 1 of 10

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on** 25.07.2018

| Sl.                                         | Quantity Measured / Instrument | Range/Frequency                                    | *Calibration Measurement Capability ( $\pm$ ) | Remarks                                                                                              |
|---------------------------------------------|--------------------------------|----------------------------------------------------|-----------------------------------------------|------------------------------------------------------------------------------------------------------|
| <b><u>ELECTRO-TECHNICAL CALIBRATION</u></b> |                                |                                                    |                                               |                                                                                                      |
| <b>I.</b>                                   | <b>MEASURE</b>                 |                                                    |                                               |                                                                                                      |
| <b>1.</b>                                   | DC Voltage <sup>\$</sup>       | 1 mV to 100 mV<br>100 mV to 1000 V                 | 1.1 % to 0.1 %<br>0.1 % to 0.05 %             | Using Fluke Source with<br>5 ½ DMM Keysight<br>34450A by Direct /<br>Comparison Method               |
| <b>2.</b>                                   | DC Current <sup>\$</sup>       | 10 uA to 100 mA<br>100 mA to 2 A                   | 0.8 % to 0.2 %<br>0.2 % to 0.5 %              | Using Fluke Source with<br>5 ½ DMM Keysight<br>34450A by Direct /<br>Comparison Method               |
|                                             |                                | 10 uA to 100 mA<br>100 mA to 10 A                  | 0.8 % to 0.2 %<br>0.2 % to 0.5 %              | Using 5 ½ DMM<br>Keysight 34450A by<br>Direct Method                                                 |
| <b>3.</b>                                   | Resistance <sup>\$</sup>       | 0.001 $\Omega$ to 1 $\Omega$                       | 0.7 % to 0.5 %                                | Using V/I Method                                                                                     |
|                                             |                                | 1 $\Omega$ to 10 M $\Omega$                        | 1.37 % to 0.2 %                               | Using Fluke Source with<br>5 ½ DMM Keysight<br>34450A and HP 3478A<br>by Direct/Comparison<br>Method |
|                                             |                                | 1M $\Omega$ to 1 G $\Omega$                        | 0.2 % to 4.65 %                               | Using HP 3456A DMM<br>Direct Method                                                                  |
| <b>4.</b>                                   | AC Voltage <sup>\$</sup>       | <b>1 kHz</b><br>10 mV to 100 mV<br>100 mV to 750 V | 2 % to 0.4 %<br>0.4 %                         | Using Fluke Source with<br>5 ½ DMM Keysight<br>34450A Direct /<br>Comparison                         |

**Vishal Shukla**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Mayuresh Associates, 4080, 4078, 3077A, Bhandup Industrial Estate,  
Pannalal Compound, LBS Marg, Bhandup West, Mumbai,  
Maharashtra

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2584

**Page** 2 of 10

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on** 25.07.2018

| Sl. | Quantity Measured / Instrument                                                 | Range/Frequency                       | *Calibration Measurement Capability ( $\pm$ ) | Remarks                                                                     |
|-----|--------------------------------------------------------------------------------|---------------------------------------|-----------------------------------------------|-----------------------------------------------------------------------------|
| 5.  | AC Current <sup>\$</sup>                                                       | 1 kHz<br>1 mA to 2 A                  | 2.6 % to 1.2 %                                | Using Fluke Source with 5 ½ DMM Keysight 34450A by Direct / Comparison      |
|     |                                                                                | 1 mA to 10 A                          | 2.6 % to 1.2 %                                | Using 5 ½ DMM Keysight 34450A by Direct Method                              |
| 6.  | Frequency <sup>\$</sup>                                                        | 20 Hz to 300 kHz<br>300 kHz to 10 MHz | 0.05 %<br>0.05 % to 0.03 %                    | Using Source with 5 ½ DMM Keysight 34450A and HP MA 130 Direct / Comparison |
| 7.  | Capacitance <sup>\$</sup>                                                      | 1 kHz<br>10 pF to 1 mF                | 3.76 % to 0.6 %                               | Using LCR meter 4262A Direct Method                                         |
| 8.  | Inductance <sup>\$</sup>                                                       | 1 kHz<br>100 uH to 10 H               | 0.5 % to 0.6 %                                | Using LCR meter 4262A Direct Method                                         |
| 9.  | AC Resistance <sup>\$</sup>                                                    | 1 kHz<br>1 $\Omega$ to 10 k $\Omega$  | 0.6 %                                         | Using LCR meter 4262A Direct Method                                         |
| 10. | Temperature Simulation <sup>#</sup><br>(Temp Indicator/ Controller / Recorder) |                                       |                                               |                                                                             |
|     | RTD                                                                            | (-) 200 °C to 600 °C                  | 0.8 °C                                        | Using Masibus Calibrator                                                    |
|     | Thermocouple                                                                   |                                       |                                               |                                                                             |
|     | J Type                                                                         | (-) 200 °C to 1000 °C                 | 1.0 °C                                        |                                                                             |
|     | K Type                                                                         | (-) 200 °C to 1200 °C                 | 1.0 °C                                        |                                                                             |
|     | R Type                                                                         | 0 °C to 1600 °C                       | 1.7 °C                                        |                                                                             |
|     | B Type                                                                         | 0 °C to 1600 °C                       | 4.0 °C                                        |                                                                             |
|     | S Type                                                                         | 0 °C to 1600 °C                       | 1.7 °C                                        |                                                                             |
|     | T Type                                                                         | (-) 200 °C to 400 °C                  | 0.9 °C                                        |                                                                             |

**Laboratory** Mayuresh Associates, 4080, 4078, 3077A, Bhandup Industrial Estate,  
Pannalal Compound, LBS Marg, Bhandup West, Mumbai,  
Maharashtra

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2584

**Page** 3 of 10

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on** 25.07.2018

| Sl. | Quantity Measured / Instrument         | Range/Frequency                     | *Calibration Measurement Capability ( $\pm$ ) | Remarks                       |
|-----|----------------------------------------|-------------------------------------|-----------------------------------------------|-------------------------------|
| 11. | Digital Timer, Stop Watch <sup>#</sup> | 5s to 2h 30min<br>2h 30 min to 24 h | 0.18 to 2.5 s<br>2.5 to 21 s                  | Using Digital Timer           |
| 12. | AC High Voltage*                       | 50 Hz<br>1 kV to 10 kV              | 5 %                                           | Using Meco HV Probe & 4 ½ DMM |

Laboratory Mayuresh Associates, 4080, 4078, 3077A, Bhandup Industrial Estate,  
Pannalal Compound, LBS Marg, Bhandup West, Mumbai,  
Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2584

Page 4 of 10

Validity 20.02.2018 to 19.02.2020

Last Amended on 25.07.2018

| Sl.                                  | Quantity Measured / Instrument                | Range/Frequency                                                                                                                              | *Calibration Measurement Capability ( $\pm$ )                                                                                                                                         | Remarks                                                                                                                                                                                      |
|--------------------------------------|-----------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b><u>MECHANICAL CALIBRATION</u></b> |                                               |                                                                                                                                              |                                                                                                                                                                                       |                                                                                                                                                                                              |
| <b>I.</b>                            | <b>MASS</b>                                   |                                                                                                                                              |                                                                                                                                                                                       |                                                                                                                                                                                              |
| <b>1.</b>                            | Weights (M1 Class and Coarser) <sup>\$</sup>  | 1 mg<br>2 mg<br>5 mg<br>10 mg<br>20 mg<br>50 mg<br>100 mg<br>200 mg<br>500 mg<br>1 g<br>2 g<br>5 g<br>10 g<br>20 g<br>50 g<br>100 g<br>200 g | 0.06 mg<br>0.06 mg<br>0.06 mg<br>0.04 mg<br>0.04 mg<br>0.08 mg<br>0.09 mg<br>0.09 mg<br>0.12 mg<br>0.12 mg<br>0.20 mg<br>0.15 mg<br>0.15 mg<br>0.18 mg<br>0.41 mg<br>0.3 mg<br>0.8 mg | Using E2 Class Weights & 1mg – 200g Weighing Balance of $d=0.01\text{mg}$ / $0.1\text{mg}$ As per OIML R-111 by substitution method through ABBA cycles                                      |
|                                      | Weights of M3-Class and Coarser <sup>\$</sup> | 500 g<br>1 kg<br>2 kg<br>5 kg<br>10kg                                                                                                        | 8 mg<br>15 mg<br>20 mg<br>70 mg<br>1.6 g                                                                                                                                              | Using F2 Class Weights & Balance of Readability $d=1\text{mg}$ upto 1 kg, $d=10\text{mg}$ upto 5 kg, $d=0.1\text{g}$ upto 15 kg As per OIML R-111 by substitution method through ABBA cycles |

**Laboratory** Mayuresh Associates, 4080, 4078, 3077A, Bhandup Industrial Estate,  
Pannalal Compound, LBS Marg, Bhandup West, Mumbai,  
Maharashtra

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2584

**Page** 5 of 10

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on** 25.07.2018

| Sl.         | Quantity Measured / Instrument                                                         | Range/Frequency                                                                                                                               | *Calibration Measurement Capability ( $\pm$ )               | Remarks                                                                                                                       |
|-------------|----------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------|
| <b>II.</b>  | <b>WEIGHING SCALE AND BALANCE</b>                                                      |                                                                                                                                               |                                                             |                                                                                                                               |
| <b>1.</b>   | Weighing Balance*<br>$d \geq 0.1\text{mg}$                                             | 10 mg to 200 g                                                                                                                                | 0.3 mg                                                      | Using Calibration of Class1 weighing balances and coarser as per R-76 using E2 class standard weights 1mg to 200g             |
|             | $d \geq 0.01\text{g}$<br>$d \geq 0.1\text{g}$<br>$d \geq 100\text{g}$                  | 500 mg to 5 kg<br>20 g to 15 kg<br>2 kg to 300 kg                                                                                             | 0.08 g<br>1.7g<br>100 g                                     | Using Calibration of Class3 weighing balances and coarser as per R-76 using F1 class & M1 class standard weights up to 300 kg |
| <b>III.</b> | <b>VOLUME</b>                                                                          |                                                                                                                                               |                                                             |                                                                                                                               |
| <b>1.</b>   | Volume Glassware (Pipette, Burette, Measuring Cylinder, Volumetric Flask) <sup>§</sup> | $1\text{ml} \leq V < 10\text{ ml}$<br>$10\text{ ml} \leq V < 1000\text{ ml}$<br>$1000\text{ ml} \leq V < 5000\text{ ml}$                      | 0.01 ml<br>0.05 ml<br>0.85 ml                               | Using Calibration of Glassware based on Gravimetric method as per ISO 4787 Method                                             |
| <b>2.</b>   | Micro Pipettes <sup>§</sup>                                                            | $10\text{ }\mu\text{l} < V \leq 100\text{ }\mu\text{l}$<br>$100\text{ }\mu\text{l} < V \leq 1\text{ml}$<br>$1\text{ml} < V \leq 10\text{ ml}$ | 0.6 $\mu\text{l}$<br>1.4 $\mu\text{l}$<br>6.6 $\mu\text{l}$ | Using Calibration of Micropipettes based on Gravimetric method as per ISO 8655 Method                                         |

**Laboratory** Mayuresh Associates, 4080, 4078, 3077A, Bhandup Industrial Estate,  
Pannalal Compound, LBS Marg, Bhandup West, Mumbai,  
Maharashtra

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2584

**Page** 6 of 10

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on** 25.07.2018

| Sl.        | Quantity Measured / Instrument                    | Range/Frequency                                            | *Calibration Measurement Capability ( $\pm$ ) | Remarks                                                                                                                                      |
|------------|---------------------------------------------------|------------------------------------------------------------|-----------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------|
| <b>IV.</b> | <b>DENSITY</b>                                    |                                                            |                                               |                                                                                                                                              |
| <b>1.</b>  | Hydrometer <sup>\$</sup>                          | 0.600 g/ml to 2.000 g/ml                                   | 0.008 g/ml                                    | Using Standard Hydrometer of Resolution 0.0005 g/ml and Appropriate Liquid by Comparison method as per Archimedes Principle based on IS 3104 |
| <b>V.</b>  | <b>ACCELERATION &amp; SPEED</b>                   |                                                            |                                               |                                                                                                                                              |
| <b>1.</b>  | Digital Tachometer <sup>\$</sup>                  | 10 RPM to 100 RPM<br>100 RPM to 100000 RPM                 | 0.15 % to 0.06 %                              | Using Digital Tachometer by Comparison Method                                                                                                |
| <b>2.</b>  | Tachometer/ Centrifuge <sup>*</sup>               | 10 RPM to 100 RPM<br>100 RPM to 40000 RPM                  | 0.15% to 0.06% RPM                            | Using Digital Tachometer by Comparison Method                                                                                                |
| <b>VI.</b> | <b>PRESSURE INDICATING DEVICES</b>                |                                                            |                                               |                                                                                                                                              |
| <b>1.</b>  | Pressure Gauge/Pressure Transmitter <sup>\$</sup> | 0 to 1 bar<br>0 to 10 bar<br>0 to 100 bar                  | 0.007 bar<br>0.061 bar<br>0.61 bar            | Using Digital Pressure Gauge by Comparison Method as per DKD R 6-1                                                                           |
| <b>2.</b>  | Pressure Gauge/Pressure Transmitter <sup>*</sup>  | 0 to 10 bar<br>0 to 100 bar<br>0 to 600 kg/cm <sup>2</sup> | 0.061 bar<br>0.61 bar<br>4.43 bar             | Using Digital Pressure Gauge by Comparison Method as per DKD R 6-1                                                                           |

**Laboratory** Mayuresh Associates, 4080, 4078, 3077A, Bhandup Industrial Estate,  
Pannalal Compound, LBS Marg, Bhandup West, Mumbai,  
Maharashtra

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2584

**Page** 7 of 10

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on** 25.07.2018

| Sl. | Quantity Measured / Instrument                        | Range/Frequency        | *Calibration Measurement Capability ( $\pm$ ) | Remarks                                                            |
|-----|-------------------------------------------------------|------------------------|-----------------------------------------------|--------------------------------------------------------------------|
| 3.  | Low Pressure Gauge/<br>Magnehelic Gauge <sup>\$</sup> | 98.06 Pa to 1961.33 Pa | 57.8 Pa                                       | Using Digital Pressure Indicator by Comparison Method as DKD R 6-1 |
| 4.  | Vacuum Gauge/<br>Vacuum Transmitter <sup>#</sup>      | (-) 0.95 o 0 bar       | 0.013 bar                                     | Using Digital Vacuum Gauge by Comparison Method as per DKD R 6-1   |

---

Vishal Shukla  
Convenor

---

Avijit Das  
Program Director

**Laboratory** Mayuresh Associates, 4080, 4078, 3077A, Bhandup Industrial Estate,  
Pannalal Compound, LBS Marg, Bhandup West, Mumbai,  
Maharashtra

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2584

**Page** 8 of 10

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on** 25.07.2018

| Sl.                               | Quantity Measured / Instrument                                                                 | Range/Frequency                                                                                | *Calibration Measurement Capability ( $\pm$ )       | Remarks                                                                                                                          |
|-----------------------------------|------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------|-----------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------|
| <b><u>THERMAL CALIBRATION</u></b> |                                                                                                |                                                                                                |                                                     |                                                                                                                                  |
| <b>I.</b>                         | <b>TEMPERATURE</b>                                                                             |                                                                                                |                                                     |                                                                                                                                  |
| <b>1.</b>                         | PRT / RTD / Thermocouple Sensors with and without Indicator, Digital Thermometer <sup>\$</sup> | (-) 80 °C to (-) 40 °C<br>(-) 40 °C to 200 °C<br>>200 °C to 600 °C<br>@~ (-) 196 °C<br>@~ 0 °C | 0.81 °C<br>0.22 °C<br>0.59 °C<br>0.3 °C<br>0.082 °C | Using PRT with DMM, Liquid Comparator, Alcohol bath, Oil bath, Dry well calibrator, LN2 Comparator, Ice bath (Comparison Method) |
| <b>2.</b>                         | Glass Thermometer / Dial Thermometer <sup>\$</sup>                                             | (-) 80 °C to 500 °C                                                                            | 0.9 °C                                              | Using PRT with DMM, Liquid Comparator, Alcohol bath, Oil bath, Liquid Tin bath (Comparison Method)                               |
| <b>3.</b>                         | Thermocouple Sensor with and without Indicator, Digital Thermometer <sup>\$</sup>              | >600 °C to <1000 °C<br>1000 °C to 1200 °C                                                      | 2.8 °C<br>3.7 °C                                    | Using R Type Thermocouple with DMM, Dry well Furnace                                                                             |
| <b>4.</b>                         | Digital Thermo Hygrometer <sup>\$</sup>                                                        | 10 °C to 60 °C<br>30 % RH to 95 % RH                                                           | 0.9 °C<br>4.6 % RH                                  | Using PRT with DMM, Humidity Indicator with Sensor, Stability Chamber                                                            |
| <b>5.</b>                         | Humidity Sensor with Indicator <sup>\$</sup>                                                   | 15 % RH to 95 % RH (@25°C Approx)                                                              | 3.4 %RH                                             | Using Humidity Indicator with Sensor, Humidity Dessicators                                                                       |

**Vishal Shukla**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Mayuresh Associates, 4080, 4078, 3077A, Bhandup Industrial Estate,  
Pannalal Compound, LBS Marg, Bhandup West, Mumbai,  
Maharashtra

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2584

**Page** 9 of 10

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on** 25.07.2018

| Sl. | Quantity Measured / Instrument                                                                                                                                  | Range/Frequency                           | *Calibration Measurement Capability ( $\pm$ ) | Remarks                                                                    |
|-----|-----------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------|-----------------------------------------------|----------------------------------------------------------------------------|
| 6.  | PRT / RTD / Thermocouple Sensors with and without Indicator, Digital Thermometer*                                                                               | 50 °C to 600 °C                           | 0.59 °C                                       | Using PRT with DMM, Dry well bath                                          |
| 7.  | Thermocouple Sensor with and without Indicator, Digital Thermometer*                                                                                            | >600 °C to <1000 °C<br>1000 °C to 1200 °C | 3.6 °C<br>3.6 °C                              | Using R type Thermocouple with DMM, Dry well Furnace                       |
| 8.  | Calibration of Temperature Indicator of Chamber (Liquid bath, Oil bath, Deep Freezer, Cooling cabinet, Incubator, Waterbath, Oven, Muffle Furnace, Autoclave) * | (-) 80 °C to 300 °C<br>300 °C to 1200 °C  | 0.9 °C<br>3.7 °C                              | Using PRT with DMM, R type Thermocouple with DMM, Single Probe Calibration |
| 9.  | Calibration of Chamber (Cooling cabinet, Incubator, Deep Freezer, Waterbath, Oven) *                                                                            | (-) 80 °C to 300 °C                       | 1.7 °C                                        | Using Datalogger with Temperature Sensors (9-probe Calibration)            |
| II. | <b>SPECIFIC HEAT AND HUMIDITY</b>                                                                                                                               |                                           |                                               |                                                                            |
| 1.  | Calibration of Humidity Indicator of Chamber*                                                                                                                   | 30 % RH to 95 % RH (@25°C Approx)         | 4.7 % RH                                      | Using Humidity Loggers, Single Probe Calibration                           |

**Vishal Shukla**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Mayuresh Associates, 4080, 4078, 3077A, Bhandup Industrial Estate,  
Pannalal Compound, LBS Marg, Bhandup West, Mumbai,  
Maharashtra

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2584

**Page** 10 of 10

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on** 25.07.2018

| Sl. | Quantity Measured / Instrument | Range/Frequency                   | *Calibration Measurement Capability ( $\pm$ ) | Remarks                                        |
|-----|--------------------------------|-----------------------------------|-----------------------------------------------|------------------------------------------------|
| 2 . | Humidity / Stability Chamber*  | 30 % RH to 95 % RH (@25°C Approx) | 7.26 % RH                                     | Using Humidity Loggers, (9- probe Calibration) |

\* Measurement Capability is expressed as an uncertainty ( $\pm$ ) at a confidence probability of 95%

§Only in Permanent Laboratory

\*Only for Site Calibration

# The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

---

Vishal Shukla  
Convenor

---

Avijit Das  
Program Director