

Laboratory **Meter Testing Laboratory, RRVNPL, 132 kV Chambal GSS Premises,
Civil Lines, Jaipur, Rajasthan**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2542 (In lieu of C-0799)**

Page **1 of 1**

Validity **22.02.2018 to 21.02.2020**

Last Amended on **-**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>ELECTRO-TECHNICAL CALIBRATION</u>				
I.	MEASURE			
1.	AC Active Energy [§] (1 Φ , 3 Φ)	50Hz 63.5 V to 230 V 1mA to 12A (0.11 W to 6.6 kW) At UPF 0.5 Lag & 0.8 Lead	0.048% to 0.075% 0.023 % to 0.055 %	Using ZERA 3- Φ Power/Energy Meter EPZ-303-5 By Direct Method
2.	AC Reactive Energy [§] (1 Φ , 3 Φ)	50Hz 40 V to 320 V 1mA to 12A (0.04 W to 3.84 kW) At UPF 0.5 Lag & 0.8 Lead	0.048% to 0.075% 0.023 % to 0.051 %	Using ZERA 3- Φ Power/Energy Meter EPZ-303-5 By Direct Method
3.	AC Voltage [§]	50 Hz 40 V to 500V (p-n)	0.064% to 0.11%	Using ZERA 3- Φ Power/Energy Meter EPZ-303-5 By Direct Method
4.	AC Current [§]	50 Hz 5 mA to 120A	0.025% to 0.58%	Using ZERA 3- Φ Power/Energy Meter EPZ-303-5 By Direct Method
5.	Frequency [§]	63.5 V, 5A 40 to 70 Hz	0.03% to 0.12 %	Using ZERA 3- Φ Power/Energy Meter EPZ-303-5 By Direct Method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

§ Only in Permanent Laboratory

Rajeshwar Kumar
Convenor

Avijit Das
Program Director