

Laboratory Mikronix Gauges Pvt. Ltd., P. O. No. 701, A-37, MIDC Shendra,
 Aurangabad, Maharashtra
Accreditation Standard ISO/IEC 17025:2005
Discipline Mechanical Calibration **Issue Date** 02.01.2015
Certificate Number C-0309 **Valid Until** 01.01.2017
Last Amended on - **Page** 1 of 1

S. No.	Quantity Measured / Instrument	Range/Frequency	* Calibration Measurement Capability (\pm)	Remarks
I. DIMENSION				
1.	PLAIN PLUG/ WIDTH GAUGE ^{\$}	Upto 100 mm	1.3 μ m	Using Slip Gauge, Comparator Stand & Dial
2.	SNAP GAUGE ^{\$}	4 mm to 100 mm	0.8 μ m	Using Slip Gauge
3.	PLAIN RING GAUGE ^{\$}	4 mm to 100 mm	1.0 μ m	Using ULM Master Setting Ring

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

^{\$}Only in Permanent Laboratory

 Naveen Jangra
 Convenor

 Avijit Das
 Program Manager