Laboratory	**NAPS Testing Laboratory, NAPS, Narora, Bulandshahar, Uttar Pradesh**		
Accreditation Standard	**ISO/IEC 17025: 2005**		
Discipline	**Thermal Calibration**	**Issue Date**	**23.09.2013**
Certificate Number	**C-0437**	**Valid Until**	**22.09.2015**
Last Amended on	**-**	**Page**	**1 of 1**

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (±)	Remarks
TEMPERATURE[$]			
1. TEMPERATURE SENSOR WITH & WITHOUT DISPLAY AND TEMPERATURE GAUGES	50°C to 500°C	0.47°C	Using Temperature Comparator Dry Block

*** Measurement Capability is expressed as an uncertainty (±) at a confidence probability of 95%**
[$]Only in Permanent Laboratory