Laboratory	Norbar Torque Tools India Pvt. Ltd., Novelty House, Plot No. A-168, 1st Floor, TTC Industrial Area, Kopar Khairane, Navi Mumbai, Maharashtra		
Accreditation Standard	ISO/IEC 17025: 2005		
Certificate Number	CC-2885	Page 1 of 1	
Validity	12.11.2018 to 11.11.2020	Last Amended on -	

SI.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks	
MECHANICAL CALIBRATION					
1.	TORQUE GENERATIN				
1.	Torque Wrenches ^{\$} Type 1 Class A, B, C, D, E Type II Class A, B, C, D, E, F and G	0.1 Nm to 3000 Nm	1.3 % rdg to 0.58% rdg	Using Torque Wrench Calibrator & Transducers as per ISO 6789 by Comparison Method	
2.	Torque Transducers ^{\$}	0.05 Nm to 300 Nm 300 Nm to 680 Nm 680 Nm to 6780 Nm	0.15 % rdg 0.15 % rdg 0.07 % rdg	Using Torque Calibration Beam and Weights as per BS 7882 by Comparison Method	

* Measurement Capability is expressed as an uncertainty (±) at a confidence probability of 95% ^{\$}Only in Permanent Laboratory