

Laboratory	Perfect Enterprises, S.No. 81/5, Kiran Industrial Estate, J Block, Shed No. 1, Shivane, Pune, Maharashtra		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	24.06.2015
Certificate Number	C-0827	Valid Until	23.06.2017
Last Amended on	-	Page	1 of 3

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
I. SOURCE			
1. AC Voltage \$	50 Hz 10 mV to 300 mV 300 mV to 30 V 30 V to 1000 V	4.10 % to 1.15 % 1.15 % to 1.50 % 1.50 %	Using Multifunction Calibrator Zeal Make & 5 ^{3/4} DMM Rish 20 Make by Comparison
2. DC Voltage \$	1 mV to 300 mV 300 mV to 30 V 30 V to 1000 V	1.20 % to 0.21 % 0.21 % to 0.50 % 0.50 % to 0.23 %	Using Multifunction Calibrator Zeal Make & 5 ^{3/4} DMM Rish 20 Make by Comparison
3. Frequency \$	45Hz to 1000 Hz	0.60 % to 0.26 %	Using Multifunction Calibrator Zeal Make & 5 ^{3/4} DMM Rish 20 Make by Comparison
4. AC Current \$	50 Hz 10 mA to 3A 3 A to 10 A 10 A to 1000 A	1.76 % to 1.79 % 1.79 % to 1.05 % 1.05 % to 1.20 %	Using Multifunction Calibrator Zeal Make & 5 ^{3/4} DMM Rish 20 Make With current coil Zeal Make by Comparison
5. DC Current \$	1 mA to 3A 3 A to 10 A 10 A to 1000 A	0.60 % to 3.00 % 3.00 % to 1.20 % 1.20 % to 1.10 %	Using Multifunction Calibrator Zeal Make & 5 ^{3/4} DMM Rish 20 Make With current coil Zeal Make by Comparison
6. DC Resistance \$	1 Ω to 1000 Ω 1000 Ω to 1 M Ω 1 M Ω to 100 M Ω 100 M Ω to 1 G Ω	5.90 % to 0.50 % 0.50 % to 0.15 % 0.15 % to 1.53 % 1.53 % to 2.00 %	Using Decade Resistance Box Zeal Make by Direct

Ram Ashray
Convenor

Avijit Das
Program Manager

Laboratory Perfect Enterprises, S.No. 81/5, Kiran Industrial Estate, J Block, Shed No. 1, Shivane, Pune, Maharashtra
Accreditation Standard ISO/IEC 17025: 2005
Discipline Electro-Technical Calibration **Issue Date** 24.06.2015
Certificate Number C-0827 **Valid Until** 23.06.2017
Last Amended on - **Page** 2 of 3

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
7. Capacitance \$	1 kHz 10 nF to 10 µF	3.00 % to 3.30 %	Using Decade Capacitance Box Multiscope Make by Direct
8. TEMPERTURE SIMULATION \$			
T/C Mode			
J - Type	30 °C to 1000 °C	2.43 °C	Using Universal Calibrator Sansui Make by Direct
K- Type	30 °C to 1000 °C	2.43 °C	
R- Type	150 °C to 950 °C	2.52 °C	
S- Type	150 °C to 950 °C	2.52 °C	
RTD	0 °C to 400 °C	1.50 °C	Using Universal Calibrator Sansui Make by Direct
II. MEASURE			
1. AC Voltage \$	50 Hz 10 mV to 300 mV 300 mV to 30 V 30 V to 1000 V	4.10 % to 1.30 % 1.30 % to 1.50 % 1.50 % to 1.40 %	Using 5 ^{3/4} DMM Rish 20 Make by Direct
2. DC Voltage \$	10 mV to 300 mV 300 mV to 30 V 30 V to 1000 V	5.59 % to 0.21 % 0.21 % to 0.06 % 0.06 %	Using 5 ^{3/4} DMM Rish 20 Make by Direct
3. AC Current \$	50 Hz 1 mA to 3 A 3 A to 10 A	2.03 % to 2.00 % 2.00 % to 1.03 %	Using 5 ^{3/4} DMM Rish 20 Make by Direct
4. DC Current \$	1 mA to 3A 3 A to 10 A	0.18 % to 1.65 % 1.65 % to 0.77 %	Using 5 ^{3/4} DMM Rish 20 Make by Direct
5. Frequency \$	45 Hz to 1000 Hz	0.24 % to 0.20 %	Using 5 ^{3/4} DMM Rish 20 Make by Direct

Ram Ashray
Convenor

Avijit Das
Program Manager

Laboratory Perfect Enterprises, S.No. 81/5, Kiran Industrial Estate, J Block, Shed No. 1, Shivane, Pune, Maharashtra
Accreditation Standard ISO/IEC 17025: 2005
Discipline Electro-Technical Calibration **Issue Date** 24.06.2015
Certificate Number C-0827 **Valid Until** 23.06.2017
Last Amended on - **Page** 3 of 3

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
6. DC Resistance \$	10 Ω to 1000 Ω 1000 Ω to 1 MΩ 1 MΩ to 30 MΩ	2.00 % to 0.40 % 0.40 % to 0.10 % 0.10 %	Using 5 ^{3/4} DMM Rish 20 Make by Direct

* Measurement Capability is expressed as an uncertainty (±) at a confidence probability of 95%
 \$Only in Permanent Laboratory

Ram Ashray
 Convenor

Avijit Das
 Program Manager