Laboratory	Photometry Laboratory, Crompton Greaves Ltd., Lighting Division, Kanjur Marg (E), Mumbai, Maharashtra		
Accreditation Standard	ISO/IEC 17025:2005		
Discipline	Optical Calibration	Issue Date	10.01.2014
Certificate Number	C-0795	Valid Until	09.01.2016
Last Amended on	27.01.2014	Page	1 of 1

	Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
<u>OP</u> 1.	<u>TICAL<sup>\$</sup></u> LUMINOUS FLUX FOR INCANDESCENT LAMPS	1 Lumen to 25000 Lumen	2.6%	Using Goniophotometer
2.	LUMINOUS FLUX FOR DISCHARGE LAMPS	1 Lumen to 50,000 Lumen	3.0%	Using Goniophotometer

\* Measurement Capability is expressed as an uncertainty (±) at a confidence probability of 95%  $^{\rm \$Only}$  in Permanent Laboratory