

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 1 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>ELECTRO TECHNICAL CALIBRATION</u>				
I.	MEASURE			
1.	DC Voltage [§]	1 mV to 10 mV 10 mV to 100 mV 100 mV to 10 V 10V to 1000 V 1 kV to 15 kV	0.053 % to 0.014 % 0.014 % to 0.002 % 0.002 % to 0.0007 % 0.0007 % to 0.001 % 2.50 %	Using 8 ½ DMM Fluke 8508A by Direct Measurement HV Probe with DMM
2.	AC Voltage [§]	50 Hz to 10kHz 1 mV to 100 V 50 Hz to 1 kHz 100 V to 1000 V 50 Hz 1 kV to 15 kV	0.83 % to 0.03% 0.02% to 0.02% 6.20%	Using 8 ½ DMM Fluke 8508A by Direct Measurement HV Probe with DMM
3.	DC Current [§]	1 μ A to 10 μ A 10 μ A to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 10 A 10 A to 20 A	0.75 % to 0.009 % 0.009 % to 0.005 % 0.005 % to 0.01 % 0.01 % to 0.025% 0.025% to 0.05% 0.050% to 0.07%	Using 8 ½ DMM Fluke 8508A by Direct Measurement
4.	AC Current [§]	50 Hz to 1 kHz 10 μ A to 100 mA 100 mA to 1 A 1 A to 10 A 10 A to 20 A	0.46 % to 0.06 % 0.06 % to 0.10 % 0.10 % to 0.12 % 0.12 %	Using 8 ½ DMM Fluke 8508A by Direct Measurement

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 2 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
5.	DC Resistance ^{\$}	1 Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 2 G Ω 2G Ω to 20 G Ω	0.06 % to 0.003 % 0.003 % to 0.005 % 0.005 % to 0.02 % 0.02 % to 0.60 % 0.60 % to 0.22%	Using 8 ½ DMM Fluke 8508A by Direct Measurement
6.	Frequency ^{\$}	5 Hz to 10 Hz 10 Hz to 1 MHz	0.05 % to 0.05 % 0.05 % to 0.06 %	Using 8 ½ DMM Fluke 8508A by Direct Measurement
7.	Capacitance ^{\$}	1 kHz 1 nF to 10 μ F	1.10 % to 0.40 %	Using Digital LCR Meter by Direct Measurement
8.	Inductance ^{\$}	1 kHz 100 μ H to 10 H	2.10 % to 1.50 %	Using Digital LCR Meter by Direct Measurement
9.	AC Resistance ^{\$}	1 kHz 1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 100 k Ω	1.60 % to 1.20 % 1.20 % to 0.70 % 0.70 % to 0.60 %	Using Digital LCR Meter by Direct Measurement
10.	Time Interval [#]	1 s to 100 s 100 s to 1000 s 1000 s to 9000 s 10 s to 1000 s 1000 s to 9900 s 9900 s to 86400 s	0.01 s to 0.02 s 0.02 s to 0.12 s 0.12 s to 1.00 s 0.15 s to 0.20 s 0.20 s to 0.50 s 0.50 s to 4.00 s	Timer Calibrator By Comparison Measurement Stop Watch By Comparison Measurement

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 3 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
11.	Temperature [§] Thermocouple (J,K,T -Type) Thermocouple (R,S,B ,N-Type) RTD-PT100 Type	(-)200 °C to 1200 °C 150 °C to 1700 °C (-)200 °C to 500 °C	0.4 °C to 0.8 °C 0.5 °C to 0.9 °C 0.02 °C to 0.06 °C	Using 8 ½ DMM Fluke 8508A by Direct Measurement / Simulation Method
12.	DC Voltage*	1 mV to 10 mV 10 mV to 100 mV 100 mV to 1000 V 1 kV to 37 kV	0.60 % to 0.04 % 0.04 % to 0.01 % 0.01 % to 0.01 % 2.6%	Using 6 ½ DMM 8846A Fluke by Direct measurement HV Probe with DMM
13.	AC Voltage*	50 Hz 10 mV to 1000 V 1 kV to 15 kV 15 kV to 100 kV	0.55 % to 0.10 % 5.80 % 4.0 % to 2.5 %	Using 6 ½ DMM 8846A Fluke HV Probe with DMM HV Divider by Direct measurement
14.	DC Current*	10 μ A to 100 mA 100 mA to 10 A	0.60 % to 0.08 % 0.08 % to 0.20 %	Using 6 ½ DMM 8846A Fluke by Direct measurement
15.	AC Current*	50 Hz 1 mA to 100 mA 100 mA to 10 A	0.60 % to 0.50 % 0.50 % to 0.45 %	Using 6 ½ DMM 8846A Fluke by Direct measurement

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 4 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

"In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020"

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
II.	SOURCE			
1.	DC Voltage #	1 mV to 32 mV 32 mV to 320 mV 320 mV to 1000 V	0.40 % to 0.02 % 0.02 % to 0.009 % 0.009 % to 0.01%	Using 5500 A MFC Fluke by Direct Sourcing
2.	AC Voltage #	50 Hz to 1 kHz 10 mV to 32 mV 32 mV to 32 V 32 V to 1000 V	0.41 % to 0.25 % 0.25 % to 0.05% 0.05% to 0.07%	Using 5500 A MFC Fluke by Direct Sourcing
3.	DC Current #	10 μ A to 320 μ A 320 μ A to 320 mA 320 mA to 2 A 2 A to 11 A 10 A to 1000 A	0.67 % to 0.04 % 0.04 % to 0.02% 0.02 % to 0.04 % 0.04 % to 0.095 % 0.90 %	Using 5500 A MFC Fluke With Current coil by Direct Sourcing
4.	AC Current #	50Hz to 1kHz 30 μ A to 320 μ A 320 μ A to 11 A 50 Hz 10 A to 1000 A	0.98 % to 0.25 % 0.25 % to 0.16 % 0.80 %	Using 5500A MFC Fluke With Current coil by Direct Sourcing
5.	Frequency #	5 Hz to 1 MHz 1 MHz to 300 MHz	0.10% to 0.03% 0.03% to 0.008%	Using 5500A MFC Fluke by Direct Sourcing

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 5 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
6.	Low Resistance #	1 m Ω 10 m Ω 100 m Ω 1 Ω to 1 k Ω	0.65 % 0.70 % 0.70 % 0.70 %	Using Standard Resistance Box (Discrete Values) by Direct Sourcing
7.	Resistance #	1 Ω to 32 Ω 32 Ω to 320 Ω 320 Ω to 3.2 M Ω 3.2 M Ω to 32 M Ω 32 M Ω to 320 M Ω	0.90 % to 0.07 % 0.07 % to 0.015% 0.015 % to 0.02 % 0.02 % to 0.12 % 0.12 % to 0.58 %	Using 5500A MFC Fluke by Direct Sourcing
8.	Temperature # Thermocouple (J,K,T -Type) Thermocouple (R,S,B ,N-Type) RTD-PT100 Type	 (-)200 $^{\circ}$ C to 1300 $^{\circ}$ C 0 $^{\circ}$ C to 1700 $^{\circ}$ C (-)200 $^{\circ}$ C to 600 $^{\circ}$ C	 0.30 $^{\circ}$ C to 0.33 $^{\circ}$ C 0.7 $^{\circ}$ C to 1.2 $^{\circ}$ C 0.11 $^{\circ}$ C to 0.18 $^{\circ}$ C	Using 5500A MFC Fluke by Direct Sourcing / Simulation Method
9.	AC Power \$ At 50Hz	0.1 W – 6 kW Voltage : (10V – 600V) Current: (0.1A – 10A) PF: UPF to \pm 0.2 PF	0.80% to 0.20%	Using 5500A MFC Fluke by Direct Sourcing
10.	DC Power \$	1 W – 10 kW Voltage : (10V – 1000V) Current: (0.1A – 10A)	0.12% to 0.15%	Using 5500A MFC Fluke by Direct Sourcing

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 6 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
11.	Power Factor [§]	\pm 0.1 - UPF	0.0032PF	Using 5500A MFC Fluke by Direct Sourcing
12.	AC Current [§]	50Hz 10A to 990 A	1.20%	Using AC High Current Source by Direct Sourcing
13.	Capacitance [§]	1 kHz 1nF to 3.2 nF 3.2 nF to 32 nF 32 nF to 320 nF 320 nF to 100 μ F	1.73% to 0.70% 0.70% to 0.40% 0.40% to 0.30% 0.30% to 0.70%	Using 5500A MFC Fluke by Direct Sourcing
14.	Inductance [§]	1 kHz 1mH , 10mH , 1 H	3.0 %	Using Fixed Inductance Box by Direct Sourcing
15.	AC Resistance [§]	1 kHz 1 Ω 10 Ω 100 Ω 1k Ω 10 k Ω 100 k Ω	1.0 % 1.0 % 1.0 % 1.0 % 1.0 % 1.10 %	Using Fixed Resistance Box by Direct Sourcing

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 7 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
16.	Oscilloscope ^{\$}	Amplitude 1 kHz (Square Wave) 5 mV to 100 V (50 Ω / 1M Ω) Time Base 5 nS to 5S Bandwidth 10 MHz to 300 MHz	2.85% to 0.3% 0.5% to 0.7% 5.2%	Using 5500A MFC Fluke by Direct Sourcing
17.	Pulse Type Counter Meter / Period Meter ^{\$}	100 to 50000 Pulse Count / Period	1.73 Pulse Count	Using Counter Calibrator by Direct Sourcing / Measurement
18.	AC High Voltage ^{\$} At 50 Hz	0 to 5 kV	2.27%	Using AC High Voltage Source , by Direct Sourcing
19.	Insulation Resistance ^{\$} (High Resistance)	2 , 20 , 200 & 500 M Ω 1 , 2 , 20 G Ω 200 G Ω	2.5% 3.5% 4.0%	Using Resistance Box (Discrete Value) By Direct Sourcing

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 8 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>MECHANICAL CALIBRATION</u>				
1.	DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)			
1.	Digital / Dial / Vernier Calipers \$ L.C. 0.01 mm ϕ Vernier Calipers # L.C. 0.02 mm	Upto 300 mm >300 mm to 600 mm >600 mm to 1000 mm >1000 mm to 2000 mm	7.0 μ m 8.0 μ m 9.0 20.0 μ m	Using Gauge Blocks / Length Bars / Caliper Checker / Slip Gauge Accessories By Comparison Method
2.	Digital / Dial / Vernier Height Gauge \$ L.C. : 0.01 mm ϕ	Upto 300 mm >300 mm to 600 mm >600 mm to 1000 mm	8 μ m 9 μ m 10 μ m	Using Gauge Blocks / Length Bar/ Caliper Checker / Surface Plate By Comparison Method
3.	Digital / Dial / Vernier Depth Gauge \$ L.C. : 0.01 mm ϕ	Upto 300 mm >300 mm to 600 mm	7 μ m 8 μ m	Using Gauge Blocks / Length Bars / Surface Plate By Comparison Method

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 9 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
4.	Digital / External Micrometer \$ L.C. 0.001 mm L.C. 0.01 mm	Upto 50 mm >50 mm to 100 mm >100 mm to 200 mm >200 mm to 300 mm Upto 300 mm >300 mm to 600 mm >600 mm to 1000 mm > 1000 mm to 2000 mm	1 μ m 2 μ m 2.5 μ m 3 μ m 5 μ m 6 μ m 7 μ m 16 μ m	Using Gauge Blocks / Length Bars By Comparison Method
5.	Digital / Internal Micrometer \$ L.C. : 0.001 mm L.C. : 0.01 mm Digital / Internal Micrometer # L.C. : 0.01 mm	Upto 200 mm >200 mm to 300 mm Upto 300 mm >300 mm to 600 mm >600 mm to 1000 mm 1000 mm to 2000 mm	3.0 μ m 3.0 μ m 6.0 μ m 7.0 μ m 8.0 μ m 17.0 μ m	Using V-Block , Dial Gauge , Gauge Blocks , Vertical Comparator By Comparison Method

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 10 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
6.	Digital / Depth Micrometer [§] L.C. : 0.001 mm L.C. : 0.01 mm	Upto 100 mm >100 mm to 200 mm >200 mm to 300 mm Upto 300 mm >300 mm to 600 mm	1.0 μ m 2.5 μ m 3.0 μ m 5.0 μ m 6.0 μ m	Using Gauge Blocks / V-Block & Surface Plate By Comparison Method
7.	Digital / Dial Indicators – Plunger Type [§] L.C. : 0.001 mm L.C. : 0.01 mm [¶]	Upto 1.00 mm >1.00 mm to 12.5 mm >12.5 mm to 25.4 mm Upto 25 mm >25 mm to 50 mm >50 mm to 100 mm	1.0 μ m 1.5 μ m 2.0 μ m 5.0 μ m 6.0 μ m 7.0 μ m	Using Dial Calibration Tester By Comparison Method Using Slip Gauge with Dial Comparator By Comparison Method
8.	Digital / Dial Indicators – Lever Type [§] L.C. : 0.001 mm	Upto 1mm	1.0 μ m	Using Dial Calibration Tester By Comparison
9.	Bore Gauge Stroke Length / Transmission [§]	Upto 2 mm	1.7 μ m	Using Dial Calibration Tester By Comparison Method
10.	Micrometer Setting Rods [§]	25 mm to 100 mm 100 mm to 300 mm 300 mm to 600 mm 600 mm o 1000 mm	3.0 μ m 5.0 μ m 7.0 μ m 8.0 μ m	Using Gauge Blocks / Surface Plate with Lever Gauge By Comparison Method

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 11 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
11.	Dial Thickness Gauge ^{\$} L.C. : 0.001 mm L.C.:0.01 mm	Upto 1 mm >1 mm to 10 mm Upto 10 mm >10 mm to 25 mm	2.5 μ m 3.5 μ m 6.0 μ m 7.0 μ m	Using Gauge Blocks By Comparison Method
12.	Metric Steel Scale ^{\$}	Upto 1000 mm	$68 \sqrt{\frac{l}{500}}$ μ m L is in mm	Using Tape Calibration Unit By Comparison Method
13.	Metric Steel Tape / Pie Tape ^{\$}	Upto 50 meter	$68 \sqrt{\frac{l}{500}}$ μ m L is in mm	Using Tape Calibration Unit By Comparison Method
14.	Bevel Protractor / Degree Protractor / Combination Set ^{\$}	0 ^o – 90 ^o – 0 ^o	4 minute of arc	Using Angle Gauges By Comparison Method
15.	Plain Plug gauge ^{\$}	3 mm to 100 mm >100 mm to 200 mm >200 mm to 300 mm	3.0 μ m 4.6 μ m 6.0 μ m	Using Gauge Blocks with Comparator By Comparison Method
16.	Plain Ring Gauge ^{\$}	Upto 100 mm >100 mm to 200 mm >200 mm to 300 mm	2.0 μ m 3.0 μ m 4.0 μ m	Using ULM100 Gauge Block with Accessories By Comparison Method
17.	Cylindrical Pins ^{\$}	Upto 5 mm >5 mm to 20 mm	2.0 μ m 2.5 μ m	Using Dial Gauge with Comparator Stand By Comparison Method

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 12 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
18.	Snap Gauge \$ Plain / Dial / Adjustable	Upto 150 mm >150 mm to 300 mm >300 mm to 450 mm	2.5 μ m 6.0 μ m 7.5 μ m	Using Gauge Blocks By Comparison Method
19.	Thread Plug Gauge \$ (Effective Diameter Only)	Upto 25 mm >25 mm to 50 mm >50 mm to 100 mm	4.0 μ m 4.5 μ m 5.0 μ m	Using Micrometer Head with Free Float Table Fixture & Three wire Unit By Comparison Method
20.	Thread Ring Gauge \$ (Effective Diameter Only)	>3 mm to 100 mm	2.5 μ m	Using ULM By Comparison Method
21.	Feeler Gauge \$	Upto 5 mm	2.0 μ m	Using Digital Micrometer By Comparison Method
22.	Angle Plate \$ Flatness Squareness	Upto 300 mm	4.0 μ m 5.0 μ m	Using Surface Gauge Set Up & Cylindrical Square By Comparison Method
23.	Engineers Square \$ Straightness Squareness	upto 600 mm	4.0 μ m 6.0 μ m	Using Surface Gauge Set & Cylindrical Square By Comparison Method
24.	V-Blocks \$ Flatness Parallelism Squareness	Upto 300 mm X 100 mm X 100 mm	4.0 μ m 5.0 μ m 6.0 μ m	Using Surface Plate / Test Mandrel & Dial Indicator , Leveling Screws By Comparison Method

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 13 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
25.	Coating Thickness Gauge [§] L.C.: 1.0 μ m L.C. : 0.1 μ m	Upto 1 mm >1 mm to 3 mm >3 mm to 5 mm	3.0 μ m 6.0 μ m 8.0 μ m	Using Thickness Foils By Comparison Method
26.	Inside / Outside Caliper / Pistol Caliper [§]	Upto 300 mm	70.0 μ m	Gauge Blocks Gauge Blocks Accessories By Comparison Method
27.	Dial Comparator Stand [§] Flatness	Upto 150 mm X 150 mm	4.0 μ m	Using Surface Plate / Dial Indicator
28.	Weld / Hi-Lo Gauge / Bridge Cam Gauge [§]	Upto 10 mm 10mm to 25 mm 25 mm to 50 mm	2.0 μ m 4.0 μ m 5.0 μ m	Using Slip Gauges / Comparator Stand By Comparison Method
29.	Width / Gap Gauge [§]	Upto 150 mm 150 mm to 300 mm	2.0 μ m 5.5 μ m	Using Slip Gauge / Digital Vernier Caliper By Comparison Method
30.	Surface Plate – Cast Iron / Granite Comparator Stand *	Upto 3000 mm	$0.5 \sqrt{\frac{L+W}{150}} \mu\text{m}$ L & W in mm	Using Straight Edge / Gauge Blocks / Spirit Level
31.	Bench Centre * Parallelism Co-axiality	Centre Distance Upto 500 mm	5.0 μ m 5.0 μ m	Using Test Mandrel / Lever Gauge By Comparison Method

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page

14 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
32.	Linear / Vernier / Digital / Electronic Height Gauge * L.C. 0.0001 mm	Upto 2000 mm	1.0 μ m	Using Gauge Block / Gauge Block Accessories Caliper Checker By Comparison Method
VI.	ACCELERATION & SPEED			
1.	Non-Contact Type # RPM Indicator / Centrifuge.	100 rpm to 12000 rpm >12000 rpm to 24000 rpm	5.14 % 0.35 %	Using Digital Tachometer
VII.	ACOUSTICS			
1.	Sound Level Meter §	1 kHz 94 dB & 114 dB	1.5 dB	Using Sound Calibrator
VIII.	PRESSURE DEVICES			
1.	Analogue / Digital Pressure Gauges, Pressure Transducer with Indicator, Magnahelic Gauges, Pressure Switches #	25 mmH ₂ O to 200 mmH ₂ O (245 Pa to 1961 Pa) 0 to 1 bar 0 to 7 bar	0.4 mmH ₂ O (3.92 Pa) 3.0 mbar 0.35 % rdg.	Using Digital Manometer / Digital Pressure Gauge with Pneumatic Comparator BY Comparison Method as per DKD R6-1.
2.	Analogue / Digital Pressure Gauges, Pressure Transducer with Indicator, Pressure Switches #	0 to 70 bar 0 to 700 bar	0.35 % rdg. 0.86 % rdg.	Using Digital Pressure Gauge with Hydraulic Comparator. BY Comparison Method as per DKD R6-1.

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 15 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
3.	Analogue / Digital Vacuum Gauge, Transducer with Indicator, Vacuum Switches #	(-) 25 mmH ₂ O to (-)140 mmH ₂ O (-245 Pa to (-)1373 Pa) 0 to -0.9 bar	0.42 mmH ₂ O [@] (4.12 Pa) 0.93 % rdg.	Using Digital Manometer / Digital Pressure/ Vacuum Gauge with Vacuum Comparator. BY Comparison Method as per DKD R6-2
IX.	WEIGHTS			
1.	Mass \$ Calibration of Weights (Conventional Mass) Of M1 Class and coarser	1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g	0.03 mg 0.03 mg 0.03 mg 0.03 mg 0.03 mg 0.03 mg 0.03 mg 0.03 mg 0.03 mg 0.03 mg 0.03 mg 0.04 mg 0.05 mg 0.05 mg 0.08 mg 0.17 mg 0.22 mg	Using F ₂ Accuracy Class Standard Weights and Digital Weighing Balance (Readability: 0.01 mg / 0.1 mg) By Comparison as per OIML R-111 by substitution method and ABBA cycle.
2.	Discrete Weight \$ (Conventional Non Std. Mass i.e. Hook / Disc / Square)	1 g to 200 g 200 g to 2 kg 2 kg to 5 kg 5 kg to 20 kg	0.3 mg 171.3 mg 328.3 mg 631.7 mg	Using F ₂ & M1 Accuracy Class Standard Weights and Digital Weighing Balance (Readability: 0.01 mg / 0.1 mg / 0.2 g / 0.5 g / 1 g)

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905 **Page** 16 of 19

Validity 03.12.2018 to 02.12.2020 **Last Amended on** -

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
X.	VOLUME			
1.	Micropipette [§] (Piston Operated Pipette)	10 µl to 100 µl 100 µl to 1000 µl 1000 µl to 5000 µl	0.10 µl 0.18 µl 0.44 µl	Using Standard weights & Semi-micro balance and distilled water of known density as per Gravimetric method based on ISO855-6
2.	Laboratory Glassware [§] (Pipette , Burette , Flask , Jar , Cylinder etc)	1 ml to 10 ml >10 ml to 100 ml >100 ml to 200 ml >200 ml to 2 liter >2 liter l to 5 liter >5 liter to 20 liter	0.0003 ml 0.0005 ml 0.0010 ml 0.42 ml 0.50 ml 0.61 ml	Using Standard weights & semi-micro balance/ Analytical balance, and distilled water of known density as per Gravimetric Method Based on ISO 4787
XI.	WEIGHING SCALE AND BALANCE			
1.	Weighing Machine * Readability :0.1 mg & Coarser	0 to 220 g	0.22 mg	Using Standard Weights of Class F2 as per Procedure based on OIML R76 (2006) calibration of Class II Weighing Balance & Coarser
	Readability: 0.1 g & Coarser	0 to 2 kg	130 mg	Using Standard Weights of Class M1 as per Procedure based on OIML R76 (2006) calibration of Class III Weighing Balance & Coarser
	Readability: 0.2 g & Coarser	Upto 10 kg	220 mg	
	Readability: 0.5 g & Coarser	Upto 60 kg	560 mg	

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 17 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

"In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020"

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>THERMAL CALIBRATION</u>				
I.	TEMPERATURE			
1.	Glass Thermometer, Digital / Dial Thermometer , RTD & Thermocouple with or without (Controller / Indicator / Data Logger / Recorder / Temperature / Transmitter) #	(-) 30°C to 50°C 50°C to 250°C	0.18 °C 0.18 °C	Using Standard PRT & 6.1/2 Digit DMM, Liquid bath / Oil bath by Comparison Method
2.	Digital Dial Thermometer , RTD & Thermocouple with or without (Controller / Indicator / Data Logger / Recorder / Temperature / Transmitter) #	250°C to 600°C 600°C to 1000°C 1000°C to 1200°C	2.5 °C 2.5 °C 2.5 °C	Using R Type Thermocouple & 6.1/2 / 5.1/2 Digit DMM , Oil bath and Dry Block Furnace by Comparison Method

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 18 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

"In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020"

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
3.	Non Contact Thermometry Optical Pyrometer , IR Thermometer \$	50°C to 500°C	4.0 °C	Using Digital pyrometer with Black Body Calibrator By Comparison Method
4.	Temperature indicator with sensor liquid bath, Dry Block Furnace , Oven , Environmental Chamber, Centrifuge , Incubator(non medical), BOD Incubator , Freezer & Deep Freezer *	(-)30 °C to 250°C 250 °C to 400°C	0.2 °C 2.5 °C	Single Point Calibration Using Std. PRT & R Type Thermocouple & 5.1/2 Digit DMM using Liquid bath / Oil bath / Dry Block . By Comparison Method
5.	Temperature Indicator with sensor of oven , Dry Block Furnace , Muffle Furnace *	50°C to 600°C 600°C to 1200°C	2.5 °C 2.5 °C	Single Point Calibration Using Type R Thermocouple & 5.1/2 Digit DMM By Comparison Method
6.	Freezer , Oven , Cold Chamber , Centrifuged Chamber , Incubator(non medical) , BOD Incubator(non medical) , Closed Chamber , Furnace *	(-)30 °C to 50°C 50 °C to 300°C	1.2 °C 2.9 °C	9 Point Calibration Using Multi point Data Logger with RTD (Pt-100) sensors By Multi Point Direct Method

Mohit Kaushik
Convenor

Avijit Das
Program Manager

Laboratory Precision Calibration and Testing Centre, 167/8 First Floor, Main Road, Sarai Jullena, New Delhi

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2905

Page 19 of 19

Validity 03.12.2018 to 02.12.2020

Last Amended on -

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
II.	SPECIFIC HEAT AND HUMIDITY			
1.	Thermo Hygrometer with sensor & Inbuilt Sensor [§]	30% RH to 95% RH @ 25 °C 15 °C to 50 °C @ 50% RH	1.40% RH @ 25 °C 0.8 °C @ 50% RH	Using Digital Thermo hygrometer with sensor & Humidity Chamber By Comparison Method
2.	Indicator of Environmental/ Conditioning Chamber/ Humidity Chamber [*]	30 % RH to 95 % RH @25 °C	1.50% RH @25 °C	Single Point Calibration Using Digital Thermo hygrometer with sensor By Comparison Method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

[§] Only in Permanent Laboratory

[^] Only for Site Calibration

[#] The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

[°] Laboratory can also calibrate instruments/devices of coarser resolution / least count within the accredited range using same reference standard/ master equipment under the scope of accreditation.

Mohit Kaushik
Convenor

Avijit Das
Program Manager