

Laboratory Procom Systems, Survey No. 30/4, Second Floor, Dhayari, Pune, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2802 Page 1 of 5

Validity 16.08.2018 to 15.08.2020 Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>ELECTRO TECHNICAL CALIBRATION</u>				
I.	SOURCE			
1.	DC Voltage [#]	1 mV to 100 mV 100 mV to 1000 V	1.43 % to 0.13 % 0.13 %	Using 5½ Digit Zeal Multi Function Calibrator By Direct method
2.	DC Current [#]	0.2 mA to 200 mA 200 mA to 10 A 10 A to 900 A	0.8 % to 0.15 % 0.15 % to 0.24 % 1.01 % to 1.2 %	Using 5½ Digit Zeal Multi Function Calibrator By Direct method Using 5½ Digit Zeal Multi Function Calibrator with Current Coil By Direct method
3.	AC Voltage [#]	50 Hz 5 mV to 200 mV 200 mV to 1000 V	1.8 % to 0.4 % 0.4 % to 0.22 %	Using 5½ Digit Zeal Multi Function Calibrator By Direct method
4.	AC Current [#]	50 Hz 0.2 mA to 200 mA 200 mA to 10 A 10 A to 1000 A	1.2 % to 0.5 % 0.5 % to 0.6 % 2.08 % to 2.50 %	Using 5½ Digit Zeal Multi Function Calibrator By Direct method Using 5½ Digit Zeal Multi Function Calibrator with Current Coil By Direct method

Shally Sharma
Convenor

Anuja Anand
Program Manager

Laboratory Procom Systems, Survey No. 30/4, Second Floor, Dhayari, Pune, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2802

Page 2 of 5

Validity 16.08.2018 to 15.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
5.	Resistance [#]	1 Ω to 1000 M Ω	0.74 % to 2.5 %	Using Decade Resistance Box By Direct method
6.	Frequency [#]	45 Hz to 1kHz	0.311 % to 0.09 %	Using 5½ Digit Zeal Multi Function Calibrator By Direct method
7.	Temperature Simulation [#]			Using Yokogawa CA71 Multi Function Calibrator By Direct method.
	RTD	(-) 200°C to 800 °C	0.85 °C	
	Thermocouple			
	K Type	(-) 200°C to 1300 °C	2.42 °C	
	J Type	(-) 200°C to 1200 °C	2.42 °C	
	T Type	(-) 100°C to 400 °C	2.41 °C	
	R Type	400°C to 1200 °C	3.71 °C	
	S Type	300°C to 1300 °C	3.0 °C	
	B Type	600°C to 1350 °C	3.28 °C	
	N Type	(-) 100°C to 1200 °C	2.41 °C	
II.	MEASURE			
1.	DC Voltage [#]	10 mV to 100 mV 100 mV to 900 V	0.6 % to 0.02 % 0.02 % to 0.05 %	Using 6½ Picotest DMM M3510A By Direct Method
2.	DC Current [#]	1 mA to 100 mA 100 mA to 9 A	0.29 % to 0.07 % 0.07 % to 0.35 %	Using 6½ Picotest DMM M3510A By Direct Method
3.	Resistance [#] 2 Wire /4 Wire	100 Ω to 400 Ω 400 Ω to 10 M Ω	1.50 % to 0.06 % 0.06 % to 1.09 %	Using 6½ Picotest DMM M3510A By Direct Method
4.	Frequency [#]	100 Hz to 1 kHz 1 kHz to 100 kHz	0.1 % to 0.02 % 0.02 %	Using 6½ Picotest DMM M3510A By Direct Method

Shally Sharma
Convenor

Anuja Anand
Program Manager

Laboratory Procom Systems, Survey No. 30/4, Second Floor, Dhayari, Pune, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2802

Page 3 of 5

Validity 16.08.2018 to 15.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
5.	Stop Watch [#]	6 sec to 60 min	0.32 s to 2.5 s	Using Digital Stop Watch By Comparison method
6.	Temperature Simulation [#]			Using 6½ DMM Picotest DMM M3510A By Direct method.
	Thermocouple			
	K Type	(-)100°C to 1100 °C	1.23 °C	
	J Type	(-)100°C to 700 °C	1.23 °C	
	T Type	(-)100°C to 400 °C	1.75 °C	
	R Type	100°C to 1600 °C	3.06 °C	
	S Type	100°C to 1600 °C	2.17 °C	
	B Type	700°C to 1350 °C	2.36 °C	
	N Type	(-)100°C to 1200 °C	1.54 °C	

Shally Sharma
Convenor

Anuja Anand
Program Manager

Laboratory Procom Systems, Survey No. 30/4, Second Floor, Dhayari, Pune, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2802 **Page** 4 of 5

Validity 16.08.2018 to 15.08.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>THERMAL CALIBRATION</u>				
I.	TEMPERATURE			
1.	RTD & T/C Temperature Sensors (With & Without Indicators / Controllers), Digital Thermometers , Glass Thermometers , Temperature Gauges [#]	50 °C to 250 °C	0.3 °C	Using 4 wire RTD sensor with Handy Digital Calibrator & Liquid bath by Comparison method.
2.	RTD & T/C Temperature Sensors (With & Without Indicators / Controllers), Digital Thermometers , Glass Thermometers , Temperature Gauges [#]	250 °C to 500 °C 500 °C to 1200 °C	1.2 °C 3.5 °C	Using R Type T/C sensor and Handy Digital Calibrator with Dry Well Bath by Comparison Method.
3.	Thermal/Climatic Chambers, Hot Air Ovens, Industrial Autoclaves* (Non-Medical Use)	(-) 80 °C to 250 °C	1.1°C	Using Multiple RTD Sensors with Data logger by Multipoint mapping method.

Shally Sharma
Convenor

Anuja Anand
Program Manager

Laboratory Procom Systems, Survey No. 30/4, Second Floor, Dhayari, Pune, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2802

Page 5 of 5

Validity 16.08.2018 to 15.08.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
4.	Hot Air Ovens, Chambers, Furnaces*	250°C to 1200°C	4.4°C	Using Multiple N Type T/C Sensors with Data logger by Multi-Point Mapping Method.

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

*Only for Site Calibration

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Shally Sharma
Convenor

Anuja Anand
Program Manager