

Laboratory	Quick Logic Controls,# 79, Industrial Area, Phase 2, Panchkula, Haryana		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Mechanical Calibration	Issue Date	16.06.2014
Certificate Number	C-1084	Valid Until	15.06.2016
Last Amended on	-	Page	1 of 1

	Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
I	PRESSURE			
1.	PRESSURE GAUGE^{\$}	1 to 25 kg/cm ²	0.1kg/cm ²	Using Digital Pressure Gauge and water based Comparator Pump. By Comparison Method

* Measurement Capability is expressed as an uncertainty (±) at a confidence probability of 95%

^{\$}Only in Permanent Laboratory