

Laboratory R. R. Electronics, Door No. 3/459, Plot No. 22, Sri Venkateshwara Nagar, 1st Street, Kottivakkam (ECR), Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration **Issue Date** 20.02.2015

Certificate Number C-0334 **Valid Until** 19.02.2017

Last Amended on 19.03.2015 **Page** 1 of 5

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
<u>SOURCE</u>			
1. DC VOLTAGE [#]	1 mV to 10 V	0.33 % to 0.0016 %	Using Multiproduct Calibrator by Direct Method
	10 V to 1000 V	0.0016 % to 0.0023 %	
2. DC CURRENT [#]	100 μ A to 1 mA	0.031 % to 0.012 %	Using Multiproduct Calibrator by Direct Method
	1 mA to 30 A	0.012 % to 0.070 %	
	20 A to 1500 A	0.82 % to 0.068 %	Using Multiproduct Calibrator with Current Coil by Direct Method
3. RESISTANCE [#]	1 Ω	0.82 %	Using Multiproduct Calibrator by Direct Method
	10 Ω	0.08 %	
	100 Ω	0.01 %	
	1 k Ω	0.004 %	
	10 k Ω	0.0028 %	
	100 k Ω	0.0039 %	
	1 M Ω	0.0048 %	
	10 M Ω	0.016 %	
	100 M Ω	0.29 %	
	1 G Ω	1.56 %	
	1 Ω to 100 Ω	1.40 % to 0.02 %	Using Multifunction Calibrator by Direct Method
100 Ω to 5 k Ω	0.02 % to 0.0086 %		
4. CAPACITANCE [#]	1nF	0.28 %	Using Multifunction Calibrator by Direct Method
	10nF	0.30 %	
	100 nF	0.30 %	
	1 μ F	0.47 %	
	10 μ F	0.70 %	
5. FREQUENCY [#]	10 Hz to 10 MHz	0.0023 %	Using Multiproduct Calibrator by Direct Method

Ranjith Kumar
Convenor

Avijit Das
Program Manager

Laboratory R. R. Electronics, Door No. 3/459, Plot No. 22, Sri Venkateshwara Nagar, 1st Street, Kottivakkam (ECR), Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration **Issue Date** 20.02.2015

Certificate Number C-0334 **Valid Until** 19.02.2017

Last Amended on 19.03.2015 **Page** 2 of 5

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
6. TEMPERATURE SIMULATION[#]			
Thermocouple			
J,K,N Type	1°C to 1200°C	0.12 °C to 0.19°C	Using Multifunction Calibrator by Direct Method
R Type	150°C to 1700°C	0.24°C to 0.33°C	
S Type	170°C to 1700°C	0.24°C to 0.33°C	
B Type	600°C to 920°C	1.93 °C to 1.52°C	Using Multifunction Calibrator by Direct Method
	920°C to 1800 °C	0.45°C to 0.56°C	
RTD (Pt-100)	(-)200°C to 800°C	0.07°C to 0.11°C	Using Multifunction Calibrator by Direct Method
7. AC VOLTAGE[#]			
	40 Hz to 1 kHz		Using Multiproduct Calibrator by Direct Method
	20 mV to 200 V	0.2 % to 0.09 %	
	200 V to 1000 V	0.09 % to 0.12 %	
8. AC CURENT[#]			
	40 Hz to 1 kHz		Using Multiproduct Calibrator by Direct Method
	20µA to 2 A	2.34 % to 0.31 %	
	2 A to 20 A	0.31 % to 0.6 %	
	50 Hz to 1 kHz		Using Multiproduct Calibrator with Current Coil by Direct Method
	20 A to 30 A	0.6 %	
	50 Hz		Using Multiproduct Calibrator with Current Coil by Direct Method
	20 A to 1500 A	0.78 % to 0.06 %	
<u>MEASURE</u>			
1. DC VOLTAGE^{\$}			
	1 mV to 10 V	0.02 % to 0.0007 %	Using Precision Multimeter by Direct Method
	10 V to 1000 V	0.0007 % to 0.001 %	

Ranjith Kumar
Convenor

Avijit Das
Program Manager

Laboratory R. R. Electronics, Door No. 3/459, Plot No. 22, Sri Venkateshwara Nagar, 1st Street, Kottivakkam (ECR), Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration **Issue Date** 20.02.2015

Certificate Number C-0334 **Valid Until** 19.02.2017

Last Amended on 19.03.2015 **Page** 3 of 5

	Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
2.	DC CURRENT ^{\$}	100 μ A to 1 A 1 A to 30 A	0.0033 % to 0.02 % 0.02 % to 0.074 %	Using Precision Multimeter by Direct Method
3.	RESISTANCE ^{\$}	1 Ω to 1k Ω 1 k Ω to 1 G Ω	0.0025 % to 0.0012 % 0.0012 % to 0.21 %	Using Precision Multimeter by Direct Method
4.	CAPACITANCE ^{\$}	1nF to 10 μ F	3.27 % to 1.86 %	Using 6 ½ digit DMM by Direct Method
5.	FREQUENCY ^{\$}	10 Hz to 300 kHz	0.58 % to 0.012 %	Using 6 ½ digit DMM by Direct Method
6.	AC VOLTAGE ^{\$}	40 Hz to 1 kHz 20 mV to 100 V 100 V to 1000 V	0.079 % to 0.035 % 0.035 %	Using Precision Multimeter by Direct Method
7.	AC CURRENT ^{\$}	40 Hz to 1 kHz 20 μ A to 100 mA 100 mA to 20 A 50 Hz to 1 kHz 20 A to 30 A	0.15 % to 0.09 % 0.09 % to 0.16 % 0.12 %	Using Precision Multimeter by Direct Method
8.	TIMER [#]	1 s to 10 s 10 s to 9999.9 s	0.9 % to 0.09 % 0.09 % to 0.0064 %	Using Time Interval Meter by Direct Method
9.	DC VOLTAGE [*]	1 mV to 1000 V	0.42 % to 0.006 %	Using 6½ digit DMM & Picotest by Direct Method
10.	DC CURRENT [*]	1 mA to 100 mA 100 mA to 10 A	0.29 % to 0.06 % 0.06 % to 0.35 %	Using 6½ digit DMM by Direct Method

Ranjith Kumar
Convenor

Avijit Das
Program Manager

Laboratory R. R. Electronics, Door No. 3/459, Plot No. 22, Sri Venkateshwara Nagar, 1st Street, Kottivakkam (ECR), Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration **Issue Date** 20.02.2015

Certificate Number C-0334 **Valid Until** 19.02.2017

Last Amended on 19.03.2015 **Page** 4 of 5

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
11. RESISTANCE*	1 Ω to 1M Ω 1 M Ω to 100 M Ω	0.48 % to 0.013 % 0.013 % to 0.94 %	Using 6½ digit DMM by Direct Method
12. CAPACITANCE*	1 nF to 10 μ F	3.27 % to 1.86 %	Using 6½ digit DMM by Direct Method
13. FREQUENCY*	10 Hz to 300 kHz	0.58 % to 0.012 %	Using 6½ digit DMM by Direct Method
14. TEMPERATURE SIMULATION#			
Thermocouple			
J,K,N Type	1°C to 1200°C	0.14 °C to 0.19°C	Using Multifunction Calibrator by Direct Method
R Type	150°C to 1700°C	0.28°C to 0.34°C	
S Type	170°C to 1700°C	0.27°C to 0.35°C	
B Type	600°C to 920°C	0.61°C	Using Precision Multimeter & Multifunction Calibrator by Direct Method
	920°C to 1800 °C	0.47°C to 0.56°C	
RTD (Pt-100)	(-)200°C to 800°C	0.07°C to 0.11°C	Using Multifunction Calibrator by Direct Method
15. AC VOLTAGE*	40 Hz to 1 kHz 100 mV to 100 V 100 V to 750 V	0.12 % to 0.10 % 0.10 % to 0.11 %	Using 6½ digit DMM by Direct Method

Ranjith Kumar
Convenor

Avijit Das
Program Manager

Laboratory R. R. Electronics, Door No. 3/459, Plot No. 22, Sri Venkateshwara Nagar, 1st Street, Kottivakkam (ECR), Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration **Issue Date** 20.02.2015

Certificate Number C-0334 **Valid Until** 19.02.2017

Last Amended on 19.03.2015 **Page** 5 of 5

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (\pm)	Remarks
16. AC CURRENT*	40 Hz to 1 kHz 0.1 A to 10 A	0.58 % to 0.96 %	Using 6½ digit DMM by Direct Method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

\$Only in Permanent Laboratory

*Only for Site Calibration

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Ranjith Kumar
Convenor

Avijit Das
Program Manager