

**Laboratory** Rectifier Station Laboratory, Hindalco Industries Limited, P.O.Renukoot,  
Dist Sonbhadra, Uttar Pradesh

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro Technical Calibration **Issue Date** 22.05.2014

**Certificate Number** C-0228 **Valid Until** 21.05.2016

**Last Amended on** - **Page** 1 of 3

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
<b>SOURCE</b>			
<b>1. DC VOLTAGE<sup>\$</sup></b>	1mV - 300mV 300mV - 3V 3V - 30V 30V - 300V 300V - 1kV	1% - 0.02% 0.02% - 0.03% 0.03% - 0.1% 0.1% - 0.015% 0.015% - 0.015%	Wavetek Calibrator 9100 Direct Method
<b>2. AC VOLTAGE<sup>\$</sup></b>	<b>50 Hz</b> 30mV - 300mV 300mV - 3V 3V - 300V 300V - 700V	0.45% - 0.16% 0.16% 0.16% - 0.1% 0.1% - 0.30%	Wavetek Calibrator 9100 Direct Method
<b>3. DC CURRENT<sup>\$</sup></b>	3mA - 30mA 30mA - 300mA 300mA - 3A 3A -10A	0.35% - 0.10% 0.10 % 0.10% - 0.11% 0.11%	Wavetek Calibrator 9100 Direct Method
<b>4. AC CURRENT<sup>\$</sup></b>	<b>50 Hz</b> 3mA-30mA 30mA -300mA 300mA -3A 3A -10A	0.35% - 0.11% 0.11% 0.11% - 0.15% 0.15 - 0.30%	Wavetek Calibrator 9100 Direct Method
<b>5. FREQUENCY<sup>\$</sup></b>	10Hz -1MHz	0.07% - 0.03%	Wavetek Calibrator 9100 Direct Method
<b>6. RESISTANCE<sup>\$</sup></b>	40Ω - 400Ω 400Ω - 4kΩ 4kΩ - 400kΩ 400kΩ - 4MΩ 4MΩ - 40MΩ 40MΩ - 300MΩ	0.07% - 0.03% 0.03% 0.03% - 0.13% 0.13% - 0.085% 0.085% - 0.24% 0.24% - 1.60%	Wavetek Calibrator 9100 Direct Method

**Laboratory** Rectifier Station Laboratory, Hindalco Industries Limited, P.O.Renukoot,  
 Dist Sonbhadra, Uttar Pradesh

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro Technical Calibration **Issue Date** 22.05.2014

**Certificate Number** C-0228 **Valid Until** 21.05.2016

**Last Amended on** - **Page** 2 of 3

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>MEASURE</b>			
<b>7. AC Active &amp; Reactive ENERGY/ POWER<sup>\$</sup> at 40 – 70Hz &amp; Pf: + 0.5 to - 0.866</b>	60 V – 240 V & 10 mA – 5 A		Precision Mesuring Instrument Make::ZERA MODEL:TOPZ303 By comparison method
	<u>ACTIVE</u> 1-Phase 0.3W – 4.8kW	0.08%	
	3- Phase 0.9W – 14.4kW	0.08%	
	<u>REACTIVE</u> 1-Phase 0.3W – 4.8kW	0.085%	
	3- Phase 0.9W – 14.4kW	0.085%	
	<b>8. AC VOLTAGE<sup>\$</sup></b>	<b>50Hz</b> 40V- 320V \$	0.035%
<b>9. AC CURRENT<sup>\$</sup></b>	50 Hz 5 mA – 20 A	0.055 – 0.035 %	Precision Mesuring Instrument Make::ZERA MODEL:TOPZ303 By comparison method

**Laboratory**                      **Rectifier Station Laboratory, Hindalco Industries Limited, P.O.Renukoot,  
Dist Sonbhadra, Uttar Pradesh**

**Accreditation Standard**   **ISO/IEC 17025:2005**

**Discipline**                      **Electro Technical Calibration**                      **Issue Date**    **22.05.2014**

**Certificate Number**        **C-0228**    **Valid Until**    **21.05.2016**

**Last Amended on**         -    **Page**             **3 of 3**

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
10. FREQUENCY <sup>\$</sup>	60 V 40Hz - 70Hz	0.075%	Precision Mesuring Instrument Make::ZERA MODEL:TOPZ303 By comparison method

\* Measurement Capability is expressed as an uncertainty (±) at a confidence probability of 95%

<sup>\$</sup>Only in Permanent Laboratory