

Laboratory **Roots Metrology & Testing Laboratory (A Unit of Roots Industries India Limited) RKG Industrial Estate, Ganapathy, Coimbatore, Tamil Nadu**

Accreditation Standard **ISO/IEC 17025:2005**

Discipline **Electro-Technical Calibration** **Issue Date** **26.06.2014**

Certificate Number **C-0826** **Valid Until** **25.06.2016**

Last Amended on **03.07.2014** **Page** **1 of 7**

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
----------------------------------	-------------------	--	---------

SOURCE

1. DC Voltage^s	10 mV to 100 mV	0.03 % to 0.004 %	Using Multi-product calibrator Make : Transmille Model : 3010 A by Direct Method
	100 mV to 1 V	0.004 % to 0.0015 %	
	1 V to 10 V	0.0015 % to 0.0012 %	
	10 V to 100 V	0.0012 % to 0.0018 %	
	100 V to 990 V	0.0018 % to 0.0017 %	
2. AC Voltage^s	40 Hz to 10kHz		Using Multi-product calibrator Make : Transmille Model : 3010 A by Direct Method
	20 mV to 100 mV	0.32 % to 0.12 %	
	100 mV to 1 V	0.12 % to 0.082 %	
	1V to 10 V	0.082 % to 0.080 %	
	10 V to 700 V	0.080 % to 0.058 %	
3. DC Current^s	56Hz		Using Multi-product calibrator Make : Transmille Model : 3010 A by Direct Method
	700 V to 990 V	0.058 % to 0.045 %	
	10 µA to 100 µA	0.13 % to 0.03 %	
	100 µA to 1 mA	0.02 % to 0.01 %	
	1 mA to 100 mA	0.01%	
4. AC Current^s	100 mA to 1 A	0.01 % to 0.02 %	Using Multi-product calibrator Make : Transmille Model : 3010 A by Direct Method
	1 A to 10 A	0.02 % to 0.06 %	
	10 A to 1500 A	0.06 % to 0.08 %	
4. AC Current^s	40 Hz to 1 kHz		Using Multi-product calibrator Make : Transmille Model : 3010 A by Direct Method
	25 µA to 100 µA	3.12 % to 1.21 %	
	100 µA to 1 mA	1.21 % to 0.61 %	
	1 mA to 100 mA	0.61 % to 0.62 %	
	100 mA to 2A	0.62 % to 0.63 %	

Laboratory **Roots Metrology & Testing Laboratory (A Unit of Roots Industries India Limited) RKG Industrial Estate, Ganapathy, Coimbatore, Tamil Nadu**

Accreditation Standard **ISO/IEC 17025:2005**

Discipline **Electro-Technical Calibration** **Issue Date** **26.06.2014**

Certificate Number **C-0826** **Valid Until** **25.06.2016**

Last Amended on **03.07.2014** **Page** **2 of 7**

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
	56Hz 1 A to 30 A 30 A to 1500 A	0.63 % to 0.36 % 0.36 % to 0.35 %	Using Current Coil
5. Resistance^s	1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 100 MΩ 1 MΩ to 1000 MΩ	1.42 % to 0.61 % 0.61 % to 0.01 % 0.01 % to 0.003 % 0.003 % to 0.0015 % 0.0015 % to 0.0028 % 0.0028 % to 0.0037 % 0.0037 % to 0.012 % 0.012 % to 0.21 % 2 % to 2.39 %	Using Muti-product calibrator Make : Transmille Model : 3010 A by Direct Method Using High Stability Decade Meg ohm box Make : Vaiseshika Model : 8400 HV by Direct Method
	10mΩ to 100mΩ 100mΩ to 1Ω	2.65 % to 0.58 % 0.58 % to 0.59 %	Using High Precision Decade Resistance Box Make : Vaiseshika Model : 7400 by Direct Method
6. Capacitance^s	1 kHz 1 nF to 10 nF 10 nF to 100 nF 100 nF to 1 μF 1 μF to 10μF 10 μF to 100μF 100 μF to 1 mF 1 mF to 10 mF	0.79 % to 0.32 % 0.32 % to 0.30 % 0.30 % to 0.47 % 0.47 % to 0.69 % 0.69 % to 0.70 % 0.70 % to 1.2 % 1.20%	Using Muti-product calibrator Make : Transmille Model : 3010 A by Direct Method

Laboratory **Roots Metrology & Testing Laboratory (A Unit of Roots Industries India Limited) RKG Industrial Estate, Ganapathy, Coimbatore, Tamil Nadu**

Accreditation Standard **ISO/IEC 17025:2005**

Discipline **Electro-Technical Calibration** **Issue Date** **26.06.2014**

Certificate Number **C-0826** **Valid Until** **25.06.2016**

Last Amended on **03.07.2014** **Page** **3 of 7**

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
7. Inductance^{\$}	1 kHz 1 mH to 10 mH 10 mH to 100 mH 100 mH to 1 H 1 H to 10 H	0.94 % to 0.70 % 0.70% 0.70% to 0.94% 0.94% to 0.67%	Using Mutiproduct calibrator Make : Transmille Model : 3010 A by Direct Method
8. Frequency^{\$}	100 Hz to 10 MHz	0.0031%	Using Mutiproduct calibrator Make : Transmille Model : 3010 A by Direct Method
9. Oscilloscope^{\$}			
a. Amplitude ^{\$} (DC)	DC,2 mV/ div to 500 mV/ div @ 1 MΩ DC,1 V/ div to 20 V/ div@ 1 MΩ	0.08 % to 0.02 % 0.03 % to 0.01 %	Using Mutiproduct calibrator Make : Transmille Model : 3010 A with scope option by Direct Method
b. Time Base^{\$}	20 ns/ div to 500 ns/ div 1 μs/ div to 500 μs/ div 1 ms/ div to 500 ms/ div 1 s/ div	0.006 % to 0.070 % 0.006 % to 0.070 % 0.006 % to 0.070 % 0.09%	
c. Band width ^{\$} (Level Flatness)	Upto 600 MHz	1.0 dB	Using Mutiproduct calibrator Make : Transmille Model : 3010 A with scope option by Direct Method

Laboratory Roots Metrology & Testing Laboratory (A Unit of Roots Industries India Limited) RKG Industrial Estate, Ganapathy, Coimbatore, Tamil Nadu

Accreditation Standard ISO/IEC 17025:2005

Discipline Electro-Technical Calibration **Issue Date** 26.06.2014

Certificate Number C-0826 **Valid Until** 25.06.2016

Last Amended on 03.07.2014 **Page** 4 of 7

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
10. Temperature Simulation[#] (Indicator/ Controller/ Recorder/Calibrator)			
a. Thermocouple Type			
K Type	-190°C to 1370°C	0.25°C to 0.42°C	Using Multiproduct Calibrator Make : Transmille Model : 3010A & T.C. Simulator Transmille option : EA001A by Direct Method
J Type	-200°C to 1200°C	0.22°C to 0.37°C	
T Type	-240°C to 400°C	0.47°C to 0.26°C	
R Type	0°C to 1700°C	0.6°C to 1.18°C	
S Type	0°C to 1700°C	0.65°C to 1.63°C	
E Type	-250°C to 1000°C	0.38°C to 0.26°C	
N Type	-190°C to 1300°C	0.33°C to 0.48°C	
B Type	600°C to 1820°C	0.46°C to 1.82°C	
L Type	-200°C to 900°C	0.26°C	
U Type	-200°C to 600°C	0.28°C to 0.26°C	
b. RTD Type			
PT100	-200° C to 800°C	0.17°C to 0.64°C	
<u>MEASURE</u>			
1. DC Voltage[§]	10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 900 V	0.013% to 0.009% 0.009% to 0.005% 0.009% to 0.005% 0.009% to 0.005% 0.009% to 0.005%	Using 6½ Digit Precision Multimeter Make : Fluke Model :8846 A by Direct Method
2. DC High Voltage[§]	1 kV to 5 kV	3.60% to 3.65%	Using Digital Multimeter HP973A and High Voltage Probe Make : Fluke Model : 80 K-6 by Direct Method

Laboratory **Roots Metrology & Testing Laboratory (A Unit of Roots Industries India Limited) RKG Industrial Estate, Ganapathy, Coimbatore, Tamil Nadu**

Accreditation Standard **ISO/IEC 17025:2005**

Discipline **Electro-Technical Calibration** **Issue Date** **26.06.2014**

Certificate Number **C-0826** **Valid Until** **25.06.2016**

Last Amended on **03.07.2014** **Page** **5 of 7**

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
3. AC Voltage^s	50 Hz		
	10 mV to 100 mV	0.62% to 0.50%	Using 6½ Digit Precision Multimeter Make : Fluke Model : 8846 A by Direct Method
	100 mV to 1 V	0.50% to 0.41%	
	1V to 10 V	0.41%	
10 V to 900 V	0.41% to 0.97%		
4. AC High Voltage^s	50Hz		
1kV to 5 kV	6.73% to 6.74%	Using Digital Multimeter HP973A and High Voltage Probe Make : Fluke Model: 80 K-6 by Direct Method	
5. DC Current^s	10 µA to 100 µA	0.35% to 0.06%	Using 6½ Digit Precision multimeter Make : Fluke Model : 8846 A
	100 µA to 1 mA	0.12% to 0.06%	
	1 mA to 100 mA	0.12% to 0.06%	
	100 mA to 1 A	0.12% to 0.08%	
	1 A to 9 A	0.12% to 0.06%	
6. AC Current^s	50 Hz		
	10 µA to 100 µA	0.81% to 0.71%	Using 6½ Digit Precision multimeter Make : Fluke Model : 8846 A
	100 µA to 1 mA	0.71% to 0.58%	
	1 mA to 100 mA	0.58% to 0.18%	
	100 mA to 9A	0.18% to 0.21%	
10 A to 1500 A	11.60% to 2.93%		
7. Resistance^s	10 Ω to 100 Ω	0.02%	Using 6½ Digit Precision multimeter Make : Fluke Model : 8846 A by Direct Method
	100 Ω to 1 kΩ	0.013%	
	1 kΩ to 10 kΩ	0.013%	
	10 kΩ to 100 kΩ	0.013%	
	100 kΩ to 1 MΩ	0.013%	
	1 MΩ to 10 MΩ	0.013% to 0.014%	
	10 MΩ to 100 MΩ	0.014% to 0.90%	

Avijit Das
Program Manager

Bibin Philip
Convenor

Laboratory **Roots Metrology & Testing Laboratory (A Unit of Roots Industries India Limited) RKG Industrial Estate, Ganapathy, Coimbatore, Tamil Nadu**

Accreditation Standard **ISO/IEC 17025:2005**

Discipline **Electro-Technical Calibration** **Issue Date** **26.06.2014**

Certificate Number **C-0826** **Valid Until** **25.06.2016**

Last Amended on **03.07.2014** **Page** **6 of 7**

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
8. Inductance^{\$}	1 kHz		
	1 mH to 10 mH	3.92%	Using LCR Meter Make : Mastech Model : MS 5300 by Direct Method
	10 mH to 100 mH	3.92%	
	100 mH to 1 H	3.92%	
	1 H to 10 H	3.92% to 4.13%	
9. Frequency^{\$}	10 Hz to 1 MHz	0.012% to 0.058%	Using 6½ digit Multimeter Make : Fluke Model : 8846 A by Direct Method
10. Temperature Simulation[#] (Indicator/ Controller/ Recorder/ Calibrator)			
a. Thermocouple Type			
K Type	-190°C to 1300°C	0.14°C to 0.13°C	Using Transmille Model : 3010 A T.C. Simulator option : EA001A by Direct Method
J Type	-210°C to 1200°C	0.14°C to 0.13°C	
T Type	-240°C to 390°C	0.27°C to 0.36°C	
R Type	0°C to 1700°C	0.28°C to 0.36°C	
S Type	0° C to 1700°C	0.27°C to 0.32°C	
E Type	-240°C to 900°C	0.19°C to 0.14°C	
N Type	-190°C to 1200°C	0.18°C to 0.11°C	
B Type	600°C to 1800°C	1°C to 4°C	

