

Laboratory S Calibration (Unit of Sakthi Powerr Solution Private Limited),
No. 44, State Bank Colony, Singanallur, Coimbatore, Tamil Nadu

Accreditation Standard ISO/IEC 17025:2005

Discipline Thermal Calibration **Issue Date** 29.09.2014

Certificate Number C-1134 **Valid Until** 28.09.2016

Last Amended on - **Page** 1 of 2

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
----------------------------------	-------------------	--	---------

I. TEMPERATURE

1. RTD's (with & without Indicator), Thermocouples, (with & without Indicator), Data Logger with sensors, Probe Thermometers, Bi- metal Thermometers, Capillary Thermometers, Thermometers, Temperature Gauges, Temperature Indicators with sensors, Recorders with sensor, Temperature Transmitter, Temperature Switches [#]	-40 °C to 0 °C	0.14 °C	By using Standard 61/2 DMM and RTD
	0 °C to 100 °C	0.18 °C	
	100 °C to 300 °C	0.42 °C	By using Standard 61/2 DMM and S-Type Thermocouple
	300 °C to 1000 °C	1.50 °C	
	1000 °C to 1200 °C	2.28 °C	
2. Indicator of Cryostatic bath, Oil bath, Temperature Baths, Furnace, Dry block calibrators ^{\$}	-40°C to 0°C	0.14°C	By using Standard 61/2 DMM and RTD
	0°C to 100°C.	0.18°C	
	100°C to 300°C	0.42°C	By using Standard 61/2 DMM and S-Type Thermocouple
	300°C to 1000°C	1.50°C	
	1000°C to 1200°C	2.28°C	

Sangeeta Kunwar
Convenor

Avijit Das
Program Manager

Laboratory S Calibration (Unit of Sakthi Powerr Solution Private Limited),
No. 44, State Bank Colony, Singanallur, Coimbatore, Tamil Nadu

Accreditation Standard ISO/IEC 17025:2005

Discipline Thermal Calibration **Issue Date** 29.09.2014

Certificate Number C-1134 **Valid Until** 28.09.2016

Last Amended on - **Page** 2 of 2

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
3. Incubators Freezers Oven Furnace* (Single Point Method)	-40°C to 100°C	1.0°C	By using Standard Multifunction calibrator and RTD
	100°C to 400°C	1.0°C	
	400°C to 1000°C	1.50°C	By using Standard Multifunction calibrator and S-Type Thermocouple
	1000°C to 1200°C	2.50°C	

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

\$Only in Permanent Laboratory

*Only for Site Calibration

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Sangeeta Kunwar
Convenor

Avijit Das
Program Manager