

**Laboratory** Sai-Meena Calibration Laboratory Pvt. Ltd., Shop No. 52, Sector-8A, Yash Paradise, Airoli, Navi Mumbai, Maharashtra  
**Location 1-** Shop No. 52, Sector-8A, Yash Paradise, Airoli, Navi Mumbai  
**Location 2-** Unit No. 113, Ashok Service Indl. Estate, Gokul Nagar, Thane (W)

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-technical Calibration **Issue Date** 01.12.2014  
**Certificate Number** C- 1167 **Valid Until** 05.08.2016  
**Last Amended on** 21.12.2015 **Page** 1 of 5

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (±)	Remarks
--------------------------------	------------------	--	---------

**LOCATION 2**

**I. SOURCE**

<b>1. DC VOLTAGE<sup>#</sup></b>	1 mV to 2 V	0.5 % to 0.0041%	Using 5½ Multifunction Calibrator with 6½ DMM by Comparison Method
	2 V to 20 V	0.041 %	
	20 V to 1000 V	0.041 %	
<b>2. DC CURRENT<sup>#</sup></b>	10 µA to 200 µA	0.49 % to 0.10 %	Using 5½ Multifunction Calibrator with 6½ DMM by Comparison Method
	200 µA to 200 mA	0.10 %	
	200 mA to 1 A	0.10% to 0.093%	
	1 A to 10 A	0.093 % to 0.213 %	Using 5½ MFC & Current Coil with 6½ DMM by Comparison Method
	10 A to 1000 A	1.1 %	
<b>3. AC VOLTAGE<sup>#</sup></b>	<b>50 Hz to 1 kHz</b>		Using 5½ Multifunction Calibrator with 6½ DMM by Comparison Method
	1 mV to 100mV	4.7 % to 0.13 %	
	100 mV to 10 V	0.13 % to 0.11 %	
	10 V to 1000 V	0.11 %	
<b>4. AC CURRENT<sup>#</sup></b>	<b>50 Hz</b>		Using 5½ Multifunction Calibrator with 6½ DMM by Comparison Method
	10 µA to 200 µA	0.89 % to 0.23%	
	200 µA to 200 mA	0.23 % to 0.19%	
	200 mA to 1 A	0.19 % to 0.20 %	
	1 A to 10 A	0.20 % to 0.27 %	
<b>5. CAPACITANCE<sup>#</sup></b>	<b>1 kHz</b>		Using Decade Capacitance Box by Direct Method
	1 nF to 1000 nF	3 % to 2.89 %	
	1000 nF to 100 µF	2.89 %	
<b>6. INDUCTANCE<sup>#</sup></b>	<b>1 kHz</b>		Using Decade Inductance Box by Direct Method
	100 µH to 1000 mH	2.93 %	

**Naveen Jangra**  
**Convenor**

**Avijit Das**  
**Program Manager**

**Laboratory** Sai-Meena Calibration Laboratory Pvt. Ltd., Shop No. 52, Sector-8A, Yash Paradise, Airoli, Navi Mumbai, Maharashtra  
**Location 1-** Shop No. 52, Sector-8A, Yash Paradise, Airoli, Navi Mumbai  
**Location 2-** Unit No. 113, Ashok Service Indl. Estate, Gokul Nagar, Thane (W)

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-technical Calibration **Issue Date** 01.12.2014  
**Certificate Number** C- 1167 **Valid Until** 05.08.2016  
**Last Amended on** 21.12.2015 **Page** 2 of 5

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (±)	Remarks
--------------------------------	------------------	--	---------

**LOCATION 2**

<b>7. FREQUENCY / PERIOD<sup>#</sup></b>	25 ns to 100 ms 0.5 Hz to 10 kHz 10 kHz to 1 MHz 1 MHz to 10 MHz 10 MHz to 160 MHz	1.4 % to 0.0021 % 0.0021 % to 0.006 % 0.006 % to 0.0022 % 0.0022 % to 0.00201 %	Using Function Generator by Direct Method
<b>8. RESISTANCE<sup>#</sup></b>	1 mΩ 10 mΩ 100 mΩ 1 Ω 1 Ω to 1 kΩ 1 kΩ to 1 MΩ 1 MΩ to 100 MΩ 100 MΩ to 1000 MΩ 1 GΩ 2 GΩ 5 GΩ 7 GΩ 10 GΩ	0.6 % 0.13 % 0.13 % 0.13 % 1.4 % to 0.12 % 0.12 % 0.12 % to 1.2 % 1.2 % 2.3 % 2.3 % 2.3 % 2.3 % 6.2 %	Using 4W Standard Resistor & Decade Resistance Box by Direct Method  Using Decade Resistance Box by Direct Method  Using High Resistance Jig and Fix Discrete Value Resistor by Direct Method
<b>9. TEMPERATURE SIMULATION<sup>#</sup></b> (Indicator / Controller / Recorder / Universal Calibrator/ Transmitter DTI Thermocouple)			
<b>RTD / PT-100</b>	(-) 200 °C to 800 °C	0.35 °C to 0.78 °C	Using Handheld Multifunction Calibrator by Simulation Method
<b>T-Type Thermocouple</b>	(-) 250 °C to 400 °C	1.5 °C	
<b>N-Type Thermocouple</b>	(-) 200 °C to 1300 °C	0.5 °C	
<b>U-Type Thermocouple</b>	(-) 200 °C to 600 °C	0.76 °C	
<b>E-Type Thermocouple</b>	(-) 200 °C to 1000 °C	0.35 °C	
<b>J-Type Thermocouple</b>	(-) 210 °C to 1200 °C	1.183 °C	
<b>K-Type Thermocouple</b>	(-) 200 °C to 1350 °C	1.267 °C	
<b>C-Type Thermocouple</b>	0 °C to 2316 °C	2.7 °C	
<b>B-Type Thermocouple</b>	600 °C to 1000 °C	1.8 °C	
<b>R-Type Thermocouple</b>	0 °C to 1767 °C	1.6 °C	
<b>S-Type Thermocouple</b>	0 °C to 1767 °C	1.6 °C	

**Naveen Jangra**  
Convenor

**Avijit Das**  
Program Manager

**Laboratory** Sai-Meena Calibration Laboratory Pvt. Ltd., Shop No. 52, Sector-8A, Yash Paradise, Airoli, Navi Mumbai, Maharashtra  
**Location 1-** Shop No. 52, Sector-8A, Yash Paradise, Airoli, Navi Mumbai  
**Location 2-** Unit No. 113, Ashok Service Indl. Estate, Gokul Nagar, Thane (W)

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-technical Calibration **Issue Date** 01.12.2014  
**Certificate Number** C- 1167 **Valid Until** 05.08.2016  
**Last Amended on** 21.12.2015 **Page** 3 of 5

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability ( $\pm$ )	Remarks
--------------------------------	------------------	--	---------

**LOCATION 2**

<b>11. AC ACTIVE POWER#</b> (UPF, 3-Phase)	<b>50 Hz</b> 50V to 300 V 0.2 A to 6 A	1.161%	Using 3 $\phi$ Power/Energy Meter Calibrator with 3 $\phi$ Load Manager by comparison method
---	--	--------	--

<b>12. OSCILLOSCOPE#</b> Amplitude (Sine Wave / Square Wave)	3 mV to 100 mV	7.18 % to 1.1 %	Using 2 Channel Arbitrary Waveform and Function Generator by Direct Method
	100 mV to 10 V	1.1 % to 1.6 %	
	Time Base Bandwidth	25 ns to 100 ms 50 kHz to 150 MHz	

**II. MEASURE**

<b>1. DC VOLTAGE#</b>	1 mV to 2 V	0.5 % – 0.0065 %	Using Fluke 6 $\frac{1}{2}$ DMM by Direct Method
	2 V to 20 V	0.0065 % to 0.01 %	
	20 V to 1000 V	0.01 % to 0.009 %	
	1 kV to 40 kV	2.0%	Using HV Probe with 6 $\frac{1}{2}$ DMM

<b>2. DC CURRENT#</b>	10 $\mu$ A to 200 $\mu$ A	0.37 % to 0.088 %	Using Fluke 6 $\frac{1}{2}$ DMM by Direct Method
	200 $\mu$ A to 200 mA	0.088 % to 0.057 %	
	200 mA to 1 A	0.057 % to 0.13 %	
	1 A to 10 A	0.13 % to 0.20 %	
	10 A to 400 A	1.0%	Using Shunt with 6 $\frac{1}{2}$ DMM Comparison Method

<b>3. AC VOLTAGE#</b>	<b>50 Hz to 1 kHz</b>		Using Fluke 6 $\frac{1}{2}$ DMM by Direct Method
	1 mV to 100 mV	5.4 % to 0.13 %	
	100 mV to 10 V	0.13 % to 0.11 %	
	10 V to 1000 V	0.11 %	
	1 kV to 28 kV (RMS)	2.4%	Using HV Probe with DMM

**Naveen Jangra**  
Convenor

**Avijit Das**  
Program Manager

**Laboratory** Sai-Meena Calibration Laboratory Pvt. Ltd., Shop No. 52, Sector-8A, Yash Paradise, Airoli, Navi Mumbai, Maharashtra  
**Location 1-** Shop No. 52, Sector-8A, Yash Paradise, Airoli, Navi Mumbai  
**Location 2-** Unit No. 113, Ashok Service Indl. Estate, Gokul Nagar, Thane (W)

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-technical Calibration **Issue Date** 01.12.2014  
**Certificate Number** C- 1167 **Valid Until** 05.08.2016  
**Last Amended on** 21.12.2015 **Page** 4 of 5

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (±)	Remarks
<b><u>LOCATION 2</u></b>			
<b>4. AC CURRENT<sup>#</sup></b>	<b>50 Hz</b>		
	10 µA to 200 µA	2.06 % to 0.39 %	Using Fluke 6½ DMM by Direct Method
	200 µA to 200 mA	0.39 % to 0.35 %	
	200 mA to 1 A	0.35 % to 0.2 %	
	1 A to 10 A	0.2 % to 0.27 %	
	10A to 1200A	1.1%	Using Shunt with 6½ DMM by Comparison Method
<b>5. RESISTANCE<sup>#</sup></b>	10 µΩ to 1 Ω	0.8 %	Using 4W Standard Resistor & two 6½ DMM by V/I Method
	1 Ω to 1 kΩ	0.4 % to 0.1 %	Using Fluke 6½ DMM by Direct Method
	1 kΩ to 1 MΩ	0.1 % to 0.1 %	
	1 MΩ to 100 MΩ	0.1 % to 1.2 %	
	100 MΩ to 1000 MΩ	1.2 % to 2.3 %	
<b>6. CAPACITANCE<sup>#</sup></b>	<b>1 kHz</b>		
	1 nF to 10 µF	0.4 % to 0.7%	Using LCR Meter by Direct Method
<b>7. INDUCTANCE<sup>#</sup></b>	<b>1 kHz</b>		
	100 µH to 1 H	0.5 %	Using LCR Meter by Direct Method
<b>8. TIME<sup>#</sup></b>	5 s to 24 hour	0.4 s to 19 s	Using Digital Timer by Comparison Method
<b>9. FREQUENCY<sup>#</sup></b>	5 Hz to 100kHz	8 ppm	Using Frequency Counter Direct Method
	100 kHz to 160 MHz	8 ppm to 6 ppm	

**Naveen Jangra**  
**Convenor**

**Avijit Das**  
**Program Manager**

**Laboratory** Sai-Meena Calibration Laboratory Pvt. Ltd., Shop No. 52, Sector-8A, Yash Paradise, Airoli, Navi Mumbai, Maharashtra  
**Location 1-** Shop No. 52, Sector-8A, Yash Paradise, Airoli, Navi Mumbai  
**Location 2-** Unit No. 113, Ashok Service Indl. Estate, Gokul Nagar, Thane (W)

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Electro-technical Calibration **Issue Date** 01.12.2014

**Certificate Number** C- 1167 **Valid Until** 05.08.2016

**Last Amended on** 21.12.2015 **Page** 5 of 5

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability ( $\pm$ )	Remarks
--------------------------------	------------------	--	---------

**LOCATION 2**

**10. TEMPERATURE SIMULATION#**

(Indicator / Controller / Recorder / Universal Calibrator/ Transmitter DTI Thermocouple)

T-Type Thermocouple	200 °C to 400 °C	1.5 °C	Using Handheld Multifunction Calibrator by Simulation Method
N-Type Thermocouple	(-) 200 °C to 1300 °C	0.5 °C	
U-Type Thermocouple	(-) 200 °C to 600 °C	0.76 °C	
E-Type Thermocouple	(-) 250 °C to 1000 °C	0.35 °C	
J-Type Thermocouple	(-) 210 °C to 1200 °C	1.22 °C to 1.184 °C	
K-Type Thermocouple	(-) 200 °C to 1350 °C	1.32 °C to 1.27 °C	
C-Type Thermocouple	0 °C to 2316 °C	2.7 °C	
B-Type Thermocouple	600 °C to 1750 °C	1.93 °C to 2.18 °C	
R-Type Thermocouple	0 °C to 1767 °C	1.6 °C	
S-Type Thermocouple	0 °C to 1767 °C	1.6 °C	
RTD PT-100	(-) 200 °C to 800 °C	0.24 °C to 0.78 °C	

\* Measurement Capability is expressed as an uncertainty ( $\pm$ ) at a confidence probability of 95%

# The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Naveen Jangra  
 Convenor

Avijit Das  
 Program Manager