

**Laboratory** Sai-Meena Calibration Laboratory Pvt. Ltd., Shop No. 52, Sector-8A, Yash Paradise, Airoli, Navi Mumbai, Maharashtra  
**Location 1-** Shop No. 52, Sector-8A, Yash Paradise, Airoli, Navi Mumbai  
**Location 2-** Unit No. 113, Ashok Service Indl. Estate, Gokul Nagar, Thane (W)

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Thermal Calibration **Issue Date** 01.12.2014  
**Certificate Number** C- 1168 **Valid Until** 05.08.2016  
**Last Amended on** 21.12.2015 **Page** 1 of 2

Quantity Measured / Instrument	Range / Frequency	* Calibration Measurement Capability ( $\pm$ )	Remarks
--------------------------------	-------------------	--	---------

**LOCATION 2**

**I. TEMPERATURE**

1. LIQUID IN GLASS THERMOMETER, DIAL / DIGITAL THERMOMETER, RECORDER WITH SENSOR, TEMPERATURE SENSORS / GAUGE (RTD, THERMOCOUPLE) WITH & WITHOUT INDICATOR / CONTROLLERS #	(-) 80 °C to 0 °C	0.39 °C to 0.38 °C	Using Oil Bath and Low Temperature Bath with PRT PT-100 and 6 ½ digit DMM by Comparison Method
	0 °C to 250 °C	0.38 °C to 0.31 °C	
2. DIAL / DIGITAL THERMOMETER, RECORDER WITH SENSOR, TEMPERATURE SENSORS / GAUGE (RTD THERMOCOUPLES WITH) & WITHOUT INDICATORS / CONTROLLER #	250 °C to 500 °C	2.1 °C	Using Dry Block Calibrator, S-Type Thermocouple & Semi-Standard PRT (PT-100) and 6½ DMM by Comparison Method
	500 °C to 1000 °C	1.91 °C to 2.9 °C	
	1000 °C to 1200 °C	2.9 °C to 3.1 °C	
3. OVEN, CHAMBERS, INCUBATORS, DEEP FREEZERS, WATER BATH, DRY WELL, OIL BATH, COOLING CABINET, AUTOCLAVE, LIQUID BATHS, MUFFLE FURNANCE, HOT AIR OVEN *	0 °C to 250 °C	1.1 °C	Using Data Logger, RTD PT-100 (9 sensors) & K and S Type Thermocouple Sensor (Single/Multipoint Calibration) by Comparison Method
4. TEMPERATURE MEASUREMENT OF ANALOG / DIGITAL HYGROMETER, HUMIDITY SENSOR WITH OR WITHOUT INDICATOR §	20 °C to 50 °C	0.35 °C to 0.67 °C	Using RTD & Rh Sensor with 6½ DMM by Comparison Method

**Naveen Jangra**  
**Convenor**

**Avijit Das**  
**Program Manager**

**Laboratory** Sai-Meena Calibration Laboratory Pvt. Ltd., Shop No. 52, Sector-8A, Yash Paradise, Airoli, Navi Mumbai, Maharashtra  
**Location 1-** Shop No. 52, Sector-8A, Yash Paradise, Airoli, Navi Mumbai  
**Location 2-** Unit No. 113, Ashok Service Indl. Estate, Gokul Nagar, Thane (W)

**Accreditation Standard** ISO/IEC 17025:2005

**Discipline** Thermal Calibration **Issue Date** 01.12.2014

**Certificate Number** C- 1168 **Valid Until** 05.08.2016

**Last Amended on** 21.12.2015 **Page** 2 of 2

Quantity Measured / Instrument	Range / Frequency	* Calibration Measurement Capability ( $\pm$ )	Remarks
--------------------------------	-------------------	--	---------

**LOCATION 2**

**II. SPECIFIC HEAT AND HUMIDITY**

1. HUMIDITY MEASUREMENT OF ANALOG / DIGITAL HYGROMETER, HUMIDITY SENSOR WITH OR WITH INDICATOR <sup>\$</sup>	20 % Rh to 95 % Rh @ 25 °C	2 % Rh	Using Rh Sensor with Indicator and Humidity Generator By Comparison Method
2. HUMIDITY CHAMBER, SPATIAL WITH TWO SENSOR *	20 % Rh to 95 % Rh @ 25 °C	2 % Rh	Using Data Logger & Humidity Sensor by Comparison Method

\* Measurement Capability is expressed as an uncertainty ( $\pm$ ) at a confidence probability of 95%

<sup>\$</sup> Only in Permanent Laboratory

\* Only for Site Calibration

# The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

**Naveen Jangra**  
**Convenor**

**Avijit Das**  
**Program Manager**