

Laboratory Scientific And Industrial Testing and Research Centre, 83 & 84, Avarampalayam Road, K.R. Puram P.O., Coimbatore, Tamil Nadu

Accreditation Standard ISO/IEC 17025: 2005

Discipline Thermal Calibration **Issue Date** 11.12.2014

Certificate Number C-0912 **Valid Until** 10.12.2016

Last Amended on - **Page** 1 of 1

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
I. TEMPERATURE			
1. LIQUID –IN-GLASS THERMOMETER, RTD, THERMOCOUPLES, TEMPERATURE SENSORS WITH AND WITHOUT INDICATOR [§]	-20°C to 150°C	0.41°C	Using Low / High Temperature Liquid Bath and Four-wire PRT with Indicator
	150°C to 600°C	0.42°C	Using High Temperature Liquid Bath and Four-wire PRT with Indicator
2. RTD, THERMOCOUPLE, TEMPERATURE SENSORS WITH AND WITHOUT INDICATOR [§]	600°C to 1000°C	1.89°C	Using Dry Block Furnace with S-type Thermocouple and Read Unit
3. DEEP FREEZER, INCUBATOR, OVEN, FURNACE, AUTOCLAVE, TEMPERATURE BATHS ETC. AT SINGLE SPECIFIED POSITION [*]	-40°C to 600°C 600°C to 1200°C	1.71°C 2.38°C	Using PRT with Indicator & S- Type Thermocouple with Read Unit
4. DEEP FREEZER, INCUBATOR, OVEN, FURNACE, TEMPERATURE BATHS ETC. AT FIVE POSITIONS INCLUDING CENTRE [*] (Spatial Mapping)	-40°C to 400°C	1.92°C	Using T-type / K-type Thermocouple/RTD and Data Logger

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

[§]Only in Permanent Laboratory

^{*}Only for Site Calibration

Naveen Jangra
Convenor

Avijit Das
Program Manager