

Laboratory Shriram Institute For Industrial Research, 14-15, Sadarmangala
Industrial Area, Whitefield Road, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2083 (in lieu of C-0216, C-0304, C-1497) Page 1 of 7

Validity 13.01.2017 to 12.01.2019 Last Amended on

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>ELECTRO-TECHNICAL CALIBRATION</u>				
I.	MEASURE			
1.	Time ^s (Digital Timer/Analog & Digital Stop Watch)	5 sec to 9900 sec	0.12 sec to 5.8 sec	Using Digital Time Totalizer by Comparison Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Shriram Institute For Industrial Research, 14-15, Sadarmangala Industrial Area, Whitefield Road, Bangalore, Karnataka**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2083 (in lieu of C-0216, C-0304, C-1497)** Page **2 of 7**

Validity **13.01.2017 to 12.01.2019** Last Amended on

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
MECHANICAL CALIBRATION				
I.	WEIGHTS			
1.	Mass ^s (F1 & Coarser)	1 mg	0.02 mg	Using Weights of Accuracy Class E1 & F1 and Precision Balances Analytical Balance by Substitution Method of Weighing and ABBA Weighing Cycle as per OIML R -111 2004
		2 mg	0.02 mg	
		5 mg	0.02 mg	
		10 mg	0.02 mg	
		20 mg	0.03 mg	
		50 mg	0.03 mg	
		100 mg	0.03 mg	
		200 mg	0.03 mg	
		500 mg	0.04 mg	
		1g	0.04 mg	
		2g	0.05 mg	
		5g	0.06 mg	
		10g	0.09 mg	
		20g	0.15 mg	
		50g	0.27 mg	
		100g	0.51 mg	
		200g	1.0 mg	
		500g	1.0 mg	
		1 kg	1.3 mg	
II.	WEIGHING SCALE AND BALANCE			
1.	Weighing Machine			
	d= 0.001 mg	0 to 200 g	0.005 mg	Using Weights of Accuracy Class E1 based on OIML R 76
	d= 0.01 mg		0.05 mg	
	d= 0.1 mg		0.4 mg	
	d= 1 mg	0 to 1 kg	5.3 mg	
				Using Weights of Accuracy Class E1 & E 2 based on OIML R 76

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Shriram Institute For Industrial Research, 14-15, Sadarmangala Industrial Area, Whitefield Road, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2083 (in lieu of C-0216, C-0304, C-1497) **Page** 3 of 7

Validity 13.01.2017 to 12.01.2019 **Last Amended on**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	d= 10 mg d= 100 mg	0 to 5 kg	90 mg 500 mg	Using Weights of Accuracy Class E2, F1 & M1 based on OIML R 76
	d= 1 g	0 to 20 kg	1.1 g	Using Weights of Accuracy Class E2, F1 & M1 based on OIML R 76
	d= 10 g	0 to 150 kg	15 kg	Using Weights of Accuracy Class F1 & M1 based on OIML R 76
III.	VOLUME			
1.	Piston Pipette ^s	>100 μ l to 1000 μ l >1000 μ l to 5000 μ l	0.28 μ l 0.28 μ l	Using Micropipette as per IS 8655-6 & ISO/TR 20461 and Digital Balance upto 200 g readability 0.01 mg and Distilled Water of known Density
2.	Glass Pipettes ^s (Graduated/ Non Graduated)	0.1 ml to 10 ml >10 ml to 50 ml	0.8 μ l 3.8 μ l	Using Digital Precision Balance and Distilled Water of known density as per ISO 4787 & ISO/TR 20461
3.	Glass Burette ^s	0.1 ml to 10 ml >10 ml to 50 ml	0.8 μ l 3.8 μ l	Using Digital Precision Balance and Distilled Water of known density as per ISO 4787 & ISO/TR 20461
4.	Measuring Cylinder/Volumetric Flask/Conical Flask / Beaker ^s	5 ml to 100 ml >100 ml to 500 ml >500 ml to 1000 ml	12.0 μ l 60.0 μ l 150.0 μ l	Using Digital Precision Balance and Distilled Water of known density as per ISO 4787 & ISO/TR 20461

Laboratory Shriram Institute For Industrial Research, 14-15, Sadarmangala Industrial Area, Whitefield Road, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2083 (in lieu of C-0216, C-0304, C-1497) **Page** 6 of 7

Validity 13.01.2017 to 12.01.2019 **Last Amended on**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>THERMAL CALIBRATION</u>				
I.	TEMPERATURE			
1.	Glass Thermometers, PRT/RTD's Thermocouples With Indicator/Without Indicator, Oven [§]	(-)80 ° C to 50 ° C 50 ° C to 300 ° C 300 ° C to 600 ° C 600 ° C to 1200 ° C	0.11 ° C 0.12 ° C 0.5 ° C 1.8 ° C	Using PRT, DTI, Low Temperature Bath & Ice Point by Comparison Method Using PRT, High Precision Bath, DTI by Comparison Method Using PRT, DTI, High Temperature Furnace by Comparison Method Using S Type Thermocouple, High Temperature Furnace, DTI by Comparison Method
2.	Oven, Thermal Chambers, Deep Freezers, Refrigerator, Spatial Mapping Furnace (Single Point) Thermocouples*	(-)80 ° C to 300 ° C 200 ° C to 1200 ° C 100 ° C to 300 ° C 200 ° C to 800 ° C 800 ° C to 1200 ° C	0.5 ° C 1.7 ° C 0.5 ° C 1.2 ° C 2.2 ° C	Using PRT, DTI, T Type with DTI by Comparison Method Using S Type Thermocouple, DTI, by Comparison Method Using R Type Thermocouple, DTI by Comparison Method

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory **Shriram Institute For Industrial Research, 14-15, Sadarmangala Industrial Area, Whitefield Road, Bangalore, Karnataka**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2083 (in lieu of C-0216, C-0304, C-1497)** Page **7 of 7**

Validity **13.01.2017 to 12.01.2019** Last Amended on

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
II.	SPECIFIC HEAT AND HUMIDITY			
1.	Temperature & Humidity Indicators [§]	30 % to 95 % Rh @ 25 ° C 5 ° C to 50 ° C @ 60 % RH	1.8 % 0.3 ° C	Using Temperature & Humidity Indicator with Probe PT 100 with Digital Indicator by Comparison Method
2.	Humidity Chambers*	30 % to 95 % Rh @ 25 ° C	1.8 % RH	Using Temperature & Humidity Indicator with Probe by Comparison Method & Single Position

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

[§]Only in Permanent Laboratory

[¶]Laboratory can also calibrate instruments/devices of coarser resolution / least count within the accredited range using same reference standard/ master equipment under the scope of accreditation.

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Dheeraj Chawla
Convenor

Avijit Das
Program Director