

Laboratory                      Shriram Institute For Industrial Research, 14-15, Sadarmangala  
Industrial Area, Whitefield Road, Bangalore, Karnataka

Accreditation Standard      ISO/IEC 17025: 2005

Certificate Number          CC-2083 (in lieu of C-0216, C-0304, C-1497)      Page                      1 of 7

Validity                          13.01.2017 to 12.01.2019                      Last Amended on

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b><u>ELECTRO-TECHNICAL CALIBRATION</u></b>				
I.	<b>MEASURE</b>			
1.	Time <sup>s</sup> (Digital Timer/Analog & Digital Stop Watch)	5 sec to 9900 sec	0.12 sec to 5.8 sec	Using Digital Time Totalizer by Comparison Method

---

Dheeraj Chawla  
Convenor

---

Avijit Das  
Program Director

Laboratory **Shriram Institute For Industrial Research, 14-15, Sadarmangala Industrial Area, Whitefield Road, Bangalore, Karnataka**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2083 (in lieu of C-0216, C-0304, C-1497)** Page **2 of 7**

Validity **13.01.2017 to 12.01.2019** Last Amended on

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>MECHANICAL CALIBRATION</b>				
<b>I.</b>	<b>WEIGHTS</b>			
1.	Mass <sup>s</sup> (F1 & Coarser)	1 mg	0.02 mg	Using Weights of Accuracy Class E1 & F1 and Precision Balances Analytical Balance by Substitution Method of Weighing and ABBA Weighing Cycle as per OIML R -111 2004
		2 mg	0.02 mg	
		5 mg	0.02 mg	
		10 mg	0.02 mg	
		20 mg	0.03 mg	
		50 mg	0.03 mg	
		100 mg	0.03 mg	
		200 mg	0.03 mg	
		500 mg	0.04 mg	
		1g	0.04 mg	
		2g	0.05 mg	
		5g	0.06 mg	
		10g	0.09 mg	
		20g	0.15 mg	
		50g	0.27 mg	
		100g	0.51 mg	
		200g	1.0 mg	
		500g	1.0 mg	
		1 kg	1.3 mg	
<b>II.</b>	<b>WEIGHING SCALE AND BALANCE</b>			
1.	Weighing Machine			
	d= 0.001 mg	0 to 200 g	0.005 mg	Using Weights of Accuracy Class E1 based on OIML R 76
	d= 0.01 mg		0.05 mg	
	d= 0.1 mg		0.4 mg	
	d= 1 mg	0 to 1 kg	5.3 mg	
				Using Weights of Accuracy Class E1 & E 2 based on OIML R 76

**Dheeraj Chawla**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Shriram Institute For Industrial Research, 14-15, Sadarmangala Industrial Area, Whitefield Road, Bangalore, Karnataka

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2083 (in lieu of C-0216, C-0304, C-1497) **Page** 3 of 7

**Validity** 13.01.2017 to 12.01.2019 **Last Amended on**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
	d= 10 mg d= 100 mg	0 to 5 kg	90 mg 500 mg	Using Weights of Accuracy Class E2, F1 & M1 based on OIML R 76
	d= 1 g	0 to 20 kg	1.1 g	Using Weights of Accuracy Class E2, F1 & M1 based on OIML R 76
	d= 10 g	0 to 150 kg	15 kg	Using Weights of Accuracy Class F1 & M1 based on OIML R 76
<b>III.</b>	<b>VOLUME</b>			
<b>1.</b>	Piston Pipette <sup>s</sup>	>100 $\mu$ l to 1000 $\mu$ l >1000 $\mu$ l to 5000 $\mu$ l	0.28 $\mu$ l 0.28 $\mu$ l	Using Micropipette as per IS 8655-6 & ISO/TR 20461 and Digital Balance upto 200 g readability 0.01 mg and Distilled Water of known Density
<b>2.</b>	Glass Pipettes <sup>s</sup> (Graduated/ Non Graduated)	0.1 ml to 10 ml >10 ml to 50 ml	0.8 $\mu$ l 3.8 $\mu$ l	Using Digital Precision Balance and Distilled Water of known density as per ISO 4787 & ISO/TR 20461
<b>3.</b>	Glass Burette <sup>s</sup>	0.1 ml to 10 ml >10 ml to 50 ml	0.8 $\mu$ l 3.8 $\mu$ l	Using Digital Precision Balance and Distilled Water of known density as per ISO 4787 & ISO/TR 20461
<b>4.</b>	Measuring Cylinder/Volumetric Flask/Conical Flask / Beaker <sup>s</sup>	5 ml to 100 ml >100 ml to 500 ml >500 ml to 1000 ml	12.0 $\mu$ l 60.0 $\mu$ l 150.0 $\mu$ l	Using Digital Precision Balance and Distilled Water of known density as per ISO 4787 & ISO/TR 20461

**Laboratory**                              **Shriram Institute For Industrial Research, 14-15, Sadarmangala  
Industrial Area, Whitefield Road, Bangalore, Karnataka**

**Accreditation Standard**      **ISO/IEC 17025: 2005**

**Certificate Number**            **CC-2083 (in lieu of C-0216, C-0304, C-1497)**      **Page**                              **4 of 7**

**Validity**                              **13.01.2017 to 12.01.2019**                              **Last Amended on**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
<b>IV.</b>	<b>VISCOSITY</b>			
<b>1.</b>	Rotational Viscometer <sup>§</sup>	10 to 100 cP 100 to 1000 cP 1000 to 15000 cP 15000 to 72000 cP	1.2 % 1.2 % 1.5 % 1.7 %	Using Dynamic/ Absolute viscosity of the standard oil is compared with the digital output of the viscometers as per standards ISO-2555, ISO 1652
<b>2.</b>	Ford Cup/ Flow cup <sup>§</sup>	10 cSt to 220 cSt 220 cSt to 1200 cSt	2.5 % 4.2 %	Using Standard Viscometer and oil as per IS 3944-1982 and ASTM D1200
<b>V</b>	<b>PRESSURE INDICATING DEVICES</b>			
<b>1.</b>	Pressure (Pneumatic) Analog/Digital Pressure Gauge/ Transducer/ Transmitters <sup>#</sup>	0.05 bar to 20 bar	0.16 % of reading	Using Pneumatic calibrator Budenberg E400 by Comparison Method as per DKD R 6.1
<b>2.</b>	Pressure (Hydraulic) Analog/Digital Pressure Gauge/ Transducer/ Transmitters <sup>§</sup>	1 bar to 60 bar 60 bar to 700 bar	0.1 % of reading 0.16 % of reading	Using Dead weight tester Budenberg HX 480 By Comparison Method as per DKD R 6.1

**Dheeraj Chawla**  
Convenor

**Avijit Das**  
Program Director

**Laboratory**                      **Shriram Institute For Industrial Research, 14-15, Sadarmangala Industrial Area, Whitefield Road, Bangalore, Karnataka**

**Accreditation Standard**    **ISO/IEC 17025: 2005**

**Certificate Number**        **CC-2083 (in lieu of C-0216, C-0304, C-1497)**    **Page**                      **5 of 7**

**Validity**                      **13.01.2017 to 12.01.2019**                      **Last Amended on**

<b>Sl.</b>	<b>Quantity Measured / Instrument</b>	<b>Range/Frequency</b>	<b>*Calibration Measurement Capability (<math>\pm</math>)</b>	<b>Remarks</b>
<b>3.</b>	Negative Pressure Analog/Digital Vacuum Gauge/ Transducer/ Transmitters <sup>#</sup>	(-) 0.85 bar to 0 bar	1.15 % of reading	Using Pneumatic calibrator Budenberg E400 by Comparison Method as per ISO 3567
<b>VI.</b>	<b>ACCELERATION AND SPEED</b>			
<b>1.</b>	Speed/Non Contact Centrifuges*	100 rpm to 25000 rpm	5.8 rpm to 58 rpm	Digital Tachometer TACH 4A by comparison as per Sanas TR 45-01

---

**Dheeraj Chawla**  
Convenor

---

**Avijit Das**  
Program Director

**Laboratory** Shriram Institute For Industrial Research, 14-15, Sadarmangala Industrial Area, Whitefield Road, Bangalore, Karnataka

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2083 (in lieu of C-0216, C-0304, C-1497) **Page** 6 of 7

**Validity** 13.01.2017 to 12.01.2019 **Last Amended on**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b><u>THERMAL CALIBRATION</u></b>				
<b>I.</b>	<b>TEMPERATURE</b>			
1.	Glass Thermometers, PRT/RTD's Thermocouples With Indicator/Without Indicator, Oven <sup>§</sup>	(-)80 ° C to 50 ° C 50 ° C to 300 ° C 300 ° C to 600 ° C 600 ° C to 1200 ° C	0.11 ° C 0.12 ° C 0.5 ° C 1.8 ° C	Using PRT, DTI, Low Temperature Bath & Ice Point by Comparison Method Using PRT, High Precision Bath, DTI by Comparison Method Using PRT, DTI, High Temperature Furnace by Comparison Method Using S Type Thermocouple, High Temperature Furnace, DTI by Comparison Method
2.	Oven, Thermal Chambers, Deep Freezers, Refrigerator, Spatial Mapping  Furnace (Single Point)  Thermocouples*	(-)80 ° C to 300 ° C  200 ° C to 1200 ° C  100 ° C to 300 ° C 200 ° C to 800 ° C 800 ° C to 1200 ° C	0.5 ° C  1.7 ° C  0.5 ° C 1.2 ° C 2.2 ° C	Using PRT, DTI, T Type with DTI by Comparison Method  Using S Type Thermocouple, DTI, by Comparison Method  Using R Type Thermocouple, DTI by Comparison Method

**Dheeraj Chawla**  
Convenor

**Avijit Das**  
Program Director

Laboratory **Shriram Institute For Industrial Research, 14-15, Sadarmangala Industrial Area, Whitefield Road, Bangalore, Karnataka**

Accreditation Standard **ISO/IEC 17025: 2005**

Certificate Number **CC-2083** (in lieu of C-0216, C-0304, C-1497) Page **7 of 7**

Validity **13.01.2017 to 12.01.2019** Last Amended on

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
II.	<b>SPECIFIC HEAT AND HUMIDITY</b>			
1.	Temperature & Humidity Indicators <sup>§</sup>	30 % to 95 % Rh @ 25 ° C  5 ° C to 50 ° C @ 60 % RH	1.8 %  0.3 ° C	Using Temperature & Humidity Indicator with Probe PT 100 with Digital Indicator by Comparison Method
2.	Humidity Chambers <sup>*</sup>	30 % to 95 % Rh @ 25 ° C	1.8 % RH	Using Temperature & Humidity Indicator with Probe by Comparison Method & Single Position

\* Measurement Capability is expressed as an uncertainty ( $\pm$ ) at a confidence probability of 95%

<sup>§</sup>Only in Permanent Laboratory

<sup>¶</sup>Laboratory can also calibrate instruments/devices of coarser resolution / least count within the accredited range using same reference standard/ master equipment under the scope of accreditation.

<sup>#</sup>The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

---

Dheeraj Chawla  
Convenor

---

Avijit Das  
Program Director