

Laboratory Star Calibration, 1st Floor, Ambala-Chandigarh Highway, Mohan Nagar, Tehsil Road, Derabassi, Punjab

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2648 (in lieu of C-1059, C-1060 & C-1061) **Page** 1 of 14

Validity 12.04.2018 to 11.04.2020 **Last Amended on** 13.04.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>ELECTRO-TECHNICAL CALIBRATION</u>				
I.	MEASURE			
1.	DC Voltage ^s	1 mV to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 1000 V	0.4 % to 0.04 % 0.04 % to 0.08 % 0.008 % to 0.09 % 0.09 % to 0.009 % 0.009 % to 0.007 %	Using Fluke 6½ DMM 8846 A
2.	AC Voltage ^s	50 Hz 10 mV to 100 mV 100 mV to 1 V 1 V to 1000 V	0.5 % to 0.12 % 0.12 % to 0.14 % 0.14 % to 0.1 %	Ussing Fluke 6½ DMM 8846 A
3.	DC Current ^s	10 µA to 1 mA 1 mA to 10 mA 10 mA to 400 mA 400 mA to 1 A 1 A to 10 A	0.35 % to 0.06 % 0.06 % to 0.08 % 0.08 % to 0.06 % 0.06 % to 0.08 % 0.08 % to 0.18 %	Using Fluke 6½ DMM 8846 A
4.	AC Current ^s	50 Hz 30 µA to 1mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 400 mA 1 A to 10 A	1.0 % to 0.16 % 0.16 % to 0.05 % 0.05 % to 0.19 % 0.24 % to 0.17 % 0.17 % to 0.24 %	Using Fluke 6½ DMM 8846 A
5.	Frequency ^s	10 Hz to 1 kHz	0.06 % to 0.01 %	Using Fluke 6½ DMM 8846 A
6.	Resistance ^s	1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 100 kΩ	0.36 % to 0.05 % 0.05 % to 0.06 % 0.02 % to 0.06 % 0.013 %	Using Fluke 6½ DMM 8846 A

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Star Calibration, 1st Floor, Ambala-Chandigarh Highway, Mohan Nagar, Tehsil Road, Derabassi, Punjab

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2648 (in lieu of C-1059, C-1060 & C-1061) **Page** 2 of 14

Validity 12.04.2018 to 11.04.2020 **Last Amended on** 13.04.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
		100 k Ω to 1 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω	0.013 % to 0.05 % 0.05 % to 0.93 % 0.93 % to 2.3 %	
7.	Temperature (Simulation) [§]			
	RTD (Pt -100)	(-) 200 °C to 850 °C	0.7 °C	Using Universal Calibrator Unical 3001 M
	E - Type	(-) 200 °C to 900 °C	0.7 °C	
	B - Type	600 °C to 1750 °C	2.5 °C	
	J - Type	(-) 200 °C to 750 °C	0.7 °C	
	K - Type	200 °C to 1300 °C	0.7 °C	
	T - Type	200 °C to 400 °C	0.7 °C	
	R - Type	200 °C to 1700 °C	1.6 °C	
	S - Type	200 °C to 1700 °C	1.6 °C	
8.	DC Voltage*	10 mV to 50 mV	1.0 % to 0.16 %	Using Universal Calibrator Unical 3001M
		50 mV to 30 V 30 V to 1000 V	0.16 % 0.16 % to 0.2 %	Using 5 ½ Digit DMM
9.	AC Voltage*	50 Hz 30 mV to 300 V 300 V to 1000 V	1.7 % to 0.4 % 0.34 % to 0.6 %	Using 5 ½ Digit DMM
10.	AC High Voltage*	1 kV to 15 kV	6.0 %	Using HV Probe with DMM
11.	DC Current*	30 μ A to 300 mA 300 mA to 10 A	0.4 % to 0.3 % 0.3 % to 0.65 %	Using 5 ½ Digit DMM
12.	AC Current *	50 Hz 30 μ A to 10 A	2.0 % to 1.0 %	Using 5 ½ Digit DMM
13.	Resistance*	30 Ω to 3 M Ω 3 M Ω to 19 M Ω	1.1 % to 0.2 % 0.2 % to 0.25 %	Using 5 ½ Digit DMM

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Star Calibration, 1st Floor, Ambala-Chandigarh Highway, Mohan Nagar, Tehsil Road, Derabassi, Punjab

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2648 (in lieu of C-1059, C-1060 & C-1061) **Page** 3 of 14

Validity 12.04.2018 to 11.04.2020 **Last Amended on** 13.04.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
14.	Timer/Stop/Watch/ Hour Meter (Mech./Digital) *	5 sec to 50 sec 50 sec to 24 hrs	1.0 % to 0.08 %	Using Time Interval Calibrator
II.	SOURCE			
1.	DC Voltage [#]	1 mV to 3 mV 3 mV to 10 mV 10 mV to 300 mV 300 mV to 100 V 100 V to 1000 V	0.35 % to 0.13 % 0.13 % to 0.04 % 0.04 % to 0.02 % 0.02 % to 0.01 % 0.01 % to 0.007 % 0.007 % to 0.006 %	Using Multi-Product Calibrator Fluke-5502A
2.	AC Voltage [#]	45 Hz to 10 kHz 1 mV to 3 mV 3 mV to 10 mV 10 V to 30 mV 30 mV to 300 mV 300 mV to 3 V 3 V to 1000 V	2.5 % to 1.1 % 1.1 % to 0.35 % 0.35 % to 0.19 % 0.19 % to 0.05 % 0.05 % to 0.04 % 0.04 % to 0.07 %	Using Multi-Product Calibrator Fluke-5502A
3.	DC Current [#]	1 μ A to 10 μ A 10 μ A to 300 μ A 300 μ A to 100 mA 300 mA to 10A 10 A to 20 A	2.3 % to 0.25 % 0.25 % to 0.03 % 0.03 % to 0.02 % 0.02 % to 0.03 % 0.03 % to 0.08 %	Using Multi-Product Calibrator Fluke-5502A & Current Coil
4.	AC Current [#]	30 μ A to 100 μ A 100 μ A to 300 μ A 300 μ A to 1 mA 1 mA to 3 mA 3 mA to 10 mA 10 mA to 1 A 1 A to 10 A 10 A to 20 A	0.59 % to 0.5 % 0.3 % to 0.21 % 0.21% to 0.28 % 0.28 % to 0.14 % 0.14 % to 0.08 % 0.08 % to 0.12 % 0.12 % to 0.14 % 0.14 % to 0.35 %	Using Multi-Product Calibrator Fluke-5502A & Current Coil

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Star Calibration, 1st Floor, Ambala-Chandigarh Highway, Mohan Nagar, Tehsil Road, Derabassi, Punjab

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2648 (in lieu of C-1059, C-1060 & C-1061) **Page** 4 of 14

Validity 12.04.2018 to 11.04.2020 **Last Amended on** 13.04.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
5.	Frequency [#]	10 Hz to 1 MHz	0.06 % to 0.01 %	Using Multi-Product Calibrator Fluke-5502A
6.	Resistance [#]	1 Ω to 3 Ω 3 Ω to 10 Ω 10 Ω to 30 Ω 30 Ω to 300 Ω 300 Ω to 30 k Ω 30 k Ω to 10 M Ω 10 M Ω to 30 M Ω 30 M Ω to 300 M Ω 300 m Ω to 1 G Ω	0.7 % to 0.25 % 0.25 % to 0.08 % 0.08 % to 0.04 % 0.04 % to 0.02 % 0.02 % to 0.03 % 0.03 % to 0.07 % 0.07 % to 0.13 % 0.13 % to 0.6 % 0.6 % to 1.8 %	Using Multi-Product Calibrator Fluke-5502A
7.	Capacitance [#]	1 nF to 3 nF 3 nF to 10 nF 10 nF to 300 nF 300 nF to 1 μ F 1 μ F to 3 μ F 3 μ F to 30 μ F 30 μ F to 100 μ F	1.7 % to 1.8 % 1.7 % to 1.0 % 1.0 % to 0.7 % 0.7 % to 0.4 % 0.4 % to 0.46 % 0.3 % to 0.35 % 0.35 % to 0.29 % 0.29 % to 0.5 %	Using Multi-Product Calibrator Fluke-5502A
8.	AC Power (UPF Single Phase) [#]	6.35 W to 6.4 kW @ 50 Hz (63.5 V to 320 V/0.1 A to 20 A)	0.5 % to 0.3 %	Using Multi-Product Calibrator Fluke-5502A
9.	Timer/Stop/Watch/ Hour Meter (Analog/Digital) [#]	5 sec to 50 sec 50 sec to 24 hrs	1.0 % to 0.08 %	Using Time Interval Calibrator
10.	Power Factor [#]	50 Hz 0.2 to 1 PF	2.0 % to 0.1 %	Using Multi-Product Calibrator Fluke-5502A

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Star Calibration, 1st Floor, Ambala-Chandigarh Highway, Mohan Nagar, Tehsil Road, Derabassi, Punjab

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2648 (in lieu of C-1059, C-1060 & C-1061) **Page** 5 of 14

Validity 12.04.2018 to 11.04.2020 **Last Amended on** 13.04.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
11.	Temperature Simulation ^s			Using Multi-Product Calibrator Fluke-5502A
	RTD (Pt -100)	(-) 200 °C to 800 °C	0.3 °C	
	E - Type	(-) 200 °C to 1000 °C	0.6 °C	
	B - Type	600 °C to 1800 °C	1.0 °C	
	J - Type	(-) 200°C to 1200 °C	0.3 °C	
	K - Type	(-) 200°C to 1350 °C	0.5 °C	
	T - Type	(-) 200 °C to 400 °C	0.8 °C to 0.20 °C	
	R - Type	0 °C to 1700 °C	1.0 °C	
	S - Type	0 °C to 1700 °C	1.1 °C	

Laboratory Star Calibration, 1st Floor, Ambala-Chandigarh Highway, Mohan Nagar, Tehsil Road, Derabassi, Punjab

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2648 (in lieu of C-1059, C-1060 & C-1061) **Page** 6 of 14

Validity 12.04.2018 to 11.04.2020 **Last Amended on** 13.04.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>MECHANICAL CALIBRATION</u>				
I.	DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)			
1.	Measuring Pin Set ^s	Up to 20 mm	6.8 μ m	Using Slip Gauge Blocks & Dial Indicator By Comparison Method as per IS: 11103
2.	Feeler Gauge/ Standard Foil ^s	Up to 1 mm	2.8 μ m	Using Digimatic Micrometer By Comparison Method as per IS: 3179
3.	V-Block ^s (Parallelism, Symmetry, Flatness)	Up to 150 mm	10.1 μ m	Using Test Mendral & Dial Gauge By Comparison Method as per IS: 2949
4.	Caliper ^s (Vernier/Dial/Digital) L.C.: 0.01 mm	0 to 1000 mm	16.2 μ m	Using Slip gauge blocks & Caliper Checker By Comparison Method as per IS: 3651
5.	External Micrometer ^s (Mechanical/Digital) L.C.: 0.001 mm	0 to 25 mm Up to 150 mm	1.5 μ m 2.7 μ m	Using Slip Gauge Blocks By Comparison Method as per IS:2967
6.	Height Gauge ^s (Vernier/Dial/Digital) L.C.: 0.01 mm	0 to 1000 mm	17.7 μ m	Using Slip gauge blocks & Caliper Checker By Comparison Method as per IS: 2921

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory

Star Calibration, 1st Floor, Ambala-Chandigarh Highway, Mohan Nagar, Tehsil Road, Derabassi, Punjab

Accreditation Standard

ISO/IEC 17025: 2005

Certificate Number

CC-2648 (in lieu of C-1059, C-1060 & C-1061)

Page

7 of 14

Validity

12.04.2018 to 11.04.2020

Last Amended on 13.04.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
7.	Internal micrometer [§] L.C.: 0.01 mm	5 mm to 35 mm	6.0 μ m	Using Slip gauge Blocks By Comparison Method as per IS: 2966
8.	Depth Micrometer [§] (Mechanical/Digital) L.C.: 0.001 mm	0 to 150 mm	2.0 μ m	Using Slip gauge Blocks By Comparison Method as per IS:JIB B 7544
9.	Depth Caliper [§] (Vernier/Dial/Digital) L.C.: 0.01 mm	0 to 150 mm	13.7 μ m	Using Slip gauge Blocks By Comparison Method as per IS:3651
10.	Plunger Type Dial Indicator(Dial/Digital) [§] L.C.: 0.001 mm	0 to 25 mm	2.5 μ m	Using Dial Calibration Tester By Comparison Method as per IS:2092
11.	Lever Dial Gauge [§] L.C.: 0.001 mm L.C.: 0.01 mm L.C.: 0.1 m	Up to 0.14 mm Up to 0.2 mm Up to 0.14 mm	2.5 μ m 3.0 μ m 6.0 μ m	Using Dial Calibration Tester By Comparison Method as per IS:11498
12.	Dial Bore Gauge [§]	Up to 1 mm Travel	2.5 μ m	Using Dial Calibration Tester By Comparison Method as per JIS B 7515
13.	Dial Thickness Gauge/ OD Caliper [§] L.C.: 0.001 mm L.C.: 0.01 mm L.C.: 0.1 mm	Up to 5 mm Up to 50 mm Up to 100 mm	6.0 μ m 7.0 μ m 60 μ m	Using Slip Gauge Blocks By Comparison Method
14.	Bevel/Angle Protractor/ Combination Set (Angle) [§] L.C.: 1 min	0 to 90°	2.9'	Using Angle gauge set By Comparison Method as per IS: 4239

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Star Calibration, 1st Floor, Ambala-Chandigarh Highway, Mohan Nagar, Tehsil Road, Derabassi, Punjab

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2648 (in lieu of C-1059, C-1060 & C-1061) **Page** 8 of 14

Validity 12.04.2018 to 11.04.2020 **Last Amended on** 13.04.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
15.	Dial snap Gauge [§] L.C.: 0.001 mm	3 mm to 100 mm	1.2 μ m	Using Slip Gauge Blocks By Comparison Method as Per IS:3455
16.	Coating Thickness Gauge [§] L.C.: 0.01 mm	0 to 1 mm	2.1 μ m	Using Standard Foils By Comparison Method
17.	Ultrasonic Thickness Gauge [§]	0 to 100 mm	6.0 μ m	Using Slip Gauge Blocks By Comparison Method
18.	Plain Plug Gauge [§]	Up to 100 mm	7.6 μ m	Using Slip Gauge Blocks with Dial Comparator Stand By Comparison Method As Per IS:3455
19.	Snap Gauge [§]	3 mm to 150 mm	2.7 μ m	Using Slip Gauge Blocks By Comparison Method As Per IS:3455
20.	Bench Centre [¶] (Parallelism, Co-Axiality)	Up to 300 mm	6.0 μ m	By Comparison Method As Per IS:5980
II.	ACOUSTICS			
1.	Sound Level Meter [§]	94 dB & 114 dB	3.4 dB	Using Sound Level Calibrator By Comparison Method
III.	ACCELERATION & SPEED			
1.	Tachometer/ Centrifuge/ RPM Measurement (Non-Contact) [§]	100 rpm to 2000 rpm 2000 rpm to 10000 rpm 10000 rpm to 40000 rpm	2.1 rpm 7.2 rpm 18.5 rpm	Using Tachometer Calibrator/ Tachometer By Comparison Method As Per IS:12508

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Star Calibration, 1st Floor, Ambala-Chandigarh Highway, Mohan Nagar, Tehsil Road, Derabassi, Punjab

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2648 (in lieu of C-1059, C-1060 & C-1061)

Page 9 of 14

Validity 12.04.2018 to 11.04.2020

Last Amended on 13.04.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
IV.	WEIGHTS			
1.	Mass/Weights [§] For Calibration of F2 class weights and coarser	1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g	0.02 mg 0.02 mg 0.02 mg 0.02 mg 0.02 mg 0.03 mg 0.03 mg 0.03 mg 0.033 mg 0.033 mg 0.033 mg 0.033 mg 0.033 mg 0.033 mg 0.04 mg 0.2 mg 0.2 mg	Using E2 class standard weights up to 5 kg with Digital weighing Balance up to 60g/220g g of d = 0.01 mg and d=0.1 mg Procedure. Calibration of Weights is based on substitution Method & ABBA Weighing Cycle as per OIML R 111
	For Calibration of M2 class weights and coarser	500 g 1 kg 2 kg 5 kg 10 kg	0.08 g 0.08 g 0.08 g 0.08 g 0.18 g	Using Digital Weighing Balance up to 12 kg with d = 100 mg
		20 kg 50 kg	0.7 g 0.8 g	Using F1 Class weights and up to 52 kg with d = 1.0 g
V.	WEIGHING SCALE AND BALANCE			
1.	Calibration of Electronic Weighing Balance of Class I and Coarser [#]	1 mg to 60 g d > 0.01 mg >60 g to 220 g d \geq 0.1 mg	0.02 mg 0.11 mg	Using E2 class standard weights up to 5 kg and F1 class above that up to 50 kg as per OIML R-76-1 :2006

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Star Calibration, 1st Floor, Ambala-Chandigarh Highway, Mohan Nagar, Tehsil Road, Derabassi, Punjab

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2648 (in lieu of C-1059, C-1060 & C-1061) **Page** 10 of 14

Validity 12.04.2018 to 11.04.2020 **Last Amended on** 13.04.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	Weighing Balance of Class II and Coarser	>220 g to 12 kg $d \geq 100$ mg	0.2 g	Using E2 class standard weights up to 5 kg and F1 class above that up to 50 kg as per OIML R-76-1 :2006
	Calibration of Class III and IV Weighing Balances and Coarser	>12 kg to 52 kg $d \geq 1$ g	3.0 g	Using F1 class standard weights as per OIML R-76-1 :2006
>52 kg to 100 kg $d \geq 5$ g		7.0 g		
>100 kg to 250 kg $d \geq 20$ g		25.0 g	Using M1 Class Standard Weights as per OIML R-76-1 :2006	
VI.	VOLUME			
1.	Piston Pipette [§] (Micro pipettes)	>10 μ l to 100 μ l	0.23 μ l	Using Digital Balance up to 60 g /200g readability 0.01/0.1 mg and distilled water of known density. Micropipette as per IS 8655-6 & ISO/TR 20461
		>100 μ l to 1000 μ l	0.24 μ l	
2.	Glass Ware [§] Glass Pipettes (Graduated / Non Graduated)	0.1 ml to 10 ml >10 ml to 50 ml	0.083 ml 0.12 ml	Using Digital Balances up to 60 g /220g readability 0.01/0.1 mg and up to 12 kg with distilled water of known density. Glass ware as per ISO 4787 & ISO/TR 20461
	Glass Burette [§]	0.1 ml to 10 ml >10 ml to 50 ml	0.083 ml 0.12 ml	

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Star Calibration, 1st Floor, Ambala-Chandigarh Highway, Mohan Nagar, Tehsil Road, Derabassi, Punjab

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2648 (in lieu of C-1059, C-1060 & C-1061)

Page 11 of 14

Validity 12.04.2018 to 11.04.2020

Last Amended on 13.04.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
	Measuring Cylinder/ Volumetric Flask/Conical Flask/Beaker [§]	1 ml to 10 ml >10 ml to 200 ml >200 ml to 1000 ml >1000 ml to 2000 ml	0.09 ml 0.1 ml 0.217 ml 0.4 ml	
VII.	PRESSURE INDICATING DEVICES			
1.	Hydraulic Pressure Dial /Digital Pressure Gauges and Calibrators, Pressure Transmitter [#]	0 to 30 bar >30 bar to 700 bar	0.12 bar 0.21 bar	Using Calibrated Digital Pressure Gauge Comparison method as per (DKD-R-6-1)
2.	Pneumatic Low Pressure - Dial /Digital Low Pressure Gauges and Manometers, Magnehelic Gauge [#]	0 to 1960 Pa	2.54 Pa	Using Calibrated Digital Manometer by Comparison method as per (DKD-R-6-1)
3.	Dial/Digital Vacuum Gauges /indicators and calibrators [#]	(-) 0.90 bar to 0 bar	0.01 bar	Using Calibrated Digital Vacuum Gauge Comparison method as per (DKD-R-6-1)

Laboratory Star Calibration, 1st Floor, Ambala-Chandigarh Highway, Mohan Nagar, Tehsil Road, Derabassi, Punjab

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2648 (in lieu of C-1059, C-1060 & C-1061) **Page** 12 of 14

Validity 12.04.2018 to 11.04.2020 **Last Amended on** 13.04.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>THERMAL CALIBRATION</u>				
I.	TEMPERATURE			
1.	RTD, Thermocouple With or Without Indicator /Data Logger/ Recorder , Temperature Gauge, Digital Thermometer , Temperature Transmitter #	(-) 40 °C to 50 °C	0.22 °C	Using Fluke 6½ DMM with RTD, Low Temperature Bath By Comparison Calibration
		50 °C to 250 °C	0.23 °C	Using Fluke 6½ DMM With RTD, Oil Bath By Comparison
2.	RTD, Thermocouple With or Without Indicator /Data Logger / Recorder , Temperature Gauge, Digital Thermometer , Temperature Transmitter #	300 °C to 1000 °C	2.3 °C	Using "S" Type Thermocouple, Universal Calibrator & Dry Block Furnace By Comparison Calibration
3.	Thermocouple With or Without Indicator / Data Logger/ Recorder, Temperature Gauge, Digital Thermometer , Temperature Transmitter #	1000 °C to 1200 °C	2.93 °C	Using S-Type Thermocouple With Universal Calibrator, Dry Block Furnace By Comparison Calibration
4.	Glass Thermometer #	50 °C to 250 °C	1.2 °C	Using Fluke 6½ DMM With RTD, Oil Bath By Comparison Calibration

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Star Calibration, 1st Floor, Ambala-Chandigarh Highway, Mohan Nagar, Tehsil Road, Derabassi, Punjab

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2648 (in lieu of C-1059, C-1060 & C-1061) **Page** 13 of 14

Validity 12.04.2018 to 11.04.2020 **Last Amended on** 13.04.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
5.	Temperature Indicator with sensor of Liquid Bath, Dry Block, Furnace, Freezer, Oven Incubator, Chamber, Muffle Furnace [#]	(-) 80 °C to 350 °C	1.77 °C	Using Fluke 6 $\frac{1}{2}$ DMM With RTD Single Position Calibration (At Measuring Location in DUC) By Comparison Calibration
		350 °C to 1000 °C	2.3 °C	Using S-Type Thermocouple , Universal Calibrator Single Position Calibration (At Measuring Location in DUC) By Comparison Calibration
6.	Calibration of Furnace Freezer/ Deep Freezer, Oven Incubator, Environmental Chamber, Cold & Hot Room [*]	(-) 30 °C to 350 °C	4.9 °C	Using Multi Position Calibration Using T Type Thermocouple (Minimum Nine) With Data Logger By Comparison Calibration
II.	SPECIFIC HEAT & HUMIDITY			
1.	Digital/Analog Hygrometer, RH Transmitter, Humidity Sensor with Indicator/ Data Logger/Recorder [§]	20 % RH to 90 % RH @25 °C	2.8 % RH	Using RH Indicator, Humidity Chamber By Comparison Calibration

Dheeraj Chawla
Convenor

Avijit Das
Program Director

Laboratory Star Calibration, 1st Floor, Ambala-Chandigarh Highway, Mohan Nagar, Tehsil Road, Derabassi, Punjab

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2648 (in lieu of C-1059, C-1060 & C-1061) **Page** 14 of 14

Validity 12.04.2018 to 11.04.2020 **Last Amended on** 13.04.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
2.	Environmental Chamber, Humidity Oven, Humidity Chamber *	20 % RH to 85 % RH @25 °C	9.8 % RH	Using Warless Data Logger (Minimum Nine) Multi Position Calibration

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

§ Only in Permanent Laboratory

* Only for Site Calibration

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Dheeraj Chawla
Convenor

Avijit Das
Program Director