

Laboratory Super Scientific International, Plot No. 12, Eicher Compound,  
Sector -23, Sanjay Colon, NIT, Faridabad, Haryana

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2536

Page 1 of 1

Validity 18.01.2018 to 17.01.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b><u>MECHANICAL CALIBRATION</u></b>				
I.	<b>UTM , TENSION CREEP AND TORSION TESTING MACHINE</b>			
1.	Digital Tensile Testing Machine. (Class-I) <sup>§</sup>	60N to 50 kN (Tension)	1.09%	Using IS:1828:2015 (Part-I) Proving Rings (Class-I)
2.	Digital Compression Testing Machine. (Class-I) <sup>§</sup>	2kN to 1000kN (Compression)	0.52%	Using IS:1828:2015 (Part-I) Proving Rings (Class-I)

<sup>§</sup>Only in Permanent Laboratory

---

Dheeraj Chawla  
Convenor

---

Avijit Das  
Program Director