

Laboratory **Superintendence Company Of India (Private) Limited, Plot Y-23, Block-EP, Sector -V, Salt Lake City, Kolkata, West Bengal**

Accreditation Standard **ISO/IEC 17025:2005**

Discipline **Electro-Technical Calibration** Issue Date **01.06.2015**

Certificate Number **C-0430** Valid Until **31.05.2017**

Last Amended on **-** Page **1 of 1**

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability ( $\pm$ )	Remarks
<b><u>SOURCE</u></b>			
1. <b>HIGH RESISTANCE<sup>§</sup></b>	0.1 M $\Omega$ to 1 G $\Omega$	1.03 % to 2.5 %	Using Decade Meg ohm Box by Direct Method
2. <b>TEMPERATURE SIMULATION<sup>§</sup></b> (For Calibration of Indicator, Controller, Recorder)			
<b>RTD</b>	(-) 200 °C to 600 °C	0.5 °C	Using Process Calibrator By Direct Method
<b>THERMOCOUPLE</b>			
<b>J Type</b>	25 °C to 1199 °C	1.8 °C to 0.5 °C	
<b>K Type</b>	25 °C to 1260 °C	0.6 °C to 0.8 °C	
<b>R Type</b>	25 °C to 1600 °C	0.65 °C to 0.8 °C	
<b>S Type</b>	25 °C to 1600 °C	1.0 °C to 0.85 °C	
<b><u>MEASURE</u></b>			
1. <b>TEMPERATURE SIMULATION<sup>§</sup></b> (For Calibration of Indicator, Controller, Recorder)			
<b>RTD</b>	(-) 200 °C to 600 °C	0.5 °C	Using Process Calibrator By Direct Method
<b>THERMOCOUPLE</b>			
<b>J Type</b>	25 °C to 1199 °C	0.5 °C	
<b>K Type</b>	25 °C to 1199 °C	0.9 °C to 0.5 °C	
<b>R Type</b>	25 °C to 1199 °C	0.9 °C to 0.62 °C	
<b>S Type</b>	25 °C to 1199 °C	1.0 °C to 0.65 °C	

\* Measurement Capability is expressed as an uncertainty ( $\pm$ ) at a confidence probability of 95%

<sup>§</sup>Only in Permanent Laboratory

**Anitha Jayanthan**  
Convenor

**Avijit Das**  
Program Manager