

Laboratory	Synchro Calibration Laboratory, P-278, C.I.T. Road, Scheme - VI (M) , Kolkata, West Bengal		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	03.02.2015
Certificate Number	C- 1183	Valid Until	02.02.2017
Last Amended on	24.04.2015	Page	1 of 3

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
<u>SOURCE</u>			
1. DC VOLTAGE^{\$}	1mV to 329 mV 329 mV to 3.29 V 3.29 V to 32 V 32 V to 300 V 300 V to 990 V	0.004 % to 0.01% 0.01% 0.01% to 0.002% 0.002% to 0.01% 0.01%	Using Calibrator by Direct Method
2. DC CURRENT^{\$}	50 µA to 3.29 mA 3.29 mA to 32 mA 32 mA to 329 mA 329 mA to 1.9 A 1.9 A to 10 A 10 A to 550 A	0.20% to 0.02% 0.02% to 0.014% 0.014% to 0.025% 0.025% to 0.10% 0.10% to 0.11% 0.14%	Using Calibrator by Direct Method Using 50 Turn Current Coil
3. RESISTANCE^{\$}	1Ω to 33 Ω 33 Ω to 1.1kΩ 1.1 kΩ to 110 kΩ 110 kΩ to 1.1MΩ 1.1 MΩ to 11 MΩ 11 MΩ to 110 MΩ 110 MΩ to 330 MΩ 1 MΩ to 1 GΩ	0.90% to 0.06% 0.06% to 0.02% 0.02 % 0.02% to 0.023% 0.023% to 0.12% 0.12% to 0.60% 0.60% 0.20%	Using Calibrator by Direct Method Using reference high Resistance Box
4. AC VOLTAGE^{\$}	50 Hz to 1 kHz 1 mV to 100 mV 100 mV to 500 mV 500 mV to 300 V 300 V to 500 V 500 V to 990 V	2.5% to 0.14% 0.14% to 0.05% 0.05 % to 0.06% 0.06% to 0.08% 0.08% to 0.07%	Using Calibrator by Direct Method

Ranjith Kumar
Convenor

Avijit Das
Program Manager

Laboratory	Synchro Calibration Laboratory, P-278, C.I.T. Road, Scheme - VI (M) , Kolkata, West Bengal		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	03.02.2015
Certificate Number	C- 1183	Valid Until	02.02.2017
Last Amended on	24.04.2015	Page	2 of 3

	Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
5.	AC CURRENT[§]	50 Hz to 1 kHz 33 µA to 100 µA 110 µA to 1 mA 1 mA to 100 mA 100 mA to 1 A 1 A to 10 A 10 A to 550 A	1.0 % to 0.32 % 0.32 % to 0.60 % 0.60 % to 0.20 % 0.20 % to 1.0 % 1.0 % to 0.10 % 0.10 % to 0.30 %	Using Calibrator by Direct Method Using 50 Turn Current Coil
6.	FREQUENCY[§] (At 3 V)	10 Hz to 100 kHz	0.10 % to 0.04 %	Using Calibrator by Direct Method
7.	TEMPERATURE SIMULATION[§] Thermocouple			
	K Type	(-) 200°C to 1200 °C	0.50°C	Using Calibrator Fluke, 5500E by Direct Method
	J Type	(-) 200°C to 1000 °C	0.31°C	
	R Type	0 °C to 1700 °C	0.70°C to 0.50°C	
	S Type	0 °C to 1700 °C	0.54°C	
	T Type	(-) 200°C to 390 °C	0.2°C to 0.20 °C	
	<u>MEASURE</u>			
1.	DC VOLTAGE[§]	1 mV to 10 mV 10 mV to 1 V 1 V to 100 V 100 V to 990 V	0.43 % to 0.05 % 0.05 % to 0.004 % 0.004 % to 0.005 % 0.005 % to 0.006 %	Using 6½ Digit Precision Multimeter by Direct Method
2.	DC CURRENT[§]	10 µA to 100 µA 100 µA to 1 A 1 A to 10 A	0.42 % to 0.095 % 0.095 % to 0.08 % 0.08 % to 0.20 %	Using 6½ Digit Precision Multimeter by Direct Method
3.	RESISTANCE[§]	1Ω to 100 Ω 100 Ω to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1MΩ 1 MΩ to 90 MΩ	0.05 % to 0.02 % 0.02 % to 0.013 % 0.013 % 0.013 % 0.013 % to 0.92 %	Using 6½ Digit Precision Multimeter by Direct Method

Ranjith Kumar
Convenor

Avijit Das
Program Manager

Laboratory	Synchro Calibration Laboratory, P-278, C.I.T. Road, Scheme - VI (M) , Kolkata, West Bengal		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	03.02.2015
Certificate Number	C- 1183	Valid Until	02.02.2017
Last Amended on	24.04.2015	Page	3 of 3

	Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
4.	AC VOLTAGE^{\$}	50 Hz to 1 kHz 1 mV to 100 mV 100 mV to 1 V 1 V to 100 V 100 V to 990 V	4.7 % to 0.14 % 0.14 % to 0.12 % 0.12 % 0.12 % to 0.10 %	Using 6½ Digit Precision Multimeter by Direct Method
5.	AC CURRENT^{\$}	50 Hz to 1 kHz 50 μ A to 100 μ A 100 μ A to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 10 A	0.50 % to 0.35 % 0.35 % to 0.25 % 0.25 % to 0.20 % 0.20 % 0.20 % to 0.25 %	Using 6½ Digit Precision Multimeter by Direct Method
6.	AC HIGH VOLTAGE^{\$}	1 kV to 5 kV	9.4 %	Using High Voltage Probe, and Digital Multimeter by Direct Method
7.	FREQUENCY^{\$} (At 3 V)	5 Hz to 10 Hz 10 Hz to 100 kHz 300 kHz to 1 MHz	0.20 % to 0.13 % 0.13 % to 0.014 % 0.014 % to 0.012 %	Using 6½ Digit Precision Multimeter by Direct Method
8.	CAPACITANCE^{\$}	10 nF to 1 μ F 1 μ F to 1 mF 1 mF to 10 mF	1.80 % 1.80 % to 2.33 % 2.33 % to 1.80 %	Using 6½ Digit Precision Multimeter by Direct Method

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

^{\$}Only in Permanent Laboratory

Ranjith Kumar
Convenor

Avijit Das
Program Manager