

Laboratory Taj Calibration Services, SCO No. 9, 2nd Floor, HUDA Market, Part-II, Sector-18, Faridabad, Haryana
Accreditation Standard ISO/IEC 17025: 2005
Certificate Number CC-2591 **Page** 1 of 7
Validity 05.03.2018 to 04.03.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>ELECTRO-TECHNICAL CALIBRATION</u>				
1.	SOURCE			
1.	DC Voltage [#]	1 mV to 1000 V	0.37 % to 0.01 %	Using Fluke 5502 A Multifunction Calibrator by Direct Method
2.	AC Voltage [#]	50 Hz to 1 kHz 10 mV to 1000 V	0.38 % to 0.06 %	Using Fluke 5502 A Multifunction Calibrator by Direct Method
3.	DC Current [#]	10 μ A to 2 A 2 A to 20 A 10 A to 1000 A	0.27 % to 0.04 % 0.04 % to 0.12 % 0.9 % to 0.60 %	Using Fluke 5502 A Multifunction Calibrator Current Coil by Direct Method
4.	AC Current [#]	50 Hz to 1 kHz 30 μ A to 2 A 2 A to 20 A 50 Hz 10 A to 1000 A	0.56 % to 0.07 % 0.07 % to 0.20 % 0.7 % to 0.65 %	Using Fluke 5502 A Multifunction Calibrator Current Coil by Direct Method
5.	DC Resistance [#]	1 Ω to 3.2 M Ω 3.2 M Ω to 1G Ω	0.93 % to 0.06 % 0.06 % to 1.8 %	Using Fluke 5502 A Multifunction Calibrator by Direct Method
6.	Low Resistance [#]	1 m Ω 10 m Ω , 100 m Ω , 1 Ω , 10 Ω , 100 Ω , 1 k Ω	0.76 % 0.60 %	Using Standard Fixed Resistance Box by Direct Method

Naveen Jangra
 Convenor

Avijit Das
 Program Director

Laboratory Taj Calibration Services, SCO No. 9, 2nd Floor, HUDA Market, Part-II, Sector-18, Faridabad, Haryana

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2591 **Page** 2 of 7

Validity 05.03.2018 to 04.03.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
7.	High Resistance [#]	1 M Ω , 5 M Ω , 10 M Ω , 20 M Ω , 50 M Ω , 100 M Ω , 190 M Ω , 500 M Ω , 1 G Ω , 2 G Ω	3.50 %	Using HV Megaohm Box By Direct Method
8.	Frequency [#]	10 Hz to 2 MHz	0.06 % to 0.003 %	Using Fluke 5502 A Multifunction Calibrator by Direct Method
9.	Temperature Simulation Method [#] (Indicator, Controller, Recorder)			
	RTD (PT-100)	(-) 200 °C to 600 °C	0.06 °C to 0.14 °C	Using Fluke 5502 A Multifunction Calibrator by Simulation Method
	J Type	(-) 200 °C to 1000 °C	0.33 °C to 0.28 °C	
	K Type	0 °C to 1300 °C	0.2 °C to 0.47 °C	
	T Type	(-) 150 °C to 400 °C	0.29 °C to 0.2 °C	
	N Type	0 °C to 1300 °C	0.23 °C to 0.32 °C	
	R Type	0 °C to 1700 °C	0.67 °C to 0.47 °C	
	S Type	0 °C to 1700 °C	0.56 °C to 0.54 °C	
10.	DC Power [#]	10 V to 600 V / 0.1 A to 20 A (1 W to 12 KW)	0.96 % to 0.23 %	Using Fluke 5502 A Multifunction Calibrator by Direct Method
11.	AC Power [#]	50 Hz / 60 Hz 10 V to 600 V 0.1 A to 20 A 0.5 PF to UPF	0.57 % to 0.20 %	Using Fluke 5502 A Multifunction Calibrator by Direct Method
12.	Power Factor [#]	50 Hz 0.2 to 1.0 PF (Lead/lag)	0.003 PF	Using Fluke 5502 A Multifunction Calibrator by Direct Method
13.	Capacitance ^{\$}	1 kHz 500 pF to 1 μ F	2.9 % to 0.42 %	Using Fluke 5502 A Multifunction Calibrator by Direct Method
		100 Hz 1 μ F to 100 μ F	0.42 % to 0.64 %	

Naveen Jangra
Convenor

Avijit Das
Program Director

Laboratory

Taj Calibration Services, SCO No. 9, 2nd Floor, HUDA Market, Part-II, Sector-18, Faridabad, Haryana

Accreditation Standard

ISO/IEC 17025: 2005

Certificate Number

CC-2591

Page

3 of 7

Validity

05.03.2018 to 04.03.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
14.	Inductance [§]	1 kHz 100 μ H, 1 mH, 10 mH, 100 mH, 1 H, 10 H	3.47 %	Using Fixed Inductance Box by Direct Method
15.	AC Resistance [§]	1 kHz 1 Ω , 10 Ω , 100 Ω , 1 k Ω , 10 k Ω , 100 k Ω	1.18 %	Using Fixed Resistance Box by Direct Method
II.	MEASURE			
1.	DC Voltage [#]	1 mV to 1000 V	0.42 % to 0.006 %	Using Fluke 8846 A DMM by Direct Method
2.	AC Voltage [#]	50 Hz to 1 kHz 10 mV to 1000 V	0.54 % to 0.1 %	Using Fluke 8846 A DMM by Direct Method
3.	DC Current [#]	10 μ A to 10 A	0.30 % to 0.20 %	Using Fluke 8846 A DMM by Direct Method
4.	AC Current [#]	50 Hz to 1 kHz 100 μ A to 1 A 1 A to 10 A	0.17 % 0.17 % to 0.26 %	Using Fluke 8846 A DMM by Direct Method
5.	DC Resistance [#]	1 Ω to 10 M Ω 10 M Ω to 1G Ω	0.4 % to 0.05 % 0.05 % to 3.0 %	Using Fluke 8846 A DMM by Direct Method
6.	Frequency [#]	10 Hz to 1 MHz	0.082 % to 0.013 %	Using Fluke 8846 A DMM by Direct Method
7.	Temperature Simulation Method [#] (Sourcing Instruments)			
	RTD (PT-100)	(-) 200 $^{\circ}$ C to 600 $^{\circ}$ C	0.16 $^{\circ}$ C	Using Fluke 5502 A Multifunction Calibrator by Simulation Method
	J Type	(-) 200 $^{\circ}$ C to 1000 $^{\circ}$ C	0.32 $^{\circ}$ C to 0.28 $^{\circ}$ C	
	K Type	0 $^{\circ}$ C to 1300 $^{\circ}$ C	0.2 $^{\circ}$ C to 0.47 $^{\circ}$ C	
	T Type	(-) 150 $^{\circ}$ C to 400 $^{\circ}$ C	0.3 $^{\circ}$ C to 0.2 $^{\circ}$ C	
	N Type	0 $^{\circ}$ C to 1300 $^{\circ}$ C	0.23 $^{\circ}$ C to 0.33 $^{\circ}$ C	
	R Type	0 $^{\circ}$ C to 1700 $^{\circ}$ C	0.67 $^{\circ}$ C to 0.48 $^{\circ}$ C	
	S Type	0 $^{\circ}$ C to 1700 $^{\circ}$ C	0.57 $^{\circ}$ C	

Naveen Jangra
Convenor

Avijit Das
Program Director

Laboratory Taj Calibration Services, SCO No. 9, 2nd Floor, HUDA Market, Part-II, Sector-18, Faridabad, Haryana
Accreditation Standard ISO/IEC 17025: 2005
Certificate Number CC-2591 **Page** 4 of 7
Validity 05.03.2018 to 04.03.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
8.	Time Interval Meter / Stop Watch [#]	1 s to 9900 s	0.015 s to 3.18 s	Using Timer Calibrator by Direct / Comparison Method
9.	DC High Voltage*	1 kV to 10 kV	3.82 %	Using High Voltage Probe and DMM by Direct Method
10.	AC High Voltage*	50 Hz 1 kV to 10 kV	2.73 %	Using High Voltage Probe and DMM by Direct Method

Naveen Jangra
Convenor

Avijit Das
Program Director

Laboratory Taj Calibration Services, SCO No. 9, 2nd Floor, HUDA Market, Part-II, Sector-18, Faridabad, Haryana
Accreditation Standard ISO/IEC 17025: 2005
Certificate Number CC-2591 **Page** 5 of 7
Validity 05.03.2018 to 04.03.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>MECHANICAL CALIBRATION</u>				
1.	PRESSURE INDICATING DEVICES			
1.	Pneumatic Pressure Analog/ Digital Pressure Gauge, Transmitter, Switch, Calibrator, Manometer [#]	0 to 2 bar 0 to 40 bar	0.002 bar 0.024 bar	Using Digital Pressure Calibrator Using Pneumatic / Vacuum Comparator Pump Based on DKD-R6-1
2.	Negative Pressure Analog/ Digital Pressure Gauge, Transmitter, Switch, Calibrator, Manometer [#]	(-) 0.85 bar to 0	0.002 bar	Using Digital Pressure Calibrator Using Pneumatic / Vacuum Comparator Pump Based on DKD-R6-1
3.	Hydraulic Pressure Analog/ Digital Gauge, Transmitter, Switch, Calibrator, Manometer [#]	0 to 700 bar	0.40 bar	Using Digital Pressure Gauge And Hydraulic Comparator Pump Based on DKD-R6-1

Naveen Jangra
Convenor

Avijit Das
Program Director

Laboratory Taj Calibration Services, SCO No. 9, 2nd Floor, HUDA Market, Part-II, Sector-18, Faridabad, Haryana
Accreditation Standard ISO/IEC 17025: 2005
Certificate Number CC-2591 **Page** 6 of 7
Validity 05.03.2018 to 04.03.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>THERMAL CALIBRATION</u>				
1.	TEMPERATURE			
1.	RTD/Thermocouple with or without Temp. Indicator / Controller / Data Logger/ Recorder, Digital / Dial Thermometer, Temperature Transmitter [#]	(-) 10 °C to 250 °C	0.30 °C	Using Standard PRT, 6½ DMM & Dry Block Bath by Comparison
2.	Thermocouple with or without Temperature Indicator / Controller / Data Logger/ Recorder, Temperature Transmitter, Digital / Dial Thermometer [#]	250 °C to 650 °C	1.42 °C	Using R Type Thermocouple, 6½ DMM & Dry Block Bath by Comparison
3.	Temperature Indicator with sensor of Freezer, Oven, Environment Chamber, Incubator, Liquid Bath, Dry Block Furnace [*]	(-) 80 °C to 50 °C 50 °C to 250 °C	0.27 °C 0.9 °C	Using Standard PRT, 6½ DMM single position calibration (at measuring location in DUC)
4.	Temperature Indicator with sensor of Oven, Chamber, Furnace, Dry Block Furnace [*]	250 °C to 1200 °C	2.66 °C	Using Standard Type R Thermocouple and 6½ DMM single position calibration (at measuring location in DUC)

Naveen Jangra
Convenor

Avijit Das
Program Director

Laboratory Taj Calibration Services, SCO No. 9, 2nd Floor, HUDA Market, Part-II, Sector-18, Faridabad, Haryana
Accreditation Standard ISO/IEC 17025: 2005
Certificate Number CC-2591 **Page** 7 of 7
Validity 05.03.2018 to 04.03.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
II.	SPECIFIC HEAT & HUMIDITY			
1.	Humidity Indicator of Humidity Chamber/ Generator/ Environment Chamber*	15 % RH to 95 % RH @ 25°C	1.6 % RH	Using standard RH sensor with Indicator single position calibration (at measuring location in DUC)

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

§ Only in Permanent Laboratory

* Only for Site Calibration

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Naveen Jangra
Convenor

Avijit Das
Program Director