

Laboratory Transcaal Engineers India Pvt. Limited, #96/1-1, 1st Floor, East Park Road, 9th Cross, Malleswaram, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2898 **Page** 1 of 9

Validity 03.12.2018 to 02.12.2020 **Last Amended on**

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

| Sl. | Quantity Measured / Instrument | Range/Frequency | *Calibration Measurement Capability (\pm) | Remarks |
|--|---|---|--|---|
| <u>MEDICAL DEVICE CALIBRATION</u> | | | | |
| 1. | Electrical Safety [#] A) Voltage B) Earth Resistance C) Equipment Current D) Leakage Current E) Insulation Resistance | 90-132 V ac rms 180-264 V ac rms 0 – 2 Ω 0- 20 A AC 0 μ A – 10 mA 0.5 – 20 MOhm 20-100 M Ω | 2.4% 3.2% 6.9% 1.2% 3.5% 8.9% | Electrical safety test, using Electrical safety analyzer IEC 62353:2014 |
| 2. | Patient Monitor [#] A) Heart rate accuracy B) Respiration rate C) NIBP leakage test D) NIBP test E) NIBP Pressure relief test F) Invasive blood pressure accuracy G) SPO2 accuracy test Heart rate | 30 - 320 bpm 0 - 100 bpm <330 mmHg 35/15 to 200/150 mmHg < 15 mmHg 35/15 to 200/150 mmHg 70 % - 100 % 30 - 240 BPM | 2.6% 8.9% 5% 4 % 4.1% 10.2 % 3.6% 1.3 % | Simulation Method – Using Vital sign simulator, SPO2 functional tester & Electrical Safety analyzer IEC 62353:2014 Instrumentation & Technical Services University of Vermont |

Vishal Shukla
Convenor

Avijit Das
Program Manager

Laboratory Transcaal Engineers India Pvt. Limited, #96/1-1, 1st Floor, East Park Road, 9th Cross, Malleswaram, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2898

Page

2 of 9

Validity 03.12.2018 to 02.12.2020

Last Amended on

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

| Sl. | Quantity Measured / Instrument | Range/Frequency | *Calibration Measurement Capability (\pm) | Remarks |
|-----|---|---|---|--|
| 3. | Syringe Pump # A) Flow B) Volume C) Occlusion | 2 ml/hr - 100 ml/hr 16 – 200 ml/hr (20ml) < 10 & >200ml/Hr (10ml) 1 ml - 20 ml 16 - 200ml/hr (20ml). < 10 & >200ml/Hr (10ml) 0 - 45 psi | 4.2% 2% 4.5% 1.3% 1.2% | Using Infusion Device Analyzer by Comparison method using IDA 1S & Electrical safety analyzer, IEC 62353:2014 |
| 4. | Infusion Pump # A) Flow B) Volume C) Occlusion | 10 - 500 ml/hr 16 – 200 ml/hr (20ml) < 10 & >200ml/Hr (10ml) 1 ml to 50 ml 16 – 200 ml/hr (20 ml) < 10 & >200ml/Hr (10ml) 0 - 45 psi | 2.5% 1.2% 2.6% 1.2% 2.5% | Using Infusion Device Analyzer by Comparison method using IDA 1S & Electrical safety analyzer IEC 62353:2014 |
| 5. | Autoclave # A) Temperature B) Pressure gauge | 110°C - 135°C 0 - 2 bar | 0.54°C 0.36% | Comparison method Using Temperature sensor, logger, Pressure Calibrator & electrical safety analyzer IEC 62353:2014, AMS 2750 DKD R6-1 |

Vishal Shukla
Convenor

Avijit Das
Program Manager

Laboratory Transcaal Engineers India Pvt. Limited, #96/1-1,1st Floor, East Park Road, 9th Cross, Malleswaram, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2898

Page

3 of 9

Validity 03.12.2018 to 02.12.2020

Last Amended on

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

| Sl. | Quantity Measured / Instrument | Range/Frequency | *Calibration Measurement Capability (±) | Remarks |
|-----|---|-----------------------|---|--|
| 6. | Defibrillator # | | | |
| | A) Heart rate accuracy | 10 to 300 bpm | 1% | Simulation & Comparison Method – Using Impulse 7000 DP, Vital sign simulator, SPO2 functional tester & Electrical safety analyzer IEC 62353:2014 Instrumentation & Technical Services University of Vermont |
| | B) Output accuracy (Energy) | 50 to 270 J | 1.3% | |
| | C) Output energy @ max setting for 10CHg cycle (at battery power) | 50 to 270 J | 1.5 % | |
| | D) Charge time after 10 discharge cycles | ≤ 15 sec | NA | |
| | E) Energy after 60 sec of full charge @ max energy | ≥ 85% | NA | |
| | F)Synchronizer Operation | ≤ 60 msec | NA | |
| | G) NIBP test | 40/20 to 200/150 mmHg | 4% | |
| | H) O2 saturation(Spo2) | 70 % to 100 % | 3.6% | |
| | Heart rate | 30 to 240 BPM | 1.3% | |
| | I) Pacer Amplitude | 0 to 25 mA | 2.6% | |
| | J) Pacer Rate accuracy | 5 to 800 ppm | 2.4% | |

Laboratory Transcaal Engineers India Pvt. Limited, #96/1-1, 1st Floor, East Park Road, 9th Cross, Malleswaram, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2898 **Page** 4 of 9

Validity 03.12.2018 to 02.12.2020 **Last Amended on**

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

| Sl. | Quantity Measured / Instrument | Range/Frequency | *Calibration Measurement Capability (±) | Remarks |
|-----|---|---|---|---|
| 7. | Suction Pump [#] (Vacuum) | 0 to 600 mmHg | 2.5% | Comparison method Using VT Mobile & electrical safety analyzer IEC 62353:2014 |
| 8. | Flow Meter [#] | 2.5 to 25 lpm | 3.6% | Comparison method Using VT mobile IEC 62353:2014 |
| 9. | BP Apparatus / Sphygmomanomete r [#] A) Leak test B) Gauge Zero C) Pressure accuracy | ≤ 15 mmHg/min ± 1 mmHg 0 - 300 mmHg | NA NA 5% | Comparison method Using Vital sign simulator |
| 10. | Digital BP Apparatus [#] A) Leak test B) Pressure accuracy C) Heart Rate accuracy | ≤ 15 mmHg/min 60/30 to 200/150 mmHg 30 to 320 bpm | NA 4.9% 1.6% | Simulation method Using Vital sign simulator Instrumentation & Technical Services University of Vermont |

Laboratory Transcaal Engineers India Pvt. Limited, #96/1-1, 1st Floor, East Park Road, 9th Cross, Malleswaram, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2898

Page

5 of 9

Validity 03.12.2018 to 02.12.2020

Last Amended on

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

| Sl. | Quantity Measured / Instrument | Range/Frequency | *Calibration Measurement Capability (\pm) | Remarks |
|-------------------------|--------------------------------------|------------------------------|---|--|
| 11. | ECG Unit # A) Heart rate accuracy | 30 to 320 BPM | 1.2 % | Simulation method Using Vital sign simulator & Electrical safety analyzer |
| | B) Amplitude accuracy | 0.05 to 5 mV | 3.9 % | IEC 62353:2014 Instrumentation & Technical Services University of Vermont |
| 12. | Ventilator # | | | |
| | A) Volume Accuracy | 0 to 1000 ml | 3.6% | Comparison method using VT Mobile & Electrical safety analyzer IEC 62353:2014 Instrumentation & Technical Services University of Vermont |
| | B) Respiration rate | 2 to 150 brpm | 1.2% | |
| | C) Inspiratory, Expiratory Time | 0.25 to 9.99 s | 2.0% | |
| | D) Pressure accuracy (PIP) | 0 to 120 cm H ₂ O | 3.6% | |
| | E) Peep | 0 to 40 cm H ₂ O | 3.7% | |
| D) Oxygen percentage | 0 to 100 % | 2.3% | | |
| 13. | Nebulizer (Flow) # | 2.5 to 25lpm | 3.7% | Comparison Method using Gas Flow Analyzer and Electrical Safety Analyzer |

Vishal Shukla
Convenor

Avijit Das
Program Manager

Laboratory Transcaal Engineers India Pvt. Limited, #96/1-1, 1st Floor, East Park Road, 9th Cross, Malleswaram, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2898 **Page** 6 of 9

Validity 03.12.2018 to 02.12.2020 **Last Amended on**

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

| Sl. | Quantity Measured / Instrument | Range/Frequency | *Calibration Measurement Capability (\pm) | Remarks |
|-----|--|---|---|--|
| 14. | Radiant Warmer [#] (Temperature) | 21- 50°C | 0.7°C | Comparison method Using temperature sensors, logger & Electrical safety analyzer IEC 62353:2014 |
| 15. | Electro Surgery Unit [#] A) Output Power B)REM Test C) Power Distribution test | 0 - 400 W < manufacturer specified resistance 0 - 400 W | 5.8% NA 6.0% | Comparison method using RF 303 & Electrical safety analyzer IEC 62353:2014 Instrumentation & Technical Services University of Vermont |
| 16. | Baby Weighing Scale [#] | 500g to 15 kg Readability: 20 g | 11.3g | Comparison method |
| 17. | External Pace Maker [#] A) Rate accuracy B) Amplitude | 5 to 800 ppm 0 - 25mA | 2.4% 2.6% | Using Impulse 7000 DP & Electrical safety analyzer by Comparison method IEC 62353:2014 |
| 18. | Fetal Monitor [#] A) Fetal Heart Rate B)Maternal Heart Rate C)TOCO | 30 to 240 bpm 60 to 160 bpm 20 to 100 IUP | 1.2% 1% 1.45% | Using PS320 & Electrical safety analyzer Simulation method IEC 62353:2014 Instrumentation & Technical Services University of Vermont |

Vishal Shukla
Convenor

Avijit Das
Program Manager

Laboratory Transcaal Engineers India Pvt. Limited, #96/1-1, 1st Floor, East Park Road, 9th Cross, Malleswaram, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2898

Page

7 of 9

Validity 03.12.2018 to 02.12.2020

Last Amended on

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

| Sl. | Quantity Measured / Instrument | Range/Frequency | *Calibration Measurement Capability (\pm) | Remarks |
|-----|---|---|---|---|
| 19. | Anesthesia Machine [#] A) Volume Accuracy B) Respiration Rate C) Inspiratory, Expiratory time D) Pressure Accuracy (PIP) E) PEEP F) Oxygen Percentage G) Hypoxic guard test | 0 to 1000 ml 2 to 150 bpm 0.25 to 9.99 s ± 120 cm H ₂ O (-5 to 40 cm H ₂ O) 0 to 100% Safety Functional Check | 3.6% 1.2% 2.0% 3.5% 3.5% 2.3% NA | using VT Mobile & Electrical safety analyzer Comparison method IEC 62353:2014 Instrumentation & Technical Services University of Vermont |
| 20. | Dialysis Machine [#] A) Dialysate Conductivity Test B) Temperature Test C) Flow Test D) pH Test E) Timer Test F) Pressure Test G) NIBP test H) PPM | 13 to 15 mS/cm 0-100°C (37°C) 200-2000 ml/min 7pH (water pH) 15min to 1 Hr ± 3 mmHg (-700 to 1900 mmHg) 180/60 mmHg 90 PPM | 0.21 mS/cm 1°C 1.9% 0.2 pH 1.2 -2.2 Sec 2.8% 3.2% 2.4% | using Dialysis meter & Electrical safety analyzer Comparison method IEC 62353:2014 Instrumentation & Technical Services University of Vermont |

Laboratory Transcaal Engineers India Pvt. Limited, #96/1-1, 1st Floor, East Park Road, 9th Cross, Malleswaram, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2898

Page

8 of 9

Validity 03.12.2018 to 02.12.2020

Last Amended on

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

| Sl. | Quantity Measured / Instrument | Range/Frequency | *Calibration Measurement Capability (\pm) | Remarks |
|-----|--|---|---|--|
| 21. | Electronic Tourniquet # A)Cuff pressure accuracy B)Timer Accuracy C)Max Cuff Pressure D)Controller Stability | 0-100 psi 15 mins \leq 550 mmHg / manufacturers Specification \pm 10 mmHg after 15 min | 2.3% 1 Sec 2.3% NA | using VT Mobile, Stop watch & Electrical safety analyzer Comparison method IEC 62353:2014 Instrumentation & Technical Services University of Vermont |
| 22. | CPAP (positive pressure) # | \pm 120 cmH ₂ O | 2.3% | using VT mobile & Electrical safety analyzer Comparison method IEC 62353:2014 |
| 23. | BiPAP # A)IPAP in cmH ₂ O B)EPAP in cmH ₂ O C)RR in bpm D)Ti in sec | \pm 120 cm H ₂ O \pm 120 cm H ₂ O 2-150 brpm 0.25-9.99s | 3% 3% 2.0% 1.2% | using VT mobile & Electrical safety analyzer Comparison method IEC 62353:2014 |

Laboratory Transcaal Engineers India Pvt. Limited, #96/1-1, 1st Floor, East Park Road, 9th Cross, Malleswaram, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2898

Page

9 of 9

Validity 03.12.2018 to 02.12.2020

Last Amended on

“In view of the transition for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020”

| Sl. | Quantity Measured / Instrument | Range/Frequency | *Calibration Measurement Capability (\pm) | Remarks |
|-----|--------------------------------|-----------------|---|--|
| 24. | Patient Warmer (Temperature) # | 32°C – 42°C | 0.2°C | using Temperature sensor with logger & electrical safety analyzer Comparison method IEC 62353:2014 |
| 25. | Lab Incubator (Temperature) # | 37°C | 0.3°C | using Temperature sensor with logger & electrical safety analyzer Comparison method IEC 62353:2014 |
| 26. | Hot Air Oven # (Temperature) | 50°C to 250°C | 1°C | using Temperature sensor with logger & electrical safety analyzer Comparison method IEC 62353:2014 |

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Vishal Shukla
Convenor

Avijit Das
Program Manager