

Laboratory WIKA Instruments India Pvt. Ltd., Plot No. 40, Gat No. 94+100,
Hi-Cliff Industrial Estate, Kesnand, Pune, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Discipline Mechanical Calibration **Issue Date** 10.02.2016

Certificate Number C-0133 **Valid Until** 09.02.2018

Last Amended on - **Page** 1 of 3

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
I. PRESSURE & VACUUM			
1. Pressure \$ (Pneumatic) Characterization of Dead Weight Tester	0.17 bar to 1.7 bar 0.7 bar to 7 bar 7 bar to 70 bar 10 bar to 100 bar	0.004 % of Rdg. 0.004 % of Rdg. 0.004 % of Rdg. 0.009 % of Rdg.	Using Cross Float method as per EAL/4 -cg-3 & NABL 122-12
2. Pressure \$ (Hydraulic) Characterization of Dead Weight Tester	6 bar to 60 bar 10 bar to 100 bar 40 bar to 400 bar 100 bar to 1200 bar 500 bar to 5000 bar	0.009 % of Rdg. 0.009 % of Rdg. 0.009 % of Rdg. 0.009 % of Rdg. 0.021 % of Rdg.	Using Cross Float method as per EAL/4 -cg-3 & NABL 122-12
3. Vacuum \$ Characterization of Dead Weight Tester	-0.03 bar to -1.0 bar	0.01 % of Rdg.	Using Cross Float method as per EAL/4 -cg-3 & NABL 122-12
4. Pressure \$ (Pneumatic) For Dial Pressure Gauges, Digital Pressure Gauges & Pressure Transmitters, Pressure Switches, Diff. Pr. Indicators (at atmospheric pressure)	0.015 bar to 0.15 bar 0.15 bar to 1.7 bar 1.7 bar to 7 bar 7 bar to 70 bar 40 bar to 400 bar	0.04 % of Rdg 0.006 % of Rdg 0.003 % of Rdg 0.003 % of Rdg 0.009 % of Rdg	By Direct method Based on DKD-R-6-1 & NABL 122-13
5. Pressure \$ (Hydraulic) For Dial Pressure Gauges, Digital Pressure Gauges & Pressure Transmitters, Diff. Pr. Indicators (at atmospheric pressure) Indicators	6 bar to 60 bar 10 bar to 100 bar 40 bar to 400 bar 100 bar to 1200 bar	0.01 % of Rdg 0.01 % of Rdg 0.009 % of Rdg 0.01 % of Rdg	By Direct method Based on DKD-R-6-1 & NABL 122-13

Srikanth R.
Convenor

Avijit Das
Program Manager

Laboratory WIKA Instruments India Pvt. Ltd., Plot No. 40, Gat No. 94+100,
Hi-Cliff Industrial Estate, Kesnand, Pune, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Discipline Mechanical Calibration **Issue Date** 10.02.2016

Certificate Number C-0133 **Valid Until** 09.02.2018

Last Amended on - **Page** 2 of 3

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
6. Pressure ^{\$} (Hydraulic) For Dial Pressure Gauges, Digital Pressure Gauges & Pressure Transmitters, Pressure Switches, Diff. Pr. Indicators ^{\$} (at atmospheric pressure)	400 bar to 5000 bar	0.021 % of Rdg	By Direct method Based on DKD-R-6-1 & NABL 122-13
7. Vacuum ^{\$} For Dial Vacuum Gauges, Digital Vacuum Gauges & Vacuum Transmitters, Switches, Diff. Indicators	-0.015 bar to -0.1 bar -0.1 bar to -0.95 bar	0.06 % of Rdg 0.005 % of Rdg	By Direct method Based on DKD-R-6-2 & NABL 122-13
8. Absolute Pressure ^{\$} For Dial Pressure Gauges, Digital Pressure Gauges & Pressure Transmitters, Barometers, Pressure Switches, Diff. Pr. Indicators (at atmospheric pressure)	0.015 bar to 0.17 bar 0.17 bar to 1.7 bar 1.7 bar to 7 bar 7 bar to 70 bar 0.0001 bar to 1 mbar (abs)	0.04 % of Rdg 0.006 % of Rdg 0.007 % of Rdg 0.003 % of Rdg 0.008 mbar	By Direct method Based on DKD-R-6-1 & NABL 122-13
9. Low Pressure ^{\$} (Pneumatic) For Dial Pressure Gauges, Digital Pressure Gauges & Pressure Transmitters, Pressure Switches Diff. Pr. Indicators (at atmospheric pressure)	0 to 1 mbar 0 to 10 mbar	0.0008 mbar 0.008 mbar	By Comparison method Based on DKD-R-6-1 & NABL 122-13

Srikanth R.
Convenor

Avijit Das
Program Manager

Laboratory WIKA Instruments India Pvt. Ltd., Plot No. 40, Gat No. 94+100,
Hi-Cliff Industrial Estate, Kesnand, Pune, Maharashtra

Accreditation Standard ISO/IEC 17025: 2005

Discipline Mechanical Calibration **Issue Date** 10.02.2016

Certificate Number C-0133 **Valid Until** 09.02.2018

Last Amended on - **Page** 3 of 3

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks	
II. MASS				
1 Mass [§] (weights) Calibration of E2 class weights and coarser	1 mg	0.002 mg	Using Mass Comparator having L.C-0.000001g E1 class 1mg – 10 kg weights Method used is as per OIML-R 111	
	2 mg	0.002 mg		
	5 mg	0.002 mg		
	10 mg	0.002 mg		
	20 mg	0.002 mg		
	50 mg	0.003 mg		
	100 mg	0.003 mg		
	200 mg	0.003 mg		
	500 mg	0.005 mg		
	1 g	0.005 mg		
	2 g	0.005 mg		
	5 g	0.006 mg		
	10 g	0.007 mg		
	20 g	0.01 mg		
Calibration of F1 class weights and coarser	50 g	0.02 mg	Using Mass Comparator having L.C - 0.00001g E1 class 1mg – 10 kg weights	
	100 g	0.02 mg		
	200 g	0.06 mg		
	500 g	0.09 mg		
	1 kg	1 mg		
	2 kg	1 mg		
	5 kg	8.23 mg		Using Mass Comparator having L.C - 0.01g E1 class 1mg – 10 kg weights Method used is as per OIML R-111
		10 mg		
		10 kg		

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

§ Only in Permanent Laboratory

Srikanth R.
Convenor

Avijit Das
Program Manager