

**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6, Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585

**Page**

**1 of 18**

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on -**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b><u>ELECTRO TECHNICAL CALIBRATION</u></b>				
<b>I.</b>	<b>SOURCE</b>			
1.	DC Voltage <sup>#</sup>	1mV to 330mV 100mV to 1000V	0.68% to 0.01% 0.01% to 0.01%	Using Fluke 5502A Multiproduct Calibrator
2.	AC Voltage <sup>#</sup>	<b>50Hz</b> 3mV to 100mV 100mV to 1000V	0.9% to 0.2% 0.2%	Using Fluke 5502A Multiproduct Calibrator
3.	DC Current <sup>#</sup>	200 $\mu$ A to 1mA 1mA to 100mA 100mA to 1A 1A to 10A 10A to 20A 20A to 1000A	0.4% to 0.21% 0.21% to 0.02% 0.02% to 0.05% 0.05% to 0.08% 0.08% to 0.13% 1.2% to 0.5%	Using Fluke 5502A Multiproduct Calibrator  with current coil
4.	AC Current <sup>#</sup>	<b>50Hz</b> 200 $\mu$ A to 1A 1A to 20A 20A to 1000A	0.2% to 0.1% 0.1 % to 0.21% 1.4% to 0.7%	Using Fluke 5502A Multiproduct Calibrator  with current coil
5.	DC Power <sup>\$</sup>	1W to 20kW (10V & 10mA, 1000V & 20A)	1.58% to 0.4%	Using Fluke 5502A Multiproduct Calibrator
6.	AC Power <sup>\$</sup>	<b>45 Hz to 65 Hz</b> 200mW to 20kW (10V & 10mA, 1000V & 20A)	0.58% to 3.01%	Using Fluke 5502A Multiproduct Calibrator with current coil

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6, Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585

**Page**

**2 of 18**

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on -**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
		100W to 120Kw (10V & 10A, 1000V & 1000A)	0.2 %	
7.	Power Factor <sup>s</sup>	45 Hz to 65 Hz 0.5 to UPF	0.33% to 0.21%	Using Fluke 5502A Multiproduct Calibrator
8.	Oscilloscope Amplitude (Square Wave) Time Base Bandwidth <sup>s</sup>	1kHz 5mV to 100 V (50 $\Omega$ /1 M $\Omega$ ) 5 ns to 5 s  1 MHz to 200 MHz	2.7% to 0.4%  0.5% to 0.8%  5.0%	Using Fluke 5502A Multiproduct Calibrator
9.	Resistance <sup>s</sup>	1m $\Omega$ to 100m $\Omega$ 1m $\Omega$ to 1 $\Omega$ 1 $\Omega$ to 100 $\Omega$ 100 $\Omega$ to 100k $\Omega$ 100k $\Omega$ to 10M $\Omega$ 10M $\Omega$ to 1000M $\Omega$  1G $\Omega$ to 200 G $\Omega$	1.05% to 0.01% 0.01% to 0.01% 0.01% to 0.01% 0.01% to 0.01% 0.01% to 0.1% 0.1% to 1.73%  4.51% to 7.02%	Using Decade Resistance Box  Using Fluke 5502A Multiproduct Calibrator  Using Decade Resistance Box
10.	Capacitance <sup>s</sup>	1kHz 220pF to 500pF 500pF to 1nF 1nF to 300nF	3.9% 3.9% to 1.7% 1.7% to 0.41%	Using Fluke 5502A Multiproduct Calibrator
11.	Inductance <sup>s</sup>	1kHz 100 $\mu$ H to 10mH 10mH to 100mH 100mH to 10H	0.63% to 0.28% 0.28% 0.28%	Using Decade Inductance Box

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6, Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585 **Page** 3 of 18

**Validity** 20.02.2018 to 19.02.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
12.	Frequency <sup>o</sup>	120Hz to 100kHz	0.004 % to 2.3%	Using Fluke 5502A Multiproduct Calibrator
13.	Temperature Simulation <sup>#</sup> (Indicator / Controller / Recorder) RTD , PT 100  J Type Thermocouple K Type Thermocouple R Type Thermocouple S Type Thermocouple T Type Thermocouple B Type Thermocouple C Type Thermocouple E Type Thermocouple L Type Thermocouple N Type Thermocouple U Type Thermocouple	(-)200 <sup>o</sup> C to 800 <sup>o</sup> C  (-)200 <sup>o</sup> C to 1200 <sup>o</sup> C (-)200 <sup>o</sup> C to 1300 <sup>o</sup> C 0 <sup>o</sup> C to 1750 <sup>o</sup> C 0 <sup>o</sup> C to 1750 <sup>o</sup> C (-)200 <sup>o</sup> C to 400 <sup>o</sup> C 0 <sup>o</sup> C to 1700 <sup>o</sup> C 0 <sup>o</sup> C to 2000 <sup>o</sup> C (-)200 <sup>o</sup> C to 1000 <sup>o</sup> C (-)200 <sup>o</sup> C to 900 <sup>o</sup> C (-)200 <sup>o</sup> C to 1300 <sup>o</sup> C (-)200 <sup>o</sup> C to 600 <sup>o</sup> C	0.1 <sup>o</sup> C  0.73 <sup>o</sup> C 0.8 <sup>o</sup> C 0.8 <sup>o</sup> C 0.8 <sup>o</sup> C 0.8 <sup>o</sup> C 0.8 <sup>o</sup> C 0.8 <sup>o</sup> C 0.8 <sup>o</sup> C 0.8 <sup>o</sup> C 0.8 <sup>o</sup> C	Using Fluke 5502A Multiproduct Calibrator
<b>II.</b>	<b>MEASURE</b>			
1.	DC Voltage <sup>#</sup>	0.1mV to 1mV 1mV to 1V 1V to 10V 10V to 1000V	0.8% 0.8% to 0.03% 0.01% 0.01%	Using 6½ DMM Fluke 8846A
2.	AC Voltage <sup>#</sup>	<b>50 Hz</b> 1.0mV to 10mV 10mV to 10V  10V to 1000V	0.8% to 0.18% 0.18% to 0.07%  0.07%	Using 6½ DMM Fluke 8846A

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6, Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585

**Page**

**4 of 18**

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on -**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
3.	DC Current <sup>#</sup>	10 $\mu$ A to 100 $\mu$ A 100 $\mu$ A to 100mA 100mA to 1A 1A to 10A 10A to 800A	0.72% to 0.14% 0.14% 0.14% 0.14% 0.14% to 1.15%	Using 6½ DMM Fluke 8846A  Using Shunt/ Current Probe
4.	AC Current <sup>#</sup>	<b>50Hz</b> 10 $\mu$ A to 100mA 100mA to 1A 1A to 10A 10A to 800A	0.24% to 0.16% 0.16% 0.16% 0.16% to 0.6%	Using 6½ DMM Fluke 8846A  Using CT/ Current Probe
5.	DC Resistance <sup>#</sup>	1m $\Omega$ to 1 $\Omega$  1 $\Omega$ to 100 $\Omega$ 100 $\Omega$ to 100M $\Omega$ 100M $\Omega$ to 1000M $\Omega$  2G $\Omega$ 200G $\Omega$	3.06% to 1.01%  0.02% to 0.06% 0.06% 0.06%  3.3% to 6.3%	Using Micro ohm meter  6½ DMM Fluke 8846A  Using Million Megaohm Meter
6.	AC Resistance <sup>§</sup>	<b>1kHz</b> 1 $\Omega$ to 10 $\Omega$ 10 $\Omega$ to 100 $\Omega$ 100 $\Omega$ to 1k $\Omega$ 1k $\Omega$ to 1M $\Omega$ 1M $\Omega$ to 900M $\Omega$	1.24% 1.24% 1.24% 1.24% 0.64%	Using LCR Meter
7.	Frequency <sup>#</sup>	10Hz to 1000kHz	0.06% to 0.01%	Using 6½ DMM Fluke 8846A

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6, Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585

**Page**

**5 of 18**

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on -**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
8.	Temp. Simulation <sup>s</sup> (Temp. Source/ Calibrator) RTD, PT 100  J Type Thermocouple K Type Thermocouple R Type Thermocouple S Type Thermocouple T Type Thermocouple B Type Thermocouple C Type Thermocouple E Type Thermocouple L Type Thermocouple N Type Thermocouple	(-)200 <sup>o</sup> C to 800 <sup>o</sup> C  (-)200 <sup>o</sup> C to 1200 <sup>o</sup> C (-)200 <sup>o</sup> C to 1300 <sup>o</sup> C 0 <sup>o</sup> C to 1750 <sup>o</sup> C 0 <sup>o</sup> C to 1750 <sup>o</sup> C (-)200 <sup>o</sup> C to 400 <sup>o</sup> C 600 <sup>o</sup> C to 1820 <sup>o</sup> C 0 <sup>o</sup> C to 2000 <sup>o</sup> C (-)250 <sup>o</sup> C to 1000 <sup>o</sup> C (-)200 <sup>o</sup> C to 900 <sup>o</sup> C (-)200 <sup>o</sup> C to 1300 <sup>o</sup> C	0.1 <sup>o</sup> C  0.8 <sup>o</sup> C	Using 6½ DMM Fluke 8846A  Using Fluke 5502A Multiproduct Calibrator
9.	Capacitance <sup>s</sup>	1kHz 220 pF to 10nF 10nF to 1000nF 1000nF to 1mF	5.01% to 2.02%	Using LCR Meter, 6½ DMM Fluke 8846A
10.	Inductance <sup>s</sup>	1kHz 1mH to 10H	1.04%	Using LCR Meter
11.	Time <sup>#</sup>	10sec to 24Hrs	2.2 s to 3.6 s	Using Time Calibrator
12.	AC Power/Energy <sup>#</sup> (1 Ø & 3Ø)	50 Hz 240 V, 1 Ø 1A to 120 A UPF-0.5 PF 415 (Ph-Ph) 3Ø 5A to 120A	0.67% to 1.54%	Using Accucheck with CT along with Power source & Load Manager

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6,  
Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585

**Page**

**6 of 18**

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on -**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
13.	DC High Voltage*	1kV TO 20kV	2.49% to 2.49%	Using HV probe with Using DMM
14.	AC High Voltage*	1kV to 28kV	4.26% to 3.5%	Using HV Probe with Using DMM
15.	Power Factor#	<b>50 Hz</b> UPF to 0.5 PF 240 V, 1A to 5A	0. to 1.54%	Using Accucheck with CT along with Power source & Load Manager

---

**Mohit Kaushik**  
Convenor

---

**Avijit Das**  
Program Director

**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6, Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585

**Page**

**7 of 18**

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on -**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b><u>MECHANICAL CALIBRATION</u></b>				
<b>I.</b>	<b>DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)</b>			
1.	External Micrometer <sup>\$</sup> L.C. 0.001mm	0 to 100 mm 0 to 300 mm 0 to 600 mm	1.6 $\mu$ m 1.7 $\mu$ m 4.6 $\mu$ m	Using Slip gauge set
2.	Internal Micrometer <sup>\$</sup> L.C. 0.001mm	0 to 100mm 100 mm to 300mm	3.6 $\mu$ m	Using Slip Gauge Set, Slip gauge Accessories, Caliper Checker
3.	Depth Micrometer, Micrometer Head <sup>\$</sup> L.C. 0.001mm	Up to 300mm	1.60 $\mu$ m	Using Slip Gauge Set / Surface Plate
4.	Caliper <sup>\$</sup> (Vernier/Dial/Digital) L.C. 0.01 mm	Up to 300mm Up to 600mm Up to 1000mm	8.0 $\mu$ m 9.0 $\mu$ m 13.0 $\mu$ m	Using Caliper Checker & Long Slip Gauges
5.	Depth Gauge <sup>\$</sup> (Vernier / Dial / Digital) L.C. 0.01mm	0 to 150 mm	6.0 $\mu$ m	Using Slip Gauge Set
6.	Inside Dial Caliper/Outside <sup>\$</sup> L.C. 0.010 mm	10 mm to 100 mm 0 to 100 mm	8.4 $\mu$ m	Using Slip Gauge set, Slip gauge Accessories

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6, Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585

**Page**

**8 of 18**

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on -**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
7.	Height Gauge <sup>s</sup> ( Vernier / Dial / Electronic ) L.C.0.001mm	0 to 600 mm 0 to 1000 mm	10.8 $\mu$ m 15.4 $\mu$ m	Using Slip Gauge Set, Caliper Checker, Long Slip Gauge/Surface Plate
8.	Height Master <sup>s</sup> L.C. 0.001mm	Up to 300 mm	3.2 $\mu$ m	Using Slip Gauge Set, Caliper Checker
9.	Caliper Checker/Step Gauge <sup>s</sup>	Up to 600 mm	7.0 $\mu$ m	Using Slip Gauge Set, Caliper Checker, Electronic Probe/CMM
10.	Dial Calibration Tester <sup>s</sup> L.C 0.0001mm	0 to 25 mm	0.6 $\mu$ m	Using Slip Gauges, Electronic Probe
11.	Slip Gauge Accessories <sup>s</sup> (Flatness)	0 to 250 mm	3.4 $\mu$ m	Using Slip Gauges Lever dial gauge
12.	Plunger Type Dial Gauge <sup>s</sup> L.C. 0.01 mm L.C. 0.001mm	0 to 25 mm 0 to 10mm	6.0 $\mu$ m 1.4 $\mu$ m	Using Dial Calibration Tester
13.	Lever Type Dial Gauge <sup>s</sup> L.C. 0.001mm	0 to 1 mm	1.4 $\mu$ m	Using Dial Calibration Tester

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Director



**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6, Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585

**Page**

**9 of 18**

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on -**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
14.	Dial Bore Gauge <sup>\$</sup> L.C. 0.001mm	Up to 150mm	1.4 $\mu$ m	Using Dial Calibration Tester
15.	Dial Thickness Gauge / Dial Snap Gauge <sup>\$</sup> L.C. 0.001mm	0 to 10mm	1.4 $\mu$ m	Using Slip Gauge
16.	Plain Plug Gauge <sup>\$</sup>	$\varnothing$ 1mm to $\varnothing$ 100 mm $\varnothing$ 100mm to $\varnothing$ 250 mm	2.4 $\mu$ m 3.0 $\mu$ m	Using ULM, Electronic Probe, Dial Gauge/Slip Gauge/CMM
17.	Flush pin Gauge <sup>\$</sup>	Step Up to 20 mm	2.4 $\mu$ m	Using Slip Gauge With Comparator Stand
18.	Extension Rod <sup>\$</sup>	25 mm to 300mm 300 mm to 600 mm	5.0 $\mu$ m 6.0 $\mu$ m	Using Long Slip Gauge, Electronic Probe
19.	Measuring Pin, Three wire Set <sup>\$</sup>	Up to 25mm	3.0 $\mu$ m	Using Slip Gauge, Electronic probe
20.	Length bar, Long Slip Gauge <sup>\$</sup>	Up to 300mm 300 to 500 mm	5.0 $\mu$ m 8.21 $\mu$ m	Using Long Slip Gauge, Electronic Probe
21.	Plain Ring Gauge/Master Ring Gauge <sup>\$</sup>	Up to 50mm 50 to 150 mm	3.28 $\mu$ m 3.3 $\mu$ m	Using ULM/CMM
22.	Snap Gauge , Gap Gauge, Dial Snap gauge <sup>\$</sup>	Up to 100mm 100 to 250 mm	4.36 $\mu$ m	Using Slip Gauge& ULM

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6, Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585

**Page** 10 of 18

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
23.	Feeler Gauge, Standard Foils <sup>§</sup>	Up to 1.0mm	2.0µm	Using Digital Micrometer
24.	V- Block <sup>§</sup> Flatness Perpendicularity Parallelism V Angle Symmetry	Up to 200mm	5.39 µm 5.39 µm 5.38 µm 5.39 µm 5.27 µm	Using Lever Type Dial Gauge, Mandrel & Surface Plate, Cylindrical Square/CMM
25.	Electronic/Spirit Level <sup>§</sup> Sensitivity 0.001mm/m	Base Length 300 mm	5.0µm/m	Using Slip Gauge Set, Sine Bar, Surface Plate
26.	Electronic Probe <sup>§</sup> L.C.0.1 µm	0 to 20 mm	0.30µm	Using Slip Gauges
27.	Steel Scale <sup>§</sup> L.C.0.5 mm	up to 300 mm	205 µm	Using Profile Projector
28.	Coating Thickness Gauge <sup>§</sup>	Up to 1.0 mm	2.0µm	Using Standard Foils
29.	Comparator Stand <sup>§</sup> (Flatness Only)	Up to 300x300 mm	3.0µm	Using Slip Gauge Set, Lever type dial gauge
31.	Thread Plug Gauge/ Wear check Plug <sup>§</sup> (Effective Dia)	Up to Ø100 mm	5.0µm	Using Three Wire Set, ULM with setting ring gauge

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6, Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585

**Page**

11 of 18

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
32.	Thread Ring Gauge/Wear Check Ring <sup>\$</sup>	Up to $\varnothing$ 100 mm	3.7 $\mu$ m	Using Universal Length measuring Machine
33.	Cylindrical Square <sup>\$</sup>	Up to 600 mm	4.0 $\mu$ m	Using Master Cylindrical Square, Slip Gauges
34.	Test Sieve <sup>\$</sup>	0.03 mm to 10 mm 10 mm to 150 mm	4.0 $\mu$ m 16 $\mu$ m	Using Profile Projector Vernier Caliper
35.	Pitch Gauge <sup>\$</sup> (Profile Angle, Pitch)	Up to 60° Up to 6.0 mm	2.1 min 4.0 $\mu$ m	Using Profile Projector
36.	Radius Gauge Set <sup>\$</sup>	0.5 mm to 15 mm 15 mm to 100 mm	4.0 $\mu$ m	Using Profile Projector
37.	Surface Roughness Tester <sup>\$</sup>	Up to Ra 10 $\mu$ m	8.1%	Using Master specimen
38.	Roughness Master <sup>\$</sup>	Up to Ra 10 $\mu$ m	10.7%	Using Surface Roughness Tester
39.	Limit gauges, Centre Distance Gauges, Pitch Circle Dia Gauges, <sup>\$</sup>	Up to 300 mm >300mm to 600mm	10 $\mu$ m 20 $\mu$ m	Using Coordinate Measuring Machine
40.	Angle Gauge, <sup>\$</sup>	Up to 90 <sup>0</sup>	15 Sec	Using Sine Bar, Slip Gauge Set, Lever Dial Gauge
41.	Taper Plug Gauge <sup>\$</sup>	Up to 60 <sup>0</sup>	15 Sec	Using MM, ULM/Sine Bar/Angle Gauge/Slip Gauge

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6, Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585

**Page**

**12 of 18**

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on -**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
42.	Taper Ring Gauge <sup>§</sup>	Up to 60 <sup>0</sup>	15 Sec	Using LM/CMM
43.	Bevel Angle Protector/ Combination Set <sup>§</sup>	0 <sup>0</sup> to 180 <sup>0</sup> to 0 <sup>0</sup>	3.0 min of arc	Using Slip Gauge Set, Lever Type Dial Gauge
44.	Sine Bar / Centre Distance Angle <sup>§</sup>	Up to 300 mm Centre to centre	4.0 $\mu$ m 10.07 "	Using Slip Gauge Set, Lever Type Dial Gauge, Angle Gauge, Surface Plate
45.	Angle Plate, Box Angle Plate <sup>§</sup> Flatness Perpendicularity Parallelism	Upto 300 mm	4.0 $\mu$ m	Using Slip Gauge Set, Lever Type Dial Gauge, Height Gauge, Cylindrical Square/CMM
46.	Try Square/ Engineers Square <sup>§</sup>	Upto 300 mm	4.0 $\mu$ m	Using Cylindrical Square, Slip Gauges
47.	3D-Coordinate Measuring Machine <sup>*</sup>	(1 x 1 x 1) m	5.2+L/500) $\mu$ m Where L in mm	Using Step Gauges/ Caliper Checker, Long Slip Gauge
48.	Universal Length Measuring Machine <sup>*</sup> L.C. 0.1 $\mu$ m	100 mm 200 mm 500 mm	1.2 $\mu$ m 1.3 $\mu$ m 3.0 $\mu$ m	Using Slip Gauge Set
49.	Surface Plate <sup>*</sup>	Up to (2500 x 1600) mm	2.0 $\sqrt{L+W}$ /150 Where L&W in mm	Using precision On Level L.C.10 $\mu$ M/M

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6, Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585

**Page**

**13 of 18**

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on -**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
50.	Bench Center* (Coaxiality)	Up to 1000 mm	3.7 $\mu$ m/300mm	Using Level, Dial Gauge, Test Mandrel (Cylindrical & Taper)
51.	Gear Rolling Tester* (Horizontal& Vertical PCD)	0-300mm (PCD of gear)	2.8 $\mu$ m	Using Master Mandrel, Slip Gauge Blocks, Cylindrical Square
52.	Straight edge* (Straightness)	Up to 2000 x 100 mm	2(L/150) <sup>1/2</sup> $\mu$ m	Using Spirit Level L.C.10 $\mu$ M/M/Cmm
53.	Air Gauge Unit* L.C. 0.001mm	0.080 mm	3.7 $\mu$ m	Using Master Ring Gauges
54.	Profile Projector* Linear X,Y Axis L.C. 0.001 mm Magnification  Angular L.C. 1sec	0 to 200 mm  10 X – 100 X  0 $^{\circ}$ to 360 $^{\circ}$	2.0 $\mu$ m  1.3 %  16 sec	Using.Glass Scale, Digital Caliper & Slip Gauge, Angle Gauge
55.	Microscope* Linear X,Y Axis L.C.0.001 Magnification	0-200mm  10 X – 100 X	2.0 $\mu$ m  1.3 %	Using Glass Scale, Gauge Blocks
II.	<b>ACCELERATION AND SPEED</b>			
1.	Speed (RPM) #	100 rpm to 40000 rpm	1.61% to 0.09%	Using Digital Tachometer with RPM Calibrator (Electric Motor with Rotating disc and Reflecting Tapes)

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6, Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585

**Page** 14 of 18

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>III.</b>	<b>ACOUSTICS</b>			
1.	Sound Level Meter <sup>s</sup>	94 dB, 114 dB	1.77dB	Using Sound Level Calibrator
<b>IV.</b>	<b>PRESSURE INDICATING DEVICE</b>			
1.	Pressure <sup>#</sup> Pressure Gauges/ Indicator, Pressure Switch Transmitter Etc.	0 to 30 bar 0 to 700 bar	0.35 bar 1.35 bar	Using Digital Pressure Gauge, DMM, Hydraulic Comparator.
2.	Vacuum (Negative Pressure) <sup>#</sup> (Vacuum Gauges, Vacuum Transducers Vacuum Transmitters & Vacuum Switches)	(-)0.9 to 0 bar	0.048 bar	Using Digital Pressure Gauge With comparator
3.	Pneumatic Pressure <sup>#</sup> Digital/Analog Pressure Gauge, Magnehelic Transducers/ Transmitters With Readout, Manometer	200 mbar	0.1 mbar	Using Differential Digital Pressure Gauge with Screw Pump

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Director



**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6, Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585

**Page**

**16 of 18**

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on -**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>VI.</b>	<b>WEIGHING SCALE AND BALANCE</b>			
1.	Weighing Balance <sup>s</sup> d $\leq$ 0.01mg d $\leq$ 0.1 g	0 to 200 g 0 to 20 kg	0.74 mg 2.6 g	Using Standard weights of Class F1/F2
2.	Weighing Balance/ Batching Plant <sup>#</sup> d $\leq$ 10g	0 to 300 kg	10 g	Using Standard SS weights of Class M1
<b>VI.</b>	<b>VOLUME</b>			
1.	Glassware <sup>s</sup> (Pipette/Burette/ Volumetric Flask/ Cylinder)	1 ml to 10 ml >10 ml to 100ml >100 ml to 1000 ml >1 L to 10 L	0.52 $\mu$ l 60 $\mu$ l 83 $\mu$ l 1.1 ml	Using Standard Weights of Class F1 Digital Balance and Distilled Water

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Director



**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6, Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585 **Page** 17 of 18

**Validity** 20.02.2018 to 19.02.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b><u>THERMAL CALIBRATION</u></b>				
<b>I.</b>	<b>TEMPERATURE</b>			
1.	Glass Thermometer <sup>s</sup>	30°C to 250°C	0.66°C	Using RTD with Temp. Indicator with oil bath
2.	Temp./Humidity Meter With Sensor, Thermo Hygrometer, Humidity Data Logger <sup>s</sup>	20% RH to 95% RH @ ~ 25°C 5°C to 45°C @ ~ 50% RH	2.1%RH 0.33°C	Using Temp. & Humidity Meter & Humidity Chamber
3.	RTD, Thermocouple with or without Indicator, Temperature Indicator/Data Logger etc with sensors <sup>#</sup>	(-)30°C to 200°C 200°C to 500°C	0.33°C 0.48°C	Using RTD with Temp. Indicator with Dry Block Calibrators
4.	Thermocouple with or without Temp. Indicator, Temperature Indicator/Data Logger etc with Sensors <sup>#</sup>	500°C to 1000°C 1000°C to 1200°C	2.0°C 3.13°C	Using 'S' Type Thermocouple with Temp. Indicator using Dry Block Calibrators

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Director

**Laboratory** Winmet Technologies Private Limited, Plot No. 140 –A, Sector -6, Faridabad, Haryana

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** CC-2585

**Page**

**18 of 18**

**Validity** 20.02.2018 to 19.02.2020

**Last Amended on -**

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
5.	Temperature Indicator with Sensor of cold Chambers, Ovens, Incubators, Furnaces, Baths*	(-)80°C to 500°C 500°C to 1200°C	0.6°C 2.7°C	Using RTD with Temp Indicator Using S-Type Thermocouple with Temp. Indicator
6.	Calibration of Freezer, Ovens, Incubators, Environmental Chambers*	(-)50°C to 300°C	3.2°C	Using Temperature data logger with RTD Sensors (Minimum Nine)
7.	Calibration of Furnaces*	300°C to 1000°C	5.0°C	Using Data Logger with K-Type Thermocouples (Minimum Nine)
8.	RH Indicator with Sensor of Environment Chambers/ RH Chambers / RH Generators *	20%RH to 95% RH @ ~ 25°C	2.7%RH	Using Digital RH Indicator with Probe
9.	Calibration Of Humidity Chamber/ Environment Chamber*	20% RH to 95% RH @ ~ 25°C 5°C to 50°C @ ~50% RH	3.2% RH 1.3°C	Using RH & Temp. Data Loggers with Inbuilt Sensors (Minimum Nine)

\* Measurement Capability is expressed as an uncertainty ( $\pm$ ) at a confidence probability of 95%

§ Only in Permanent Laboratory

\* Only for Site Calibration

# The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

**Mohit Kaushik**  
Convenor

**Avijit Das**  
Program Director