

**Laboratory** Yadav Measurements Private Limited, Plot No. 373-375, RIICO, Bhamashah Industrial Area, Kaladwas, Udaipur, Rajasthan

**Accreditation Standard** ISO/IEC 17025: 2005

**Discipline** Electro – Technical Calibration **Issue Date** 24.07.2014

**Certificate Number** C-0035 **Valid Until** 23.07.2016

**Last Amended on** 11.09.2015 **Page** 1 of 14

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>I. SOURCE</b>			
<b>1. DC VOLTAGE</b> \$	1mV to 100 mV	0.37 % to 0.013 %	Direct Method Using Fluke5500A
	100 mV to 1000 V	0.013 % to 0.008 %	
<b>2. DC CURRENT</b> \$	1 mA to 100 mA	0.02 % to 0.03 %	Direct Method/ Using Fluke5500A
	100 mA to 1 A	0.03 % to 0.05 %	
	1 A to 11 A	0.05 % to 0.58 %	
<b>3. RESISTANCE</b> \$	1 m $\Omega$ to 100 m $\Omega$	0.6 % to 0.2 %	Comparison Method Using $\mu\Omega$ meter Using 8½ DMM HP3458
	0.1 $\Omega$ to 1 $\Omega$	0.18 % to 0.06 %	
	1 $\Omega$ to 10 $\Omega$	0.06 % to 0.004 %	
	10 $\Omega$ to 10 M $\Omega$	0.004 % to 0.015 %	
	10 M $\Omega$ to 100 M $\Omega$	0.015 % to 0.12 %	
	100 M $\Omega$ to 330 M $\Omega$ 330 M $\Omega$ to 1 G $\Omega$	0.12 % to 1.14 % 1.5 %	
<b>4. AC VOLTAGE</b> \$	<b>50 Hz</b> 10 V to 480 V	0.011 %	Direct Method/ Using TS/VCS/COM3003
	<b>40 Hz to 1 kHz</b> 1 V to 700 V	0.25 % to 0.20 %	
	<b>1kHz</b> 10 mV to 1 V	0.64 % to 0.25 %	
	<b>50Hz</b> 700 V to 90 kV	0.50 %	Using Capacitive Divider + DMM
	<b>5. AC CURRENT</b> \$	<b>40 Hz to 70 Hz</b> 1 mA to 120 A	0.021 % to 0.012 %
<b>40 Hz to 1 kHz</b> 1 mA to 1 A		0.35 %	Using Fluke5500A

Sangeeta Kunwar  
Convenor

Avijit Das  
Program Manager

**Laboratory** Yadav Measurements Private Limited, Plot No. 373-375, RIICO,  
 Bhamashah Industrial Area, Kaladwas, Udaipur, Rajasthan  
**Accreditation Standard** ISO/IEC 17025: 2005  
**Discipline** Electro – Technical Calibration **Issue Date** 24.07.2014  
**Certificate Number** C-0035 **Valid Until** 23.07.2016  
**Last Amended on** 11.09.2015 **Page** 2 of 14

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
6. FREQUENCY \$	10 Hz to 225 MHz	0.015 %	Direct Method/ Using Fluke + Signal generator by Direct Method
7. POWER FACTOR #	40 Hz to 70 Hz +1 to -1	0.005 PF	Direct Method/ Using MTS/VCS/COM3003
8. AC POWER / ENERGY # 1 PHASE & 3 PHASE ACTIVE COS Ø ± 0.1 TO 1 REACTIVE SIN Ø ± 0.10 TO 1 APPARENT	45 Hz to 65 Hz 40 V to 240 V 1 mA to 10 mA 11 mA to 20 mA 21 mA to 100 mA 101 mA to 120 A Active / Reactive 0.004 to 7.2 W/VAr 0.044 to 14.4 W/VAr 0.084 to 72 W /VAr 0.404 W/VAr to 86.4 kW/kVAr	0.036 % 0.016 % 0.009 % 0.007 %	Direct Method/ Using COM3003 Make:- ZERA
	Apparent 0.004 to 7.2 VA 0.044 to 14.4 VA 0.084 to 72 VA 0.404 VA to 86.4 kVA	Apparent 0.052 % 0.023 % 0.013 % 0.008 %	Using Com 3003 Zear By Direct Method
	45 Hz to 65 Hz 241 V to 320 V 1 mA to 10 mA 11 mA to 20 mA 21 mA to 100 mA 101 mA to 120 A		

<b>Laboratory</b>	<b>Yadav Measurements Private Limited, Plot No. 373-375, RIICO, Bhamashah Industrial Area, Kaladwas, Udaipur, Rajasthan</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro – Technical Calibration</b>	<b>Issue Date</b>	<b>24.07.2014</b>
<b>Certificate Number</b>	<b>C-0035</b>	<b>Valid Until</b>	<b>23.07.2016</b>
<b>Last Amended on</b>	<b>11.09.2015</b>	<b>Page</b>	<b>3 of 14</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
	<b>Active / Reactive</b>		
	0.0241 to 9.6 W/VAr	0.036 %	
	0.265 to 19.2 W /VAr	0.016 %	
	0.506 to 96 W /VAr	0.008 %	
	2.434 W/VAr to 115.2 kW/kVAr	0.009 %	
	<b>Apparent</b>		
	0.0241 to 9.6 VA	0.052 %	
	0.265 to 19.2 VA	0.023 %	
	0.506 to 96 VA	0.011 %	
	2.434 VA to 115.2 kVA	0.011 %	
<b>AC POWER</b>	<b>45 Hz to 65 Hz</b>		
<b>1 PHASE &amp; 3 PHASE</b>	10 V to 240 V	0.9%	
<b>ACTIVE</b>	0.1 A to 10 A		
<b>COS <math>\theta</math> <math>\pm</math>0.01 to 0.1</b>	10 mW to 720 W		
<b>9. HARMONIC #</b>	<b>2 to 40th harmonic</b>		Direct Method/ Using MTS/ MT3000
<b>VOLTAGE</b>	1 V to 240 V	0.6 %	
<b>CURRENT</b>	50 mA to 24 A	0.6 %	
<b>10. FREQUENCY *</b>			Direct Method / Using 3 $\Phi$ cal system ZERA TPZ/MTE
<b>40 to 240V</b>	40 Hz to 70 Hz	0.03 %	
<b>11. AC VOLTAGE *</b>	<b>50 Hz</b>		
	10 V to 550 V	0.011 %	Direct Method / Using MTS / VCS/ COM3003/CALMET
	<b>40 Hz to 1 kHz</b>		
	1 V to 700 V	0.25 % to 0.20 %	Using Fluke5500A
	<b>1 kHz</b>		
	10 mV to 1 V	0.64 % to 0.25 %	
	<b>50Hz</b>		
	700V to 90kV	0.50%	Using Capacitve Divider & DMM

**Sangeeta Kunwar**  
Convenor

**Avijit Das**  
Program Manager

**Laboratory** Yadav Measurements Private Limited, Plot No. 373-375, RIICO,  
 Bhamashah Industrial Area, Kaladwas, Udaipur, Rajasthan  
**Accreditation Standard** ISO/IEC 17025: 2005  
**Discipline** Electro – Technical Calibration **Issue Date** 24.07.2014  
**Certificate Number** C-0035 **Valid Until** 23.07.2016  
**Last Amended on** 11.09.2015 **Page** 4 of 14

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
12. AC CURRENT *	45 Hz to 70Hz 1mA to 120A	0.04%	Direct Method/ Using COM3003
	50Hz 120A to 2000A	0.09%	Using ELTEL CT set up & MTE PRS 400.3 by Direct Method
<b>II. MEASURE</b>			
1. DC VOLTAGE \$	1 mV to 5 mV	0.13 % to 0.03 %	Direct Method/ Using 8½ DMM HP3458
	5 mV to 10 mV	0.03 % to 0.014 %	
	10 mV to 40 mV	0.014 % to 0.005 %	
	40 mV to 1000 V	0.005 % to 0.004 %	
2. DC CURRENT \$	1 mA to 100 mA	0.01 %	Direct Method/ Using 8½ DMM HP3458 Using shunt
	100 mA to 1 A	0.01 % to 0.03 %	
	1 A to 11 A	0.03 % to 0.58 %	
3. DC RESISTANCE \$	1 mΩ to 100 mΩ	0.5 % to 0.2 %	Direct Method/ Using μΩ meter
	0.1 Ω to 1 Ω	0.18 % to 0.06 %	
	1 Ω to 10 Ω	0.06 % to 0.004 %	
	10 Ω to 10 MΩ	0.004 % to 0.015 %	Using 8½ DMM HP3458
	10 MΩ to 100 MΩ	0.015 % to 0.12 %	
	100 MΩ to 330 MΩ 330 MΩ to 1 GΩ	0.12 % to 1.14 % 1.2 %	
4. AC VOLTAGE \$	50 Hz 10 V to 480 V	0.011 %	Using COM3003
	480 to 33 kV	0.2 %	Using 6 ½ DMM Standard PT (6.6kV), 33 kV Capacitor &EPD& EMVT & MTE/MT3000
	33 kV to 90 kV	0.2 %	Using 6 ½ DMM Standard PT, 132 kV Capacitor &EPD & EMVT & MTE/MT3000

**Sangeeta Kunwar**  
 Convenor

**Avijit Das**  
 Program Manager

**Laboratory** Yadav Measurements Private Limited, Plot No. 373-375, RIICO, Bhamashah Industrial Area, Kaladwas, Udaipur, Rajasthan

**Accreditation Standard** ISO/IEC 17025: 2005

**Discipline** Electro – Technical Calibration **Issue Date** 24.07.2014

**Certificate Number** C-0035 **Valid Until** 23.07.2016

**Last Amended on** 11.09.2015 **Page** 5 of 14

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
AC VOLTAGE \$	40Hz to 1 kHz 1 V to 120 V 120 V to 700 V	0.15 % to 0.10 % 0.10 % to 0.13 %	Using 8½ DMM HP3458
	At 1 kHz 10 mV to 1 V	0.07 % to 0.15 %	Using 8½ DMM HP3458
	1 kHz to 300 kHz 0.5 V to 10 V	0.37 % to 0.6 %	Using 8½ DMM HP3458
5. AC CURRENT \$	40 Hz to 70 Hz 1 mA to 120 A	0.021 % to 0.012 %	Using COM3003
	40 Hz to 1 kHz 1 mA to 1 A	0.30 %	Using 8½ DMM HP3458
6. FREQUENCY #	10 Hz to 225 MHz	0.0020 %	Direct Method Using 53131A Frequency Counter
7. POWER FACTOR #	40 Hz to 70 Hz +1 to -1	0.005 PF	Direct Method/ Using COM3003
8. TIME \$	5 s to 6000 s	0.03 s to 0.60 s	Comparison Method/ Using 53131A Frequency Counter
9. AC POWER / ENERGY # 1 phase & 3 phase Active Cos $\phi \pm 0.1$ to 1 Reactive Sin $\phi \pm 0.10$ to 1 Apparent	45 Hz to 65 Hz 40 V to 240 V 1 mA to 10 mA 11 mA to 20 mA 21 mA to 100 mA 101 mA to 120 A Active / Reactive		Using COM 3003 Zera By Comparison Method
	0.004 to 7.2 W/VAr 0.044 to 14.4 W/VAr	0.036 % 0.016 %	

Sangeeta Kunwar  
Convenor

Avijit Das  
Program Manager

<b>Laboratory</b>	<b>Yadav Measurements Private Limited, Plot No. 373-375, RIICO, Bhamashah Industrial Area, Kaladwas, Udaipur, Rajasthan</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro – Technical Calibration</b>	<b>Issue Date</b>	<b>24.07.2014</b>
<b>Certificate Number</b>	<b>C-0035</b>	<b>Valid Until</b>	<b>23.07.2016</b>
<b>Last Amended on</b>	<b>11.09.2015</b>	<b>Page</b>	<b>6 of 14</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
	0.084 to 72 W /VAr 0.404 W/VAr to 86.4 kW/kVAr	0.009 % 0.007 %	
	Apparent 0.004 to 7.2 VA 0.044 to 14.4 VA 0.084 to 72 VA 0.404 VA to 86.4 kVA	Apparent 0.052 % 0.023 % 0.013 % 0.008 %	
	<b>45 Hz to 65 Hz</b> 241 V to 320 V 1 mA to 10 mA 11 mA to 20 mA 21 mA to 100 mA 101 mA to 120 A		
	Active / Reactive 0.0241 to 9.6 W/VAr 0.265 to 19.2 W /VAr 0.506 to 96 W /VAr 2.434 W/VAr to 115.2 kW/kVAr	Active /Reactive 0.036 % 0.016 % 0.008 % 0.009 %	
	Apparent 0.0241 to 9.6 VA 0.265 to 19.2 VA 0.506 to 96 VA 2.434 VA to 115.2 kVA	Apparent 0.052 % 0.023 % 0.011 % 0.011 %	
<b>10. AC POWER / ENERGY \$ 1 phase &amp; 3 phase, Active Cos Ø :± 0.01 to ±0.1</b>	<b>45 Hz to 65 Hz</b> 10 V to 240 V 0.1 A to 10 A 10 mW to 720 W	0.9 %	Using COM3003 by Comparison Method

**Sangeeta Kunwar**  
Convenor

**Avijit Das**  
Program Manager

**Laboratory** Yadav Measurements Private Limited, Plot No. 373-375, RIICO, Bhamashah Industrial Area, Kaladwas, Udaipur, Rajasthan

**Accreditation Standard** ISO/IEC 17025: 2005

**Discipline** Electro – Technical Calibration **Issue Date** 24.07.2014

**Certificate Number** C-0035 **Valid Until** 23.07.2016

**Last Amended on** 11.09.2015 **Page** 7 of 14

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
11. <b>CURRENT TRANSFORMER # (Primary Injection) Ratio error Phase error</b>	5 A to 2000 A (Primary) 1 A, 5 A (Secondary)	0.01 % 0.7 Minute	Using Precision current transformer & Automatic Instrument transformer test set (AIITS) By Direct Method
	2000 A to 10000 A (Primary) 1 A, 5 A (Secondary)	0.05 % 2.5 Minute	
12. <b>CURRENT TRANSFORMER # - Ratio error - Phase error</b>	10 A to 10000 A (Primary) 1 A, 5 A. (Secondary) & 5 A (Primary) 5 A (secondary)	0.09 % 2.8 Minute	Using Portable CT/VT Calibrator  Using Portable CT Calibrator
13. <b>VOLTAGE TRANSFORMER / CAPACITOR VOLTAGE TRANSFORMER/ VOLTAGE DIVIDER # Ratio error Phase Error</b>	2.2 kV to 33 kV (Primary) 50 V to 144 V (Secondary)	0.04 % 4 Minute	Using Electronic Potential Divider 33kV & Standard Capacitor & EMVT By Direct Method
	33 kV to 132 kV (Primary) 50 V to 144 V (Secondary)	0.043 % 4 Minute	Using Electronic Potential Divider 132kV + Standard Capacitor + RE & Automatic instrument transformer test set (AIITS) & EMVT By Direct Method
14. <b>VOLTAGE TRANSFORMER / CAPACITOR VOLTAGE TRANSFORMER/ VOLTAGE DIVIDER RATIO ERROR PHASE ERROR #</b>	2.2 kV to 33 kV (Primary) 50 V to 144 V (Secondary)	0.15 % 5.6 Minutes	Using Portable CT/VT Calibrator By Direct Method

Sangeeta Kunwar  
Convenor

Avijit Das  
Program Manager

**Laboratory** Yadav Measurements Private Limited, Plot No. 373-375, RIICO,  
 Bhamashah Industrial Area, Kaladwas, Udaipur, Rajasthan  
**Accreditation Standard** ISO/IEC 17025: 2005  
**Discipline** Electro – Technical Calibration **Issue Date** 24.07.2014  
**Certificate Number** C-0035 **Valid Until** 23.07.2016  
**Last Amended on** 11.09.2015 **Page** 8 of 14

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
15. CT-VT COMPARATOR / BRIDGE #	Current 1 A , 5 A		Using Precision current transformer & Automatic Instrument transformer test set (AIITS) by Comparison Method
	Ratio error Phase error	RE 0.005 % PE 0.3 min	
	Voltage 63.5 V, 110 V		Using Electronic Potential Divider 33kV + Standard Capacitor & EM by Comparison Method
16. HARMONIC #  VOLTAGE  CURRENT	2 to 40th harmonic		Using MTS/ MT3000 by Direct Method
	1 V to 240 V	0.6 %	
	50 mA to 24 A	0.6 %	
17. AC VOLTAGE *	<b>50 Hz</b> 10 V to 480 V	0.03 %	Using COM3003 By Direct/ Comparison Method
	<b>50 Hz</b> 480 V to 33 kV	0.20 %	Using 6 ½ DMM Standard PT (6.6kV), 33 kV Capacitor +EPD +EMVT & MTE/MT3000 By Direct/ Comparison Method
	<b>50 Hz</b> 33 kV to 90 kV	0.20 %	Using 6 ½ DMM Standard PT, 132 kV Capacitor +EPD +EMVT & MTE/MT3000 By Direct/ Comparison Method

**Sangeeta Kunwar**  
 Convenor

**Avijit Das**  
 Program Manager



**Laboratory** Yadav Measurements Private Limited, Plot No. 373-375, RIICO,  
 Bhamashah Industrial Area, Kaladwas, Udaipur, Rajasthan  
**Accreditation Standard** ISO/IEC 17025: 2005  
**Discipline** Electro – Technical Calibration **Issue Date** 24.07.2014  
**Certificate Number** C-0035 **Valid Until** 23.07.2016  
**Last Amended on** 11.09.2015 **Page** 9 of 14

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
18. AC CURRENT *	45 Hz to 70 Hz 1 mA to 120 A	0.04 %	Using COM3003 by Direct/ Comparison Method
	50 Hz 120 A to 2000 A	0.09 %	Using ELTEL CT set up + MTE PRS 400.3 by Direct/ Comparison Method
19. DC RESISTANCE *	1 $\Omega$ to 10 $\Omega$	0.5 %	Using Micro ohm meter
	10 $\Omega$ to 100 M $\Omega$	1.1 %	Using 6½ DMM
	100 M $\Omega$ to 1 G $\Omega$	1.5 %	By Direct Method
20. Voltage Transformer / Capacitor Voltage Transformer/Voltage Divider -Ratio Error -Phase Error \$	6.6 kV to 11 kV (Primary) 50 V to 144 V (Secondary)	0.017 % to 0.013 % 1.5 min	Using EMVT & Automatic Instrument Transformer Test Set (AIITS) by Direct Method
	11 kV to 33 kV (Primary) 50 V to 144 V (Secondary)	0.013 % 1.5 min	
	33 kV to 132 kV (Primary) 50 V to 144 V (Secondary)	0.013 % to 0.017 % 2 min	
	2.2 kV to 6.6 kV (Primary) 50 V to 144 V (Secondary)	0.04 % 4 min	Using EPD 33 kV & Standard Capacitor & EMVT by Direct Method
21. Voltage Transformer / Capacitor Voltage Transformer/Voltage Divider -Ratio Error -Phase Error #	2.2 kV to 6.6 kV (Primary) 50 V to 144 V (Secondary)	0.04 % 4 min	Using EPD 33 kV & Standard Capacitor & EMVT by Direct Method

**Sangeeta Kunwar**  
 Convenor

**Avijit Das**  
 Program Manager

<b>Laboratory</b>	<b>Yadav Measurements Private Limited, Plot No. 373-375, RIICO, Bhamashah Industrial Area, Kaladwas, Udaipur, Rajasthan</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro – Technical Calibration</b>	<b>Issue Date</b>	<b>24.07.2014</b>
<b>Certificate Number</b>	<b>C-0035</b>	<b>Valid Until</b>	<b>23.07.2016</b>
<b>Last Amended on</b>	<b>11.09.2015</b>	<b>Page</b>	<b>10 of 14</b>

	<b>Quantity Measured/ Instrument</b>	<b>Range / Frequency</b>	<b>*Calibration Measurement Capability (±)</b>	<b>Remarks</b>
<b>22.</b>	<b>Voltage Transformer/Capacitor Voltage Transformer/ Voltage Divide3r -Ratio Error -Phase Error</b>	6.6 kV to 33 kV (Priamary) 50 V to 144 V (Secondary)	0.04 % 4 min	Using EPD 33kV & Standard Capacitor & EMVT by Director Method
		33 kV to 132 kV (Priamary) 50 V to 144 V (Secondary)	0.043 % 4 min	Using EPD 132 kV & Standard Capacitor & RE & AITTS & EMVT by Director Method
<b>23.</b>	<b>Total harmonic Distortion Voltage Current</b>	2 <sup>nd</sup> to 40 th Harmonic  1 V to 240 V 50 mA to 24 A	  0. 0.6 % 0. 0.6 %	Using Zero MT 3000 Power/Energy Meter with Source MTS by Direct/Comparison Method
<b>24.</b>	<b>Distrotion Factor Voltage Current</b>	2 <sup>nd</sup> to 40 th Harmonic 1 V to 240 V 50 mA to 24 A	  0. 0.6 % 0. 0.6 %	Using Zero MT 3000 Power/Energy Meter with Source MTS by Direct/Comparison Method

#### MOBILE CALIBRATION FACILITY

##### MEASURE

<b>1.</b>	<b>FREQUENCY</b>	10 Hz to 225 MHz	0.002 %	Using Frequency Counter HP 53131A By Comparison Method
<b>2.</b>	<b>AC VOLTAGE</b>	<b>40 to 70 Hz</b> 10 V to 480 V	0.03 %	Using Ref. Power / Energy Calibrator, COMM 3003 Make:- Zera By Comparison Method

**Sangeeta Kunwar**  
Convenor

**Avijit Das**  
Program Manager

<b>Laboratory</b>	<b>Yadav Measurements Private Limited, Plot No. 373-375, RIICO, Bhamashah Industrial Area, Kaladwas, Udaipur, Rajasthan</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro – Technical Calibration</b>	<b>Issue Date</b>	<b>24.07.2014</b>
<b>Certificate Number</b>	<b>C-0035</b>	<b>Valid Until</b>	<b>23.07.2016</b>
<b>Last Amended on</b>	<b>11.09.2015</b>	<b>Page</b>	<b>11 of 14</b>

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
3. AC CURRENT	45 Hz to 70 Hz 1 mA to 120 mA	0.04 %	Using Ref. Power / Energy Calibrator, COMM 3003 Make:- Zera by Comparison Method
4. AC POWER / ENERGY 1 phase & 3 phase Active Cos $\phi \pm 0.1$ to 1 Reactive Sin $\phi \pm 0.10$ to 1 Apparent	45 Hz to 65 Hz 40 V to 320 V Active 1 mA to 20 mA 0.004 W to 19.2 W 20 mA to 120 A 19.2 W to 115.2 kW	0.042 % 0.011 %	Using Ref. power / energy calibrator, COMM 3003 Make:- Zera By Direct Method
	Reactive 1 mA to 20 mA 0.004 VAr to 19.2 VAr 20 mA to 120 A 19.2 VAr to 115.2 kVAr	0.043 % 0.011 %	
	Apparent 1 mA to 20 mA 0.004 VA to 19.2 VA 20 mA to 120 A 19.2 VA to 115.2 kVA	0.056 % 0.015 %	
AC POWER / ENERGY 1 phase & 3 phase Active Cos $\phi : \pm 0.01$ to $\pm 0.1$	Active 45 Hz to 65 Hz 10 V to 240 V 0.1 A to 10 A 10 mW to 720 W	0.9 %	

Sangeeta Kunwar  
Convenor

Avijit Das  
Program Manager

<b>Laboratory</b>	<b>Yadav Measurements Private Limited, Plot No. 373-375, RIICO, Bhamashah Industrial Area, Kaladwas, Udaipur, Rajasthan</b>		
<b>Accreditation Standard</b>	<b>ISO/IEC 17025: 2005</b>		
<b>Discipline</b>	<b>Electro – Technical Calibration</b>	<b>Issue Date</b>	<b>24.07.2014</b>
<b>Certificate Number</b>	<b>C-0035</b>	<b>Valid Until</b>	<b>23.07.2016</b>
<b>Last Amended on</b>	<b>11.09.2015</b>	<b>Page</b>	<b>12 of 14</b>

	<b>Quantity Measured/ Instrument</b>	<b>Range / Frequency</b>	<b>*Calibration Measurement Capability (±)</b>	<b>Remarks</b>
5.	<b>POWER FACTOR</b>	PF: -1 to +1	0.008 PF	Using Ref. power / energy calibrator, COMM 3003 Make:- Zera By Comparison Method
6.	<b>CURRENT TRANSFORMER (Primary Injection) RATIO ERROR PHASE ERROR</b>	5 A to 2000 A (Primary) 1 A , 5 A (Secondary)	0.015 % 0.65 min	Using Precision Current Transformer & Automatic Instrument Transformer Test Set (AIITS) By Direct Method
7.	<b>CURRENT TRANSFORMER RATIO ERROR PHASE ERROR</b>	10 A to 10000 A (Primary) 1 A, 5 A (Secondary) 5 A (Primary) 5 A (Secondary)	0.09 % 2.8 min	Using Portable Ct/VT Calibrator Using Portable CT Calibrator By Direct Method
8.	<b>VOLTAGE TRANSFORMER/ CAPACITOR VOLTAGE TRANSFORMER/ VOLTAGE DIVIDER RATIO ERROR PHASE ERROR</b>	2.2 k V to 33 kV (Primary) 50 V to 144 V (Secondary)	0.04 % 4 min	Using Electronic Potential Divider 33 kV & Standard Capacitor & EMVT By Direct Method
9.	<b>VOLTAGE TRANSFORMER/ CAPACITOR VOLTAGE TRANSFORMER/ VOLTAGE DIVIDER RATIO ERROR PHASE ERROR</b>	2.2. kV to 33 kV (Primary) 50 V to 144 V (Secondary)	0.15 % 5.6 min	Using Portable CT/VT Calibrator By Direct Method

**Sangeeta Kunwar**  
Convenor

**Avijit Das**  
Program Manager

**Laboratory** Yadav Measurements Private Limited, Plot No. 373-375, RIICO,  
 Bhamashah Industrial Area, Kaladwas, Udaipur, Rajasthan  
**Accreditation Standard** ISO/IEC 17025: 2005  
**Discipline** Electro – Technical Calibration **Issue Date** 24.07.2014  
**Certificate Number** C-0035 **Valid Until** 23.07.2016  
**Last Amended on** 11.09.2015 **Page** 13 of 14

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (±)	Remarks
<b>10. CT-VT COMPARATOR/ BRIDGE</b>	Current 1 A, 5 A Ratio Error Phase Error Voltage 63.5 V, 110 V Ratio Error Phase Error	0.005 % 0.3 min  0.005 % 0.3 min	Using Precision Current  Transformer & Automatic Instrument Transformer Test Set (AIITS) Using Electronic Potential Divider 33 kV & Standard Capacitor & EMVT By Direct Method
<b>SOURCE</b>			
<b>1. FREQUENCY</b>	40 Hz to 225 Hz	0.002 %	Using Ref. power / energy calibrator, COMM 3003 Make:- Zera By Comparison Method
<b>2. AC VOLTAGE</b>	<b>40 Hz to 70 Hz</b> 10 V to 480 V	0.03 %	Using Ref. power / energy calibrator, COMM 3003 Make:- Zera By Comparison Method
<b>3. AC CURRENT</b>	<b>45 Hz to 70 Hz</b> 1 mA to 120 mA	0.04 %	Using Ref. power / energy calibrator, COMM 3003 Make:- Zera By Comparison Method

Sangeeta Kunwar  
 Convenor

Avijit Das  
 Program Manager

